



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Zuccaro Inc.

SDG No.: Q3032

Contract: ZUCC01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB22	Mercury	0.076	+/-0.2	U	0.20	CV	09/08/2025	14:36	LB137106

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Zuccaro Inc.

SDG No.: Q3032

Contract: ZUCC01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB63	Mercury	0.076	+/-0.2	U	0.20	CV	09/08/2025	14:43	LB137106
CCB64	Mercury	0.076	+/-0.2	U	0.20	CV	09/08/2025	15:14	LB137106
CCB65	Mercury	0.076	+/-0.2	U	0.20	CV	09/08/2025	15:41	LB137106
CCB66	Mercury	0.076	+/-0.2	U	0.20	CV	09/08/2025	15:58	LB137106

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Zuccaro Inc.

SDG No.: Q3032

Contract: ZUCC01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	09/08/2025	12:18	LB137113
	Antimony	6.76	+/-25	U	50.0	P	09/08/2025	12:18	LB137113
	Arsenic	5.12	+/-10	U	20.0	P	09/08/2025	12:18	LB137113
	Barium	14.6	+/-50	U	100	P	09/08/2025	12:18	LB137113
	Beryllium	0.56	+/-3	U	6.00	P	09/08/2025	12:18	LB137113
	Cadmium	0.50	+/-3	U	6.00	P	09/08/2025	12:18	LB137113
	Calcium	234	+/-1000	U	2000	P	09/08/2025	12:18	LB137113
	Chromium	2.12	+/-5	U	10.0	P	09/08/2025	12:18	LB137113
	Cobalt	2.26	+/-15	U	30.0	P	09/08/2025	12:18	LB137113
	Copper	4.60	+/-10	U	20.0	P	09/08/2025	12:18	LB137113
	Iron	23.4	+/-50	U	100	P	09/08/2025	12:18	LB137113
	Lead	2.30	+/-6	U	12.0	P	09/08/2025	12:18	LB137113
	Magnesium	244	+/-1000	U	2000	P	09/08/2025	12:18	LB137113
	Manganese	5.94	+/-10	U	20.0	P	09/08/2025	12:18	LB137113
	Nickel	3.06	+/-20	U	40.0	P	09/08/2025	12:18	LB137113
	Potassium	918	+/-1000	U	2000	P	09/08/2025	12:18	LB137113
	Selenium	9.64	+/-10	U	20.0	P	09/08/2025	12:18	LB137113
	Silver	1.62	+/-5	U	10.0	P	09/08/2025	12:18	LB137113
	Sodium	868	+/-1000	U	2000	P	09/08/2025	12:18	LB137113
	Thallium	4.38	+/-20	U	40.0	P	09/08/2025	12:18	LB137113
	Vanadium	6.26	+/-20	U	40.0	P	09/08/2025	12:18	LB137113
	Zinc	16.7	+/-20	U	40.0	P	09/08/2025	12:18	LB137113

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Zuccaro Inc.

SDG No.: Q3032

Contract: ZUCC01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	09/08/2025	12:47	LB137113
	Antimony	6.76	+/-25	U	50.0	P	09/08/2025	12:47	LB137113
	Arsenic	5.12	+/-10	U	20.0	P	09/08/2025	12:47	LB137113
	Barium	14.6	+/-50	U	100	P	09/08/2025	12:47	LB137113
	Beryllium	0.56	+/-3	U	6.00	P	09/08/2025	12:47	LB137113
	Cadmium	0.50	+/-3	U	6.00	P	09/08/2025	12:47	LB137113
	Calcium	234	+/-1000	U	2000	P	09/08/2025	12:47	LB137113
	Chromium	2.12	+/-5	U	10.0	P	09/08/2025	12:47	LB137113
	Cobalt	2.26	+/-15	U	30.0	P	09/08/2025	12:47	LB137113
	Copper	4.60	+/-10	U	20.0	P	09/08/2025	12:47	LB137113
	Iron	23.4	+/-50	U	100	P	09/08/2025	12:47	LB137113
	Lead	2.30	+/-6	U	12.0	P	09/08/2025	12:47	LB137113
	Magnesium	244	+/-1000	U	2000	P	09/08/2025	12:47	LB137113
	Manganese	6.65	+/-10	J	20.0	P	09/08/2025	12:47	LB137113
	Nickel	3.06	+/-20	U	40.0	P	09/08/2025	12:47	LB137113
	Potassium	918	+/-1000	U	2000	P	09/08/2025	12:47	LB137113
	Selenium	9.64	+/-10	U	20.0	P	09/08/2025	12:47	LB137113
	Silver	1.62	+/-5	U	10.0	P	09/08/2025	12:47	LB137113
	Sodium	868	+/-1000	U	2000	P	09/08/2025	12:47	LB137113
	Thallium	4.38	+/-20	U	40.0	P	09/08/2025	12:47	LB137113
	Vanadium	6.26	+/-20	U	40.0	P	09/08/2025	12:47	LB137113
	Zinc	16.7	+/-20	U	40.0	P	09/08/2025	12:47	LB137113
CCB02	Aluminum	11.3	+/-50	U	100	P	09/08/2025	13:40	LB137113
	Antimony	6.76	+/-25	U	50.0	P	09/08/2025	13:40	LB137113
	Arsenic	5.12	+/-10	U	20.0	P	09/08/2025	13:40	LB137113
	Barium	14.6	+/-50	U	100	P	09/08/2025	13:40	LB137113
	Beryllium	0.56	+/-3	U	6.00	P	09/08/2025	13:40	LB137113
	Cadmium	0.50	+/-3	U	6.00	P	09/08/2025	13:40	LB137113
	Calcium	234	+/-1000	U	2000	P	09/08/2025	13:40	LB137113
	Chromium	2.12	+/-5	U	10.0	P	09/08/2025	13:40	LB137113
	Cobalt	2.26	+/-15	U	30.0	P	09/08/2025	13:40	LB137113
	Copper	4.60	+/-10	U	20.0	P	09/08/2025	13:40	LB137113
	Iron	23.4	+/-50	U	100	P	09/08/2025	13:40	LB137113
	Lead	2.30	+/-6	U	12.0	P	09/08/2025	13:40	LB137113
	Magnesium	244	+/-1000	U	2000	P	09/08/2025	13:40	LB137113
	Manganese	5.94	+/-10	U	20.0	P	09/08/2025	13:40	LB137113
	Nickel	3.06	+/-20	U	40.0	P	09/08/2025	13:40	LB137113
	Potassium	918	+/-1000	U	2000	P	09/08/2025	13:40	LB137113
	Selenium	9.64	+/-10	U	20.0	P	09/08/2025	13:40	LB137113

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Zuccaro Inc.

SDG No.: Q3032

Contract: ZUCC01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	09/08/2025	13:40	LB137113
	Sodium	868	+/-1000	U	2000	P	09/08/2025	13:40	LB137113
	Thallium	4.38	+/-20	U	40.0	P	09/08/2025	13:40	LB137113
	Vanadium	6.26	+/-20	U	40.0	P	09/08/2025	13:40	LB137113
	Zinc	16.7	+/-20	U	40.0	P	09/08/2025	13:40	LB137113
CCB03	Aluminum	11.3	+/-50	U	100	P	09/08/2025	14:54	LB137113
	Antimony	6.76	+/-25	U	50.0	P	09/08/2025	14:54	LB137113
	Arsenic	5.12	+/-10	U	20.0	P	09/08/2025	14:54	LB137113
	Barium	14.6	+/-50	U	100	P	09/08/2025	14:54	LB137113
	Beryllium	0.56	+/-3	U	6.00	P	09/08/2025	14:54	LB137113
	Cadmium	0.50	+/-3	U	6.00	P	09/08/2025	14:54	LB137113
	Calcium	234	+/-1000	U	2000	P	09/08/2025	14:54	LB137113
	Chromium	2.12	+/-5	U	10.0	P	09/08/2025	14:54	LB137113
	Cobalt	2.26	+/-15	U	30.0	P	09/08/2025	14:54	LB137113
	Copper	4.60	+/-10	U	20.0	P	09/08/2025	14:54	LB137113
	Iron	23.4	+/-50	U	100	P	09/08/2025	14:54	LB137113
	Lead	2.30	+/-6	U	12.0	P	09/08/2025	14:54	LB137113
	Magnesium	244	+/-1000	U	2000	P	09/08/2025	14:54	LB137113
	Manganese	5.94	+/-10	U	20.0	P	09/08/2025	14:54	LB137113
	Nickel	3.06	+/-20	U	40.0	P	09/08/2025	14:54	LB137113
	Potassium	918	+/-1000	U	2000	P	09/08/2025	14:54	LB137113
	Selenium	9.64	+/-10	U	20.0	P	09/08/2025	14:54	LB137113
	Silver	1.62	+/-5	U	10.0	P	09/08/2025	14:54	LB137113
	Sodium	868	+/-1000	U	2000	P	09/08/2025	14:54	LB137113
	Thallium	4.38	+/-20	U	40.0	P	09/08/2025	14:54	LB137113
	Vanadium	6.26	+/-20	U	40.0	P	09/08/2025	14:54	LB137113
	Zinc	16.7	+/-20	U	40.0	P	09/08/2025	14:54	LB137113
CCB04	Aluminum	11.3	+/-50	U	100	P	09/08/2025	15:52	LB137113
	Antimony	6.76	+/-25	U	50.0	P	09/08/2025	15:52	LB137113
	Arsenic	5.12	+/-10	U	20.0	P	09/08/2025	15:52	LB137113
	Barium	14.6	+/-50	U	100	P	09/08/2025	15:52	LB137113
	Beryllium	0.56	+/-3	U	6.00	P	09/08/2025	15:52	LB137113
	Cadmium	0.50	+/-3	U	6.00	P	09/08/2025	15:52	LB137113
	Calcium	234	+/-1000	U	2000	P	09/08/2025	15:52	LB137113
	Chromium	2.12	+/-5	U	10.0	P	09/08/2025	15:52	LB137113
	Cobalt	2.26	+/-15	U	30.0	P	09/08/2025	15:52	LB137113
	Copper	4.60	+/-10	U	20.0	P	09/08/2025	15:52	LB137113
	Iron	23.4	+/-50	U	100	P	09/08/2025	15:52	LB137113
	Lead	2.30	+/-6	U	12.0	P	09/08/2025	15:52	LB137113

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Zuccaro Inc.

SDG No.: Q3032

Contract: ZUCC01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	09/08/2025	15:52	LB137113
	Manganese	5.94	+/-10	U	20.0	P	09/08/2025	15:52	LB137113
	Nickel	3.06	+/-20	U	40.0	P	09/08/2025	15:52	LB137113
	Potassium	918	+/-1000	U	2000	P	09/08/2025	15:52	LB137113
	Selenium	9.64	+/-10	U	20.0	P	09/08/2025	15:52	LB137113
	Silver	1.62	+/-5	U	10.0	P	09/08/2025	15:52	LB137113
	Sodium	868	+/-1000	U	2000	P	09/08/2025	15:52	LB137113
	Thallium	4.38	+/-20	U	40.0	P	09/08/2025	15:52	LB137113
	Vanadium	6.26	+/-20	U	40.0	P	09/08/2025	15:52	LB137113
	Zinc	16.7	+/-20	U	40.0	P	09/08/2025	15:52	LB137113
CCB05	Aluminum	11.3	+/-50	U	100	P	09/08/2025	16:17	LB137113
	Antimony	6.76	+/-25	U	50.0	P	09/08/2025	16:17	LB137113
	Arsenic	5.12	+/-10	U	20.0	P	09/08/2025	16:17	LB137113
	Barium	14.6	+/-50	U	100	P	09/08/2025	16:17	LB137113
	Beryllium	0.56	+/-3	U	6.00	P	09/08/2025	16:17	LB137113
	Cadmium	0.50	+/-3	U	6.00	P	09/08/2025	16:17	LB137113
	Calcium	234	+/-1000	U	2000	P	09/08/2025	16:17	LB137113
	Chromium	2.12	+/-5	U	10.0	P	09/08/2025	16:17	LB137113
	Cobalt	2.26	+/-15	U	30.0	P	09/08/2025	16:17	LB137113
	Copper	4.60	+/-10	U	20.0	P	09/08/2025	16:17	LB137113
	Iron	23.4	+/-50	U	100	P	09/08/2025	16:17	LB137113
	Lead	2.30	+/-6	U	12.0	P	09/08/2025	16:17	LB137113
	Magnesium	244	+/-1000	U	2000	P	09/08/2025	16:17	LB137113
	Manganese	5.94	+/-10	U	20.0	P	09/08/2025	16:17	LB137113
	Nickel	3.06	+/-20	U	40.0	P	09/08/2025	16:17	LB137113
	Potassium	918	+/-1000	U	2000	P	09/08/2025	16:17	LB137113
	Selenium	9.64	+/-10	U	20.0	P	09/08/2025	16:17	LB137113
	Silver	1.62	+/-5	U	10.0	P	09/08/2025	16:17	LB137113
	Sodium	868	+/-1000	U	2000	P	09/08/2025	16:17	LB137113
	Thallium	4.38	+/-20	U	40.0	P	09/08/2025	16:17	LB137113
	Vanadium	6.26	+/-20	U	40.0	P	09/08/2025	16:17	LB137113
	Zinc	16.7	+/-20	U	40.0	P	09/08/2025	16:17	LB137113

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Zuccaro Inc.

SDG No.: Q3032

Contract: ZUCC01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	------	---	------------------	------------------	---------------
