



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Zuccaro Inc.

SDG No.: Q3032

Contract: ZUCC01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	16.4	20.0	82	65 - 135	P	09/09/2025	12:36	LB137122
	Barium	92.2	100	92	65 - 135	P	09/09/2025	12:36	LB137122
	Cadmium	5.53	6.0	92	65 - 135	P	09/09/2025	12:36	LB137122
	Chromium	9.50	10.0	95	65 - 135	P	09/09/2025	12:36	LB137122
	Lead	9.44	12.0	79	65 - 135	P	09/09/2025	12:36	LB137122
	Selenium	19.9	20.0	100	65 - 135	P	09/09/2025	12:36	LB137122
	Silver	9.64	10.0	96	65 - 135	P	09/09/2025	12:36	LB137122
CRA	Mercury	0.21	0.2	107	70 - 130	CV	09/10/2025	10:24	LB137126