



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Zuccaro Inc.

SDG No.: Q3032

Contract: ZUCC01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB24	Mercury	0.076	+/-0.2	U	0.20	CV	09/10/2025	09:56	LB137126

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Zuccaro Inc.

SDG No.: Q3032

Contract: ZUCC01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB70	Mercury	0.076	+/-0.2	U	0.20	CV	09/10/2025	10:09	LB137126
CCB71	Mercury	0.076	+/-0.2	U	0.20	CV	09/10/2025	11:07	LB137126
CCB72	Mercury	0.076	+/-0.2	U	0.20	CV	09/10/2025	12:31	LB137126
CCB73	Mercury	0.076	+/-0.2	U	0.20	CV	09/10/2025	13:26	LB137126

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Zuccaro Inc.

SDG No.: Q3032

Contract: ZUCC01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	09/09/2025	12:31	LB137122
	Barium	14.6	+/-50	U	100	P	09/09/2025	12:31	LB137122
	Cadmium	0.50	+/-3	U	6.00	P	09/09/2025	12:31	LB137122
	Chromium	2.12	+/-5	U	10.0	P	09/09/2025	12:31	LB137122
	Lead	2.30	+/-6	U	12.0	P	09/09/2025	12:31	LB137122
	Selenium	9.64	+/-10	U	20.0	P	09/09/2025	12:31	LB137122
	Silver	1.62	+/-5	U	10.0	P	09/09/2025	12:31	LB137122

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Zuccaro Inc.

SDG No.: Q3032

Contract: ZUCC01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	09/09/2025	13:26	LB137122
	Barium	14.6	+/-50	U	100	P	09/09/2025	13:26	LB137122
	Cadmium	0.50	+/-3	U	6.00	P	09/09/2025	13:26	LB137122
	Chromium	2.12	+/-5	U	10.0	P	09/09/2025	13:26	LB137122
	Lead	2.30	+/-6	U	12.0	P	09/09/2025	13:26	LB137122
	Selenium	9.64	+/-10	U	20.0	P	09/09/2025	13:26	LB137122
	Silver	1.62	+/-5	U	10.0	P	09/09/2025	13:26	LB137122
CCB02	Arsenic	5.12	+/-10	U	20.0	P	09/09/2025	14:19	LB137122
	Barium	14.6	+/-50	U	100	P	09/09/2025	14:19	LB137122
	Cadmium	0.50	+/-3	U	6.00	P	09/09/2025	14:19	LB137122
	Chromium	2.12	+/-5	U	10.0	P	09/09/2025	14:19	LB137122
	Lead	2.30	+/-6	U	12.0	P	09/09/2025	14:19	LB137122
	Selenium	9.64	+/-10	U	20.0	P	09/09/2025	14:19	LB137122
	Silver	1.62	+/-5	U	10.0	P	09/09/2025	14:19	LB137122
CCB03	Arsenic	5.12	+/-10	U	20.0	P	09/09/2025	15:25	LB137122
	Barium	14.6	+/-50	U	100	P	09/09/2025	15:25	LB137122
	Cadmium	0.50	+/-3	U	6.00	P	09/09/2025	15:25	LB137122
	Chromium	2.12	+/-5	U	10.0	P	09/09/2025	15:25	LB137122
	Lead	2.30	+/-6	U	12.0	P	09/09/2025	15:25	LB137122
	Selenium	9.64	+/-10	U	20.0	P	09/09/2025	15:25	LB137122
	Silver	1.62	+/-5	U	10.0	P	09/09/2025	15:25	LB137122
CCB04	Arsenic	5.12	+/-10	U	20.0	P	09/09/2025	16:28	LB137122
	Barium	14.6	+/-50	U	100	P	09/09/2025	16:28	LB137122
	Cadmium	0.50	+/-3	U	6.00	P	09/09/2025	16:28	LB137122
	Chromium	2.12	+/-5	U	10.0	P	09/09/2025	16:28	LB137122
	Lead	2.30	+/-6	U	12.0	P	09/09/2025	16:28	LB137122
	Selenium	9.64	+/-10	U	20.0	P	09/09/2025	16:28	LB137122
	Silver	1.62	+/-5	U	10.0	P	09/09/2025	16:28	LB137122
CCB05	Arsenic	5.12	+/-10	U	20.0	P	09/09/2025	17:17	LB137122
	Barium	14.6	+/-50	U	100	P	09/09/2025	17:17	LB137122
	Cadmium	0.50	+/-3	U	6.00	P	09/09/2025	17:17	LB137122
	Chromium	2.12	+/-5	U	10.0	P	09/09/2025	17:17	LB137122
	Lead	2.30	+/-6	U	12.0	P	09/09/2025	17:17	LB137122
	Selenium	9.64	+/-10	U	20.0	P	09/09/2025	17:17	LB137122
	Silver	1.62	+/-5	U	10.0	P	09/09/2025	17:17	LB137122
CCB06	Arsenic	5.12	+/-10	U	20.0	P	09/09/2025	17:50	LB137122
	Barium	14.6	+/-50	U	100	P	09/09/2025	17:50	LB137122
	Cadmium	0.50	+/-3	U	6.00	P	09/09/2025	17:50	LB137122
	Chromium	2.12	+/-5	U	10.0	P	09/09/2025	17:50	LB137122

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Zuccaro Inc.

SDG No.: Q3032

Contract: ZUCC01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	2.30	+/-6	U	12.0	P	09/09/2025	17:50	LB137122
	Selenium	9.64	+/-10	U	20.0	P	09/09/2025	17:50	LB137122
	Silver	1.62	+/-5	U	10.0	P	09/09/2025	17:50	LB137122