

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Zuccaro Inc.</u>	SDG No.: <u>Q3032</u>
Contract: <u>ZUCC01</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	-0.30			-20	20	09/09/2025	12:42	LB137122
	Barium	-8.71	6.0	145	-94	106	09/09/2025	12:42	LB137122
	Cadmium	-1.23	1.0	123	-5	7	09/09/2025	12:42	LB137122
	Chromium	50.6	52.0	97	42	62	09/09/2025	12:42	LB137122
	Lead	-3.71			-12	12	09/09/2025	12:42	LB137122
	Selenium	4.10			-20	20	09/09/2025	12:42	LB137122
	Silver	0.80			-10	10	09/09/2025	12:42	LB137122
ICSAB01	Arsenic	107	104	103	88.4	120	09/09/2025	12:46	LB137122
	Barium	496	537	92	437	637	09/09/2025	12:46	LB137122
	Cadmium	1010	972	104	826	1120	09/09/2025	12:46	LB137122
	Chromium	606	542	112	460	624	09/09/2025	12:46	LB137122
	Lead	44.4	49.0	91	37	61	09/09/2025	12:46	LB137122
	Selenium	56.1	46.0	122	26	66	09/09/2025	12:46	LB137122
	Silver	221	201	110	170	232	09/09/2025	12:46	LB137122