

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Zuccaro Inc.</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3032</u>
Contract:	<u>ZUCC01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3051-02</u>	Client ID:	<u>VNJ-223DUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3051-02DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	204	J	175	J	15		P
Cadmium	ug/L	20	30.0	U	30.0	U			P
Chromium	ug/L	20	50.0	U	50.0	U			P
Lead	ug/L	20	60.0	U	60.0	U			P
Mercury	ug/L	20	2.00	U	2.00	U			CV
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Zuccaro Inc.</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3032</u>
Contract:	<u>ZUCC01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3051-02MS</u>	Client ID:	<u>VNJ-223MSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3051-02MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	3750		3910		4		P
Barium	ug/L	20	1160		1430		21	*	P
Cadmium	ug/L	20	887		917		3		P
Chromium	ug/L	20	1850		1910		3		P
Lead	ug/L	20	4280		4390		3		P
Mercury	ug/L	20	45.0		39.9		12		CV
Selenium	ug/L	20	9480		9960		5		P
Silver	ug/L	20	327		336		3		P