

CALIBRATION VERIFICATION SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>ZUCC01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3032</u>
Continuing Calib Date:	<u>09/08/2025</u>	Initial Calibration Date(s):	<u>08/19/2025</u> <u>08/19/2025</u>
Continuing Calib Time:	<u>12:11</u>	Initial Calibration Time(s):	<u>10:02</u> <u>17:36</u>

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	5.81	5.81	5.71	5.91	0.00
Aroclor-1242-2 (2)	5.83	5.83	5.73	5.93	0.00
Aroclor-1242-3 (3)	5.90	5.89	5.79	5.99	-0.01
Aroclor-1242-4 (4)	5.99	5.99	5.89	6.09	0.00
Aroclor-1242-5 (5)	6.72	6.72	6.62	6.82	0.00
Tetrachloro-m-xylene	4.66	4.66	4.56	4.76	0.00
Decachlorobiphenyl	10.44	10.43	10.33	10.53	-0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ZUCC01

Lab Code: ACE

SDG NO.: Q3032

Continuing Calib Date: 09/08/2025

Initial Calibration Date(s): 08/19/2025

08/19/2025

Continuing Calib Time: 12:11

Initial Calibration Time(s): 10:02

17:36

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.90	4.90	4.80	5.00	0.00
Aroclor-1242-2 (2)	4.96	4.96	4.86	5.06	0.01
Aroclor-1242-3 (3)	5.08	5.08	4.98	5.18	0.01
Aroclor-1242-4 (4)	5.16	5.16	5.06	5.26	0.00
Aroclor-1242-5 (5)	5.68	5.68	5.58	5.78	0.00
Tetrachloro-m-xylene	3.80	3.80	3.70	3.90	0.00
Decachlorobiphenyl	8.81	8.81	8.71	8.91	0.00