

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ZUCC01

Lab Code: ACE

SDG NO.: Q3032

Continuing Calib Date: 09/09/2025

Initial Calibration Date(s): 08/19/2025

08/19/2025

Continuing Calib Time: 01:06

Initial Calibration Time(s): 10:02

17:36

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.80	5.81	5.71	5.91	0.01
Aroclor-1016-2	(2)	5.82	5.83	5.73	5.93	0.01
Aroclor-1016-3	(3)	5.89	5.89	5.79	5.99	0.00
Aroclor-1016-4	(4)	5.98	5.99	5.89	6.09	0.01
Aroclor-1016-5	(5)	6.27	6.28	6.18	6.38	0.01
Aroclor-1260-1	(1)	7.39	7.40	7.30	7.50	0.01
Aroclor-1260-2	(2)	7.64	7.65	7.55	7.75	0.01
Aroclor-1260-3	(3)	8.00	8.01	7.91	8.11	0.01
Aroclor-1260-4	(4)	8.23	8.24	8.14	8.34	0.01
Aroclor-1260-5	(5)	8.55	8.56	8.46	8.66	0.01
Tetrachloro-m-xylene		4.65	4.65	4.55	4.75	0.00
Decachlorobiphenyl		10.42	10.43	10.33	10.53	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ZUCC01

Lab Code: ACE

SDG NO.: Q3032

Continuing Calib Date: 09/09/2025

Initial Calibration Date(s): 08/19/2025

08/19/2025

Continuing Calib Time: 01:06

Initial Calibration Time(s): 10:02

17:36

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.89	4.90	4.80	5.00	0.01
Aroclor-1016-2	(2)	4.95	4.96	4.86	5.06	0.01
Aroclor-1016-3	(3)	5.07	5.08	4.98	5.18	0.01
Aroclor-1016-4	(4)	5.11	5.12	5.02	5.22	0.01
Aroclor-1016-5	(5)	5.33	5.33	5.23	5.43	0.00
Aroclor-1260-1	(1)	6.36	6.36	6.26	6.46	0.00
Aroclor-1260-2	(2)	6.54	6.55	6.45	6.65	0.01
Aroclor-1260-3	(3)	6.70	6.70	6.60	6.80	0.00
Aroclor-1260-4	(4)	7.16	7.17	7.07	7.27	0.01
Aroclor-1260-5	(5)	7.40	7.41	7.31	7.51	0.01
Tetrachloro-m-xylene		3.80	3.80	3.70	3.90	0.00
Decachlorobiphenyl		8.80	8.82	8.72	8.92	0.02