

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ZUCC01
Lab Code: ACE **SDG NO.:** Q3032
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/08/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074963.D **Time Analyzed:** 16:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.807	5.706	5.906	495.470	500.000	-0.9
Aroclor-1016-2	5.828	5.727	5.927	508.330	500.000	1.7
Aroclor-1016-3	5.891	5.790	5.990	457.120	500.000	-8.6
Aroclor-1016-4	5.988	5.888	6.088	484.050	500.000	-3.2
Aroclor-1016-5	6.280	6.180	6.380	498.190	500.000	-0.4
Aroclor-1260-1	7.397	7.297	7.497	464.500	500.000	-7.1
Aroclor-1260-2	7.649	7.550	7.750	463.230	500.000	-7.4
Aroclor-1260-3	8.007	7.908	8.108	488.340	500.000	-2.3
Aroclor-1260-4	8.235	8.136	8.336	459.660	500.000	-8.1
Aroclor-1260-5	8.560	8.462	8.662	456.930	500.000	-8.6
Decachlorobiphenyl	10.427	10.333	10.533	46.690	50.000	-6.6
Tetrachloro-m-xylene	4.656	4.554	4.754	49.480	50.000	-1.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ZUCC01
Lab Code: ACE **SDG NO.:** Q3032
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/08/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074963.D **Time Analyzed:** 16:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.898	4.799	4.999	498.850	500.000	-0.2
Aroclor-1016-2	4.956	4.855	5.055	509.890	500.000	2.0
Aroclor-1016-3	5.076	4.976	5.176	485.380	500.000	-2.9
Aroclor-1016-4	5.117	5.017	5.217	485.820	500.000	-2.8
Aroclor-1016-5	5.331	5.231	5.431	449.260	500.000	-10.1
Aroclor-1260-1	6.359	6.261	6.461	459.180	500.000	-8.2
Aroclor-1260-2	6.545	6.448	6.648	440.890	500.000	-11.8
Aroclor-1260-3	6.700	6.602	6.802	403.430	500.000	-19.3
Aroclor-1260-4	7.169	7.072	7.272	429.010	500.000	-14.2
Aroclor-1260-5	7.408	7.311	7.511	431.040	500.000	-13.8
Decachlorobiphenyl	8.808	8.715	8.915	41.540	50.000	-16.9
Tetrachloro-m-xylene	3.801	3.699	3.899	54.430	50.000	8.9