

Surrogate Summary

SDG No.: **Q3032**

Client: **Zuccaro Inc.**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD089961.D	PIBLK-PD089961.D	Decachlorobiphen	1	20	19.3	96		30 (57)	150 (171)
		Tetrachloro-m-xyl	1	20	17.4	87		30 (61)	150 (148)
		Decachlorobiphen	2	20	18.3	91		30 (57)	150 (171)
		Tetrachloro-m-xyl	2	20	17.5	88		30 (61)	150 (148)
I.BLK-PD090182.D	PIBLK-PD090182.D	Decachlorobiphen	1	20	17.9	89		30 (57)	150 (171)
		Tetrachloro-m-xyl	1	20	18.6	93		30 (61)	150 (148)
		Decachlorobiphen	2	20	16.0	80		30 (57)	150 (171)
		Tetrachloro-m-xyl	2	20	20.7	104		30 (61)	150 (148)
Q3032-04	SOIL-PILE-2	Decachlorobiphen	1	20	8.29	41		30 (20)	150 (144)
		Tetrachloro-m-xyl	1	20	10.3	52		30 (19)	150 (148)
		Decachlorobiphen	2	20	4.86	24	*	30 (20)	150 (144)
		Tetrachloro-m-xyl	2	20	11.4	57		30 (19)	150 (148)
PB169592BL	PB169592BL	Decachlorobiphen	1	20	15.7	78		30 (20)	150 (144)
		Tetrachloro-m-xyl	1	20	17.0	85		30 (19)	150 (148)
		Decachlorobiphen	2	20	12.2	61		30 (20)	150 (144)
		Tetrachloro-m-xyl	2	20	18.4	92		30 (19)	150 (148)
I.BLK-PD090194.D	PIBLK-PD090194.D	Decachlorobiphen	1	20	18.1	90		30 (57)	150 (171)
		Tetrachloro-m-xyl	1	20	19.0	95		30 (61)	150 (148)
		Decachlorobiphen	2	20	14.9	74		30 (57)	150 (171)
		Tetrachloro-m-xyl	2	20	20.6	103		30 (61)	150 (148)
I.BLK-PD090202.D	PIBLK-PD090202.D	Decachlorobiphen	1	20	19.0	95		30 (57)	150 (171)
		Tetrachloro-m-xyl	1	20	18.8	94		30 (61)	150 (148)
		Decachlorobiphen	2	20	18.2	91		30 (57)	150 (171)
		Tetrachloro-m-xyl	2	20	20.6	103		30 (61)	150 (148)
Q3032-01	SOIL-PILE-1	Decachlorobiphen	1	20	14.3	71		30 (20)	150 (144)
		Tetrachloro-m-xyl	1	20	17.1	86		30 (19)	150 (148)
		Decachlorobiphen	2	20	12.7	64		30 (20)	150 (144)
		Tetrachloro-m-xyl	2	20	17.9	90		30 (19)	150 (148)
Q3032-01MS	SOIL-PILE-1MS	Decachlorobiphen	1	20	13.7	69		30 (20)	150 (144)
		Tetrachloro-m-xyl	1	20	17.4	87		30 (19)	150 (148)
		Decachlorobiphen	2	20	13.2	66		30 (20)	150 (144)
		Tetrachloro-m-xyl	2	20	18.8	94		30 (19)	150 (148)
Q3032-01MSD	SOIL-PILE-1MSD	Decachlorobiphen	1	20	13.4	67		30 (20)	150 (144)
		Tetrachloro-m-xyl	1	20	17.5	87		30 (19)	150 (148)
		Decachlorobiphen	2	20	12.9	65		30 (20)	150 (144)
		Tetrachloro-m-xyl	2	20	18.8	94		30 (19)	150 (148)
I.BLK-PD090208.D	PIBLK-PD090208.D	Decachlorobiphen	1	20	17.8	89		30 (57)	150 (171)
		Tetrachloro-m-xyl	1	20	18.8	94		30 (61)	150 (148)
		Decachlorobiphen	2	20	16.6	83		30 (57)	150 (171)
		Tetrachloro-m-xyl	2	20	20.6	103		30 (61)	150 (148)
I.BLK-PD090211.D	PIBLK-PD090211.D	Decachlorobiphen	1	20	19.6	98		30 (57)	150 (171)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: Q3032

Client: Zuccaro Inc.

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD090211.D	PIBLK-PD090211.D	Tetrachloro-m-xyl	1	20	19.4	97		30 (61)	150 (148)
		Decachlorobiphen	2	20	19.8	99		30 (57)	150 (171)
		Tetrachloro-m-xyl	2	20	22.2	111		30 (61)	150 (148)
PB169592BS	PB169592BS	Decachlorobiphen	1	20	24.2	121		30 (20)	150 (144)
		Tetrachloro-m-xyl	1	20	24.9	124		30 (19)	150 (148)
		Decachlorobiphen	2	20	24.3	122		30 (20)	150 (144)
		Tetrachloro-m-xyl	2	20	24.9	125		30 (19)	150 (148)
		Decachlorobiphen	1	20	20.8	104		30 (57)	150 (171)
I.BLK-PD090216.D	PIBLK-PD090216.D	Tetrachloro-m-xyl	1	20	20.2	101		30 (61)	150 (148)
		Decachlorobiphen	2	20	21.2	106		30 (57)	150 (171)
		Tetrachloro-m-xyl	2	20	21.8	109		30 (61)	150 (148)
		Decachlorobiphen	1	20	20.8	104		30 (57)	150 (171)