

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: ZUCC01

Lab Code: ACE

SDG NO.: Q3032

Instrument ID: ECD_S

Calibration Date(s): 09/09/2025 09/09/2025

Calibration Times: 11:32 13:08

GC Column: RTX-CLP

ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS031668.D</u>	CF 500 =	<u>PS031669.D</u>		
CF 750 =		<u>PS031670.D</u>	CF 1000 =	<u>PS031671.D</u>	CF 1500 =	<u>PS031672.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	23314400000	21668300000	20672900000	19997800000	19215100000	20973700000	8
2,4-D	3623310000	3886470000	3301510000	3166900000	3150010000	3425640000	9
2,4-DCAA	3985360000	3774560000	3531130000	3388800000	3327800000	3601530000	8

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: ZUCC01

Lab Code: ACE

SDG NO.: Q3032

Instrument ID: ECD_S

Calibration Date(s): 09/09/2025 09/09/2025

Calibration Times: 11:32 13:08

GC Column: RTX-CLP2

ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS031668.D</u>	CF 500 =	<u>PS031669.D</u>		
CF 750 =		<u>PS031670.D</u>	CF 1000 =	<u>PS031671.D</u>	CF 1500 =	<u>PS031672.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	16378200000	15425200000	14913400000	14552800000	14002000000	15054300000	6
2,4-D	1798080000	1663870000	1619090000	1589690000	1554820000	1645110000	6
2,4-DCAA	975058000	912799000	892983000	889945000	873713000	908899000	4