

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ZUCC01
Lab Code: ACE **SDG NO.:** Q3032
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/11/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031705.D **Time Analyzed:** 18:00

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.288	9.185	9.385	762.010	712.500	6.9
2,4-D	8.416	8.314	8.514	755.300	705.000	7.1
2,4-DCAA	7.287	7.185	7.385	801.560	750.000	6.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ZUCC01
Lab Code: ACE **SDG NO.:** Q3032
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/11/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031705.D **Time Analyzed:** 18:00

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.895	9.793	9.993	742.490	712.500	4.2
2,4-D	9.003	8.900	9.100	709.910	705.000	0.7
2,4-DCAA	7.744	7.643	7.843	756.270	750.000	0.8