

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ZUCC01
Lab Code: ACE **SDG NO.:** Q3032
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/11/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031714.D **Time Analyzed:** 21:44

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.287	9.185	9.385	763.930	712.500	7.2
2,4-D	8.415	8.314	8.514	744.760	705.000	5.6
2,4-DCAA	7.286	7.185	7.385	804.150	750.000	7.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ZUCC01
Lab Code: ACE **SDG NO.:** Q3032
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/11/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031714.D **Time Analyzed:** 21:44

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.894	9.793	9.993	735.030	712.500	3.2
2,4-D	9.002	8.900	9.100	708.410	705.000	0.5
2,4-DCAA	7.744	7.643	7.843	749.090	750.000	-0.1