

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ZUCC01

Lab Code: ACE

SDG NO.: Q3032

Continuing Calib Date: 09/11/2025

Initial Calibration Date(s): 08/20/2025

08/20/2025

Continuing Calib Time: 22:10

Initial Calibration Time(s): 15:53

17:20

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.00	9.00	8.90	9.10	0.00
Tetrachloro-m-xylene	3.53	3.53	3.43	3.63	0.00
gamma-BHC (Lindane)	4.30	4.31	4.21	4.41	0.01
Heptachlor	4.89	4.90	4.80	5.00	0.01
Heptachlor epoxide	5.65	5.65	5.55	5.75	0.00
Endrin	6.53	6.53	6.43	6.63	0.00
Methoxychlor	7.45	7.45	7.35	7.55	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ZUCC01

Lab Code: ACE

SDG NO.: Q3032

Continuing Calib Date: 09/11/2025

Initial Calibration Date(s): 08/20/2025

08/20/2025

Continuing Calib Time: 22:10

Initial Calibration Time(s): 15:53

17:20

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	7.98	7.99	7.89	8.09	0.01
Tetrachloro-m-xylene	2.82	2.83	2.73	2.93	0.01
gamma-BHC (Lindane)	3.66	3.66	3.56	3.76	0.00
Heptachlor	4.01	4.01	3.91	4.11	0.00
Heptachlor epoxide	4.79	4.80	4.70	4.90	0.01
Endrin	5.70	5.70	5.60	5.80	0.00
Methoxychlor	6.67	6.68	6.58	6.78	0.01