

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ZUCC01
Lab Code: ACE **SDG NO.:** Q3032
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/20/2025 08/20/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/11/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097150.D **Time Analyzed:** 17:24

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.999	8.904	9.104	44.940	50.000	-10.1
Endrin	6.530	6.434	6.634	48.200	50.000	-3.6
gamma-BHC (Lindane)	4.302	4.205	4.405	49.670	50.000	-0.7
Heptachlor	4.892	4.796	4.996	50.160	50.000	0.3
Heptachlor epoxide	5.649	5.554	5.754	46.910	50.000	-6.2
Methoxychlor	7.450	7.354	7.554	46.120	50.000	-7.8
Tetrachloro-m-xylene	3.527	3.430	3.630	48.670	50.000	-2.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ZUCC01
Lab Code: ACE **SDG NO.:** Q3032
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/20/2025 08/20/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/11/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097150.D **Time Analyzed:** 17:24

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.983	7.887	8.087	48.950	50.000	-2.1
Endrin	5.702	5.604	5.804	53.190	50.000	6.4
gamma-BHC (Lindane)	3.659	3.562	3.762	52.100	50.000	4.2
Heptachlor	4.007	3.911	4.111	50.010	50.000	0.0
Heptachlor epoxide	4.791	4.695	4.895	50.970	50.000	1.9
Methoxychlor	6.673	6.577	6.777	49.940	50.000	-0.1
Tetrachloro-m-xylene	2.823	2.725	2.925	47.830	50.000	-4.3