

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ZUCC01
Lab Code: ACE **SDG NO.:** Q3032
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/20/2025 08/20/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/11/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097159.D **Time Analyzed:** 19:27

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.999	8.904	9.104	44.360	50.000	-11.3
Endrin	6.529	6.434	6.634	47.490	50.000	-5.0
gamma-BHC (Lindane)	4.301	4.205	4.405	49.190	50.000	-1.6
Heptachlor	4.892	4.796	4.996	49.570	50.000	-0.9
Heptachlor epoxide	5.649	5.554	5.754	45.860	50.000	-8.3
Methoxychlor	7.451	7.354	7.554	44.200	50.000	-11.6
Tetrachloro-m-xylene	3.526	3.430	3.630	48.080	50.000	-3.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ZUCC01
Lab Code: ACE **SDG NO.:** Q3032
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/20/2025 08/20/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/11/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097159.D **Time Analyzed:** 19:27

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	7.983	7.887	8.087	49.050	50.000	-1.9
Endrin	5.701	5.604	5.804	52.880	50.000	5.8
gamma-BHC (Lindane)	3.659	3.562	3.762	52.150	50.000	4.3
Heptachlor	4.007	3.911	4.111	50.030	50.000	0.1
Heptachlor epoxide	4.791	4.695	4.895	51.080	50.000	2.2
Methoxychlor	6.673	6.577	6.777	49.750	50.000	-0.5
Tetrachloro-m-xylene	2.823	2.725	2.925	47.800	50.000	-4.4