

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL082025

Review By	Abdul	Review On	8/22/2025 9:15:22 AM
Supervise By		Supervise On	
SubDirectory	PL082025	HP Acquire Method	HP Processing Method pl082025 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24890,PP24651
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL096920.D	20 Aug 2025 14:58		AR\AJ	Ok
2	I.BLK	I.BLK	PL096921.D	20 Aug 2025 15:12		AR\AJ	Ok
3	PEM	PEM	PL096922.D	20 Aug 2025 15:25		AR\AJ	Ok,M
4	RESCHK	RESCHK	PL096923.D	20 Aug 2025 15:39		AR\AJ	Ok
5	PSTDICC100	PSTDICC100	PL096924.D	20 Aug 2025 15:53		AR\AJ	Ok
6	PSTDICC075	PSTDICC075	PL096925.D	20 Aug 2025 16:06		AR\AJ	Ok
7	PSTDICC050	PSTDICC050	PL096926.D	20 Aug 2025 16:20	Comp#18 kept on L.R. for 2nd column	AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PL096927.D	20 Aug 2025 16:33		AR\AJ	Ok
9	PSTDICC005	PSTDICC005	PL096928.D	20 Aug 2025 17:20		AR\AJ	Ok,M
10	PCHLORICC1000	PCHLORICC1000	PL096929.D	20 Aug 2025 17:33		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PL096930.D	20 Aug 2025 17:47		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PL096931.D	20 Aug 2025 18:01		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PL096932.D	20 Aug 2025 18:14		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PL096933.D	20 Aug 2025 18:28		AR\AJ	Ok
15	PTOXICC1000	PTOXICC1000	PL096934.D	20 Aug 2025 18:42		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PL096935.D	20 Aug 2025 18:55		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PL096936.D	20 Aug 2025 19:09		AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL082025

Review By	Abdul	Review On	8/22/2025 9:15:22 AM
Supervise By		Supervise On	
SubDirectory	PL082025	HP Acquire Method	HP Processing Method pl082025 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24890,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	PTOXICC250	PTOXICC250	PL096937.D	20 Aug 2025 19:23		AR\AJ	Ok
19	PTOXICC100	PTOXICC100	PL096938.D	20 Aug 2025 19:36		AR\AJ	Ok,M
20	PSTDICV050	ICVPL082025	PL096939.D	20 Aug 2025 19:50		AR\AJ	Ok,M
21	PCHLORICV500	ICVPL082025CHLOR	PL096940.D	20 Aug 2025 20:03		AR\AJ	Ok
22	PTOXICV500	ICVPL082025TOX	PL096941.D	20 Aug 2025 20:17	ICV Fail on higher side	AR\AJ	Not Ok
23	PTOXICV500	ICVPL082025	PL096942.D	21 Aug 2025 13:18		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL091325

Review By	Review On
Supervise By	Supervise On
SubDirectory	PL091325
HP Acquire Method	HP Processing Method
	p1082025 8081
STD. NAME	STD REF.#
Tune/Reschk	PP24890,PP24651
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL097147.D	11 Sep 2025 16:14		AR\AJ	Ok
2	I.BLK	I.BLK	PL097148.D	11 Sep 2025 16:28		AR\AJ	Ok
3	PEM	PEM	PL097149.D	11 Sep 2025 17:11		AR\AJ	Ok
4	PSTDCCC050	PSTDCCC050	PL097150.D	11 Sep 2025 17:24		AR\AJ	Ok,NS
5	PB169663BL	PB169663BL	PL097151.D	11 Sep 2025 17:38		AR\AJ	Ok
6	PB169663BS	PB169663BS	PL097152.D	11 Sep 2025 17:52		AR\AJ	Ok,NS
7	PB169589TB	PB169589TB	PL097153.D	11 Sep 2025 18:05		AR\AJ	Ok
8	Q3032-02	SOIL-PILE-1	PL097154.D	11 Sep 2025 18:19		AR\AJ	Ok
9	Q3032-02MS	SOIL-PILE-1MS	PL097155.D	11 Sep 2025 18:32		AR\AJ	Ok,NS
10	Q3032-02MSD	SOIL-PILE-1MSD	PL097156.D	11 Sep 2025 18:46		AR\AJ	Ok,NS
11	Q3032-05	SOIL-PILE-2	PL097157.D	11 Sep 2025 19:00		AR\AJ	Ok
12	I.BLK	I.BLK	PL097158.D	11 Sep 2025 19:13		AR\AJ	Ok
13	PSTDCCC050	PSTDCCC050	PL097159.D	11 Sep 2025 19:27		AR\AJ	Ok,NS
14	PB169648BL	PB169648BL	PL097160.D	11 Sep 2025 19:54		AR\AJ	Ok
15	PB169648BS	PB169648BS	PL097161.D	11 Sep 2025 20:08		AR\AJ	Ok,NS
16	Q3065-01	VNJ-204	PL097162.D	11 Sep 2025 20:21	need dilution	AR\AJ	Dilution
17	Q3065-01MS	VNJ-204MS	PL097163.D	11 Sep 2025 20:35	some compound recovery fail , F flag in comp#12,16 in 1st column	AR\AJ	Not Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL091325

Review By	Review On
Supervise By	Supervise On
SubDirectory	PL091325
HP Acquire Method	HP Processing Method
	pl082025 8081
STD. NAME	STD REF.#
Tune/Reschk	PP24890,PP24651
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	Q3065-01MSD	VNJ-204MSD	PL097164.D	11 Sep 2025 20:48	some compound recovery fail , F flag in comp#5,12,16 in 1st column	AR\AJ	Not Ok
19	Q3066-01	354	PL097165.D	11 Sep 2025 21:02		AR\AJ	Ok,NS
20	Q3066-03	VNJ-242	PL097166.D	11 Sep 2025 21:16		AR\AJ	Ok,NS
21	Q3072-01	30-GARFIELD-PL-COM	PL097167.D	11 Sep 2025 21:29		AR\AJ	Ok,NS
22	I.BLK	I.BLK	PL097168.D	11 Sep 2025 21:43		AR\AJ	Ok,NS
23	PEM	PEM	PL097169.D	11 Sep 2025 21:56		AR\AJ	Ok
24	PSTDCCC050	PSTDCCC050	PL097170.D	11 Sep 2025 22:10		AR\AJ	Ok,NS
25	Q3073-02	BEL-25-0023	PL097171.D	11 Sep 2025 22:37		AR\AJ	Ok,NS
26	Q3074-01	MOO-25-0141	PL097172.D	11 Sep 2025 23:04		AR\AJ	Ok
27	Q3075-02	MOO-25-0254-0256	PL097173.D	11 Sep 2025 23:18		AR\AJ	Ok,NS
28	Q3075-04	MOO-25-0258	PL097174.D	11 Sep 2025 23:32	TCMX low both column	AR\AJ	ReRun
29	Q3075-06	MOO-25-0259	PL097175.D	11 Sep 2025 23:45		AR\AJ	Ok
30	Q3075-08	MOO-25-0260	PL097176.D	11 Sep 2025 23:59	F Flag in DCB in 2nd column	AR\AJ	Not Ok
31	Q3075-11	MOO-25-02561-0263	PL097177.D	12 Sep 2025 00:12		AR\AJ	Ok
32	Q3077-01	LAW-25-0137	PL097178.D	12 Sep 2025 00:26		AR\AJ	Ok,NS
33	Q3077-03	LAW-25-0139	PL097179.D	12 Sep 2025 00:40	TCMX low in 2nd column	AR\AJ	Ok,NS
34	I.BLK	I.BLK	PL097180.D	12 Sep 2025 00:53		AR\AJ	Ok
35	PSTDCCC050	PSTDCCC050	PL097181.D	12 Sep 2025 02:01		AR\AJ	Ok,NS

M : Manual Integration