

Instrument ID: BNA_G

Daily Analysis Runlog For Sequence/QC Batch ID # BG082825

Review By	Rahul	Review On	8/29/2025 10:48:53 AM
Supervise By	Jagrut	Supervise On	8/29/2025 10:59:39 AM
SubDirectory	BG082825	HP Acquire Method	BNA_G
		HP Processing Method	bg082825

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13173,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BG064217.D	28 Aug 2025 10:11		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BG064218.D	28 Aug 2025 10:51		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BG064219.D	28 Aug 2025 11:32		RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BG064220.D	28 Aug 2025 12:12		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BG064221.D	28 Aug 2025 12:53		RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BG064222.D	28 Aug 2025 13:33	compound #54 kept on LR	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BG064223.D	28 Aug 2025 14:55		RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BG064224.D	28 Aug 2025 15:35		RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BG064225.D	28 Aug 2025 16:16	Compound #9 revoved from 80PPM	RC/JU	Ok,M
10	SSTDICV040	ICVBG082825	BG064226.D	28 Aug 2025 18:19		RC/JU	Ok,M
11	PB169325TB	PB169325TB	BG064227.D	28 Aug 2025 19:00		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_G

Daily Analysis Runlog For Sequence/QC Batch ID # BG090825

Review By	Rahul	Review On	9/9/2025 8:28:19 AM
Supervise By	Jagrut	Supervise On	9/9/2025 12:09:44 PM
SubDirectory	BG090825	HP Acquire Method	BNA_G
		HP Processing Method	BG082825

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13174,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BG064300.D	8 Sep 2025 9:04		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BG064301.D	8 Sep 2025 10:26		RC/JU	Ok
3	PB169566BL	PB169566BL	BG064302.D	8 Sep 2025 11:12		RC/JU	Ok
4	PB169566BS	PB169566BS	BG064303.D	8 Sep 2025 11:52	Recovery fail high for a few compounds.	RC/JU	Not Ok
5	Q3023-01	TP-1	BG064304.D	8 Sep 2025 12:55		RC/JU	Ok
6	Q3023-01MS	TP-1MS	BG064305.D	8 Sep 2025 13:35		RC/JU	Ok,M
7	Q3023-01MSD	TP-1MSD	BG064306.D	8 Sep 2025 14:16		RC/JU	Ok,M
8	Q3023-04	TP-2	BG064307.D	8 Sep 2025 14:57		RC/JU	Ok
9	Q3022-01	1413	BG064308.D	8 Sep 2025 15:38		RC/JU	Ok
10	Q3032-01	SOIL-PILE-1	BG064309.D	8 Sep 2025 16:19		RC/JU	Ok
11	Q3032-01MS	SOIL-PILE-1MS	BG064310.D	8 Sep 2025 17:00		RC/JU	Ok,M
12	Q3032-01MSD	SOIL-PILE-1MSD	BG064311.D	8 Sep 2025 17:41		RC/JU	Ok,M
13	Q3032-04	SOIL-PILE-2	BG064312.D	8 Sep 2025 18:22		RC/JU	Ok,M
14	Q3031-01	PL-01-09052025	BG064313.D	8 Sep 2025 19:03	Need Further 5X Dilution	RC/JU	Dilution
15	Q3023-02	TP-1	BG064314.D	8 Sep 2025 19:44		RC/JU	Ok
16	Q3023-05	TP-2	BG064315.D	8 Sep 2025 20:25		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_G

Daily Analysis Runlog For Sequence/QC Batch ID # BG091025

Review By	Rahul	Review On	9/11/2025 8:50:20 AM
Supervise By	Jagrut	Supervise On	9/11/2025 10:34:49 AM
SubDirectory	BG091025	HP Acquire Method	BNA_G
		HP Processing Method	BG082825

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13174,10ul/1000ul sample
ICV/I.BLK	SP6796
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BG064328.D	10 Sep 2025 10:50		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BG064329.D	10 Sep 2025 12:37		RC/JU	Ok
3	PB169593BL	PB169593BL	BG064330.D	10 Sep 2025 13:18		RC/JU	Ok
4	PB169593BS	PB169593BS	BG064331.D	10 Sep 2025 13:58		RC/JU	Ok,M
5	PB169566BS	PB169566BS	BG064332.D	10 Sep 2025 14:39		RC/JU	Ok,M
6	PB169633BL	PB169633BL	BG064333.D	10 Sep 2025 15:19		RC/JU	Ok
7	PB169633BS	PB169633BS	BG064334.D	10 Sep 2025 16:00		RC/JU	Ok,M
8	Q3051-01	VNJ-215	BG064335.D	10 Sep 2025 16:44		RC/JU	Ok
9	Q3054-01	MH-385	BG064336.D	10 Sep 2025 17:25		RC/JU	Ok
10	Q3048-01	VNJ-258	BG064337.D	10 Sep 2025 18:07	CCC fail for comp#70 from high side	RC/JU	ReRun
11	Q3048-01MS	VNJ-258MS	BG064338.D	10 Sep 2025 18:48		RC/JU	Ok,M
12	Q3048-01MSD	VNJ-258MSD	BG064339.D	10 Sep 2025 19:29		RC/JU	Ok,M
13	Q3046-01	IRV-CONCRETE	BG064340.D	10 Sep 2025 20:10		RC/JU	Ok
14	Q3054-05	MH-384	BG064341.D	10 Sep 2025 20:51		RC/JU	Ok
15	Q3059-01	TR-06-09092025	BG064342.D	10 Sep 2025 21:32		RC/JU	Ok
16	Q3056-01	OK-02-090925	BG064343.D	10 Sep 2025 22:13		RC/JU	Ok
17	Q3055-01	OR-03-090925	BG064344.D	10 Sep 2025 22:53	Out of Tune	RC/JU	Not Ok

M : Manual Integration