

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3032

Client ID : SSTDCCC040

Date Analyzed: 09/10/2025

Lab File ID: BG064329.D

Time Analyzed: 12:37

Instrument ID: BNA_G

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	27684	7.846	118232	10.63	85027	14.47
UPPER LIMIT	55368	8.346	236464	11.131	170054	14.968
LOWER LIMIT	13842	7.346	59116	10.131	42513.5	13.968
EPA SAMPLE NO.						
01 PB169593BL	23634	7.84	95435	10.63	68731	14.47
02 PB169593BS	23229	7.85	98242	10.63	69463	14.47

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.