

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BG064217.D
Instrument ID: BNA_G

Contract: ZUCC01
SDG NO.: Q3032
DFTPP Injection Date: 08/28/2025
DFTPP Injection Time: 10:11

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	34.7
70	Less than 2.0% of mass 69	0.1 (0.4) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7.1
365	Greater than 1% of mass 198	5.2
441	Present, but less than mass 443	15
442	Greater than 50% of mass 198	90.8
443	15.0 - 24.0% of mass 442	17 (18.7) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BG064218.D	08/28/2025	10:51
SSTDICC005	SSTDICC005	BG064219.D	08/28/2025	11:32
SSTDICC010	SSTDICC010	BG064220.D	08/28/2025	12:12
SSTDICC020	SSTDICC020	BG064221.D	08/28/2025	12:53
SSTDICCC040	SSTDICCC040	BG064222.D	08/28/2025	13:33
SSTDICC050	SSTDICC050	BG064223.D	08/28/2025	14:55
SSTDICC060	SSTDICC060	BG064224.D	08/28/2025	15:35
SSTDICC080	SSTDICC080	BG064225.D	08/28/2025	16:16