

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: BF091225

SequenceID : BF091225

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)	_____	_____	_____✓_____
2. GC/MS Tuning Specifications. DFTPP Meet Criteria Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)	_____	_____	_____✓_____
3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series	_____	_____	_____✓_____
4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series	_____	_____	_____✓_____
5. GC/MS Calibration Met:	_____	_____	_____✓_____
a. Initial calibration Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓_____
b. Continuous Calibration(CCC) Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____✓_____	_____
6. Blank Contamination - If yes, list compounds and concentrations in each blank:	_____	_____	_____✓_____
a. B/N Fraction			
d. Acid Fraction			

7. Surrogate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

a. B/N Fraction

d. Acid Fraction

8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable range.

a. B/N Fraction

d. Acid Fraction

9. Internal Standard Area/Retention Time Shift Meet Criteria

Comments:

10. Extraction Holding Time Met

If not met, list number of days exceeded for each sample:

11. Analysis Holding Time Met

If not met, list number of days exceeded for each sample:

ADDITIONAL COMMENTS:

Internal standard failed in the samples Q3077-03 & Q3075-11 due to matrix interference as evidenced by the chromatograms. The sample Q3077-03 was analyzed straight as well as with 10X dilution. The sample Q3073-02 was analyzed with 20X dilution due to oily viscous matrix. Internal standards failed even in 5X dilution sample Q3080-01. This was analyzed with 5X dilution as the matrix was viscous. The sample Q3082-09 with concentrated matrix was analyzed with 5X dilution.

dhruv

Analyst

09/12/2025

Date

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: BF091225

SequenceID : BF091225 NA NO YES