

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF091025

Review By	Rahul	Review On	9/10/2025 9:03:37 AM
Supervise By	Jagrut	Supervise On	9/10/2025 11:02:08 AM
SubDirectory	BF091025	HP Acquire Method	BNA_F
		HP Processing Method	bf091025

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13174,10ul/1000ul sample
ICV/I.BLK	SP6796
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF143675.D	09 Sep 2025 08:44		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF143676.D	09 Sep 2025 09:12		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF143677.D	09 Sep 2025 09:41		RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF143678.D	09 Sep 2025 10:10		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF143679.D	09 Sep 2025 10:39		RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BF143680.D	09 Sep 2025 11:08		RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BF143681.D	09 Sep 2025 11:37		RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BF143682.D	09 Sep 2025 12:07		RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BF143683.D	09 Sep 2025 13:54		RC/JU	Ok,M
10	SSTDICV040	ICVBF091025	BF143684.D	09 Sep 2025 15:03		RC/JU	Ok,M
11	PB169490BL	PB169490BL	BF143685.D	09 Sep 2025 15:32		RC/JU	Ok
12	Q2893-02	LOQ-SOIL-02-QT3-202	BF143686.D	09 Sep 2025 16:01	LOQ-SOIL-10PPM (For Benzidine only)	RC/JU	Ok
13	Q2893-08	LOQ-WATER-02-QT3-2	BF143687.D	09 Sep 2025 16:30	LOQ-WATER-10PPM (For Benzidine only)	RC/JU	Ok
14	Q2893-01	LOD-MDL-SOIL-01-QT	BF143688.D	09 Sep 2025 16:59	LOD-MDL-SOIL-8PPM (For Benzidine only)	RC/JU	Ok
15	Q2893-07	LOD-MDL-WATER-01-Q	BF143689.D	09 Sep 2025 17:28	LOD-MDL-WATER-8 PPM (For Benzidine only)	RC/JU	Ok,M
16	PB169499BL	PB169499BL	BF143690.D	09 Sep 2025 17:58		RC/JU	Ok
17	SSTDCCC040	SSTDCCC040EC	BF143691.D	09 Sep 2025 18:27		RC/JU	Ok,M

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF091025

Review By	Rahul	Review On	9/10/2025 9:03:37 AM		
Supervise By	Jagrut	Supervise On	9/10/2025 11:02:08 AM		
SubDirectory	BF091025	HP Acquire Method	BNA_F	HP Processing Method	bf091025
STD. NAME		STD REF.#			
Tune/Reschk		SP6757			
Initial Calibration Stds		SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864			
CCC		SP6860			
Internal Standard/PEM		S13174,10ul/1000ul sample			
ICV/I.BLK		SP6796			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	DFTPP	DFTPP	BF143692.D	10 Sep 2025 09:13		RC/JU	Ok
19	SSTDCCC040	SSTDCCC040	BF143693.D	10 Sep 2025 09:41		RC/JU	Ok,M
20	PB169528BL	PB169528BL	BF143694.D	10 Sep 2025 10:10		RC/JU	Ok
21	Q2893-03	MDL-SOIL-03-QT3-202	BF143695.D	10 Sep 2025 10:49	MDL-SOIL 8PPM (For Benzidine only)	RC/JU	Ok,M
22	Q2893-09	MDL-WATER-03-QT3-2	BF143696.D	10 Sep 2025 11:18	MDL-WATER 8PPM (For Benzidine only)	RC/JU	Ok,M
23	PB169526BL	PB169526BL	BF143697.D	10 Sep 2025 11:49		RC/JU	Ok
24	SP6869	SP6869	BF143698.D	10 Sep 2025 12:41		RC/JU	Ok
25	SSTDCCC040	SSTDCCC040EC	BF143699.D	10 Sep 2025 13:47		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF091125

Review By	Rahul	Review On	9/12/2025 10:17:16 AM
Supervise By	Jagrut	Supervise On	9/12/2025 11:51:04 AM
SubDirectory	BF091125	HP Acquire Method	BNA_F
		HP Processing Method	bf091025

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13174,10ul/1000ul sample
ICV/I.BLK	SP6796
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF143700.D	11 Sep 2025 12:02		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF143701.D	11 Sep 2025 12:31		RC/JU	Ok
3	PB169649BL	PB169649BL	BF143702.D	11 Sep 2025 13:00		RC/JU	Ok
4	PB169649BS	PB169649BS	BF143703.D	11 Sep 2025 13:29		RC/JU	Ok,M
5	Q3048-01	VNJ-258	BF143704.D	11 Sep 2025 14:03		RC/JU	Ok,M
6	Q3048-01MS	VNJ-258MS	BF143705.D	11 Sep 2025 14:33		RC/JU	Ok,M
7	Q3048-01MSD	VNJ-258MSD	BF143706.D	11 Sep 2025 15:02		RC/JU	Ok,M
8	Q3055-01	OR-03-090925	BF143707.D	11 Sep 2025 15:32		RC/JU	Ok,M
9	Q3065-01	VNJ-204	BF143708.D	11 Sep 2025 16:01		RC/JU	Ok,M
10	Q3066-01	354	BF143709.D	11 Sep 2025 16:31		RC/JU	Ok,M
11	Q3066-03	VNJ-242	BF143710.D	11 Sep 2025 17:00		RC/JU	Ok,M
12	Q3032-02	SOIL-PILE-1	BF143711.D	11 Sep 2025 17:30		RC/JU	Ok
13	Q3032-02MS	SOIL-PILE-1MS	BF143712.D	11 Sep 2025 17:59		RC/JU	Ok
14	Q3032-02MSD	SOIL-PILE-1MSD	BF143713.D	11 Sep 2025 18:29		RC/JU	Ok
15	Q3032-05	SOIL-PILE-2	BF143714.D	11 Sep 2025 18:59		RC/JU	Ok
16	Q3075-04	MOO-25-0258	BF143715.D	11 Sep 2025 19:28		RC/JU	Ok
17	Q3074-01	MOO-25-0141	BF143716.D	11 Sep 2025 19:57		RC/JU	Ok
18	Q3075-06	MOO-25-0259	BF143717.D	11 Sep 2025 20:27		RC/JU	Ok,M

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF091125

Review By	Rahul	Review On	9/12/2025 10:17:16 AM		
Supervise By	Jagrut	Supervise On	9/12/2025 11:51:04 AM		
SubDirectory	BF091125	HP Acquire Method	BNA_F	HP Processing Method	bf091025
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864				
CCC	SP6860				
Internal Standard/PEM	S13174,10ul/1000ul sample				
ICV/I.BLK	SP6796				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q3077-03	LAW-25-0139	BF143718.D	11 Sep 2025 20:56	Surrogate , Internal fail, Need 20X, then decide further dilution.	RC/JU	Dilution
20	Q3075-02	MOO-25-0254-0256	BF143719.D	11 Sep 2025 21:25		RC/JU	Ok,M
21	Q3075-02MS	MOO-25-0254-0256MS	BF143720.D	11 Sep 2025 21:54		RC/JU	Ok
22	Q3075-02MSD	MOO-25-0254-0256MS	BF143721.D	11 Sep 2025 22:24		RC/JU	Ok
23	Q3072-01	30-GARFIELD-PL-COM	BF143722.D	11 Sep 2025 22:53		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF091225

Review By	Rahul	Review On	9/12/2025 10:25:15 AM
Supervise By	Jagrut	Supervise On	9/15/2025 10:16:27 AM
SubDirectory	BF091225	HP Acquire Method	BNA_F
		HP Processing Method	bf091025

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13174,10ul/1000ul sample
ICV/I.BLK	SP6796
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF143723.D	12 Sep 2025 10:14		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF143724.D	12 Sep 2025 10:43		RC/JU	Ok
3	PB169661BL	PB169661BL	BF143725.D	12 Sep 2025 11:12		RC/JU	Ok
4	PB169661BS	PB169661BS	BF143726.D	12 Sep 2025 11:42		RC/JU	Ok,M
5	PB169589TB	PB169589TB	BF143727.D	12 Sep 2025 12:11		RC/JU	Ok
6	Q3077-03DL	LAW-25-0139DL	BF143728.D	12 Sep 2025 12:46	Internal standard fail	RC/JU	Ok,M
7	Q3075-11	MOO-25-0261-0263	BF143729.D	12 Sep 2025 13:15	Internal standard fail	RC/JU	Ok,M
8	Q3082-01	WC-29	BF143730.D	12 Sep 2025 13:44		RC/JU	Ok,M
9	Q3075-08	MOO-25-0260	BF143731.D	12 Sep 2025 14:13	Need 2X Dilution	RC/JU	Dilution
10	Q3082-03	WC-30	BF143732.D	12 Sep 2025 14:42		RC/JU	Ok,M
11	Q3082-05	WC-31	BF143733.D	12 Sep 2025 15:12		RC/JU	Ok,M
12	Q3082-05MS	WC-31MS	BF143734.D	12 Sep 2025 15:41		RC/JU	Ok,M
13	Q3082-05MSD	WC-31MSD	BF143735.D	12 Sep 2025 16:11		RC/JU	Ok,M
14	Q3054-04	MH-385	BF143736.D	12 Sep 2025 16:40		RC/JU	Ok
15	Q3054-08	MH-384	BF143737.D	12 Sep 2025 17:09	Internal standard fail	RC/JU	ReRun
16	Q3072-02	30-GARFIELD-PL-COM	BF143738.D	12 Sep 2025 17:39		RC/JU	Ok
17	Q3072-02MS	30-GARFIELD-PL-COM	BF143739.D	12 Sep 2025 18:08		RC/JU	Ok,M
18	Q3072-02MSD	30-GARFIELD-PL-COM	BF143740.D	12 Sep 2025 18:38		RC/JU	Ok,M

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF091225

Review By	Rahul	Review On	9/12/2025 10:25:15 AM		
Supervise By	Jagrut	Supervise On	9/15/2025 10:16:27 AM		
SubDirectory	BF091225	HP Acquire Method	BNA_F	HP Processing Method	bf091025
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864				
CCC	SP6860				
Internal Standard/PEM	S13174,10ul/1000ul sample				
ICV/I.BLK	SP6796				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q3082-07	WC-32	BF143741.D	12 Sep 2025 19:07		RC/JU	Ok,M
20	Q3077-01	LAW-25-0137	BF143742.D	12 Sep 2025 19:36		RC/JU	Ok
21	Q3075-08DL	MOO-25-0260DL	BF143743.D	12 Sep 2025 20:06		RC/JU	Ok,M
22	Q3082-09	WC-33	BF143744.D	12 Sep 2025 20:35	Need Further 10X Dilution	RC/JU	Dilution
23	Q3080-01	EO-03-091125	BF143745.D	12 Sep 2025 21:04	Internal Standard Fail	RC/JU	Ok,M
24	Q3073-02	BEL-25-0023	BF143746.D	12 Sep 2025 21:34		RC/JU	Ok

M : Manual Integration