

SOP ID : M1311-TCLP-16
SDG No : N/A
Weigh By : JP
Balance ID : WC SC-7
pH Meter ID : WC PH METER-1
Extraction By : JP
Filter By : JP
Pipette ID : WC
Tumbler ID : ZHE-1
TCLP Filter ID : 60828725

Start Prep Date : 09/05/2025 **Time :** 16:00
End Prep Date : 09/06/2025 **Time :** 09:25
Combination Ratio : 20
ZHE Cleaning Batch : N/A
Initial Room Temperature: 23 °C
Final Room Temperature: 21 °C
TCLP Technician Signature : 
Supervisor By : 12

Standard Name	MLS USED	STD REF. # FROM LOG
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE U	Lot Number
TCLP-FLUID-1	N/A	WP114525
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
40ml VOA Vials	4127	N/A

Extraction Conformance/Non-Conformance Comments:

Matrix spikes are added after filtration and before preservation. TUMBLER ZHE-1 checked, 30 rpm. Leak checked after 10 minutes of tumbling.

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
09/08/25 08:45	JP TCLP Room	S-Y IVOC Lab
	Preparation Group	Analysis Group

Sample ID	ClientID	ZHE Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
PB169581TB	LEB581	05	N/A	500	N/A	N/A	N/A	4.93	N/A	ZHE-1
Q3023-02	TP-1	01	25.02	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
Q3023-05	TP-2	02	25.03	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
Q3032-02	SOIL-PILE-1	03	25.02	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
Q3032-05	SOIL-PILE-2	04	25.03	500	N/A	N/A	N/A	N/A	N/A	ZHE-1

SampleID	ClientID	Sample Weight (g)	Filter Weight (g)	Filtrate (mL)	Filter + Solid (After 100°C)	% solids	% Dry Solids
PB169581TB	LEB581	N/A	N/A	N/A	N/A	N/A	N/A
Q3023-02	TP-1	N/A	N/A	N/A	N/A	100	N/A
Q3023-05	TP-2	N/A	N/A	N/A	N/A	100	N/A
Q3032-02	SOIL-PILE-1	N/A	N/A	N/A	N/A	100	N/A
Q3032-05	SOIL-PILE-2	N/A	N/A	N/A	N/A	100	N/A