

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN082125

Review By	John Carlone	Review On	8/22/2025 9:17:52 AM
Supervise By	Maresh Dadoda	Supervise On	8/25/2025 8:11:38 AM
SubDirectory	VN082125	HP Acquire Method	HP Processing Method 82N082125W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135214 VP135215,VP135216,VP135217,VP135218,VP135219,VP135220 VP135221

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087614.D	21 Aug 2025 09:30		JC\MD	Ok
2	VSTDCCC020	VSTDCCC020	VN087615.D	21 Aug 2025 10:04	Need ICAL	JC\MD	Not Ok
3	VN0821WBS01	VN0821WBS01	VN087616.D	21 Aug 2025 10:37		JC\MD	Not Ok
4	VN0821WBSD01	VN0821WBSD01	VN087617.D	21 Aug 2025 11:11		JC\MD	Not Ok
5	VN0821WBL01	VN0821WBL01	VN087618.D	21 Aug 2025 11:36		JC\MD	Not Ok
6	VSTDICC001	VSTDICC001	VN087619.D	21 Aug 2025 12:09	LR-10,20	JC\MD	Ok,M
7	VSTDICC005	VSTDICC005	VN087620.D	21 Aug 2025 13:08		JC\MD	Ok,M
8	VSTDICC020	VSTDICC020	VN087621.D	21 Aug 2025 13:41	method not used for #N-amyle Acetate	JC\MD	Ok,M
9	VSTDICCC050	VSTDICCC050	VN087622.D	21 Aug 2025 14:02		JC\MD	Ok,M
10	VSTDICC100	VSTDICC100	VN087623.D	21 Aug 2025 14:23		JC\MD	Ok,M
11	VSTDICC150	VSTDICC150	VN087624.D	21 Aug 2025 14:44		JC\MD	Ok,M
12	IBLK	IBLK	VN087625.D	21 Aug 2025 15:05		JC\MD	Ok
13	VSTDICV050	ICVVN082125	VN087626.D	21 Aug 2025 15:47	Icv Failed for Com.# 64 for DOD	JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN090825

Review By	Semsettin Yesilyurt	Review On	9/9/2025 3:14:54 PM
Supervise By		Supervise On	
SubDirectory	VN090825	HP Acquire Method	MSVOA_N HP Processing Method VN082125W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135400		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135401,VP135402 VP134956		

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087748.D	08 Sep 2025 07:42		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN087749.D	08 Sep 2025 08:58		JC\MD	Ok,M
3	VN0908MBL01	VN0908MBL01	VN087750.D	08 Sep 2025 09:39		JC\MD	Ok
4	VN0908WBL01	VN0908WBL01	VN087751.D	08 Sep 2025 09:59		JC\MD	Ok
5	VN0908WBS01	VN0908WBS01	VN087752.D	08 Sep 2025 10:59		JC\MD	Ok,M
6	VN0908WBSD01	VN0908WBSD01	VN087753.D	08 Sep 2025 11:20		JC\MD	Ok,M
7	Q3036-03	EGR-2	VN087754.D	08 Sep 2025 12:30		JC\MD	Ok
8	VIBLK	VIBLK	VN087755.D	08 Sep 2025 12:51		JC\MD	Ok
9	PB169581TB	PB169581TB	VN087756.D	08 Sep 2025 13:19	pH#5.0 A	JC\MD	Ok
10	Q3023-02	TP-1	VN087757.D	08 Sep 2025 13:40	pH#5.0 A	JC\MD	Ok
11	Q3023-05	TP-2	VN087758.D	08 Sep 2025 14:01	pH#5.0 A	JC\MD	Ok
12	Q3032-02	SOIL-PILE-1	VN087759.D	08 Sep 2025 14:21	pH#5.0 A	JC\MD	Ok,M
13	Q3032-05	SOIL-PILE-2	VN087760.D	08 Sep 2025 14:42	pH#5.0 A	JC\MD	Ok,M
14	VIBLK	VIBLK	VN087761.D	08 Sep 2025 15:46		JC\MD	Ok
15	VSTDCCC050	VSTDCCC050EC	VN087762.D	08 Sep 2025 16:07		JC\MD	Ok,M

M : Manual Integration