

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following "Results Qualifiers" are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. "10 U". This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as "12 B".
E	Indicates the analyte 's concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a "P".
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

LAB CHRONICLE

OrderID:	Q3032	OrderDate:	9/5/2025 12:41:17 PM
Client:	Zuccaro Inc.	Project:	3 Coal Ave., Morristown, NJ
Contact:	Sal Zuccaro	Location:	D31,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3032-02	SOIL-PILE-1	TCLP	TCLP VOA	8260D	09/05/25		09/08/25	09/05/25
Q3032-05	SOIL-PILE-2	TCLP	TCLP VOA	8260D	09/05/25		09/08/25	09/05/25

Hit Summary Sheet
SW-846

SDG No.: Q3032
Client: Zuccaro Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID: Q3032-02	SOIL-PILE-1 SOIL-PILE-1	TCLP	2-Butanone	11.2	J	0.98	25.0	ug/L
			Total Voc :	11.2				
			Total Concentration:	11.2				
Client ID: Q3032-05	SOIL-PILE-2 SOIL-PILE-2	TCLP	2-Butanone	13.1	J	0.98	25.0	ug/L
			Total Voc :	13.1				
			Total Concentration:	13.1				



QC SUMMARY

Surrogate Summary

SDG No.: **Q3032**

Client: **Zuccaro Inc.**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3032-02	SOIL-PILE-1	1,2-Dichloroethane-d4	50	51.9	104		70 (74)	130 (125)
		Dibromofluoromethane	50	50.3	101		70 (75)	130 (124)
		Toluene-d8	50	50.1	100		70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.0	96		70 (77)	130 (121)
Q3032-05	SOIL-PILE-2	1,2-Dichloroethane-d4	50	55.5	111		70 (74)	130 (125)
		Dibromofluoromethane	50	51.8	103		70 (75)	130 (124)
		Toluene-d8	50	49.5	99		70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.3	97		70 (77)	130 (121)
VN0908WBL01	VN0908WBL01	1,2-Dichloroethane-d4	50	52.3	105		70 (74)	130 (125)
		Dibromofluoromethane	50	48.8	98		70 (75)	130 (124)
		Toluene-d8	50	48.4	97		70 (86)	130 (113)
		4-Bromofluorobenzene	50	46.0	92		70 (77)	130 (121)
VN0908WBS01	VN0908WBS01	1,2-Dichloroethane-d4	50	41.3	83		70 (74)	130 (125)
		Dibromofluoromethane	50	45.2	90		70 (75)	130 (124)
		Toluene-d8	50	45.1	90		70 (86)	130 (113)
		4-Bromofluorobenzene	50	44.1	88		70 (77)	130 (121)
VN0908WBSD01	VN0908WBSD01	1,2-Dichloroethane-d4	50	40.4	81		70 (74)	130 (125)
		Dibromofluoromethane	50	45.0	90		70 (75)	130 (124)
		Toluene-d8	50	44.3	89		70 (86)	130 (113)
		4-Bromofluorobenzene	50	44.5	89		70 (77)	130 (121)

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3032</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Zuccaro Inc.</u>	Datafile :	<u>VN087752.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0908WBS01	Vinyl chloride	20	19.6	ug/L	98			70 (65)	130 (117)	
	1,1-Dichloroethene	20	18.7	ug/L	94			70 (74)	130 (110)	
	2-Butanone	100	84.1	ug/L	84			40 (65)	160 (122)	
	Carbon Tetrachloride	20	20.1	ug/L	101			70 (77)	130 (113)	
	Chloroform	20	18.0	ug/L	90			70 (79)	130 (113)	
	Benzene	20	18.0	ug/L	90			70 (82)	130 (109)	
	1,2-Dichloroethane	20	17.7	ug/L	89			70 (80)	130 (115)	
	Trichloroethene	20	18.4	ug/L	92			70 (77)	130 (113)	
	Tetrachloroethene	20	18.4	ug/L	92			70 (67)	130 (123)	
	Chlorobenzene	20	18.5	ug/L	93			70 (82)	130 (109)	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3032</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Zuccaro Inc.</u>	Datafile :	<u>VN087753.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0908WBSD01	Vinyl chloride	20	19.1	ug/L	96	2		70 (65)	130 (117)	20 (19)
	1,1-Dichloroethene	20	19.0	ug/L	95	1		70 (74)	130 (110)	20 (20)
	2-Butanone	100	83.3	ug/L	83	1		40 (65)	160 (122)	20 (26)
	Carbon Tetrachloride	20	21.0	ug/L	105	4		70 (77)	130 (113)	20 (15)
	Chloroform	20	17.8	ug/L	89	1		70 (79)	130 (113)	20 (20)
	Benzene	20	18.6	ug/L	93	3		70 (82)	130 (109)	20 (15)
	1,2-Dichloroethane	20	18.5	ug/L	93	4		70 (80)	130 (115)	20 (20)
	Trichloroethene	20	18.5	ug/L	93	1		70 (77)	130 (113)	20 (15)
	Tetrachloroethene	20	19.1	ug/L	96	4		70 (67)	130 (123)	20 (15)
	Chlorobenzene	20	18.8	ug/L	94	1		70 (82)	130 (109)	20 (15)

VOLATILE METHOD BLANK SUMMARY

Client ID

VN0908WBL01

Lab Name: Alliance

Contract: ZUCC01

Lab Code: ACE

SDG NO.: Q3032

Lab File ID: VN087751.D

Lab Sample ID: VN0908WBL01

Date Analyzed: 09/08/2025

Time Analyzed: 09:59

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_N

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VN0908WBS01	VN0908WBS01	VN087752.D	09/08/2025
VN0908WBSD01	VN0908WBSD01	VN087753.D	09/08/2025
SOIL-PILE-1	Q3032-02	VN087759.D	09/08/2025
SOIL-PILE-2	Q3032-05	VN087760.D	09/08/2025

COMMENTS: _____



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: ZUCC01

Lab Code: ACE

SDG NO.: Q3032

Lab File ID: VN087614.D

BFB Injection Date: 08/21/2025

Instrument ID: MSVOA_N

BFB Injection Time: 09:30

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	25.8
75	30.0 - 60.0% of mass 95	58.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.3
173	Less than 2.0% of mass 174	1.1 (1.6) 1
174	50.0 - 100.0% of mass 95	69
175	5.0 - 9.0% of mass 174	5.6 (8.1) 1
176	95.0 - 101.0% of mass 174	67.7 (98.1) 1
177	5.0 - 9.0% of mass 176	4.8 (7.1) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VN087619.D	08/21/2025	12:09
VSTDIC005	VSTDIC005	VN087620.D	08/21/2025	13:08
VSTDIC020	VSTDIC020	VN087621.D	08/21/2025	13:41
VSTDIC050	VSTDIC050	VN087622.D	08/21/2025	14:02
VSTDIC100	VSTDIC100	VN087623.D	08/21/2025	14:23
VSTDIC150	VSTDIC150	VN087624.D	08/21/2025	14:44



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: ZUCC01

Lab Code: ACE

SDG NO.: Q3032

Lab File ID: VN087748.D

BFB Injection Date: 09/08/2025

Instrument ID: MSVOA_N

BFB Injection Time: 07:42

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	23.6
75	30.0 - 60.0% of mass 95	59.4
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.4
173	Less than 2.0% of mass 174	0.9 (1.2) 1
174	50.0 - 100.0% of mass 95	71.2
175	5.0 - 9.0% of mass 174	5.1 (7.1) 1
176	95.0 - 101.0% of mass 174	67.6 (94.9) 1
177	5.0 - 9.0% of mass 176	5.3 (7.8) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VN087749.D	09/08/2025	08:58
VN0908WBL01	VN0908WBL01	VN087751.D	09/08/2025	09:59
VN0908WBS01	VN0908WBS01	VN087752.D	09/08/2025	10:59
VN0908WBSD01	VN0908WBSD01	VN087753.D	09/08/2025	11:20
SOIL-PILE-1	Q3032-02	VN087759.D	09/08/2025	14:21
SOIL-PILE-2	Q3032-05	VN087760.D	09/08/2025	14:42

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ZUCC01
 Lab Code: ACE SDG NO.: Q3032
 Lab File ID: VN087749.D Date Analyzed: 09/08/2025
 Instrument ID: MSVOA_N Time Analyzed: 08:58
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	264779	8.21	492400	9.08	447402	11.84
UPPER LIMIT	529558	8.706	984800	9.577	894804	12.341
LOWER LIMIT	132390	7.706	246200	8.577	223701	11.341
EPA SAMPLE NO.						
SOIL-PILE-1	173684	8.21	396253	9.08	398856	11.84
SOIL-PILE-2	159239	8.21	389686	9.08	380890	11.84
VN0908WBL01	246345	8.21	567426	9.08	531276	11.84
VN0908WBS01	258577	8.20	485156	9.08	441331	11.84
VN0908WBSD01	249385	8.21	457673	9.08	416132	11.84

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>ZUCC01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3032</u>
Lab File ID:	<u>VN087749.D</u>	Date Analyzed:	<u>09/08/2025</u>
Instrument ID:	<u>MSVOA_N</u>	Time Analyzed:	<u>08:58</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	232401	13.77				
UPPER LIMIT	464802	14.27				
LOWER LIMIT	116201	13.27				
EPA SAMPLE NO.						
SOIL-PILE-1	185969	13.77				
SOIL-PILE-2	180770	13.77				
VN0908WBL01	244144	13.77				
VN0908WBS01	222091	13.77				
VN0908WBSD01	211983	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.



SAMPLE DATA

Report of Analysis

Client:	Zuccaro Inc.	Date Collected:	09/05/25
Project:	3 Coal Ave., Morristown, NJ	Date Received:	09/05/25
Client Sample ID:	SOIL-PILE-1	SDG No.:	Q3032
Lab Sample ID:	Q3032-02	Matrix:	TCLP
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	TCLP VOA
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :	SW5035		

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087759.D	1	09/08/25 14:21	VN090825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	0.26	U	0.26	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	5.00	ug/L
78-93-3	2-Butanone	11.2	J	0.98	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	5.00	ug/L
71-43-2	Benzene	0.15	U	0.15	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	5.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	5.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	5.00	ug/L
108-90-7	Chlorobenzene	0.12	U	0.12	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.9		70 (74) - 130 (125)	104%	SPK: 50
1868-53-7	Dibromofluoromethane	50.3		70 (75) - 130 (124)	101%	SPK: 50
2037-26-5	Toluene-d8	50.1		70 (86) - 130 (113)	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.0		70 (77) - 130 (121)	96%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	174000	8.206			
540-36-3	1,4-Difluorobenzene	396000	9.082			
3114-55-4	Chlorobenzene-d5	399000	11.841			
3855-82-1	1,4-Dichlorobenzene-d4	186000	13.77			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090825\
 Data File : VN087759.D
 Acq On : 08 Sep 2025 14:21
 Operator : JC\MD
 Sample : Q3032-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SOIL-PILE-1

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 09/09/2025
 Supervised By : Mahesh Dadoda 09/09/2025

Quant Time: Sep 09 02:29:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:55:42 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	173684	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	396253	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	398856	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	185969	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	184702	51.903	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.800%			
35) Dibromofluoromethane	8.147	113	143909	50.301	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.600%			
50) Toluene-d8	10.547	98	536840	50.135	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.260%			
62) 4-Bromofluorobenzene	12.823	95	182937	48.039	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 96.080%			
Target Compounds						Qvalue
11) Tert butyl alcohol	5.541	59	4186	6.776	ug/l #	84
16) Acetone	4.430	43	41838	21.597	ug/l	90
18) Methyl Acetate	5.024	43	31478	7.761	ug/l	100
20) Methylene Chloride	5.247	84	10376m	2.524	ug/l	
25) 2-Butanone	7.471	43	28484	11.176	ug/l	94
37) Ethyl Acetate	7.553	43	20641	3.442	ug/l #	94
43) Isopropyl Acetate	8.677	43	13801	1.474	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090825\
Data File : VN087759.D
Acq On : 08 Sep 2025 14:21
Operator : JC\MD
Sample : Q3032-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
SOIL-PILE-1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 09/09/2025
Supervised By :Mahesh Dadoda 09/09/2025


Quant Time: Sep 09 02:29:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:55:42 2025
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.206 min Scan# 1063
Delta R.T. 0.000 min
Lab File: VN087759.D
Acq: 08 Sep 2025 14:21




Instrument :
MSVOA_N
ClientSampleId :
SOIL-PILE-1



Tgt Ion:168 Resp: 173684
Ion Ratio Lower Upper
168 100
99 66.8 57.2 85.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 09/09/2025
Supervised By :Mahesh Dadoda 09/09/2025



#11
Tert butyl alcohol
Concen: 6.776 ug/l
RT: 5.541 min Scan# 610
Delta R.T. 0.023 min
Lab File: VN087759.D
Acq: 08 Sep 2025 14:21


Tgt Ion: 59 Resp: 4186
Ion Ratio Lower Upper
59 100
57 5.0 9.0 13.4#





#16
Acetone
Concen: 21.597 ug/l
RT: 4.430 min Scan# 4183
Delta R.T. 0.006 min
Lab File: VN087759.D
Acq: 08 Sep 2025 14:21




Instrument :
MSVOA_N
ClientSampleId :
SOIL-PILE-1



Tgt Ion: 43 Resp: 4183
Ion Ratio Lower Upper
43 100
58 33.4 22.6 33.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 09/09/2025
Supervised By :Mahesh Dadoda 09/09/2025



#18
Methyl Acetate
Concen: 7.761 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.012 min
Lab File: VN087759.D
Acq: 08 Sep 2025 14:21

Tgt Ion: 43 Resp: 31478
Ion Ratio Lower Upper
43 100
74 22.2 17.6 26.4





#20

Methylene Chloride

Concen: 2.524 ug/l m

RT: 5.247 min Scan# 51

Delta R.T. -0.018 min

Lab File: VN087759.D


Acq: 08 Sep 2025 14:21

Instrument :

MSVOA_N

ClientSampleId :

SOIL-PILE-1



Tgt Ion: 84 Resp: 10370

Ion Ratio Lower Upper

84 100

49 124.7 103.4 155.2

51 24.3 33.8 50.8




86 40.4 49.6 74.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 09/09/2025

Supervised By :Mahesh Dadoda 09/09/2025



#25

2-Butanone

Concen: 11.176 ug/l

RT: 7.471 min Scan# 938

Delta R.T. 0.006 min

Lab File: VN087759.D




Acq: 08 Sep 2025 14:21


Tgt Ion: 43 Resp: 28484

Ion Ratio Lower Upper

43 100

72 21.8 19.9 29.9









#33
1,2-Dichloroethane-d4
Concen: 51.903 ug/l
RT: 8.559 min Scan# 1123
Delta R.T. -0.000 min
Lab File: VN087759.D
Acq: 08 Sep 2025 14:21

Instrument :
MSVOA_N
ClientSampleId :
SOIL-PILE-1

Tgt Ion: 65 Resp: 184702
Ion Ratio Lower Upper
65 100
67 53.0 0.0 100.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 09/09/2025
Supervised By :Mahesh Dadoda 09/09/2025



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.082 min Scan# 1212
Delta R.T. -0.000 min
Lab File: VN087759.D
Acq: 08 Sep 2025 14:21

Tgt Ion:114 Resp: 396253
Ion Ratio Lower Upper
114 100
63 21.9 0.0 45.2
88 15.4 0.0 33.4





#35

Dibromofluoromethane

Concen: 50.301 ug/l

RT: 8.147 min Scan# 1053

Delta R.T. -0.006 min

Lab File: VN087759.D


Acq: 08 Sep 2025 14:21

Instrument :

MSVOA_N

ClientSampleId :

SOIL-PILE-1



Tgt Ion:113 Resp: 143909

Ion Ratio Lower Upper

113 100

111 104.4 82.8 124.2




192 16.1 12.8 19.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 09/09/2025

Supervised By :Mahesh Dadoda 09/09/2025



#37

Ethyl Acetate

Concen: 3.442 ug/l

RT: 7.553 min Scan# 952

Delta R.T. 0.006 min

Lab File: VN087759.D

Acq: 08 Sep 2025 14:21




Tgt Ion: 43 Resp: 20641


Ion Ratio Lower Upper

43 100

61 14.2 9.3 13.9#


70 11.0 7.0 10.6#





#43
Isopropyl Acetate
Concen: 1.474 ug/l
RT: 8.677 min Scan# 1143
Delta R.T. 0.006 min
Lab File: VN087759.D
Acq: 08 Sep 2025 14:21




Instrument :
MSVOA_N
ClientSampleId :
SOIL-PILE-1



Tgt Ion: 43 Resp: 1380
Ion Ratio Lower Upper
43 100
61 18.5 18.8 28.2
87 11.1 9.6 14.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 09/09/2025
Supervised By :Mahesh Dadoda 09/09/2025



#50
Toluene-d8
Concen: 50.135 ug/l
RT: 10.547 min Scan# 1461
Delta R.T. -0.000 min
Lab File: VN087759.D
Acq: 08 Sep 2025 14:21

Tgt Ion: 98 Resp: 536840
Ion Ratio Lower Upper
98 100
100 63.9 52.8 79.2









#62
4-Bromofluorobenzene
Concen: 48.039 ug/l
RT: 12.823 min Scan# 1848
Delta R.T. -0.000 min
Lab File: VN087759.D
Acq: 08 Sep 2025 14:21

Instrument :
MSVOA_N
ClientSampleId :
SOIL-PILE-1

Tgt Ion: 95 Resp: 18293
Ion Ratio Lower Upper
95 100
174 71.2 0.0 138.4
176 69.2 0.0 132.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 09/09/2025
Supervised By :Mahesh Dadoda 09/09/2025



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.841 min Scan# 1681
Delta R.T. -0.000 min
Lab File: VN087759.D
Acq: 08 Sep 2025 14:21


Tgt Ion:117 Resp: 398856
Ion Ratio Lower Upper
117 100
82 57.7 47.4 71.2
119 32.6 24.3 36.5





#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.770 min Scan# 20
Delta R.T. 0.006 min
Lab File: VN087759.D
Acq: 08 Sep 2025 14:21



Instrument :
MSVOA_N
ClientSampleId :
SOIL-PILE-1



Tgt Ion	Ratio	Lower	Upper
152	100		
115	63.3	32.3	96.8
150	156.4	0.0	351.0

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 09/09/2025
Supervised By :Mahesh Dadoda 09/09/2025



Report of Analysis

Client:	Zuccaro Inc.	Date Collected:	09/05/25
Project:	3 Coal Ave., Morristown, NJ	Date Received:	09/05/25
Client Sample ID:	SOIL-PILE-2	SDG No.:	Q3032
Lab Sample ID:	Q3032-05	Matrix:	TCLP
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	TCLP VOA
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :	SW5035		

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087760.D	1	09/08/25 14:42	VN090825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	0.26	U	0.26	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	5.00	ug/L
78-93-3	2-Butanone	13.1	J	0.98	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	5.00	ug/L
71-43-2	Benzene	0.15	U	0.15	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	5.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	5.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	5.00	ug/L
108-90-7	Chlorobenzene	0.12	U	0.12	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	55.5		70 (74) - 130 (125)	111%	SPK: 50
1868-53-7	Dibromofluoromethane	51.7		70 (75) - 130 (124)	103%	SPK: 50
2037-26-5	Toluene-d8	49.5		70 (86) - 130 (113)	99%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.3		70 (77) - 130 (121)	97%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	159000	8.206			
540-36-3	1,4-Difluorobenzene	390000	9.082			
3114-55-4	Chlorobenzene-d5	381000	11.841			
3855-82-1	1,4-Dichlorobenzene-d4	181000	13.77			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090825\
 Data File : VN087760.D
 Acq On : 08 Sep 2025 14:42
 Operator : JC\MD
 Sample : Q3032-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SOIL-PILE-2

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 09/09/2025
 Supervised By : Mahesh Dadoda 09/09/2025

Quant Time: Sep 09 02:29:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:55:42 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	159239	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	389686	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	380890	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	180770	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	181221	55.545	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 111.080%			
35) Dibromofluoromethane	8.147	113	145598	51.749	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.500%			
50) Toluene-d8	10.547	98	521403	49.514	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.020%			
62) 4-Bromofluorobenzene	12.829	95	181031	48.340	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 96.680%			
Target Compounds					Qvalue	
11) Tert butyl alcohol	5.542	59	4626m	8.167	ug/l	
16) Acetone	4.424	43	61220	34.469	ug/l	92
18) Methyl Acetate	5.024	43	30099	8.094	ug/l	100
20) Methylene Chloride	5.259	84	9689	2.595	ug/l	95
25) 2-Butanone	7.471	43	30576	13.085	ug/l	95
37) Ethyl Acetate	7.547	43	14840	2.516	ug/l	98
43) Isopropyl Acetate	8.671	43	10190	1.107	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090825\
Data File : VN087760.D
Acq On : 08 Sep 2025 14:42
Operator : JC\MD
Sample : Q3032-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
SOIL-PILE-2

Quant Time: Sep 09 02:29:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:55:42 2025
Response via : Initial Calibration

Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 09/09/2025
Supervised By :Mahesh Dadoda 09/09/2025





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.206 min Scan# 1063
Delta R.T. 0.000 min
Lab File: VN087760.D
Acq: 08 Sep 2025 14:42




Instrument :
MSVOA_N
ClientSampleId :
SOIL-PILE-2



Tgt Ion:168 Resp: 159239
Ion Ratio Lower Upper
168 100
99 68.8 57.2 85.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 09/09/2025
Supervised By :Mahesh Dadoda 09/09/2025



#11
Tert butyl alcohol
Concen: 8.167 ug/l m
RT: 5.542 min Scan# 610
Delta R.T. 0.023 min
Lab File: VN087760.D
Acq: 08 Sep 2025 14:42


Tgt Ion: 59 Resp: 4626
Ion Ratio Lower Upper
59 100
57 1.4 9.0 13.4#





#16
Acetone
Concen: 34.469 ug/l
RT: 4.424 min Scan# 41
Delta R.T. -0.000 min
Lab File: VN087760.D
Acq: 08 Sep 2025 14:42




Instrument :
MSVOA_N
ClientSampleId :
SOIL-PILE-2



Tgt Ion: 43 Resp: 61220
Ion Ratio Lower Upper
43 100
58 32.6 22.6 33.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 09/09/2025
Supervised By :Mahesh Dadoda 09/09/2025



#18
Methyl Acetate
Concen: 8.094 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.012 min
Lab File: VN087760.D
Acq: 08 Sep 2025 14:42

Tgt Ion: 43 Resp: 30099
Ion Ratio Lower Upper
43 100
74 21.9 17.6 26.4





#20

Methylene Chloride

Concen: 2.595 ug/l

RT: 5.259 min Scan# 51

Delta R.T. -0.006 min

Lab File: VN087760.D

Acq: 08 Sep 2025 14:42

Instrument :

MSVOA_N

ClientSampleId :

SOIL-PILE-2

Tgt Ion: 84 Resp: 9689

Ion Ratio Lower Upper

84 100

49 134.6 103.4 155.2

51 44.8 33.8 50.8





86 58.1 49.6 74.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 09/09/2025

Supervised By :Mahesh Dadoda 09/09/2025



#25

2-Butanone

Concen: 13.085 ug/l

RT: 7.471 min Scan# 938

Delta R.T. 0.006 min

Lab File: VN087760.D




Acq: 08 Sep 2025 14:42


Tgt Ion: 43 Resp: 30576

Ion Ratio Lower Upper

43 100

72 27.3 19.9 29.9





#33

1,2-Dichloroethane-d4

Concen: 55.545 ug/l

RT: 8.559 min Scan# 1123

Delta R.T. -0.000 min

Lab File: VN087760.D

Acq: 08 Sep 2025 14:42

Instrument :

MSVOA_N

ClientSampleId :

SOIL-PILE-2

Tgt Ion: 65 Resp: 18122

Ion Ratio Lower Upper

65 100





67 53.1 0.0 100.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 09/09/2025

Supervised By :Mahesh Dadoda 09/09/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.082 min Scan# 1212

Delta R.T. -0.000 min

Lab File: VN087760.D

Acq: 08 Sep 2025 14:42




Tgt Ion:114 Resp: 389686


Ion Ratio Lower Upper

114 100

63 22.5 0.0 45.2

88 15.5 0.0 33.4





#35

Dibromofluoromethane

Concen: 51.749 ug/l

RT: 8.147 min Scan# 1053

Delta R.T. -0.006 min

Lab File: VN087760.D

Acq: 08 Sep 2025 14:42

Instrument :

MSVOA_N

ClientSampleId :





SOIL-PILE-2

Tgt Ion:	113	Resp:	145598
Ion	Ratio	Lower	Upper
113	100		
111	99.4	82.8	124.2
192	16.1	12.8	19.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 09/09/2025
Supervised By :Mahesh Dadoda 09/09/2025



#37

Ethyl Acetate

Concen: 2.516 ug/l




RT: 7.547 min Scan# 951


Delta R.T. -0.000 min

Lab File: VN087760.D

Acq: 08 Sep 2025 14:42


Tgt Ion:	43	Resp:	14840
Ion	Ratio	Lower	Upper
43	100		
61	12.7	9.3	13.9
70	8.2	7.0	10.6





#43
Isopropyl Acetate
Concen: 1.107 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. -0.000 min
Lab File: VN087760.D
Acq: 08 Sep 2025 14:42




Instrument :
MSVOA_N
ClientSampleId :
SOIL-PILE-2



Tgt Ion: 43 Resp: 10190
Ion Ratio Lower Upper
43 100
61 17.1 18.8 28.2
87 9.2 9.6 14.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 09/09/2025
Supervised By :Mahesh Dadoda 09/09/2025



#50
Toluene-d8
Concen: 49.514 ug/l
RT: 10.547 min Scan# 1461
Delta R.T. -0.000 min
Lab File: VN087760.D
Acq: 08 Sep 2025 14:42


Tgt Ion: 98 Resp: 521403
Ion Ratio Lower Upper
98 100
100 63.5 52.8 79.2





#62
4-Bromofluorobenzene
Concen: 48.340 ug/l
RT: 12.829 min Scan# 1848
Delta R.T. 0.006 min
Lab File: VN087760.D
Acq: 08 Sep 2025 14:42




Instrument :
MSVOA_N
ClientSampleId :
SOIL-PILE-2




Tgt Ion: 95 Resp: 18103
Ion Ratio Lower Upper
95 100
174 72.1 0.0 138.4
176 67.6 0.0 132.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 09/09/2025
Supervised By :Mahesh Dadoda 09/09/2025




#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.841 min Scan# 1681
Delta R.T. -0.000 min
Lab File: VN087760.D
Acq: 08 Sep 2025 14:42



Tgt Ion:117 Resp: 380890
Ion Ratio Lower Upper
117 100
82 62.6 47.4 71.2
119 32.4 24.3 36.5





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.770 min Scan# 20

Delta R.T. 0.006 min

Lab File: VN087760.D


Acq: 08 Sep 2025 14:42

Instrument :

MSVOA_N

ClientSampleId :

SOIL-PILE-2



Tgt Ion:152 Resp: 180770

Ion Ratio Lower Upper

152 100

115 61.7 32.3 96.8



150 159.7 0.0 351.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 09/09/2025

Supervised By :Mahesh Dadoda 09/09/2025





CALIBRATION SUMMARY



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance Contract: ZUCC01
 Lab Code: ACE SDG No.: Q3032
 Instrument ID: MSVOA_N Calibration Date(s): 08/21/2025 08/21/2025
 Heated Purge: (Y/N) N Calibration Time(s): 12:09 14:44
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF001 = VN087619.D		RRF005 = VN087620.D		RRF020 = VN087621.D			
		RRF050 = VN087622.D		RRF100 = VN087623.D		RRF150 = VN087624.D			
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD	
Vinyl Chloride	0.909	0.739	0.832	0.903	0.847	0.846	0.846	7.3	
1,1-Dichloroethene	0.892	0.649	0.730	0.700	0.667	0.687	0.721	12.3	
2-Butanone	0.783	0.722	0.757	0.732	0.700	0.707	0.734	4.3	
Carbon Tetrachloride	0.571	0.509	0.568	0.542	0.549	0.542	0.547	4.1	
Chloroform	1.677	1.535	1.629	1.579	1.519	1.528	1.578	4	
Benzene	1.853	1.637	1.727	1.684	1.639	1.652	1.699	4.9	
1,2-Dichloroethane	0.714	0.627	0.675	0.650	0.624	0.626	0.653	5.5	
Trichloroethene	0.435	0.387	0.390	0.377	0.367	0.370	0.388	6.5	
Tetrachloroethene	0.447	0.353	0.340	0.313	0.315	0.313	0.347	14.9	
Chlorobenzene	1.408	1.164	1.253	1.210	1.216	1.212	1.244	6.9	
1,2-Dichloroethane-d4		1.025	1.016	0.959	1.035	1.089	1.024	4.5	
Dibromofluoromethane		0.369	0.344	0.334	0.373	0.386	0.361	6	
Toluene-d8		1.384	1.266	1.223	1.408	1.475	1.351	7.7	
4-Bromofluorobenzene		0.446	0.454	0.450	0.514	0.540	0.481	9	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

Method Path : Z:\voasrv\HPCHEM1\MSVOA_N\methods\
 Method File : 82N082125W.M
 Title : SW846 8260
 Last Update : Fri Aug 22 02:55:42 2025
 Response Via : Initial Calibration

Calibration Files

1 =VN087619.D 5 =VN087620.D 20 =VN087621.D 50 =VN087622.D 100 =VN087623.D 150 =VN087624.D

	Compound	1	5	20	50	100	150	Avg	%RSD
1) I	Pentafluorobenzene	-----ISTD-----							
2) T	Dichlorodifluo...	0.707	0.579	0.606	0.810	0.777	0.782	0.710	13.74
3) P	Chloromethane	0.739	0.637	0.710	0.795	0.752	0.747	0.730	7.27
4) C	Vinyl Chloride	0.909	0.739	0.832	0.903	0.847	0.846	0.846	7.29#
5) T	Bromomethane		0.388	0.393	0.448	0.466	0.488	0.437	10.17
6) T	Chloroethane	0.727	0.528	0.599	0.603	0.558	0.554	0.595	11.88
7) T	Trichlorofluor...	1.323	1.139	1.242	1.274	1.227	1.233	1.240	4.91
8) T	Diethyl Ether	0.637	0.501	0.544	0.529	0.510	0.519	0.540	9.22
9) T	1,1,2-Trichlor...	0.879	0.648	0.677	0.695	0.654	0.656	0.701	12.62
10) T	Methyl Iodide		0.265	0.355	0.481	0.562	0.586	0.450	30.46
11) T	Tert butyl alc...		0.170	0.187	0.184	0.174	0.174	0.178	4.02
12) CM	1,1-Dichloroet...	0.892	0.649	0.730	0.700	0.667	0.687	0.721	12.28#
13) T	Acrolein		0.251	0.198	0.215	0.196	0.233	0.219	10.78
14) T	Allyl chloride	1.577	1.184	1.254	1.240	1.208	1.376	1.306	11.34
15) T	Acrylonitrile	0.555	0.466	0.510	0.498	0.487	0.493	0.502	6.00
16) T	Acetone	0.633	0.634	0.584	0.522	0.484	0.489	0.558	12.34
17) T	Carbon Disulfide	2.272	1.746	1.970	2.003	1.937	1.979	1.985	8.51
18) T	Methyl Acetate	1.450	1.043	1.097	1.157	1.117	1.141	1.168	12.33
19) T	Methyl tert-bu...	2.859	2.508	2.740	2.733	2.643	2.742	2.704	4.37
20) T	Methylene Chlo...	1.494	0.916	0.875	0.823	0.779	0.799	0.948	28.76
21) T	trans-1,2-Dich...	0.952	0.734	0.768	0.750	0.737	0.745	0.781	10.83
22) T	Diisopropyl ether	2.977	2.591	2.903	2.870	2.766	2.807	2.819	4.74
23) T	Vinyl Acetate	2.629	2.379	2.484	2.692	2.576	2.629	2.565	4.46
24) P	1,1-Dichloroet...	1.798	1.500	1.549	1.501	1.454	1.471	1.546	8.28
25) T	2-Butanone	0.783	0.722	0.757	0.732	0.700	0.707	0.734	4.31
26) T	2,2-Dichloropr...	1.500	1.243	1.326	1.317	1.277	1.296	1.327	6.78
27) T	cis-1,2-Dichlo...	0.997	0.916	0.937	0.941	0.908	0.905	0.934	3.67
28) T	Bromochloromet...	0.800	0.781	0.722	0.701	0.748	0.731	0.747	4.95
29) T	Tetrahydrofuran	0.520	0.432	0.479	0.474	0.461	0.468	0.472	6.01
30) C	Chloroform	1.677	1.535	1.629	1.579	1.519	1.528	1.578	4.03#
31) T	Cyclohexane		1.490	1.310	1.234	1.201	1.207	1.289	9.38
32) T	1,1,1-Trichlor...	1.510	1.297	1.359	1.303	1.265	1.288	1.337	6.75
33) S	1,2-Dichloroet...		1.025	1.016	0.959	1.035	1.089	1.024	4.53
34) I	1,4-Difluorobenzene	-----ISTD-----							
35) S	Dibromofluorom...		0.369	0.344	0.334	0.373	0.386	0.361	5.97
36) T	1,1-Dichloropr...	0.700	0.508	0.550	0.528	0.522	0.514	0.554	13.21
37) T	Ethyl Acetate	0.877	0.700	0.761	0.742	0.731	0.731	0.757	8.20
38) T	Carbon Tetrach...	0.571	0.509	0.568	0.542	0.549	0.542	0.547	4.12
39) T	Methylcyclohexane	0.686	0.580	0.610	0.590	0.597	0.596	0.610	6.33
40) TM	Benzene	1.853	1.637	1.727	1.684	1.639	1.652	1.699	4.88
41) T	Methacrylonitrile	0.466	0.346	0.388	0.389	0.384	0.379	0.392	10.06
42) TM	1,2-Dichloroet...	0.714	0.627	0.675	0.650	0.624	0.626	0.653	5.48
43) T	Isopropyl Acetate	1.236	1.076	1.215	1.198	1.187	1.178	1.182	4.71
44) TM	Trichloroethene	0.435	0.387	0.390	0.377	0.367	0.370	0.388	6.46
45) C	1,2-Dichloropr...	0.544	0.435	0.447	0.433	0.415	0.412	0.448	10.96#
46) T	Dibromomethane	0.351	0.294	0.327	0.318	0.310	0.312	0.319	6.06
47) T	Bromodichlorom...	0.686	0.649	0.667	0.638	0.638	0.639	0.653	3.01
48) T	Methyl methacr...	0.598	0.493	0.572	0.558	0.565	0.567	0.559	6.26
49) T	1,4-Dioxane	0.006	0.007	0.008	0.008	0.008	0.007	0.007	11.46
50) S	Toluene-d8		1.384	1.266	1.223	1.408	1.475	1.351	7.68
51) T	4-Methyl-2-Pen...	0.701	0.675	0.744	0.736	0.718	0.704	0.713	3.55
52) CM	Toluene	1.148	0.987	1.062	1.050	1.037	1.033	1.053	5.03#
53) T	t-1,3-Dichloro...	0.620	0.628	0.691	0.703	0.699	0.715	0.676	6.08
54) T	cis-1,3-Dichlo...	0.717	0.640	0.712	0.718	0.714	0.710	0.702	4.32
55) T	1,1,2-Trichlor...	0.439	0.383	0.417	0.408	0.400	0.397	0.407	4.69
56) T	Ethyl methacry...	0.532	0.546	0.681	0.705	0.715	0.727	0.651	13.55

Method Path : Z:\voasrv\HPCHEM1\MSVOA_N\methods\


Method File : 82N082125W.M

57)	T	1,3-Dichloropr...	0.781	0.656	0.743	0.731	0.708	0.706	0.721	5.80
58)	T	2-Chloroethyl ...	0.360	0.333	0.305	0.354	0.374	0.388	0.352	8.40
59)	T	2-Hexanone	0.295	0.390	0.493	0.510	0.508	0.507	0.451	19.77
60)	T	Dibromochlorom...	0.535	0.440	0.458	0.466	0.465	0.468	0.472	6.85
61)	T	1,2-Dibromoethane	0.412	0.433	0.444	0.432	0.429	0.427	0.430	2.44
62)	S	4-Bromofluorob...	0.446	0.454	0.450	0.514	0.540	0.481		8.99
63)	I	Chlorobenzene-d5	-----ISTD-----							
64)	T	Tetrachloroethene	0.447	0.353	0.340	0.313	0.315	0.313	0.347	14.89
65)	PM	Chlorobenzene	1.408	1.164	1.253	1.210	1.216	1.212	1.244	6.85
66)	T	1,1,1,2-Tetrac...	0.391	0.404	0.427	0.421	0.416	0.418	0.413	3.14
67)	C	Ethyl Benzene	2.232	1.993	2.224	2.144	2.152	2.178	2.154	4.03#
68)	T	m/p-Xylenes	0.759	0.735	0.811	0.812	0.808	0.814	0.790	4.33
69)	T	o-Xylene	0.777	0.707	0.794	0.796	0.788	0.784	0.774	4.35
70)	T	Styrene	1.142	1.086	1.391	1.382	1.386	1.393	1.297	10.99
71)	P	Bromoform	0.303	0.285	0.320	0.320	0.336	0.336	0.317	6.27
72)	I	1,4-Dichlorobenzen...	-----ISTD-----							
73)	T	Isopropylbenzene	4.328	3.674	4.096	4.026	3.998	4.120	4.040	5.29
74)	T	N-amyl acetate	1.569	1.538	1.265	1.555	1.527	1.733	1.531	9.86
75)	P	1,1,2,2-Tetrac...	1.488	1.381	1.421	1.363	1.322	1.371	1.391	4.12
76)	T	1,2,3-Trichlor...	1.274	1.171	1.188	1.073	1.074	1.256	1.173	7.36
77)	T	Bromobenzene	1.019	0.901	0.988	0.941	0.929	0.938	0.953	4.51
78)	T	n-propylbenzene	5.137	4.545	5.125	5.024	4.965	5.064	4.977	4.44
79)	T	2-Chlorotoluene	2.911	2.953	3.116	2.976	2.975	3.037	2.995	2.41
80)	T	1,3,5-Trimethy...	3.586	3.159	3.498	3.471	3.398	3.452	3.428	4.24
81)	T	trans-1,4-Dich...	0.332	0.416	0.451	0.493	0.523	0.443		16.70
82)	T	4-Chlorotoluene	3.352	3.001	3.225	3.153	3.069	3.093	3.149	3.98
83)	T	tert-Butylbenzene	2.913	2.674	2.899	2.911	2.857	2.884	2.856	3.22
84)	T	1,2,4-Trimethy...	3.502	3.091	3.551	3.496	3.471	3.476	3.431	4.93
85)	T	sec-Butylbenzene	4.464	4.019	4.367	4.225	4.163	4.194	4.239	3.71
86)	T	p-Isopropyltol...	3.656	3.249	3.550	3.530	3.502	3.512	3.500	3.85
87)	T	1,3-Dichlorobe...	2.087	1.770	1.940	1.827	1.806	1.826	1.876	6.29
88)	T	1,4-Dichlorobe...	2.348	1.884	1.956	1.867	1.818	1.841	1.952	10.22
89)	T	n-Butylbenzene	3.764	3.208	3.589	3.461	3.398	3.435	3.476	5.40
90)	T	Hexachloroethane	0.742	0.606	0.680	0.655	0.649	0.669	0.667	6.69
91)	T	1,2-Dichlorobe...	1.942	1.662	1.853	1.755	1.723	1.743	1.780	5.66
92)	T	1,2-Dibromo-3-...	0.370	0.323	0.329	0.317	0.313	0.330	0.330	6.21
93)	T	1,2,4-Trichlor...	1.192	0.995	1.081	1.026	1.040	1.079	1.069	6.43
94)	T	Hexachlorobuta...	0.469	0.371	0.385	0.355	0.346	0.361	0.381	11.80
95)	T	Naphthalene	3.989	3.240	3.724	3.754	3.917	4.089	3.786	7.95
96)	T	1,2,3-Trichlor...	1.140	0.972	1.059	1.006	1.019	1.072	1.045	5.68

(#) = Out of Range

Methylene Chloride

Response Ratio



$$\text{Response} = 7.866\text{e-}001 * \text{Amt} + 2.003\text{e-}002$$


Coef of Det (r^2) = 0.999518 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N082125W.M

Calibration Table Last Updated: Fri Aug 22 02:55:42 2025

Methyl Iodide

Response Ratio



$$\text{Response} = 6.069\text{e-}001 * \text{Amt} - 8.267\text{e-}002$$

Coef of Det (r^2) = 0.997633 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N082125W.M

Calibration Table Last Updated: Fri Aug 22 02:55:42 2025

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082125\
 Data File : VN087619.D
 Acq On : 21 Aug 2025 12:09
 Operator : JC\MD
 Sample : VSTDIC001
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC001

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/22/2025
 Supervised By : Mahesh Dadoda 08/25/2025

Quant Time: Aug 22 02:28:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:09:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	214194	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	425894	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	384310	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	174637	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	0.000	65	0d	0.000	ug/l	
Spiked Amount	50.000	Range	74 - 125	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0d	0.000	ug/l	
Spiked Amount	50.000	Range	75 - 124	Recovery	=	0.000%#
50) Toluene-d8	0.000	98	0d	0.000	ug/l	
Spiked Amount	50.000	Range	86 - 113	Recovery	=	0.000%#
62) 4-Bromofluorobenzene	0.000	95	0d	0.000	ug/l	
Spiked Amount	50.000	Range	77 - 121	Recovery	=	0.000%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.142	85	3030	0.996	ug/l	75
3) Chloromethane	2.389	50	3165	1.012	ug/l #	87
4) Vinyl Chloride	2.542	62	3896	1.075	ug/l	98
6) Chloroethane	3.142	64	3113	1.222	ug/l	89
7) Trichlorofluoromethane	3.518	101	5667	1.067	ug/l	88
8) Diethyl Ether	3.977	74	2728m	1.179	ug/l	
9) 1,1,2-Trichlorotrifluo...	4.371	101	3764	1.253	ug/l #	68
12) 1,1-Dichloroethene	4.353	96	3823	1.238	ug/l #	55
14) Allyl chloride	5.024	41	6754	1.207	ug/l #	70
15) Acrylonitrile	5.706	53	11893	5.535	ug/l #	90
16) Acetone	4.424	43	13568	5.679	ug/l #	75
17) Carbon Disulfide	4.700	76	9734	1.145	ug/l #	90
18) Methyl Acetate	5.012	43	6213m	1.242	ug/l	
19) Methyl tert-butyl Ether	5.794	73	12249	1.057	ug/l	91
20) Methylene Chloride	5.277	84	6402	0.627	ug/l #	57
21) trans-1,2-Dichloroethene	5.777	96	4079	1.219	ug/l	93
22) Diisopropyl ether	6.665	45	12751	1.056	ug/l #	76
23) Vinyl Acetate	6.588	43	56320	5.126	ug/l #	90
24) 1,1-Dichloroethane	6.541	63	7704	1.163	ug/l #	94
25) 2-Butanone	7.471	43	16779	5.338	ug/l	95
26) 2,2-Dichloropropane	7.477	77	6425	1.131	ug/l	77
27) cis-1,2-Dichloroethene	7.477	96	4272	1.068	ug/l	94
28) Bromochloromethane	7.794	49	3426	1.070	ug/l #	91
29) Tetrahydrofuran	7.830	42	11129	5.502	ug/l	98
30) Chloroform	7.953	83	7183	1.063	ug/l	97
32) 1,1,1-Trichloroethane	8.159	97	6468	1.129	ug/l #	52
36) 1,1-Dichloropropene	8.353	75	5965	1.264	ug/l	90
37) Ethyl Acetate	7.547	43	7468m	1.159	ug/l	
38) Carbon Tetrachloride	8.347	117	4866	1.045	ug/l	90
39) Methylcyclohexane	9.582	83	5843	1.125	ug/l #	87
40) Benzene	8.588	78	15783	1.091	ug/l	100
41) Methacrylonitrile	7.771	41	3967	1.188	ug/l #	71
42) 1,2-Dichloroethane	8.659	62	6078	1.093	ug/l	95
43) Isopropyl Acetate	8.677	43	10524	1.046	ug/l #	64
44) Trichloroethene	9.329	130	3709	1.123	ug/l	96
45) 1,2-Dichloropropane	9.594	63	4636	1.216	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082125\
 Data File : VN087619.D
 Acq On : 21 Aug 2025 12:09
 Operator : JC\MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/22/2025
 Supervised By : Mahesh Dadoda 08/25/2025

Quant Time: Aug 22 02:28:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:09:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) Dibromomethane	9.694	93	2989	1.101	ug/l	96
47) Bromodichloromethane	9.871	83	5845	1.051	ug/l #	90
48) Methyl methacrylate	9.671	41	5091	1.069	ug/l	94
49) 1,4-Dioxane	9.700	88	1008	16.337	ug/l #	13
51) 4-Methyl-2-Pentanone	10.423	43	29834	4.913	ug/l	99
52) Toluene	10.606	92	9776	1.090	ug/l	92
53) t-1,3-Dichloropropene	10.818	75	5278	0.917	ug/l #	80
54) cis-1,3-Dichloropropene	10.300	75	6105	1.021	ug/l #	80
55) 1,1,2-Trichloroethane	10.994	97	3739	1.078	ug/l #	87
56) Ethyl methacrylate	10.865	69	4529	0.817	ug/l #	89
57) 1,3-Dichloropropane	11.141	76	6652	1.083	ug/l	90
58) 2-Chloroethyl Vinyl ether	10.141	63	15333	5.109	ug/l	100
59) 2-Hexanone	11.200	43	12564	3.274	ug/l	94
60) Dibromochloromethane	11.347	129	4554	1.133	ug/l	93
61) 1,2-Dibromoethane	11.447	107	3511	0.960	ug/l	93
64) Tetrachloroethene	11.076	164	3433	1.288	ug/l #	84
65) Chlorobenzene	11.876	112	10823	1.132	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.941	131	3009	0.948	ug/l #	64
67) Ethyl Benzene	11.941	91	17159	1.037	ug/l	92
68) m/p-Xylenes	12.047	106	11663	1.921	ug/l	90
69) o-Xylene	12.376	106	5970	1.003	ug/l	100
70) Styrene	12.394	104	8781	0.881	ug/l	97
71) Bromoform	12.559	173	2328	0.956	ug/l #	84
73) Isopropylbenzene	12.676	105	15118	1.071	ug/l	92
74) N-amyl acetate	13.012	43	5481m	1.010	ug/l	
75) 1,1,2,2-Tetrachloroethane	12.917	83	5198	1.070	ug/l	95
76) 1,2,3-Trichloropropane	12.970	75	4451m	1.084	ug/l	
77) Bromobenzene	12.970	156	3560	1.070	ug/l	96
78) n-propylbenzene	13.017	91	17943	1.032	ug/l	98
79) 2-Chlorotoluene	13.112	91	10167	0.972	ug/l	85
80) 1,3,5-Trimethylbenzene	13.147	105	12525	1.046	ug/l	92
82) 4-Chlorotoluene	13.200	91	11708	1.065	ug/l	96
83) tert-Butylbenzene	13.417	119	10175	1.020	ug/l	98
84) 1,2,4-Trimethylbenzene	13.459	105	12231	1.021	ug/l	99
85) sec-Butylbenzene	13.588	105	15591	1.053	ug/l	95
86) p-Isopropyltoluene	13.706	119	12768	1.044	ug/l	98
87) 1,3-Dichlorobenzene	13.717	146	7290	1.113	ug/l	89
88) 1,4-Dichlorobenzene	13.782	146	8201m	1.203	ug/l	
89) n-Butylbenzene	14.029	91	13148	1.083	ug/l	96
90) Hexachloroethane	14.311	117	2591	1.113	ug/l	100
91) 1,2-Dichlorobenzene	14.094	146	6783	1.091	ug/l	93
92) 1,2-Dibromo-3-Chloropr...	14.694	75	1293	1.120	ug/l	83
93) 1,2,4-Trichlorobenzene	15.376	180	4163	1.115	ug/l	88
94) Hexachlorobutadiene	15.470	225	1637	1.230	ug/l	97
95) Naphthalene	15.617	128	13934	1.054	ug/l	98
96) 1,2,3-Trichlorobenzene	15.811	180	3983	1.091	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED

Reviewed By :John Carlone 08/22/2025
Supervised By :Mahesh Dadoda 08/25/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082125\
 Data File : VN087620.D
 Acq On : 21 Aug 2025 13:08
 Operator : JC\MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/22/2025
 Supervised By : Mahesh Dadoda 08/25/2025

Quant Time: Aug 22 02:29:54 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:09:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	230389	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	446380	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	411161	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	194457	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	23607	5.001	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	10.000%	#	
35) Dibromofluoromethane	8.141	113	16452	5.105	ug/l	-0.01
Spiked Amount 50.000	Range 75 - 124		Recovery =	10.200%	#	
50) Toluene-d8	10.547	98	61777	5.121	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	10.240%	#	
62) 4-Bromofluorobenzene	12.829	95	19896	4.638	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	9.280%	#	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.142	85	13341	4.077	ug/l	91
3) Chloromethane	2.383	50	14670	4.363	ug/l #	87
4) Vinyl Chloride	2.542	62	17021	4.366	ug/l	96
5) Bromomethane	2.971	94	8950	4.449	ug/l #	67
6) Chloroethane	3.130	64	12160	4.437	ug/l	99
7) Trichlorofluoromethane	3.512	101	26231	4.592	ug/l #	69
8) Diethyl Ether	3.959	74	11532	4.635	ug/l	91
9) 1,1,2-Trichlorotrifluo...	4.371	101	14929	4.619	ug/l #	85
10) Methyl Iodide	4.583	142	6109	8.995	ug/l	90
11) Tert butyl alcohol	5.530	59	19630	23.954	ug/l #	71
12) 1,1-Dichloroethene	4.336	96	14950	4.501	ug/l	89
13) Acrolein	4.183	56	28918	28.704	ug/l	96
14) Allyl chloride	5.012	41	27277	4.532	ug/l	94
15) Acrylonitrile	5.706	53	53640	23.209	ug/l	99
16) Acetone	4.418	43	73025	28.418	ug/l	97
17) Carbon Disulfide	4.694	76	40223	4.399	ug/l #	94
18) Methyl Acetate	5.012	43	24038	4.468	ug/l #	61
19) Methyl tert-butyl Ether	5.783	73	57783	4.637	ug/l	98
20) Methylene Chloride	5.253	84	21101	4.549	ug/l #	80
21) trans-1,2-Dichloroethene	5.777	96	16922	4.701	ug/l	96
22) Diisopropyl ether	6.653	45	59705	4.596	ug/l	99
23) Vinyl Acetate	6.588	43	274045	23.189	ug/l	98
24) 1,1-Dichloroethane	6.565	63	34563	4.853	ug/l	98
25) 2-Butanone	7.477	43	83189	24.607	ug/l	93
26) 2,2-Dichloropropane	7.482	77	28641	4.685	ug/l	97
27) cis-1,2-Dichloroethene	7.465	96	21105	4.904	ug/l	95
28) Bromochloromethane	7.788	49	17986	5.224	ug/l #	98
29) Tetrahydrofuran	7.830	42	49801	22.890	ug/l	97
30) Chloroform	7.953	83	35366	4.865	ug/l	89
31) Cyclohexane	8.241	56	34337	5.783	ug/l #	96
32) 1,1,1-Trichloroethane	8.147	97	29878	4.850	ug/l	90
36) 1,1-Dichloropropene	8.353	75	22697	4.591	ug/l	96
37) Ethyl Acetate	7.547	43	31230m	4.622	ug/l	
38) Carbon Tetrachloride	8.341	117	22716	4.653	ug/l #	84
39) Methylcyclohexane	9.576	83	25902	4.758	ug/l	91
40) Benzene	8.582	78	73071	4.818	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082125\
 Data File : VN087620.D
 Acq On : 21 Aug 2025 13:08
 Operator : JC\MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/22/2025
 Supervised By : Mahesh Dadoda 08/25/2025

Quant Time: Aug 22 02:29:54 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:09:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.759	41	15464	4.419	ug/l #	95
42) 1,2-Dichloroethane	8.653	62	27989	4.803	ug/l	95
43) Isopropyl Acetate	8.671	43	48032	4.554	ug/l	99
44) Trichloroethene	9.329	130	17267	4.987	ug/l	91
45) 1,2-Dichloropropane	9.606	63	19397	4.853	ug/l	95
46) Dibromomethane	9.694	93	13115	4.611	ug/l	91
47) Bromodichloromethane	9.865	83	28992	4.973	ug/l	99
48) Methyl methacrylate	9.659	41	22006	4.411	ug/l	94
49) 1,4-Dioxane	9.682	88	6079	94.002	ug/l #	89
51) 4-Methyl-2-Pentanone	10.423	43	150608	23.665	ug/l	97
52) Toluene	10.606	92	44071	4.688	ug/l	98
53) t-1,3-Dichloropropene	10.812	75	28043	4.647	ug/l	92
54) cis-1,3-Dichloropropene	10.288	75	28580	4.561	ug/l	95
55) 1,1,2-Trichloroethane	10.994	97	17105	4.703	ug/l	96
56) Ethyl methacrylate	10.853	69	24379	4.194	ug/l #	92
57) 1,3-Dichloropropane	11.141	76	29300	4.552	ug/l	96
58) 2-Chloroethyl Vinyl ether	10.141	63	74236	23.598	ug/l	100
59) 2-Hexanone	11.182	43	87097	21.654	ug/l	98
60) Dibromochloromethane	11.335	129	19653	4.664	ug/l	100
61) 1,2-Dibromoethane	11.447	107	19347	5.045	ug/l	93
64) Tetrachloroethene	11.082	164	14530	5.094	ug/l #	82
65) Chlorobenzene	11.870	112	47879	4.681	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.935	131	16599	4.888	ug/l	88
67) Ethyl Benzene	11.941	91	81935	4.626	ug/l	98
68) m/p-Xylenes	12.053	106	60444	9.306	ug/l	96
69) o-Xylene	12.376	106	29064	4.566	ug/l	98
70) Styrene	12.388	104	44658	4.188	ug/l	93
71) Bromoform	12.559	173	11710	4.496	ug/l #	92
73) Isopropylbenzene	12.670	105	71448	4.547	ug/l	98
74) N-amyl acetate	12.670	43	29903m	4.946	ug/l	
75) 1,1,2,2-Tetrachloroethane	12.917	83	26850	4.963	ug/l	95
76) 1,2,3-Trichloropropane	12.976	75	22762m	4.980	ug/l	
77) Bromobenzene	12.959	156	17527	4.731	ug/l	97
78) n-propylbenzene	13.012	91	88388	4.567	ug/l	99
79) 2-Chlorotoluene	13.100	91	57432	4.931	ug/l	94
80) 1,3,5-Trimethylbenzene	13.153	105	61434	4.609	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.717	75	6458	3.749	ug/l	87
82) 4-Chlorotoluene	13.200	91	58355	4.765	ug/l	96
83) tert-Butylbenzene	13.417	119	51989	4.680	ug/l	98
84) 1,2,4-Trimethylbenzene	13.464	105	60099	4.504	ug/l	98
85) sec-Butylbenzene	13.594	105	78154	4.741	ug/l	98
86) p-Isopropyltoluene	13.706	119	63187	4.642	ug/l	98
87) 1,3-Dichlorobenzene	13.711	146	34424	4.718	ug/l	100
88) 1,4-Dichlorobenzene	13.788	146	36629	4.824	ug/l	95
89) n-Butylbenzene	14.035	91	62374	4.614	ug/l	99
90) Hexachloroethane	14.306	117	11787	4.545	ug/l	98
91) 1,2-Dichlorobenzene	14.088	146	32320	4.669	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.694	75	6280	4.887	ug/l	86
93) 1,2,4-Trichlorobenzene	15.364	180	19345	4.654	ug/l	97
94) Hexachlorobutadiene	15.470	225	7207	4.864	ug/l	96
95) Naphthalene	15.611	128	63009	4.280	ug/l	99
96) 1,2,3-Trichlorobenzene	15.811	180	18896	4.650	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082125\
Data File : VN087620.D
Acq On : 21 Aug 2025 13:08
Operator : JC\MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/22/2025
Supervised By :Mahesh Dadoda 08/25/2025

Quant Time: Aug 22 02:29:54 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:09:28 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED

Reviewed By :John Carlone 08/22/2025
Supervised By :Mahesh Dadoda 08/25/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082125\
 Data File : VN087621.D
 Acq On : 21 Aug 2025 13:41
 Operator : JC\MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC020

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/22/2025
 Supervised By : Mahesh Dadoda 08/25/2025

Quant Time: Aug 22 02:30:54 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:09:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	225514	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	446006	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	409187	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.771	152	204594	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	91607	19.826	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	39.660%#		
35) Dibromofluoromethane	8.153	113	61315	19.041	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	38.080%#		
50) Toluene-d8	10.547	98	225931	18.746	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	37.500%#		
62) 4-Bromofluorobenzene	12.829	95	80936	18.883	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	37.760%#		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.142	85	54651	17.063	ug/l	96
3) Chloromethane	2.383	50	64072	19.466	ug/l	95
4) Vinyl Chloride	2.536	62	75059	19.667	ug/l	99
5) Bromomethane	2.971	94	35415	17.984	ug/l	90
6) Chloroethane	3.136	64	54039	20.146	ug/l	98
7) Trichlorofluoromethane	3.506	101	112067	20.042	ug/l	92
8) Diethyl Ether	3.954	74	49068	20.150	ug/l	97
9) 1,1,2-Trichlorotrifluo...	4.365	101	61080	19.305	ug/l	96
10) Methyl Iodide	4.583	142	32047	18.518	ug/l #	97
11) Tert butyl alcohol	5.524	59	84156	104.914	ug/l	98
12) 1,1-Dichloroethene	4.336	96	65870	20.259	ug/l	93
13) Acrolein	4.171	56	89266	90.522	ug/l	99
14) Allyl chloride	5.012	41	113139	19.202	ug/l	96
15) Acrylonitrile	5.706	53	230008	101.673	ug/l	99
16) Acetone	4.418	43	263348	104.700	ug/l	97
17) Carbon Disulfide	4.701	76	177734	19.857	ug/l	100
18) Methyl Acetate	5.012	43	98949	18.789	ug/l	97
19) Methyl tert-butyl Ether	5.789	73	247157	20.264	ug/l	99
20) Methylene Chloride	5.265	84	78920	20.972	ug/l	96
21) trans-1,2-Dichloroethene	5.777	96	69321	19.672	ug/l	95
22) Diisopropyl ether	6.653	45	261888	20.597	ug/l	97
23) Vinyl Acetate	6.589	43	1120284	96.845	ug/l	99
24) 1,1-Dichloroethane	6.553	63	139765	20.048	ug/l	99
25) 2-Butanone	7.465	43	341538	103.210	ug/l	99
26) 2,2-Dichloropropane	7.477	77	119616	19.991	ug/l	99
27) cis-1,2-Dichloroethene	7.471	96	84502	20.059	ug/l	97
28) Bromochloromethane	7.800	49	65173	19.338	ug/l	97
29) Tetrahydrofuran	7.830	42	215847	101.354	ug/l	98
30) Chloroform	7.947	83	146915	20.647	ug/l	99
31) Cyclohexane	8.241	56	118194	20.337	ug/l	98
32) 1,1,1-Trichloroethane	8.147	97	122634	20.336	ug/l	99
36) 1,1-Dichloropropene	8.353	75	98127	19.863	ug/l	98
37) Ethyl Acetate	7.547	43	135691	20.101	ug/l	97
38) Carbon Tetrachloride	8.341	117	101311	20.771	ug/l	92
39) Methylcyclohexane	9.577	83	108801	20.005	ug/l	97
40) Benzene	8.588	78	308160	20.338	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082125\
 Data File : VN087621.D
 Acq On : 21 Aug 2025 13:41
 Operator : JC\MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/22/2025

Supervised By :Mahesh Dadoda 08/25/2025

Quant Time: Aug 22 02:30:54 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M

Quant Title : SW846 8260

QLast Update : Fri Aug 22 02:09:28 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.765	41	69274	19.811	ug/l	98
42) 1,2-Dichloroethane	8.653	62	120495	20.694	ug/l	98
43) Isopropyl Acetate	8.671	43	216674	20.559	ug/l	99
44) Trichloroethene	9.336	130	69650	20.134	ug/l	92
45) 1,2-Dichloropropane	9.600	63	79779	19.975	ug/l	99
46) Dibromomethane	9.688	93	58410	20.551	ug/l	99
47) Bromodichloromethane	9.871	83	118959	20.423	ug/l	90
48) Methyl methacrylate	9.665	41	101988	20.458	ug/l	99
49) 1,4-Dioxane	9.688	88	29266	452.929	ug/l #	99
51) 4-Methyl-2-Pentanone	10.424	43	663566	104.353	ug/l	100
52) Toluene	10.612	92	189511	20.175	ug/l	100
53) t-1,3-Dichloropropene	10.818	75	123317	20.453	ug/l	98
54) cis-1,3-Dichloropropene	10.294	75	127086	20.300	ug/l	97
55) 1,1,2-Trichloroethane	10.994	97	74360	20.463	ug/l	96
56) Ethyl methacrylate	10.859	69	121534	20.928	ug/l	97
57) 1,3-Dichloropropane	11.141	76	132575	20.615	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.141	63	272481	86.689	ug/l	100
59) 2-Hexanone	11.177	43	439624	109.392	ug/l	99
60) Dibromochloromethane	11.335	129	81730	19.411	ug/l	99
61) 1,2-Dibromoethane	11.447	107	79247	20.682	ug/l	97
64) Tetrachloroethene	11.082	164	55688	19.616	ug/l	97
65) Chlorobenzene	11.871	112	205088	20.146	ug/l	92
66) 1,1,1,2-Tetrachloroethane	11.935	131	69837	20.664	ug/l	99
67) Ethyl Benzene	11.941	91	364023	20.653	ug/l	98
68) m/p-Xylenes	12.053	106	265461	41.067	ug/l	99
69) o-Xylene	12.377	106	129958	20.515	ug/l	99
70) Styrene	12.388	104	227729	21.458	ug/l	100
71) Bromoform	12.559	173	52443	20.233	ug/l #	98
73) Isopropylbenzene	12.676	105	335219	20.276	ug/l	99
74) N-amyl acetate	12.547	43	103508m	16.273	ug/l	
75) 1,1,2,2-Tetrachloroethane	12.918	83	116302	20.433	ug/l	99
76) 1,2,3-Trichloropropane	12.976	75	97260m	20.223	ug/l	
77) Bromobenzene	12.959	156	80835	20.737	ug/l	98
78) n-propylbenzene	13.012	91	419422	20.596	ug/l	99
79) 2-Chlorotoluene	13.106	91	255043	20.812	ug/l	99
80) 1,3,5-Trimethylbenzene	13.153	105	286272	20.412	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.718	75	34072	18.801	ug/l	88
82) 4-Chlorotoluene	13.200	91	263916	20.483	ug/l	99
83) tert-Butylbenzene	13.418	119	237268	20.302	ug/l	99
84) 1,2,4-Trimethylbenzene	13.459	105	290636	20.701	ug/l	99
85) sec-Butylbenzene	13.594	105	357394	20.607	ug/l	99
86) p-Isopropyltoluene	13.706	119	290561	20.288	ug/l	99
87) 1,3-Dichlorobenzene	13.712	146	158749	20.680	ug/l	97
88) 1,4-Dichlorobenzene	13.788	146	160067	20.037	ug/l	98
89) n-Butylbenzene	14.035	91	293738	20.652	ug/l	99
90) Hexachloroethane	14.312	117	55637	20.391	ug/l	98
91) 1,2-Dichlorobenzene	14.082	146	151644	20.823	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.694	75	26895	19.891	ug/l	96
93) 1,2,4-Trichlorobenzene	15.365	180	88494	20.236	ug/l	97
94) Hexachlorobutadiene	15.476	225	31498	20.203	ug/l	97
95) Naphthalene	15.612	128	304789	19.677	ug/l	99
96) 1,2,3-Trichlorobenzene	15.812	180	86694	20.278	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082125\
Data File : VN087621.D
Acq On : 21 Aug 2025 13:41
Operator : JC\MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/22/2025
Supervised By :Mahesh Dadoda 08/25/2025

Quant Time: Aug 22 02:30:54 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:09:28 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations

Reviewed By :John Carlone 08/22/2025
Supervised By :Mahesh Dadoda 08/25/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082125\
 Data File : VN087622.D
 Acq On : 21 Aug 2025 14:02
 Operator : JC\MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/22/2025
 Supervised By : Mahesh Dadoda 08/25/2025

Quant Time: Aug 22 02:32:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:09:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.206	168	234546	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	468145	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	438262	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.765	152	220862	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	224856	46.791	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	93.580%		
35) Dibromofluoromethane	8.153	113	156332	46.252	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.500%		
50) Toluene-d8	10.547	98	572648	45.266	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	90.540%		
62) 4-Bromofluorobenzene	12.823	95	210592	46.809	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	93.620%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.142	85	189901	57.007	ug/l	100
3) Chloromethane	2.383	50	186373	54.441	ug/l	100
4) Vinyl Chloride	2.542	62	211909	53.387	ug/l	100
5) Bromomethane	2.971	94	105105	51.318	ug/l	100
6) Chloroethane	3.136	64	141335	50.661	ug/l	100
7) Trichlorofluoromethane	3.512	101	298853	51.389	ug/l	100
8) Diethyl Ether	3.953	74	124074	48.989	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.365	101	162971	49.526	ug/l	100
10) Methyl Iodide	4.577	142	112901	46.467	ug/l	100
11) Tert butyl alcohol	5.518	59	216189	259.137	ug/l	100
12) 1,1-Dichloroethene	4.336	96	164165	48.547	ug/l	100
13) Acrolein	4.171	56	251949	245.655	ug/l	100
14) Allyl chloride	5.006	41	290757	47.448	ug/l	100
15) Acrylonitrile	5.706	53	584519	248.432	ug/l	100
16) Acetone	4.424	43	612147	234.001	ug/l	100
17) Carbon Disulfide	4.700	76	469891	50.475	ug/l	100
18) Methyl Acetate	5.012	43	271363	49.542	ug/l	100
19) Methyl tert-butyl Ether	5.789	73	641051	50.534	ug/l	100
20) Methylene Chloride	5.265	84	193144	51.071	ug/l	100
21) trans-1,2-Dichloroethene	5.777	96	175989	48.020	ug/l	100
22) Diisopropyl ether	6.659	45	673069	50.897	ug/l	100
23) Vinyl Acetate	6.589	43	3156445	262.358	ug/l	100
24) 1,1-Dichloroethane	6.547	63	352088	48.560	ug/l	100
25) 2-Butanone	7.465	43	858584	249.467	ug/l	100
26) 2,2-Dichloropropane	7.477	77	308944	49.645	ug/l	100
27) cis-1,2-Dichloroethene	7.465	96	220668	50.366	ug/l	100
28) Bromochloromethane	7.800	49	164506	46.931	ug/l	100
29) Tetrahydrofuran	7.824	42	555506	250.802	ug/l	100
30) Chloroform	7.947	83	370301	50.036	ug/l	100
31) Cyclohexane	8.236	56	289400	47.878	ug/l	100
32) 1,1,1-Trichloroethane	8.153	97	305533	48.715	ug/l	100
36) 1,1-Dichloropropene	8.353	75	247135	47.660	ug/l	100
37) Ethyl Acetate	7.547	43	347439	49.034	ug/l	100
38) Carbon Tetrachloride	8.341	117	253643	49.543	ug/l	100
39) Methylcyclohexane	9.583	83	276006	48.348	ug/l	100
40) Benzene	8.588	78	788238	49.562	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082125\
 Data File : VN087622.D
 Acq On : 21 Aug 2025 14:02
 Operator : JC\MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By :John Carlone 08/22/2025
 Supervised By :Mahesh Dadoda 08/25/2025

Quant Time: Aug 22 02:32:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:09:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.759	41	182028	49.594	ug/l	100
42) 1,2-Dichloroethane	8.653	62	304323	49.794	ug/l	100
43) Isopropyl Acetate	8.671	43	560899	50.703	ug/l	100
44) Trichloroethene	9.330	130	176493	48.607	ug/l	100
45) 1,2-Dichloropropane	9.600	63	202893	48.398	ug/l	100
46) Dibromomethane	9.688	93	148909	49.915	ug/l	100
47) Bromodichloromethane	9.871	83	298888	48.886	ug/l	100
48) Methyl methacrylate	9.665	41	261445	49.964	ug/l	100
49) 1,4-Dioxane	9.683	88	72780	1073.097	ug/l #	100
51) 4-Methyl-2-Pentanone	10.424	43	1722397	258.056	ug/l	100
52) Toluene	10.606	92	491725	49.872	ug/l	100
53) t-1,3-Dichloropropene	10.818	75	329003	51.987	ug/l	100
54) cis-1,3-Dichloropropene	10.288	75	335961	51.127	ug/l	100
55) 1,1,2-Trichloroethane	10.994	97	190842	50.035	ug/l	100
56) Ethyl methacrylate	10.853	69	330271	54.182	ug/l	100
57) 1,3-Dichloropropane	11.141	76	342170	50.690	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.141	63	829191	251.329	ug/l	100
59) 2-Hexanone	11.177	43	1194193	283.100	ug/l	100
60) Dibromochloromethane	11.335	129	218275	49.389	ug/l	100
61) 1,2-Dibromoethane	11.447	107	202283	50.296	ug/l	100
64) Tetrachloroethene	11.082	164	137281	45.149	ug/l	100
65) Chlorobenzene	11.871	112	530421	48.648	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.941	131	184496	50.970	ug/l	100
67) Ethyl Benzene	11.941	91	939520	49.767	ug/l	100
68) m/p-Xylenes	12.047	106	711871	102.820	ug/l	100
69) o-Xylene	12.376	106	348655	51.386	ug/l	100
70) Styrene	12.388	104	605599	53.277	ug/l	100
71) Bromoform	12.559	173	140399	50.575	ug/l	100
73) Isopropylbenzene	12.671	105	889165	49.821	ug/l	100
74) N-amyl acetate	12.500	43	343435m	50.016	ug/l	
75) 1,1,2,2-Tetrachloroethane	12.918	83	301018	48.990	ug/l	100
76) 1,2,3-Trichloropropane	12.971	75	236877m	45.626	ug/l	
77) Bromobenzene	12.959	156	207789	49.379	ug/l	100
78) n-propylbenzene	13.012	91	1109533	50.472	ug/l	100
79) 2-Chlorotoluene	13.100	91	657210	49.679	ug/l	100
80) 1,3,5-Trimethylbenzene	13.153	105	766722	50.642	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.718	75	99551	50.885	ug/l	100
82) 4-Chlorotoluene	13.200	91	696270	50.059	ug/l	100
83) tert-Butylbenzene	13.418	119	642869	50.955	ug/l	100
84) 1,2,4-Trimethylbenzene	13.459	105	772176	50.948	ug/l	100
85) sec-Butylbenzene	13.594	105	933105	49.838	ug/l	100
86) p-Isopropyltoluene	13.706	119	779657	50.430	ug/l	100
87) 1,3-Dichlorobenzene	13.712	146	403466	48.688	ug/l	100
88) 1,4-Dichlorobenzene	13.788	146	412346	47.815	ug/l	100
89) n-Butylbenzene	14.029	91	764482	49.789	ug/l	100
90) Hexachloroethane	14.312	117	144635	49.105	ug/l	100
91) 1,2-Dichlorobenzene	14.082	146	387612	49.303	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.700	75	70059	47.997	ug/l	100
93) 1,2,4-Trichlorobenzene	15.365	180	226580	47.995	ug/l	100
94) Hexachlorobutadiene	15.476	225	78424	46.597	ug/l	100
95) Naphthalene	15.612	128	829102	49.583	ug/l	100
96) 1,2,3-Trichlorobenzene	15.812	180	222211	48.148	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082125\
Data File : VN087622.D
Acq On : 21 Aug 2025 14:02
Operator : JC\MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/22/2025
Supervised By :Mahesh Dadoda 08/25/2025

Quant Time: Aug 22 02:32:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:09:28 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 08/22/2025
Supervised By :Mahesh Dadoda 08/25/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082125\
 Data File : VN087623.D
 Acq On : 21 Aug 2025 14:23
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/22/2025
 Supervised By : Mahesh Dadoda 08/25/2025

Quant Time: Aug 22 02:33:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:09:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	247255	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	481199	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	447994	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	226864	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	511669	101.001	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 202.000%#			
35) Dibromofluoromethane	8.147	113	358663	103.235	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 206.460%#			
50) Toluene-d8	10.547	98	1354664	104.177	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 208.360%#			
62) 4-Bromofluorobenzene	12.829	95	494494	106.930	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 213.860%#			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.142	85	384461	109.480	ug/l	97
3) Chloromethane	2.383	50	371688	102.992	ug/l	99
4) Vinyl Chloride	2.536	62	418975	100.128	ug/l	97
5) Bromomethane	2.959	94	230328	106.679	ug/l	97
6) Chloroethane	3.124	64	276071	93.871	ug/l	99
7) Trichlorofluoromethane	3.506	101	606894	98.995	ug/l	100
8) Diethyl Ether	3.959	74	252092	94.419	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.365	101	323480	93.250	ug/l	99
10) Methyl Iodide	4.577	142	277692	99.335	ug/l	96
11) Tert butyl alcohol	5.530	59	430345	489.322	ug/l	99
12) 1,1-Dichloroethene	4.330	96	329881	92.539	ug/l	95
13) Acrolein	4.171	56	484654	448.258	ug/l	100
14) Allyl chloride	5.012	41	597335	92.467	ug/l	99
15) Acrylonitrile	5.706	53	1204622	485.671	ug/l	99
16) Acetone	4.424	43	1195609	433.545	ug/l	99
17) Carbon Disulfide	4.700	76	957812	97.598	ug/l	98
18) Methyl Acetate	5.012	43	552409	95.669	ug/l	98
19) Methyl tert-butyl Ether	5.783	73	1306892	97.726	ug/l	98
20) Methylene Chloride	5.259	84	385191	97.752	ug/l	97
21) trans-1,2-Dichloroethene	5.777	96	364392	94.317	ug/l	95
22) Diisopropyl ether	6.659	45	1368059	98.135	ug/l	98
23) Vinyl Acetate	6.588	43	6369332	502.195	ug/l	98
24) 1,1-Dichloroethane	6.553	63	719127	94.084	ug/l	98
25) 2-Butanone	7.471	43	1730205	476.881	ug/l	98
26) 2,2-Dichloropropane	7.471	77	631604	96.278	ug/l	98
27) cis-1,2-Dichloroethene	7.471	96	449130	97.241	ug/l	100
28) Bromochloromethane	7.794	49	369878	100.097	ug/l	99
29) Tetrahydrofuran	7.824	42	1140137	488.295	ug/l	100
30) Chloroform	7.947	83	751186	96.286	ug/l	97
31) Cyclohexane	8.235	56	593829	93.193	ug/l	98
32) 1,1,1-Trichloroethane	8.153	97	625473	94.602	ug/l	98
36) 1,1-Dichloropropene	8.353	75	502476	94.275	ug/l	99
37) Ethyl Acetate	7.547	43	703230	96.555	ug/l	99
38) Carbon Tetrachloride	8.341	117	528406	100.412	ug/l	93
39) Methylcyclohexane	9.582	83	574374	97.883	ug/l	98
40) Benzene	8.588	78	1577387	96.490	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082125\
 Data File : VN087623.D
 Acq On : 21 Aug 2025 14:23
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/22/2025
 Supervised By : Mahesh Dadoda 08/25/2025

Quant Time: Aug 22 02:33:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:09:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.765	41	369145	97.845	ug/l	99
42) 1,2-Dichloroethane	8.653	62	600726	95.625	ug/l	99
43) Isopropyl Acetate	8.671	43	1142494	100.474	ug/l	99
44) Trichloroethene	9.329	130	353426	94.694	ug/l	99
45) 1,2-Dichloropropane	9.600	63	399209	92.644	ug/l	99
46) Dibromomethane	9.688	93	298076	97.207	ug/l	99
47) Bromodichloromethane	9.871	83	613683	97.651	ug/l	95
48) Methyl methacrylate	9.665	41	543991	101.141	ug/l	99
49) 1,4-Dioxane	9.682	88	148991	2137.188	ug/l #	99
51) 4-Methyl-2-Pentanone	10.423	43	3455510	503.673	ug/l	99
52) Toluene	10.612	92	998331	98.507	ug/l	100
53) t-1,3-Dichloropropene	10.818	75	672690	103.410	ug/l	99
54) cis-1,3-Dichloropropene	10.294	75	686982	101.709	ug/l	98
55) 1,1,2-Trichloroethane	10.994	97	385129	98.234	ug/l	96
56) Ethyl methacrylate	10.853	69	688119	109.827	ug/l	99
57) 1,3-Dichloropropane	11.141	76	681646	98.241	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.141	63	1799387	530.602	ug/l	99
59) 2-Hexanone	11.176	43	2445031	563.903	ug/l	100
60) Dibromochloromethane	11.335	129	447781	98.572	ug/l	100
61) 1,2-Dibromoethane	11.447	107	412721	99.836	ug/l	98
64) Tetrachloroethene	11.082	164	282055	90.747	ug/l	95
65) Chlorobenzene	11.870	112	1089532	97.756	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.935	131	373160	100.851	ug/l	99
67) Ethyl Benzene	11.941	91	1928084	99.914	ug/l	99
68) m/p-Xylenes	12.047	106	1448543	204.677	ug/l	100
69) o-Xylene	12.376	106	705872	101.775	ug/l	99
70) Styrene	12.388	104	1242271	106.913	ug/l	99
71) Bromoform	12.559	173	300919	106.042	ug/l #	98
73) Isopropylbenzene	12.670	105	1813905	98.947	ug/l	99
74) N-amyl acetate	12.488	43	692783	98.225	ug/l #	92
75) 1,1,2,2-Tetrachloroethane	12.917	83	599955	95.058	ug/l	100
76) 1,2,3-Trichloropropane	12.970	75	487205m	91.360	ug/l	
77) Bromobenzene	12.959	156	421382	97.488	ug/l	100
78) n-propylbenzene	13.011	91	2252725	99.764	ug/l	100
79) 2-Chlorotoluene	13.100	91	1350026	99.349	ug/l	99
80) 1,3,5-Trimethylbenzene	13.153	105	1541753	99.138	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.717	75	223478	111.207	ug/l	85
82) 4-Chlorotoluene	13.200	91	1392464	97.464	ug/l	98
83) tert-Butylbenzene	13.417	119	1296218	100.022	ug/l	100
84) 1,2,4-Trimethylbenzene	13.459	105	1574771	101.155	ug/l	100
85) sec-Butylbenzene	13.594	105	1888785	98.213	ug/l	100
86) p-Isopropyltoluene	13.706	119	1589055	100.064	ug/l	99
87) 1,3-Dichlorobenzene	13.711	146	819384	96.262	ug/l	98
88) 1,4-Dichlorobenzene	13.788	146	824944	93.128	ug/l	99
89) n-Butylbenzene	14.029	91	1541853	97.761	ug/l	99
90) Hexachloroethane	14.306	117	294494	97.338	ug/l	97
91) 1,2-Dichlorobenzene	14.082	146	781928	96.828	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.700	75	142185	94.833	ug/l	96
93) 1,2,4-Trichlorobenzene	15.364	180	471693	97.272	ug/l	99
94) Hexachlorobutadiene	15.476	225	157139	90.896	ug/l	99
95) Naphthalene	15.611	128	1777069	103.462	ug/l	100
96) 1,2,3-Trichlorobenzene	15.811	180	462335	97.526	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082125\
Data File : VN087623.D
Acq On : 21 Aug 2025 14:23
Operator : JC\MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/22/2025
Supervised By :Mahesh Dadoda 08/25/2025

Quant Time: Aug 22 02:33:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:09:28 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082125\  
Data File : VN087623.D  
Acq On    : 21 Aug 2025  14:23  
Operator  : JC\MD  
Sample    : VSTDICC100  
Misc      : 5.0mL/MSVOA_N/WATER  
ALS Vial  : 11   Sample Multiplier: 1
```

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED

Reviewed By :John Carlone 08/22/2025
Supervised By :Mahesh Dadoda 08/25/2025

Quant Time: Aug 22 02:33:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:09:28 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082125\
 Data File : VN087624.D
 Acq On : 21 Aug 2025 14:44
 Operator : JC\MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC150

Quant Time: Aug 22 02:36:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:09:28 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 08/22/2025
 Supervised By : Mahesh Dadoda 08/25/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	245979	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	485482	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	448278	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.771	152	220585	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.559	65	803343	159.399	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 318.800%#	
35) Dibromofluoromethane	8.147	113	562370	160.440	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 320.880%#	
50) Toluene-d8	10.547	98	2147647	163.702	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 327.400%#	
62) 4-Bromofluorobenzene	12.823	95	785776	168.419	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 336.840%#	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.142	85	576704	165.075	ug/l	97
3) Chloromethane	2.383	50	550985	153.466	ug/l	98
4) Vinyl Chloride	2.536	62	624259	149.961	ug/l	96
5) Bromomethane	2.942	94	360172	167.683	ug/l	96
6) Chloroethane	3.124	64	408766	139.711	ug/l	99
7) Trichlorofluoromethane	3.506	101	909991	149.205	ug/l	94
8) Diethyl Ether	3.959	74	383278	144.299	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.365	101	484239	140.317	ug/l	98
10) Methyl Iodide	4.577	142	432564	151.685	ug/l	97
11) Tert butyl alcohol	5.530	59	641433	733.123	ug/l	98
12) 1,1-Dichloroethene	4.336	96	506728	142.886	ug/l	97
13) Acrolein	4.171	56	861137	800.600	ug/l	99
14) Allyl chloride	5.012	41	1015124	157.956	ug/l	93
15) Acrylonitrile	5.706	53	1818844	737.113	ug/l	99
16) Acetone	4.424	43	1805178	657.978	ug/l	98
17) Carbon Disulfide	4.700	76	1460081	149.550	ug/l	100
18) Methyl Acetate	5.012	43	842183	146.610	ug/l	99
19) Methyl tert-butyl Ether	5.783	73	2023752	152.117	ug/l	97
20) Methylene Chloride	5.259	84	589373	151.029	ug/l	99
21) trans-1,2-Dichloroethene	5.771	96	549970	143.089	ug/l	98
22) Diisopropyl ether	6.659	45	2071435	149.361	ug/l	99
23) Vinyl Acetate	6.589	43	9699326	768.719	ug/l	100
24) 1,1-Dichloroethane	6.553	63	1085255	142.721	ug/l	98
25) 2-Butanone	7.471	43	2610414	723.217	ug/l	99
26) 2,2-Dichloropropane	7.471	77	956513	146.562	ug/l	99
27) cis-1,2-Dichloroethene	7.471	96	667748	145.324	ug/l	98
28) Bromochloromethane	7.794	49	539576	146.779	ug/l	99
29) Tetrahydrofuran	7.824	42	1725889	742.994	ug/l	100
30) Chloroform	7.947	83	1127250	145.239	ug/l	98
31) Cyclohexane	8.241	56	890974	140.551	ug/l	98
32) 1,1,1-Trichloroethane	8.153	97	950731	144.543	ug/l	97
36) 1,1-Dichloropropene	8.353	75	748756	139.242	ug/l	99
37) Ethyl Acetate	7.547	43	1064446	144.861	ug/l	98
38) Carbon Tetrachloride	8.341	117	789246	148.656	ug/l	98
39) Methylcyclohexane	9.583	83	867837	146.589	ug/l	98
40) Benzene	8.588	78	2405823	145.868	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082125\
 Data File : VN087624.D
 Acq On : 21 Aug 2025 14:44
 Operator : JC\MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC150

Manual Integrations
 APPROVED

Reviewed By :John Carlone 08/22/2025
 Supervised By :Mahesh Dadoda 08/25/2025

Quant Time: Aug 22 02:36:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:09:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.765	41	552329	145.108	ug/l	97
42) 1,2-Dichloroethane	8.653	62	912127	143.914	ug/l	99
43) Isopropyl Acetate	8.671	43	1715458	149.531	ug/l	98
44) Trichloroethene	9.335	130	538839	143.098	ug/l	99
45) 1,2-Dichloropropane	9.600	63	600437	138.113	ug/l	100
46) Dibromomethane	9.688	93	454128	146.791	ug/l	99
47) Bromodichloromethane	9.865	83	931257	146.876	ug/l	96
48) Methyl methacrylate	9.665	41	826047	152.226	ug/l	99
49) 1,4-Dioxane	9.683	88	204489	2907.396	ug/l #	99
51) 4-Methyl-2-Pentanone	10.424	43	5127143	740.736	ug/l	99
52) Toluene	10.612	92	1505069	147.197	ug/l	100
53) t-1,3-Dichloropropene	10.818	75	1040899	158.602	ug/l	99
54) cis-1,3-Dichloropropene	10.294	75	1034308	151.780	ug/l	99
55) 1,1,2-Trichloroethane	10.994	97	578828	146.338	ug/l	97
56) Ethyl methacrylate	10.853	69	1058252	167.411	ug/l	99
57) 1,3-Dichloropropane	11.141	76	1028391	146.908	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.141	63	2825105	825.716	ug/l	99
59) 2-Hexanone	11.177	43	3690702	843.686	ug/l	100
60) Dibromochloromethane	11.335	129	680951	148.578	ug/l	100
61) 1,2-Dibromoethane	11.447	107	621238	148.950	ug/l	98
64) Tetrachloroethene	11.082	164	421013	135.369	ug/l	96
65) Chlorobenzene	11.871	112	1629434	146.105	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.941	131	562761	151.997	ug/l	99
67) Ethyl Benzene	11.941	91	2928525	151.661	ug/l	98
68) m/p-Xylenes	12.047	106	2189611	309.193	ug/l	99
69) o-Xylene	12.376	106	1053696	151.829	ug/l	99
70) Styrene	12.388	104	1873074	161.099	ug/l	99
71) Bromoform	12.559	173	451851	159.129	ug/l #	98
73) Isopropylbenzene	12.671	105	2726116	152.940	ug/l	100
74) N-amyl acetate	12.482	43	1146975	167.250	ug/l #	90
75) 1,1,2,2-Tetrachloroethane	12.918	83	907151	147.822	ug/l	100
76) 1,2,3-Trichloropropane	12.971	75	830843m	160.233	ug/l	
77) Bromobenzene	12.959	156	620711	147.691	ug/l	98
78) n-propylbenzene	13.012	91	3351021	152.627	ug/l	100
79) 2-Chlorotoluene	13.100	91	2010067	152.133	ug/l	99
80) 1,3,5-Trimethylbenzene	13.153	105	2284564	151.084	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.712	75	345953	177.054	ug/l	84
82) 4-Chlorotoluene	13.200	91	2047130	147.365	ug/l	98
83) tert-Butylbenzene	13.418	119	1908253	151.441	ug/l	99
84) 1,2,4-Trimethylbenzene	13.459	105	2300224	151.960	ug/l	100
85) sec-Butylbenzene	13.594	105	2775113	148.408	ug/l	100
86) p-Isopropyltoluene	13.706	119	2324153	150.520	ug/l	99
87) 1,3-Dichlorobenzene	13.712	146	1208472	146.014	ug/l	98
88) 1,4-Dichlorobenzene	13.788	146	1218359	141.456	ug/l	99
89) n-Butylbenzene	14.035	91	2273344	148.244	ug/l	99
90) Hexachloroethane	14.312	117	442763	150.511	ug/l	97
91) 1,2-Dichlorobenzene	14.082	146	1153653	146.927	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.694	75	218577	149.934	ug/l	97
93) 1,2,4-Trichlorobenzene	15.365	180	714001	151.432	ug/l	99
94) Hexachlorobutadiene	15.476	225	238563	141.923	ug/l	99
95) Naphthalene	15.612	128	2705715	162.013	ug/l	100
96) 1,2,3-Trichlorobenzene	15.812	180	709644	153.956	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082125\
Data File : VN087624.D
Acq On : 21 Aug 2025 14:44
Operator : JC\MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/22/2025
Supervised By :Mahesh Dadoda 08/25/2025

Quant Time: Aug 22 02:36:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:09:28 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082125\  
Data File : VN087624.D  
Acq On    : 21 Aug 2025   14:44  
Operator  : JC\MD  
Sample    : VSTDICC150  
Misc      : 5.0mL/MSVOA_N/WATER  
ALS Vial  : 12   Sample Multiplier: 1
```

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED

Reviewed By :John Carlone 08/22/2025
Supervised By :Mahesh Dadoda 08/25/2025

Quant Time: Aug 22 02:36:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:09:28 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082125\
 Data File : VN087626.D
 Acq On : 21 Aug 2025 15:47
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN082125

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/22/2025
 Supervised By : Mahesh Dadoda 08/25/2025

Quant Time: Aug 22 02:59:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:54:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	236538	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	462145	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	426678	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	215492	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	211874	43.718	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	87.440%		
35) Dibromofluoromethane	8.153	113	144206	43.218	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	86.440%		
50) Toluene-d8	10.547	98	538430	43.114	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	86.220%		
62) 4-Bromofluorobenzene	12.829	95	197985	44.578	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	89.160%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.142	85	166005	49.414	ug/l	98
3) Chloromethane	2.383	50	164410	47.621	ug/l	99
4) Vinyl Chloride	2.536	62	181981	45.461	ug/l	97
5) Bromomethane	2.971	94	104036	50.369	ug/l	90
6) Chloroethane	3.136	64	119563	42.496	ug/l	99
7) Trichlorofluoromethane	3.512	101	258055	44.000	ug/l	99
8) Diethyl Ether	3.959	74	106479	41.688	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.365	101	138202	41.645	ug/l	99
10) Methyl Iodide	4.577	142	101477	42.154	ug/l	96
11) Tert butyl alcohol	5.518	59	188337	223.850	ug/l	99
12) 1,1-Dichloroethene	4.342	96	143630	42.117	ug/l	94
13) Acrolein	4.177	56	214366	207.251	ug/l	99
14) Allyl chloride	5.012	41	251818	40.748	ug/l	98
15) Acrylonitrile	5.712	53	512416	215.953	ug/l	98
16) Acetone	4.424	43	534765	202.699	ug/l	100
17) Carbon Disulfide	4.700	76	403340	42.961	ug/l	99
18) Methyl Acetate	5.012	43	239686	43.391	ug/l	99
19) Methyl tert-butyl Ether	5.783	73	554938	43.377	ug/l	95
20) Methylene Chloride	5.259	84	171230	44.741	ug/l	99
21) trans-1,2-Dichloroethene	5.771	96	155031	41.945	ug/l	95
22) Diisopropyl ether	6.659	45	570242	42.758	ug/l	98
23) Vinyl Acetate	6.588	43	2684129	221.221	ug/l	100
24) 1,1-Dichloroethane	6.553	63	308870	42.240	ug/l	97
25) 2-Butanone	7.471	43	748996	215.792	ug/l	98
26) 2,2-Dichloropropane	7.471	77	266418	42.451	ug/l	98
27) cis-1,2-Dichloroethene	7.477	96	187583	42.454	ug/l	98
28) Bromochloromethane	7.794	49	162438	45.951	ug/l	98
29) Tetrahydrofuran	7.824	42	468943	209.937	ug/l	99
30) Chloroform	7.947	83	321565	43.085	ug/l	99
31) Cyclohexane	8.241	56	255575	41.926	ug/l	99
32) 1,1,1-Trichloroethane	8.153	97	268667	42.477	ug/l	98
36) 1,1-Dichloropropene	8.353	75	213398	41.688	ug/l	99
37) Ethyl Acetate	7.547	43	286649	40.980	ug/l	97
38) Carbon Tetrachloride	8.341	117	220428	43.614	ug/l	96
39) Methylcyclohexane	9.582	83	234965	41.693	ug/l	99
40) Benzene	8.588	78	677138	43.129	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082125\
 Data File : VN087626.D
 Acq On : 21 Aug 2025 15:47
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN082125

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/22/2025
 Supervised By : Mahesh Dadoda 08/25/2025

Quant Time: Aug 22 02:59:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:54:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.765	41	154175	42.550	ug/l	97
42) 1,2-Dichloroethane	8.653	62	260576	43.189	ug/l	98
43) Isopropyl Acetate	8.671	43	475910	43.578	ug/l	98
44) Trichloroethene	9.329	130	150412	41.962	ug/l	98
45) 1,2-Dichloropropane	9.600	63	170592	41.221	ug/l	97
46) Dibromomethane	9.688	93	127636	43.340	ug/l	99
47) Bromodichloromethane	9.871	83	260671	43.189	ug/l	95
48) Methyl methacrylate	9.665	41	224953	43.548	ug/l	99
49) 1,4-Dioxane	9.682	88	62435	932.518	ug/l #	98
51) 4-Methyl-2-Pentanone	10.423	43	1463763	222.154	ug/l	100
52) Toluene	10.612	92	421282	43.282	ug/l	99
53) t-1,3-Dichloropropene	10.818	75	278049	44.506	ug/l	95
54) cis-1,3-Dichloropropene	10.294	75	284560	43.867	ug/l	96
55) 1,1,2-Trichloroethane	10.994	97	165061	43.838	ug/l	99
56) Ethyl methacrylate	10.853	69	283473	47.109	ug/l	98
57) 1,3-Dichloropropane	11.141	76	287664	43.169	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.141	63	842734	258.751	ug/l	100
59) 2-Hexanone	11.176	43	1001252	240.442	ug/l	100
60) Dibromochloromethane	11.335	129	187315	42.934	ug/l	98
61) 1,2-Dibromoethane	11.447	107	173643	43.736	ug/l	99
64) Tetrachloroethene	11.082	164	117611	39.730	ug/l	94
65) Chlorobenzene	11.870	112	449967	42.389	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.941	131	156161	44.313	ug/l	100
67) Ethyl Benzene	11.941	91	803704	43.729	ug/l	100
68) m/p-Xylenes	12.047	106	605035	89.762	ug/l	100
69) o-Xylene	12.376	106	289138	43.771	ug/l	96
70) Styrene	12.388	104	506439	45.763	ug/l	99
71) Bromoform	12.559	173	121336	44.894	ug/l #	98
73) Isopropylbenzene	12.670	105	761746	43.745	ug/l	98
74) N-amyl acetate	12.500	43	294763m	44.668	ug/l	
75) 1,1,2,2-Tetrachloroethane	12.917	83	264712	44.155	ug/l	99
76) 1,2,3-Trichloropropane	12.970	75	246217m	48.723	ug/l	
77) Bromobenzene	12.959	156	177807	43.307	ug/l	98
78) n-propylbenzene	13.012	91	946175	44.113	ug/l	99
79) 2-Chlorotoluene	13.100	91	559855	43.374	ug/l	99
80) 1,3,5-Trimethylbenzene	13.153	105	645684	43.710	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.717	75	85496	44.790	ug/l	86
82) 4-Chlorotoluene	13.200	91	598781	44.123	ug/l	99
83) tert-Butylbenzene	13.417	119	547166	44.450	ug/l	98
84) 1,2,4-Trimethylbenzene	13.459	105	666050	45.041	ug/l	98
85) sec-Butylbenzene	13.594	105	802373	43.924	ug/l	99
86) p-Isopropyltoluene	13.706	119	658963	43.685	ug/l	99
87) 1,3-Dichlorobenzene	13.711	146	346038	42.798	ug/l	99
88) 1,4-Dichlorobenzene	13.788	146	359327	42.705	ug/l	99
89) n-Butylbenzene	14.035	91	654734	43.704	ug/l	100
90) Hexachloroethane	14.311	117	119987	41.752	ug/l	96
91) 1,2-Dichlorobenzene	14.082	146	329254	42.924	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.694	75	60532	42.504	ug/l	98
93) 1,2,4-Trichlorobenzene	15.364	180	197383	42.852	ug/l	99
94) Hexachlorobutadiene	15.476	225	67769	41.269	ug/l	99
95) Naphthalene	15.611	128	706917	43.329	ug/l	99
96) 1,2,3-Trichlorobenzene	15.811	180	191126	42.444	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082125\
Data File : VN087626.D
Acq On : 21 Aug 2025 15:47
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN082125

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/22/2025
Supervised By :Mahesh Dadoda 08/25/2025

Quant Time: Aug 22 02:59:20 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:54:13 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082125\  
Data File : VN087626.D  
Acq On    : 21 Aug 2025   15:47  
Operator  : JC\MD  
Sample    : VSTDICV050  
Misc      : 5.0mL/MSVOA_N/WATER  
ALS Vial  : 15   Sample Multiplier: 1
```

Instrument :
MSVOA_N
ClientSampleId :
ICVVN082125

Manual Integrations APPROVED

Reviewed By :John Carlone 08/22/2025
Supervised By :Mahesh Dadoda 08/25/2025

Quant Time: Aug 22 02:59:20 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:54:13 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082125\
 Data File : VN087626.D
 Acq On : 21 Aug 2025 15:47
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN082125

Quant Time: Aug 22 02:59:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:54:13 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	0.710	0.702	1.1	87	0.00
3 P	Chloromethane	0.730	0.695	4.8	88	0.00
4 C	Vinyl Chloride	0.846	0.769	9.1#	86	0.00
5 T	Bromomethane	0.437	0.440	-0.7	99	0.00
6 T	Chloroethane	0.595	0.505	15.1	85	0.00
7 T	Trichlorofluoromethane	1.240	1.091	12.0	86	0.00
8 T	Diethyl Ether	0.540	0.450	16.7	86	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.701	0.584	16.7	85	0.00
10 T	Methyl Iodide	0.450	0.429	4.7	90	0.00
11 T	Tert butyl alcohol	0.178	0.159	10.7	87	0.00
12 CM	1,1-Dichloroethene	0.721	0.607	15.8#	87	0.00
13 T	Acrolein	0.219	0.181	17.4	85	0.00
14 T	Allyl chloride	1.306	1.065	18.5	87	0.00
15 T	Acrylonitrile	0.502	0.433	13.7	88	0.00
16 T	Acetone	0.558	0.452	19.0	87	0.00
17 T	Carbon Disulfide	1.985	1.705	14.1	86	0.00
18 T	Methyl Acetate	1.168	1.013	13.3	88	0.00
19 T	Methyl tert-butyl Ether	2.704	2.346	13.2	87	0.00
20 T	Methylene Chloride	0.948	0.724	23.6	89	0.00
21 T	trans-1,2-Dichloroethene	0.781	0.655	16.1	88	0.00
22 T	Diisopropyl ether	2.819	2.411	14.5	85	0.00
23 T	Vinyl Acetate	2.565	2.270	11.5	85	0.00
24 P	1,1-Dichloroethane	1.546	1.306	15.5	88	0.00
25 T	2-Butanone	0.734	0.633	13.8	87	0.00
26 T	2,2-Dichloropropane	1.327	1.126	15.1	86	0.00
27 T	cis-1,2-Dichloroethene	0.934	0.793	15.1	85	0.01
28 T	Bromochloromethane	0.747	0.687	8.0	99	0.00
29 T	Tetrahydrofuran	0.472	0.397	15.9	84	0.00
30 C	Chloroform	1.578	1.359	13.9#	87	0.00
31 T	Cyclohexane	1.289	1.080	16.2	88	0.00
32 T	1,1,1-Trichloroethane	1.337	1.136	15.0	88	0.00
33 S	1,2-Dichloroethane-d4	1.024	0.896	12.5	94	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	99	0.00
35 S	Dibromofluoromethane	0.361	0.312	13.6	92	0.00
36 T	1,1-Dichloropropene	0.554	0.462	16.6	86	0.00
37 T	Ethyl Acetate	0.757	0.620	18.1	83	0.00
38 T	Carbon Tetrachloride	0.547	0.477	12.8	87	0.00
39 T	Methylcyclohexane	0.610	0.508	16.7	85	0.00
40 TM	Benzene	1.699	1.465	13.8	86	0.00
41 T	Methacrylonitrile	0.392	0.334	14.8	85	0.00
42 TM	1,2-Dichloroethane	0.653	0.564	13.6	86	0.00
43 T	Isopropyl Acetate	1.182	1.030	12.9	85	0.00
44 TM	Trichloroethene	0.388	0.325	16.2	85	0.00
45 C	1,2-Dichloropropane	0.448	0.369	17.6#	84	0.00
46 T	Dibromomethane	0.319	0.276	13.5	86	0.00
47 T	Bromodichloromethane	0.653	0.564	13.6	87	0.00
48 T	Methyl methacrylate	0.559	0.487	12.9	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082125\
 Data File : VN087626.D
 Acq On : 21 Aug 2025 15:47
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN082125

Quant Time: Aug 22 02:59:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:54:13 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.007	0.0	86	0.00
50 S	Toluene-d8	1.351	1.165	13.8	94	0.00
51 T	4-Methyl-2-Pentanone	0.713	0.633	11.2	85	0.00
52 CM	Toluene	1.053	0.912	13.4#	86	0.00
53 T	t-1,3-Dichloropropene	0.676	0.602	10.9	85	0.00
54 T	cis-1,3-Dichloropropene	0.702	0.616	12.3	85	0.00
55 T	1,1,2-Trichloroethane	0.407	0.357	12.3	86	0.00
56 T	Ethyl methacrylate	0.651	0.613	5.8	86	0.00
57 T	1,3-Dichloropropane	0.721	0.622	13.7	84	0.00
58 T	2-Chloroethyl Vinyl ether	0.352	0.365	-3.7	102	0.00
59 T	2-Hexanone	0.451	0.433	4.0	84	0.00
60 T	Dibromochloromethane	0.472	0.405	14.2	86	0.00
61 T	1,2-Dibromoethane	0.430	0.376	12.6	86	0.00
62 S	4-Bromofluorobenzene	0.481	0.428	11.0	94	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.00
64 T	Tetrachloroethene	0.347	0.276	20.5	86	0.00
65 PM	Chlorobenzene	1.244	1.055	15.2	85	0.00
66 T	1,1,1,2-Tetrachloroethane	0.413	0.366	11.4	85	0.00
67 C	Ethyl Benzene	2.154	1.884	12.5#	86	0.00
68 T	m/p-Xylenes	0.790	0.709	10.3	85	0.00
69 T	o-Xylene	0.774	0.678	12.4	83	0.00
70 T	Styrene	1.297	1.187	8.5	84	0.00
71 P	Bromoform	0.317	0.284	10.4	86	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
73 T	Isopropylbenzene	4.040	3.535	12.5	86	0.00
74 T	N-amyl acetate	1.531	1.368	10.6	86	0.00
75 P	1,1,2,2-Tetrachloroethane	1.391	1.228	11.7	88	0.00
76 T	1,2,3-Trichloropropane	1.173	1.143	2.6	104	0.00
77 T	Bromobenzene	0.953	0.825	13.4	86	0.00
78 T	n-propylbenzene	4.977	4.391	11.8	85	0.00
79 T	2-Chlorotoluene	2.995	2.598	13.3	85	0.00
80 T	1,3,5-Trimethylbenzene	3.428	2.996	12.6	84	0.00
81 T	trans-1,4-Dichloro-2-butene	0.443	0.397	10.4	86	0.00
82 T	4-Chlorotoluene	3.149	2.779	11.7	86	0.00
83 T	tert-Butylbenzene	2.856	2.539	11.1	85	0.00
84 T	1,2,4-Trimethylbenzene	3.431	3.091	9.9	86	0.00
85 T	sec-Butylbenzene	4.239	3.723	12.2	86	0.00
86 T	p-Isopropyltoluene	3.500	3.058	12.6	85	0.00
87 T	1,3-Dichlorobenzene	1.876	1.606	14.4	86	0.00
88 T	1,4-Dichlorobenzene	1.952	1.667	14.6	87	0.00
89 T	n-Butylbenzene	3.476	3.038	12.6	86	0.00
90 T	Hexachloroethane	0.667	0.557	16.5	83	0.00
91 T	1,2-Dichlorobenzene	1.780	1.528	14.2	85	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.330	0.281	14.8	86	0.00
93 T	1,2,4-Trichlorobenzene	1.069	0.916	14.3	87	0.00
94 T	Hexachlorobutadiene	0.381	0.314	17.6	86	0.00
95 T	Naphthalene	3.786	3.280	13.4	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082125\
Data File : VN087626.D
Acq On : 21 Aug 2025 15:47
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN082125

Quant Time: Aug 22 02:59:20 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:54:13 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.045	0.887	15.1	86	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082125\
 Data File : VN087626.D
 Acq On : 21 Aug 2025 15:47
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN082125

Quant Time: Aug 22 02:59:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:54:13 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	50.000	49.414	1.2	87	0.00
3 P	Chloromethane	50.000	47.621	4.8	88	0.00
4 C	Vinyl Chloride	50.000	45.461	9.1#	86	0.00
5 T	Bromomethane	50.000	50.369	-0.7	99	0.00
6 T	Chloroethane	50.000	42.496	15.0	85	0.00
7 T	Trichlorofluoromethane	50.000	44.000	12.0	86	0.00
8 T	Diethyl Ether	50.000	41.688	16.6	86	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	41.645	16.7	85	0.00
10 T	Methyl Iodide	50.000	42.154	15.7	90	0.00
11 T	Tert butyl alcohol	250.000	223.850	10.5	87	0.00
12 CM	1,1-Dichloroethene	50.000	42.117	15.8#	87	0.00
13 T	Acrolein	250.000	207.251	17.1	85	0.00
14 T	Allyl chloride	50.000	40.748	18.5	87	0.00
15 T	Acrylonitrile	250.000	215.953	13.6	88	0.00
16 T	Acetone	250.000	202.699	18.9	87	0.00
17 T	Carbon Disulfide	50.000	42.961	14.1	86	0.00
18 T	Methyl Acetate	50.000	43.391	13.2	88	0.00
19 T	Methyl tert-butyl Ether	50.000	43.377	13.2	87	0.00
20 T	Methylene Chloride	50.000	44.741	10.5	89	0.00
21 T	trans-1,2-Dichloroethene	50.000	41.945	16.1	88	0.00
22 T	Diisopropyl ether	50.000	42.758	14.5	85	0.00
23 T	Vinyl Acetate	250.000	221.221	11.5	85	0.00
24 P	1,1-Dichloroethane	50.000	42.240	15.5	88	0.00
25 T	2-Butanone	250.000	215.792	13.7	87	0.00
26 T	2,2-Dichloropropane	50.000	42.451	15.1	86	0.00
27 T	cis-1,2-Dichloroethene	50.000	42.454	15.1	85	0.01
28 T	Bromochloromethane	50.000	45.951	8.1	99	0.00
29 T	Tetrahydrofuran	250.000	209.937	16.0	84	0.00
30 C	Chloroform	50.000	43.085	13.8#	87	0.00
31 T	Cyclohexane	50.000	41.926	16.1	88	0.00
32 T	1,1,1-Trichloroethane	50.000	42.477	15.0	88	0.00
33 S	1,2-Dichloroethane-d4	50.000	43.718	12.6	94	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	99	0.00
35 S	Dibromofluoromethane	50.000	43.218	13.6	92	0.00
36 T	1,1-Dichloropropene	50.000	41.688	16.6	86	0.00
37 T	Ethyl Acetate	50.000	40.980	18.0	83	0.00
38 T	Carbon Tetrachloride	50.000	43.614	12.8	87	0.00
39 T	Methylcyclohexane	50.000	41.693	16.6	85	0.00
40 TM	Benzene	50.000	43.129	13.7	86	0.00
41 T	Methacrylonitrile	50.000	42.550	14.9	85	0.00
42 TM	1,2-Dichloroethane	50.000	43.189	13.6	86	0.00
43 T	Isopropyl Acetate	50.000	43.578	12.8	85	0.00
44 TM	Trichloroethene	50.000	41.962	16.1	85	0.00
45 C	1,2-Dichloropropane	50.000	41.221	17.6#	84	0.00
46 T	Dibromomethane	50.000	43.340	13.3	86	0.00
47 T	Bromodichloromethane	50.000	43.189	13.6	87	0.00
48 T	Methyl methacrylate	50.000	43.548	12.9	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082125\
 Data File : VN087626.D
 Acq On : 21 Aug 2025 15:47
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN082125

Quant Time: Aug 22 02:59:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:54:13 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	932.518	6.7	86	0.00
50 S	Toluene-d8	50.000	43.114	13.8	94	0.00
51 T	4-Methyl-2-Pentanone	250.000	222.154	11.1	85	0.00
52 CM	Toluene	50.000	43.282	13.4#	86	0.00
53 T	t-1,3-Dichloropropene	50.000	44.506	11.0	85	0.00
54 T	cis-1,3-Dichloropropene	50.000	43.867	12.3	85	0.00
55 T	1,1,2-Trichloroethane	50.000	43.838	12.3	86	0.00
56 T	Ethyl methacrylate	50.000	47.109	5.8	86	0.00
57 T	1,3-Dichloropropane	50.000	43.169	13.7	84	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	258.751	-3.5	102	0.00
59 T	2-Hexanone	250.000	240.442	3.8	84	0.00
60 T	Dibromochloromethane	50.000	42.934	14.1	86	0.00
61 T	1,2-Dibromoethane	50.000	43.736	12.5	86	0.00
62 S	4-Bromofluorobenzene	50.000	44.578	10.8	94	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	97	0.00
64 T	Tetrachloroethene	50.000	39.730	20.5	86	0.00
65 PM	Chlorobenzene	50.000	42.389	15.2	85	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	44.313	11.4	85	0.00
67 C	Ethyl Benzene	50.000	43.729	12.5#	86	0.00
68 T	m/p-Xylenes	100.000	89.762	10.2	85	0.00
69 T	o-Xylene	50.000	43.771	12.5	83	0.00
70 T	Styrene	50.000	45.763	8.5	84	0.00
71 P	Bromoform	50.000	44.894	10.2	86	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	98	0.00
73 T	Isopropylbenzene	50.000	43.745	12.5	86	0.00
74 T	N-amyl acetate	50.000	44.668	10.7	86	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	44.155	11.7	88	0.00
76 T	1,2,3-Trichloropropane	50.000	48.723	2.6	104	0.00
77 T	Bromobenzene	50.000	43.307	13.4	86	0.00
78 T	n-propylbenzene	50.000	44.113	11.8	85	0.00
79 T	2-Chlorotoluene	50.000	43.374	13.3	85	0.00
80 T	1,3,5-Trimethylbenzene	50.000	43.710	12.6	84	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	44.790	10.4	86	0.00
82 T	4-Chlorotoluene	50.000	44.123	11.8	86	0.00
83 T	tert-Butylbenzene	50.000	44.450	11.1	85	0.00
84 T	1,2,4-Trimethylbenzene	50.000	45.041	9.9	86	0.00
85 T	sec-Butylbenzene	50.000	43.924	12.2	86	0.00
86 T	p-Isopropyltoluene	50.000	43.685	12.6	85	0.00
87 T	1,3-Dichlorobenzene	50.000	42.798	14.4	86	0.00
88 T	1,4-Dichlorobenzene	50.000	42.705	14.6	87	0.00
89 T	n-Butylbenzene	50.000	43.704	12.6	86	0.00
90 T	Hexachloroethane	50.000	41.752	16.5	83	0.00
91 T	1,2-Dichlorobenzene	50.000	42.924	14.2	85	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	42.504	15.0	86	0.00
93 T	1,2,4-Trichlorobenzene	50.000	42.852	14.3	87	0.00
94 T	Hexachlorobutadiene	50.000	41.269	17.5	86	0.00
95 T	Naphthalene	50.000	43.329	13.3	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082125\
Data File : VN087626.D
Acq On : 21 Aug 2025 15:47
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN082125

Quant Time: Aug 22 02:59:20 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:54:13 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	42.444	15.1	86	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>ZUCC01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3032</u>
Instrument ID:	<u>MSVOA_N</u>	Calibration Date/Time:	<u>09/08/2025</u> <u>08:58</u>
Lab File ID:	<u>VN087749.D</u>	Init. Calib. Date(s):	<u>08/21/2025</u> <u>08/21/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>12:09</u> <u>14:44</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.846	0.817		-3.43	20
1,1-Dichloroethene	0.721	0.665		-7.77	20
2-Butanone	0.734	0.601		-18.12	20
Carbon Tetrachloride	0.547	0.558		2.01	20
Chloroform	1.578	1.393		-11.72	20
Benzene	1.699	1.562		-8.06	20
1,2-Dichloroethane	0.653	0.573		-12.25	20
Trichloroethene	0.388	0.360		-7.22	20
Tetrachloroethene	0.347	0.323		-6.92	20
Chlorobenzene	1.244	1.164	0.3	-6.43	20
1,2-Dichloroethane-d4	1.024	0.897		-12.4	20
Dibromofluoromethane	0.361	0.354		-1.94	20
Toluene-d8	1.351	1.337		-1.04	20
4-Bromofluorobenzene	0.481	0.470		-2.29	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090825\
 Data File : VN087749.D
 Acq On : 08 Sep 2025 08:58
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/09/2025
 Supervised By :Mahesh Dadoda 09/09/2025

Quant Time: Sep 09 02:23:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:55:42 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	264779	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.077	114	492400	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	447402	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	232401	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.559	65	237572	43.792	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.580%
35) Dibromofluoromethane	8.147	113	174258	49.016	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.040%
50) Toluene-d8	10.547	98	658372	49.479	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.960%
62) 4-Bromofluorobenzene	12.829	95	231565	48.935	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.880%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.142	85	193839	51.545	ug/l	99
3) Chloromethane	2.383	50	173873	44.990	ug/l	100
4) Vinyl Chloride	2.542	62	216226	48.254	ug/l	98
5) Bromomethane	2.977	94	116326	50.312	ug/l	98
6) Chloroethane	3.136	64	137302	43.596	ug/l	98
7) Trichlorofluoromethane	3.512	101	335019	51.030	ug/l	97
8) Diethyl Ether	3.959	74	120880	42.278	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.365	101	187772	50.547	ug/l	97
10) Methyl Iodide	4.583	142	133351	48.301	ug/l	96
11) Tert butyl alcohol	5.518	59	183597	194.942	ug/l	97
12) 1,1-Dichloroethene	4.342	96	176167	46.148	ug/l	96
13) Acrolein	4.171	56	259964	224.528	ug/l	99
14) Allyl chloride	5.006	41	279641	40.423	ug/l	99
15) Acrylonitrile	5.706	53	547710	206.207	ug/l	99
16) Acetone	4.418	43	619041	209.617	ug/l	96
17) Carbon Disulfide	4.706	76	440714	41.935	ug/l	98
18) Methyl Acetate	5.012	43	289633	46.840	ug/l	98
19) Methyl tert-butyl Ether	5.783	73	612958	42.802	ug/l	97
20) Methylene Chloride	5.259	84	185071	43.156	ug/l	98
21) trans-1,2-Dichloroethene	5.771	96	175170	42.339	ug/l	99
22) Diisopropyl ether	6.653	45	646592	43.312	ug/l	98
23) Vinyl Acetate	6.589	43	2892767	212.987	ug/l	99
24) 1,1-Dichloroethane	6.553	63	355491	43.431	ug/l	98
25) 2-Butanone	7.465	43	795768	204.814	ug/l	98
26) 2,2-Dichloropropane	7.471	77	333020	47.404	ug/l	99
27) cis-1,2-Dichloroethene	7.471	96	212784	43.021	ug/l	97
28) Bromochloromethane	7.794	49	179324	45.317	ug/l	99
29) Tetrahydrofuran	7.824	42	474593	189.805	ug/l	97
30) Chloroform	7.947	83	368853	44.150	ug/l	100
31) Cyclohexane	8.236	56	314346	46.067	ug/l	95
32) 1,1,1-Trichloroethane	8.147	97	316535	44.707	ug/l	99
36) 1,1-Dichloropropene	8.353	75	253170	46.419	ug/l	98
37) Ethyl Acetate	7.541	43	303404	40.710	ug/l	96
38) Carbon Tetrachloride	8.341	117	274734	51.020	ug/l	92
39) Methylcyclohexane	9.583	83	311292	51.843	ug/l	98
40) Benzene	8.588	78	769311	45.989	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090825\
 Data File : VN087749.D
 Acq On : 08 Sep 2025 08:58
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/09/2025
 Supervised By :Mahesh Dadoda 09/09/2025

Quant Time: Sep 09 02:23:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:55:42 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.765	41	168250	43.582	ug/l	99
42) 1,2-Dichloroethane	8.653	62	282348	43.923	ug/l	100
43) Isopropyl Acetate	8.671	43	516340	44.375	ug/l	99
44) Trichloroethene	9.330	130	177123	46.377	ug/l	95
45) 1,2-Dichloropropane	9.600	63	196650	44.598	ug/l	100
46) Dibromomethane	9.688	93	140094	44.647	ug/l	99
47) Bromodichloromethane	9.865	83	303159	47.142	ug/l	95
48) Methyl methacrylate	9.659	41	242139	43.995	ug/l	98
49) 1,4-Dioxane	9.682	88	64163	899.444	ug/l #	95
51) 4-Methyl-2-Pentanone	10.424	43	1577008	224.635	ug/l	100
52) Toluene	10.606	92	480446	46.328	ug/l	100
53) t-1,3-Dichloropropene	10.818	75	315264	47.362	ug/l	98
54) cis-1,3-Dichloropropene	10.294	75	325135	47.042	ug/l	96
55) 1,1,2-Trichloroethane	10.994	97	186467	46.480	ug/l	97
56) Ethyl methacrylate	10.853	69	310723	48.465	ug/l	98
57) 1,3-Dichloropropane	11.141	76	323072	45.503	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.141	63	797294	229.757	ug/l	100
59) 2-Hexanone	11.177	43	1096766	247.196	ug/l	99
60) Dibromochloromethane	11.335	129	214894	46.229	ug/l	99
61) 1,2-Dibromoethane	11.447	107	190435	45.018	ug/l	100
64) Tetrachloroethene	11.082	164	144466	46.542	ug/l	97
65) Chlorobenzene	11.871	112	520802	46.790	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.941	131	180431	48.828	ug/l	99
67) Ethyl Benzene	11.941	91	955458	49.578	ug/l	98
68) m/p-Xylenes	12.047	106	701710	99.282	ug/l	100
69) o-Xylene	12.376	106	339075	48.953	ug/l	100
70) Styrene	12.388	104	581549	50.116	ug/l	98
71) Bromoform	12.553	173	144267	50.906	ug/l #	96
73) Isopropylbenzene	12.671	105	915828	48.767	ug/l	99
74) N-amyl acetate	12.500	43	323885m	45.510	ug/l	
75) 1,1,2,2-Tetrachloroethane	12.918	83	290676	44.958	ug/l	99
76) 1,2,3-Trichloropropane	12.971	75	267426m	49.070	ug/l	
77) Bromobenzene	12.959	156	202161	45.656	ug/l	98
78) n-propylbenzene	13.012	91	1149579	49.697	ug/l	100
79) 2-Chlorotoluene	13.106	91	658088	47.275	ug/l	100
80) 1,3,5-Trimethylbenzene	13.147	105	770431	48.360	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.718	75	102397	49.741	ug/l	83
82) 4-Chlorotoluene	13.200	91	685064	46.808	ug/l	100
83) tert-Butylbenzene	13.412	119	676989	50.995	ug/l	98
84) 1,2,4-Trimethylbenzene	13.459	105	772911	48.465	ug/l	99
85) sec-Butylbenzene	13.594	105	1015392	51.541	ug/l	99
86) p-Isopropyltoluene	13.706	119	832888	51.198	ug/l	100
87) 1,3-Dichlorobenzene	13.712	146	405313	46.482	ug/l	98
88) 1,4-Dichlorobenzene	13.788	146	405975	44.739	ug/l	99
89) n-Butylbenzene	14.035	91	843275	52.194	ug/l	100
90) Hexachloroethane	14.312	117	162454	52.416	ug/l	98
91) 1,2-Dichlorobenzene	14.082	146	373721	45.176	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.694	75	61788	40.229	ug/l	98
93) 1,2,4-Trichlorobenzene	15.364	180	231388	46.580	ug/l	99
94) Hexachlorobutadiene	15.476	225	90786	51.263	ug/l	98
95) Naphthalene	15.612	128	784176	44.567	ug/l	100
96) 1,2,3-Trichlorobenzene	15.817	180	219378	45.174	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090825\
Data File : VN087749.D
Acq On : 08 Sep 2025 08:58
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 09/09/2025
Supervised By :Mahesh Dadoda 09/09/2025

Quant Time: Sep 09 02:23:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:55:42 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090825\
Data File : VN087749.D
Acq On    : 08 Sep 2025  08:58
Operator  : JC\MD
Sample    : VSTDCCC050
Misc      : 5.0mL/MSVOA_N/WATER
ALS Vial  : 1      Sample Multiplier: 1
```

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations

Reviewed By :Semsettin Yesilyurt 09/09/2025
Supervised By :Mahesh Dadoda 09/09/2025

Quant Time: Sep 09 02:23:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:55:42 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090825\
 Data File : VN087749.D
 Acq On : 08 Sep 2025 08:58
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 09 02:23:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:55:42 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	113	0.00
2 T	Dichlorodifluoromethane	0.710	0.732	-3.1	102	0.00
3 P	Chloromethane	0.730	0.657	10.0	93	0.00
4 C	Vinyl Chloride	0.846	0.817	3.4#	102	0.00
5 T	Bromomethane	0.437	0.439	-0.5	111	0.00
6 T	Chloroethane	0.595	0.519	12.8	97	0.00
7 T	Trichlorofluoromethane	1.240	1.265	-2.0	112	0.00
8 T	Diethyl Ether	0.540	0.457	15.4	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.701	0.709	-1.1	115	0.00
10 T	Methyl Iodide	0.450	0.504	-12.0	118	0.00
11 T	Tert butyl alcohol	0.178	0.139	21.9	85	0.00
12 CM	1,1-Dichloroethene	0.721	0.665	7.8#	107	0.00
13 T	Acrolein	0.219	0.196	10.5	103	0.00
14 T	Allyl chloride	1.306	1.056	19.1	96	0.00
15 T	Acrylonitrile	0.502	0.414	17.5	94	0.00
16 T	Acetone	0.558	0.468	16.1	101	0.00
17 T	Carbon Disulfide	1.985	1.664	16.2	94	0.00
18 T	Methyl Acetate	1.168	1.094	6.3	107	0.00
19 T	Methyl tert-butyl Ether	2.704	2.315	14.4	96	0.00
20 T	Methylene Chloride	0.948	0.699	26.3#	96	0.00
21 T	trans-1,2-Dichloroethene	0.781	0.662	15.2	100	0.00
22 T	Diisopropyl ether	2.819	2.442	13.4	96	0.00
23 T	Vinyl Acetate	2.565	2.185	14.8	92	0.00
24 P	1,1-Dichloroethane	1.546	1.343	13.1	101	0.00
25 T	2-Butanone	0.734	0.601	18.1	93	0.00
26 T	2,2-Dichloropropane	1.327	1.258	5.2	108	0.00
27 T	cis-1,2-Dichloroethene	0.934	0.804	13.9	96	0.00
28 T	Bromochloromethane	0.747	0.677	9.4	109	0.00
29 T	Tetrahydrofuran	0.472	0.358	24.2	85	0.00
30 C	Chloroform	1.578	1.393	11.7#	100	0.00
31 T	Cyclohexane	1.289	1.187	7.9	109	0.00
32 T	1,1,1-Trichloroethane	1.337	1.195	10.6	104	0.00
33 S	1,2-Dichloroethane-d4	1.024	0.897	12.4	106	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	105	0.00
35 S	Dibromofluoromethane	0.361	0.354	1.9	111	0.00
36 T	1,1-Dichloropropene	0.554	0.514	7.2	102	0.00
37 T	Ethyl Acetate	0.757	0.616	18.6	87	0.00
38 T	Carbon Tetrachloride	0.547	0.558	-2.0	108	0.00
39 T	Methylcyclohexane	0.610	0.632	-3.6	113	0.00
40 TM	Benzene	1.699	1.562	8.1	98	0.00
41 T	Methacrylonitrile	0.392	0.342	12.8	92	0.00
42 TM	1,2-Dichloroethane	0.653	0.573	12.3	93	0.00
43 T	Isopropyl Acetate	1.182	1.049	11.3	92	0.00
44 TM	Trichloroethene	0.388	0.360	7.2	100	0.00
45 C	1,2-Dichloropropane	0.448	0.399	10.9#	97	0.00
46 T	Dibromomethane	0.319	0.285	10.7	94	0.00
47 T	Bromodichloromethane	0.653	0.616	5.7	101	0.00
48 T	Methyl methacrylate	0.559	0.492	12.0	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090825\
 Data File : VN087749.D
 Acq On : 08 Sep 2025 08:58
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 09 02:23:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:55:42 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.007	0.0	88	0.00
50 S	Toluene-d8	1.351	1.337	1.0	115	0.00
51 T	4-Methyl-2-Pentanone	0.713	0.641	10.1	92	0.00
52 CM	Toluene	1.053	0.976	7.3#	98	0.00
53 T	t-1,3-Dichloropropene	0.676	0.640	5.3	96	0.00
54 T	cis-1,3-Dichloropropene	0.702	0.660	6.0	97	0.00
55 T	1,1,2-Trichloroethane	0.407	0.379	6.9	98	0.00
56 T	Ethyl methacrylate	0.651	0.631	3.1	94	0.00
57 T	1,3-Dichloropropane	0.721	0.656	9.0	94	0.00
58 T	2-Chloroethyl Vinyl ether	0.352	0.324	8.0	96	0.00
59 T	2-Hexanone	0.451	0.445	1.3	92	0.00
60 T	Dibromochloromethane	0.472	0.436	7.6	98	0.00
61 T	1,2-Dibromoethane	0.430	0.387	10.0	94	0.00
62 S	4-Bromofluorobenzene	0.481	0.470	2.3	110	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	102	0.00
64 T	Tetrachloroethene	0.347	0.323	6.9	105	0.00
65 PM	Chlorobenzene	1.244	1.164	6.4	98	0.00
66 T	1,1,1,2-Tetrachloroethane	0.413	0.403	2.4	98	0.00
67 C	Ethyl Benzene	2.154	2.136	0.8#	102	0.00
68 T	m/p-Xylenes	0.790	0.784	0.8	99	0.00
69 T	o-Xylene	0.774	0.758	2.1	97	0.00
70 T	Styrene	1.297	1.300	-0.2	96	0.00
71 P	Bromoform	0.317	0.322	-1.6	103	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
73 T	Isopropylbenzene	4.040	3.941	2.5	103	0.00
74 T	N-amyl acetate	1.531	1.394	8.9	94	0.00
75 P	1,1,2,2-Tetrachloroethane	1.391	1.251	10.1	97	0.00
76 T	1,2,3-Trichloropropane	1.173	1.151	1.9	113	0.00
77 T	Bromobenzene	0.953	0.870	8.7	97	0.00
78 T	n-propylbenzene	4.977	4.947	0.6	104	0.00
79 T	2-Chlorotoluene	2.995	2.832	5.4	100	0.00
80 T	1,3,5-Trimethylbenzene	3.428	3.315	3.3	100	0.00
81 T	trans-1,4-Dichloro-2-butene	0.443	0.441	0.5	103	0.00
82 T	4-Chlorotoluene	3.149	2.948	6.4	98	0.00
83 T	tert-Butylbenzene	2.856	2.913	-2.0	105	0.00
84 T	1,2,4-Trimethylbenzene	3.431	3.326	3.1	100	0.00
85 T	sec-Butylbenzene	4.239	4.369	-3.1	109	0.00
86 T	p-Isopropyltoluene	3.500	3.584	-2.4	107	0.00
87 T	1,3-Dichlorobenzene	1.876	1.744	7.0	100	0.00
88 T	1,4-Dichlorobenzene	1.952	1.747	10.5	98	0.00
89 T	n-Butylbenzene	3.476	3.629	-4.4	110	0.00
90 T	Hexachloroethane	0.667	0.699	-4.8	112	0.00
91 T	1,2-Dichlorobenzene	1.780	1.608	9.7	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.330	0.266	19.4	88	0.00
93 T	1,2,4-Trichlorobenzene	1.069	0.996	6.8	102	0.00
94 T	Hexachlorobutadiene	0.381	0.391	-2.6	116	0.00
95 T	Naphthalene	3.786	3.374	10.9	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090825\
Data File : VN087749.D
Acq On : 08 Sep 2025 08:58
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Sep 09 02:23:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:55:42 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.045	0.944	9.7	99	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090825\
 Data File : VN087749.D
 Acq On : 08 Sep 2025 08:58
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 09 02:23:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:55:42 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	113	0.00
2 T	Dichlorodifluoromethane	50.000	51.545	-3.1	102	0.00
3 P	Chloromethane	50.000	44.990	10.0	93	0.00
4 C	Vinyl Chloride	50.000	48.254	3.5#	102	0.00
5 T	Bromomethane	50.000	50.312	-0.6	111	0.00
6 T	Chloroethane	50.000	43.596	12.8	97	0.00
7 T	Trichlorofluoromethane	50.000	51.030	-2.1	112	0.00
8 T	Diethyl Ether	50.000	42.278	15.4	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	50.547	-1.1	115	0.00
10 T	Methyl Iodide	50.000	48.301	3.4	118	0.00
11 T	Tert butyl alcohol	250.000	194.942	22.0	85	0.00
12 CM	1,1-Dichloroethene	50.000	46.148	7.7#	107	0.00
13 T	Acrolein	250.000	224.528	10.2	103	0.00
14 T	Allyl chloride	50.000	40.423	19.2	96	0.00
15 T	Acrylonitrile	250.000	206.207	17.5	94	0.00
16 T	Acetone	250.000	209.617	16.2	101	0.00
17 T	Carbon Disulfide	50.000	41.935	16.1	94	0.00
18 T	Methyl Acetate	50.000	46.840	6.3	107	0.00
19 T	Methyl tert-butyl Ether	50.000	42.802	14.4	96	0.00
20 T	Methylene Chloride	50.000	43.156	13.7	96	0.00
21 T	trans-1,2-Dichloroethene	50.000	42.339	15.3	100	0.00
22 T	Diisopropyl ether	50.000	43.312	13.4	96	0.00
23 T	Vinyl Acetate	250.000	212.987	14.8	92	0.00
24 P	1,1-Dichloroethane	50.000	43.431	13.1	101	0.00
25 T	2-Butanone	250.000	204.814	18.1	93	0.00
26 T	2,2-Dichloropropane	50.000	47.404	5.2	108	0.00
27 T	cis-1,2-Dichloroethene	50.000	43.021	14.0	96	0.00
28 T	Bromochloromethane	50.000	45.317	9.4	109	0.00
29 T	Tetrahydrofuran	250.000	189.805	24.1	85	0.00
30 C	Chloroform	50.000	44.150	11.7#	100	0.00
31 T	Cyclohexane	50.000	46.067	7.9	109	0.00
32 T	1,1,1-Trichloroethane	50.000	44.707	10.6	104	0.00
33 S	1,2-Dichloroethane-d4	50.000	43.792	12.4	106	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	105	0.00
35 S	Dibromofluoromethane	50.000	49.016	2.0	111	0.00
36 T	1,1-Dichloropropene	50.000	46.419	7.2	102	0.00
37 T	Ethyl Acetate	50.000	40.710	18.6	87	0.00
38 T	Carbon Tetrachloride	50.000	51.020	-2.0	108	0.00
39 T	Methylcyclohexane	50.000	51.843	-3.7	113	0.00
40 TM	Benzene	50.000	45.989	8.0	98	0.00
41 T	Methacrylonitrile	50.000	43.582	12.8	92	0.00
42 TM	1,2-Dichloroethane	50.000	43.923	12.2	93	0.00
43 T	Isopropyl Acetate	50.000	44.375	11.3	92	0.00
44 TM	Trichloroethene	50.000	46.377	7.2	100	0.00
45 C	1,2-Dichloropropane	50.000	44.598	10.8#	97	0.00
46 T	Dibromomethane	50.000	44.647	10.7	94	0.00
47 T	Bromodichloromethane	50.000	47.142	5.7	101	0.00
48 T	Methyl methacrylate	50.000	43.995	12.0	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090825\
 Data File : VN087749.D
 Acq On : 08 Sep 2025 08:58
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 09 02:23:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:55:42 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	899.444	10.1	88	0.00
50 S	Toluene-d8	50.000	49.479	1.0	115	0.00
51 T	4-Methyl-2-Pentanone	250.000	224.635	10.1	92	0.00
52 CM	Toluene	50.000	46.328	7.3#	98	0.00
53 T	t-1,3-Dichloropropene	50.000	47.362	5.3	96	0.00
54 T	cis-1,3-Dichloropropene	50.000	47.042	5.9	97	0.00
55 T	1,1,2-Trichloroethane	50.000	46.480	7.0	98	0.00
56 T	Ethyl methacrylate	50.000	48.465	3.1	94	0.00
57 T	1,3-Dichloropropane	50.000	45.503	9.0	94	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	229.757	8.1	96	0.00
59 T	2-Hexanone	250.000	247.196	1.1	92	0.00
60 T	Dibromochloromethane	50.000	46.229	7.5	98	0.00
61 T	1,2-Dibromoethane	50.000	45.018	10.0	94	0.00
62 S	4-Bromofluorobenzene	50.000	48.935	2.1	110	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	102	0.00
64 T	Tetrachloroethene	50.000	46.542	6.9	105	0.00
65 PM	Chlorobenzene	50.000	46.790	6.4	98	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	48.828	2.3	98	0.00
67 C	Ethyl Benzene	50.000	49.578	0.8#	102	0.00
68 T	m/p-Xylenes	100.000	99.282	0.7	99	0.00
69 T	o-Xylene	50.000	48.953	2.1	97	0.00
70 T	Styrene	50.000	50.116	-0.2	96	0.00
71 P	Bromoform	50.000	50.906	-1.8	103	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	105	0.00
73 T	Isopropylbenzene	50.000	48.767	2.5	103	0.00
74 T	N-amyl acetate	50.000	45.510	9.0	94	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	44.958	10.1	97	0.00
76 T	1,2,3-Trichloropropane	50.000	49.070	1.9	113	0.00
77 T	Bromobenzene	50.000	45.656	8.7	97	0.00
78 T	n-propylbenzene	50.000	49.697	0.6	104	0.00
79 T	2-Chlorotoluene	50.000	47.275	5.5	100	0.00
80 T	1,3,5-Trimethylbenzene	50.000	48.360	3.3	100	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.741	0.5	103	0.00
82 T	4-Chlorotoluene	50.000	46.808	6.4	98	0.00
83 T	tert-Butylbenzene	50.000	50.995	-2.0	105	0.00
84 T	1,2,4-Trimethylbenzene	50.000	48.465	3.1	100	0.00
85 T	sec-Butylbenzene	50.000	51.541	-3.1	109	0.00
86 T	p-Isopropyltoluene	50.000	51.198	-2.4	107	0.00
87 T	1,3-Dichlorobenzene	50.000	46.482	7.0	100	0.00
88 T	1,4-Dichlorobenzene	50.000	44.739	10.5	98	0.00
89 T	n-Butylbenzene	50.000	52.194	-4.4	110	0.00
90 T	Hexachloroethane	50.000	52.416	-4.8	112	0.00
91 T	1,2-Dichlorobenzene	50.000	45.176	9.6	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	40.229	19.5	88	0.00
93 T	1,2,4-Trichlorobenzene	50.000	46.580	6.8	102	0.00
94 T	Hexachlorobutadiene	50.000	51.263	-2.5	116	0.00
95 T	Naphthalene	50.000	44.567	10.9	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090825\
Data File : VN087749.D
Acq On : 08 Sep 2025 08:58
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Sep 09 02:23:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:55:42 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	45.174	9.7	99	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6





QC SAMPLE DATA

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082125\
Data File : VN087614.D
Acq On : 21 Aug 2025 09:30
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Title : SW846 8260
Last Update : Fri Aug 22 02:55:42 2025



AutoFind: Scans 1848, 1849, 1850; Background Corrected with Scan 1840

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	25.8	13089	PASS
75	95	30	60	58.1	29435	PASS
95	95	100	100	100.0	50640	PASS
96	95	5	9	7.3	3712	PASS
173	174	0.00	2	1.6	548	PASS
174	95	50	100	69.0	34952	PASS
175	174	5	9	8.1	2829	PASS
176	174	95	101	98.1	34291	PASS
177	176	5	9	7.1	2426	PASS

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
35.85	675	49.00	2529	59.95	672	69.85	591
37.00	3688	50.00	13089	61.00	3695	71.85	411
38.00	3178	51.00	3882	61.95	3294	72.95	297
38.95	1555	51.90	219	63.00	2681	73.95	1069
39.85	60	52.10	64	63.85	251	75.00	29435
43.95	374	54.65	128	64.10	153	76.00	2714
44.85	476	55.00	70	64.70	121	77.10	352
45.10	237	55.85	685	64.95	324	77.80	133
46.20	55	56.10	397	66.85	193	78.85	2691
46.80	616	56.90	2340	68.00	6681	79.80	662
47.95	526	57.95	176	69.05	6065	80.85	2901

Average of 12.823 to 12.835 min.: VN087614.D\data.ms

BFB

Modified:subtracted



m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
81.80	684	103.60	98	127.85	257	171.40	114
82.90	109	103.85	252	129.70	113	171.90	297
85.60	55	104.95	136	129.95	176	172.10	131
86.90	1959	105.65	343	130.90	79	172.85	548
87.90	1858	106.10	71	134.80	197	173.90	34952
90.90	194	112.70	85	136.75	152	174.95	2829
91.95	1755	114.90	88	140.85	705	175.90	34291
93.00	2573	115.80	249	142.85	649	176.95	2426
93.95	6967	116.85	391	148.10	51	281.10	79
95.00	50640	117.70	342	154.70	64		
96.00	3712	118.95	421	170.20	52		

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090825\
Data File : VN087748.D
Acq On : 08 Sep 2025 07:42
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Title : SW846 8260
Last Update : Fri Aug 22 02:55:42 2025



AutoFind: Scans 1848, 1849, 1850; Background Corrected with Scan 1838

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.6	12076	PASS
75	95	30	60	59.4	30448	PASS
95	95	100	100	100.0	51237	PASS
96	95	5	9	6.4	3271	PASS
173	174	0.00	2	1.2	436	PASS
174	95	50	100	71.2	36485	PASS
175	174	5	9	7.1	2597	PASS
176	174	95	101	94.9	34635	PASS
177	176	5	9	7.8	2708	PASS

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
35.70	108	45.10	205	54.85	271	68.00	648
36.05	526	46.80	279	55.95	1206	69.00	654
36.95	3799	47.05	583	56.90	1964	69.95	63
38.00	3302	47.90	297	60.00	765	71.85	40
39.00	1377	48.10	114	60.95	3189	73.00	2808
40.10	260	48.95	3042	62.00	3134	74.00	10519
40.70	69	50.00	12076	62.90	2719	75.00	30448
42.10	98	51.05	3834	63.90	484	75.90	1593
43.00	123	51.90	118	64.95	399	76.10	898
43.85	347	52.10	72	66.00	77	76.70	92
44.90	361	53.00	63	66.75	136	77.00	214

Average of 12.823 to 12.835 min.: VN087748.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
77.75	346	92.95	2509	116.90	480	141.70	57
78.90	2745	93.95	6984	117.85	253	142.80	574
79.90	802	95.00	51237	118.80	311	143.00	209
80.85	2727	95.95	3271	127.80	219	144.80	191
81.85	639	96.90	57	129.00	103	148.00	224
85.90	207	103.00	52	129.80	50	148.90	63
86.90	2424	103.90	334	131.10	53	155.00	63
87.80	845	105.80	384	134.90	108	160.90	52
87.95	1356	110.90	70	136.90	57	171.90	755
90.85	380	112.90	63	140.50	118	172.70	187
91.90	2367	115.80	290	140.95	770	172.95	436

Average of 12.823 to 12.835 min.: VN087748.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
173.95	36485						
174.95	2597						
175.90	34635						
176.90	2708						
192.80	54						

Report of Analysis

Client:	Zuccaro Inc.	Date Collected:	
Project:	3 Coal Ave., Morristown, NJ	Date Received:	
Client Sample ID:	VN0908WBL01	SDG No.:	Q3032
Lab Sample ID:	VN0908WBL01	Matrix:	TCLP
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	TCLP VOA
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087751.D	1	09/08/25 09:59	VN090825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	0.26	U	0.26	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	5.00	ug/L
71-43-2	Benzene	0.15	U	0.15	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	5.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	5.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	5.00	ug/L
108-90-7	Chlorobenzene	0.12	U	0.12	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	52.3		70 (74) - 130 (125)	105%	SPK: 50
1868-53-7	Dibromofluoromethane	48.8		70 (75) - 130 (124)	98%	SPK: 50
2037-26-5	Toluene-d8	48.3		70 (86) - 130 (113)	97%	SPK: 50
460-00-4	4-Bromofluorobenzene	46.0		70 (77) - 130 (121)	92%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	246000	8.206			
540-36-3	1,4-Difluorobenzene	567000	9.082			
3114-55-4	Chlorobenzene-d5	531000	11.841			
3855-82-1	1,4-Dichlorobenzene-d4	244000	13.77			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of OC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090825\
 Data File : VN087751.D
 Acq On : 08 Sep 2025 09:59
 Operator : JC\MD
 Sample : VN0908WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0908WBL01

Quant Time: Sep 09 02:24:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:55:42 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	246345	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	567426	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	531276	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	244144	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	263783	52.262	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.520%
35) Dibromofluoromethane	8.147	113	199858	48.784	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.560%
50) Toluene-d8	10.547	98	741361	48.349	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.700%
62) 4-Bromofluorobenzene	12.829	95	250915	46.013	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	92.020%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090825\
Data File : VN087751.D
Acq On : 08 Sep 2025 09:59
Operator : JC\MD
Sample : VN0908WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0908WBL01

Quant Time: Sep 09 02:24:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:55:42 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.206 min Scan# 1063
Delta R.T. 0.000 min
Lab File: VN087751.D
Acq: 08 Sep 2025 09:59




Instrument :
MSVOA_N
ClientSampleId :
VN0908WBL01


Tgt Ion:168 Resp: 246345
Ion Ratio Lower Upper
168 100
99 65.2 57.2 85.8



#33
1,2-Dichloroethane-d4
Concen: 52.262 ug/l
RT: 8.559 min Scan# 1123
Delta R.T. -0.000 min
Lab File: VN087751.D
Acq: 08 Sep 2025 09:59

Tgt Ion: 65 Resp: 263783
Ion Ratio Lower Upper
65 100
67 50.9 0.0 100.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.082 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN087751.D

Acq: 08 Sep 2025 09:59

Instrument :

MSVOA_N

ClientSampleId :

VN0908WBL01





Tgt Ion:114 Resp: 567426

Ion Ratio Lower Upper

114 100

63 22.7 0.0 45.2

88 15.5 0.0 33.4



#35

Dibromofluoromethane

Concen: 48.784 ug/l

RT: 8.147 min Scan# 1053

Delta R.T. -0.006 min

Lab File: VN087751.D

Acq: 08 Sep 2025 09:59




Tgt Ion:113 Resp: 199858


Ion Ratio Lower Upper

113 100

111 106.0 82.8 124.2

192 16.2 12.8 19.2









#50
Toluene-d8
Concen: 48.349 ug/l
RT: 10.547 min Scan# 1461
Delta R.T. -0.000 min
Lab File: VN087751.D
Acq: 08 Sep 2025 09:59




Instrument :
MSVOA_N
ClientSampleId :
VN0908WBL01


Tgt Ion: 98 Resp: 741361
Ion Ratio Lower Upper
98 100
100 63.9 52.8 79.2



#62
4-Bromofluorobenzene
Concen: 46.013 ug/l
RT: 12.829 min Scan# 1849
Delta R.T. 0.006 min
Lab File: VN087751.D
Acq: 08 Sep 2025 09:59

Tgt Ion: 95 Resp: 250915
Ion Ratio Lower Upper
95 100
174 70.0 0.0 138.4
176 65.5 0.0 132.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.841 min Scan# 117
Delta R.T. -0.000 min
Lab File: VN087751.D
Acq: 08 Sep 2025 09:59




Instrument :
MSVOA_N
ClientSampleId :
VN0908WBL01

Tgt Ion	Ratio	Lower	Upper
117	100		
82	60.1	47.4	71.2
119	31.3	24.3	36.5



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.770 min Scan# 2009
Delta R.T. 0.006 min
Lab File: VN087751.D
Acq: 08 Sep 2025 09:59

Tgt Ion	Ratio	Lower	Upper
152	100		
115	63.7	32.3	96.8
150	157.5	0.0	351.0



Report of Analysis

Client:	Zuccaro Inc.	Date Collected:	
Project:	3 Coal Ave., Morristown, NJ	Date Received:	
Client Sample ID:	VN0908WBS01	SDG No.:	Q3032
Lab Sample ID:	VN0908WBS01	Matrix:	TCLP
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	TCLP VOA
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087752.D	1	09/08/25 10:59	VN090825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	19.6		0.26	5.00	ug/L
75-35-4	1,1-Dichloroethene	18.7		0.23	5.00	ug/L
78-93-3	2-Butanone	84.1		0.98	25.0	ug/L
56-23-5	Carbon Tetrachloride	20.1		0.25	5.00	ug/L
67-66-3	Chloroform	18.0		0.25	5.00	ug/L
71-43-2	Benzene	18.0		0.15	5.00	ug/L
107-06-2	1,2-Dichloroethane	17.7		0.22	5.00	ug/L
79-01-6	Trichloroethene	18.4		0.090	5.00	ug/L
127-18-4	Tetrachloroethene	18.4		0.23	5.00	ug/L
108-90-7	Chlorobenzene	18.5		0.12	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	41.3		70 (74) - 130 (125)	83%	SPK: 50
1868-53-7	Dibromofluoromethane	45.2		70 (75) - 130 (124)	90%	SPK: 50
2037-26-5	Toluene-d8	45.1		70 (86) - 130 (113)	90%	SPK: 50
460-00-4	4-Bromofluorobenzene	44.1		70 (77) - 130 (121)	88%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	259000	8.2			
540-36-3	1,4-Difluorobenzene	485000	9.082			
3114-55-4	Chlorobenzene-d5	441000	11.841			
3855-82-1	1,4-Dichlorobenzene-d4	222000	13.77			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of OC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090825\
 Data File : VN087752.D
 Acq On : 08 Sep 2025 10:59
 Operator : JC\MD
 Sample : VN0908WBS01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0908WBS01

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/09/2025
 Supervised By :Mahesh Dadoda 09/09/2025

Quant Time: Sep 09 02:25:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:55:42 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.200	168	258577	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	485156	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	441331	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	222091	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	218750	41.290	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	82.580%		
35) Dibromofluoromethane	8.153	113	158477	45.243	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	90.480%		
50) Toluene-d8	10.541	98	590898	45.071	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	90.140%		
62) 4-Bromofluorobenzene	12.829	95	205499	44.075	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	88.160%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.142	85	73365	19.977	ug/l	100
3) Chloromethane	2.383	50	69513	18.418	ug/l	97
4) Vinyl Chloride	2.536	62	85838	19.616	ug/l	95
5) Bromomethane	2.977	94	45484	20.144	ug/l	99
6) Chloroethane	3.136	64	55192	17.945	ug/l	91
7) Trichlorofluoromethane	3.512	101	136167	21.239	ug/l	98
8) Diethyl Ether	3.965	74	47810	17.123	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.365	101	72668	20.031	ug/l	97
10) Methyl Iodide	4.577	142	41396	19.999	ug/l	94
11) Tert butyl alcohol	5.524	59	79929	86.904	ug/l	99
12) 1,1-Dichloroethene	4.336	96	69614	18.673	ug/l	93
13) Acrolein	4.171	56	93977	83.114	ug/l	100
14) Allyl chloride	5.018	41	113894	16.859	ug/l	97
15) Acrylonitrile	5.700	53	217682	83.921	ug/l	98
16) Acetone	4.418	43	238653	82.750	ug/l	100
17) Carbon Disulfide	4.700	76	174549	17.007	ug/l #	93
18) Methyl Acetate	5.006	43	119785	19.837	ug/l	98
19) Methyl tert-butyl Ether	5.783	73	235653	16.850	ug/l	95
20) Methylene Chloride	5.265	84	74389	17.014	ug/l	96
21) trans-1,2-Dichloroethene	5.777	96	65787	16.282	ug/l	96
22) Diisopropyl ether	6.659	45	257486	17.661	ug/l	95
23) Vinyl Acetate	6.583	43	1118229	84.307	ug/l	99
24) 1,1-Dichloroethane	6.553	63	139411	17.441	ug/l	98
25) 2-Butanone	7.465	43	319032	84.082	ug/l	99
26) 2,2-Dichloropropane	7.465	77	127888	18.641	ug/l	98
27) cis-1,2-Dichloroethene	7.465	96	80740	16.716	ug/l	95
28) Bromochloromethane	7.794	49	69833	18.071	ug/l	100
29) Tetrahydrofuran	7.824	42	210586	86.240	ug/l	97
30) Chloroform	7.953	83	147251	18.048	ug/l	98
31) Cyclohexane	8.236	56	127627	19.152	ug/l	95
32) 1,1,1-Trichloroethane	8.147	97	132533	19.168	ug/l	92
36) 1,1-Dichloropropene	8.347	75	102079	18.996	ug/l	98
37) Ethyl Acetate	7.541	43	124877	17.006	ug/l	99
38) Carbon Tetrachloride	8.347	117	106764	20.123	ug/l	95
39) Methylcyclohexane	9.582	83	118230	19.984	ug/l	95
40) Benzene	8.583	78	297160	18.029	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090825\
 Data File : VN087752.D
 Acq On : 08 Sep 2025 10:59
 Operator : JC\MD
 Sample : VN0908WBS01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0908WBS01

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/09/2025
 Supervised By :Mahesh Dadoda 09/09/2025

Quant Time: Sep 09 02:25:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:55:42 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.759	41	67081	17.635	ug/l	98
42) 1,2-Dichloroethane	8.653	62	112275	17.726	ug/l	100
43) Isopropyl Acetate	8.671	43	210401	18.352	ug/l	99
44) Trichloroethene	9.330	130	69127	18.370	ug/l	93
45) 1,2-Dichloropropane	9.600	63	78756	18.128	ug/l	100
46) Dibromomethane	9.682	93	56820	18.379	ug/l	99
47) Bromodichloromethane	9.865	83	118102	18.639	ug/l	98
48) Methyl methacrylate	9.659	41	94316	17.393	ug/l	99
49) 1,4-Dioxane	9.682	88	27406	389.917	ug/l #	94
51) 4-Methyl-2-Pentanone	10.424	43	646677	93.490	ug/l	100
52) Toluene	10.606	92	185975	18.201	ug/l	100
53) t-1,3-Dichloropropene	10.818	75	118986	18.142	ug/l	96
54) cis-1,3-Dichloropropene	10.294	75	125482	18.426	ug/l	93
55) 1,1,2-Trichloroethane	10.994	97	72632	18.375	ug/l	92
56) Ethyl methacrylate	10.859	69	114157	18.071	ug/l	97
57) 1,3-Dichloropropane	11.141	76	126458	18.077	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.141	63	282264	82.555	ug/l	99
59) 2-Hexanone	11.176	43	423656	96.912	ug/l	100
60) Dibromochloromethane	11.335	129	82036	17.912	ug/l	97
61) 1,2-Dibromoethane	11.447	107	75577	18.133	ug/l	99
64) Tetrachloroethene	11.082	164	56389	18.416	ug/l	94
65) Chlorobenzene	11.871	112	203637	18.547	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.941	131	71951	19.739	ug/l	97
67) Ethyl Benzene	11.941	91	358884	18.878	ug/l	97
68) m/p-Xylenes	12.053	106	268431	38.502	ug/l	100
69) o-Xylene	12.376	106	127085	18.600	ug/l	99
70) Styrene	12.388	104	220351	19.250	ug/l	98
71) Bromoform	12.559	173	53814	19.250	ug/l #	96
73) Isopropylbenzene	12.671	105	345651	19.260	ug/l	99
74) N-amyl acetate	12.518	43	133540m	19.635	ug/l	
75) 1,1,2,2-Tetrachloroethane	12.918	83	116185	18.804	ug/l	99
76) 1,2,3-Trichloropropane	12.970	75	108953m	20.920	ug/l	
77) Bromobenzene	12.959	156	78290	18.502	ug/l	98
78) n-propylbenzene	13.012	91	436855	19.762	ug/l	99
79) 2-Chlorotoluene	13.106	91	246272	18.513	ug/l	98
80) 1,3,5-Trimethylbenzene	13.153	105	285968	18.784	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.718	75	36306	18.455	ug/l	89
82) 4-Chlorotoluene	13.200	91	254916	18.226	ug/l	95
83) tert-Butylbenzene	13.412	119	253599	19.989	ug/l	98
84) 1,2,4-Trimethylbenzene	13.459	105	289694	19.008	ug/l	99
85) sec-Butylbenzene	13.594	105	383799	20.386	ug/l	100
86) p-Isopropyltoluene	13.706	119	312580	20.106	ug/l	99
87) 1,3-Dichlorobenzene	13.712	146	151885	18.227	ug/l	97
88) 1,4-Dichlorobenzene	13.788	146	155616	17.945	ug/l	99
89) n-Butylbenzene	14.029	91	318818	20.649	ug/l	98
90) Hexachloroethane	14.306	117	62932	21.248	ug/l	98
91) 1,2-Dichlorobenzene	14.082	146	146989	18.593	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	14.694	75	24273	16.537	ug/l	91
93) 1,2,4-Trichlorobenzene	15.364	180	87432	18.418	ug/l	98
94) Hexachlorobutadiene	15.476	225	37129	21.939	ug/l	98
95) Naphthalene	15.611	128	286160	17.018	ug/l	99
96) 1,2,3-Trichlorobenzene	15.806	180	83735	18.043	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090825\
Data File : VN087752.D
Acq On : 08 Sep 2025 10:59
Operator : JC\MD
Sample : VN0908WBS01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0908WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 09/09/2025
Supervised By :Mahesh Dadoda 09/09/2025

Quant Time: Sep 09 02:25:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:55:42 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090825\
Data File : VN087752.D
Acq On : 08 Sep 2025 10:59
Operator : JC\MD
Sample : VN0908WBS01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0908WBS01

Manual Integrations

Reviewed By :Semsettin Yesilyurt 09/09/2025
Supervised By :Mahesh Dadoda 09/09/2025

Quant Time: Sep 09 02:25:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:55:42 2025
Response via : Initial Calibration



Report of Analysis

Client:	Zuccaro Inc.	Date Collected:	
Project:	3 Coal Ave., Morristown, NJ	Date Received:	
Client Sample ID:	VN0908WBSD01	SDG No.:	Q3032
Lab Sample ID:	VN0908WBSD01	Matrix:	TCLP
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	TCLP VOA
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087753.D	1	09/08/25 11:20	VN090825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	19.1		0.26	5.00	ug/L
75-35-4	1,1-Dichloroethene	19.0		0.23	5.00	ug/L
78-93-3	2-Butanone	83.3		0.98	25.0	ug/L
56-23-5	Carbon Tetrachloride	21.0		0.25	5.00	ug/L
67-66-3	Chloroform	17.8		0.25	5.00	ug/L
71-43-2	Benzene	18.6		0.15	5.00	ug/L
107-06-2	1,2-Dichloroethane	18.5		0.22	5.00	ug/L
79-01-6	Trichloroethene	18.5		0.090	5.00	ug/L
127-18-4	Tetrachloroethene	19.1		0.23	5.00	ug/L
108-90-7	Chlorobenzene	18.8		0.12	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	40.4		70 (74) - 130 (125)	81%	SPK: 50
1868-53-7	Dibromofluoromethane	45.0		70 (75) - 130 (124)	90%	SPK: 50
2037-26-5	Toluene-d8	44.3		70 (86) - 130 (113)	89%	SPK: 50
460-00-4	4-Bromofluorobenzene	44.5		70 (77) - 130 (121)	89%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	249000	8.206			
540-36-3	1,4-Difluorobenzene	458000	9.083			
3114-55-4	Chlorobenzene-d5	416000	11.841			
3855-82-1	1,4-Dichlorobenzene-d4	212000	13.771			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090825\
 Data File : VN087753.D
 Acq On : 08 Sep 2025 11:20
 Operator : JC\MD
 Sample : VN0908WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0908WBSD01

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/09/2025
 Supervised By :Mahesh Dadoda 09/09/2025

Quant Time: Sep 09 02:25:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:55:42 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	249385	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	457673	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	416132	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.771	152	211983	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	206596	40.433	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	80.860%		
35) Dibromofluoromethane	8.147	113	148695	44.999	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	90.000%		
50) Toluene-d8	10.541	98	548220	44.327	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	88.660%		
62) 4-Bromofluorobenzene	12.824	95	195567	44.464	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	88.920%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.142	85	72836	20.564	ug/l	100
3) Chloromethane	2.383	50	68923	18.935	ug/l	100
4) Vinyl Chloride	2.536	62	80749	19.133	ug/l	95
5) Bromomethane	2.983	94	45587	20.934	ug/l	93
6) Chloroethane	3.142	64	52824	17.808	ug/l	98
7) Trichlorofluoromethane	3.512	101	127701	20.652	ug/l	98
8) Diethyl Ether	3.959	74	46202	17.157	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.371	101	71165	20.340	ug/l	98
10) Methyl Iodide	4.577	142	43325	21.123	ug/l	97
11) Tert butyl alcohol	5.518	59	74221	83.672	ug/l	100
12) 1,1-Dichloroethene	4.342	96	68169	18.960	ug/l	91
13) Acrolein	4.183	56	92142	84.495	ug/l	98
14) Allyl chloride	5.012	41	103832	15.936	ug/l	96
15) Acrylonitrile	5.706	53	212443	84.920	ug/l	100
16) Acetone	4.418	43	215166	77.356	ug/l	99
17) Carbon Disulfide	4.706	76	167947	16.967	ug/l	97
18) Methyl Acetate	5.018	43	117921	20.248	ug/l	97
19) Methyl tert-butyl Ether	5.783	73	237911	17.638	ug/l	96
20) Methylene Chloride	5.265	84	69247	16.377	ug/l	96
21) trans-1,2-Dichloroethene	5.777	96	64185	16.471	ug/l	98
22) Diisopropyl ether	6.659	45	249101	17.716	ug/l	98
23) Vinyl Acetate	6.589	43	1099462	85.948	ug/l	100
24) 1,1-Dichloroethane	6.553	63	136523	17.709	ug/l	98
25) 2-Butanone	7.471	43	304929	83.327	ug/l	100
26) 2,2-Dichloropropane	7.477	77	122562	18.523	ug/l	98
27) cis-1,2-Dichloroethene	7.471	96	79529	17.072	ug/l	98
28) Bromochloromethane	7.794	49	65165	17.484	ug/l	99
29) Tetrahydrofuran	7.824	42	196862	83.591	ug/l	98
30) Chloroform	7.947	83	140240	17.822	ug/l	91
31) Cyclohexane	8.236	56	119832	18.645	ug/l	95
32) 1,1,1-Trichloroethane	8.147	97	121426	18.209	ug/l	97
36) 1,1-Dichloropropene	8.347	75	93562	18.456	ug/l	97
37) Ethyl Acetate	7.547	43	119316	17.224	ug/l	99
38) Carbon Tetrachloride	8.341	117	105004	20.979	ug/l	93
39) Methylcyclohexane	9.583	83	111909	20.051	ug/l	97
40) Benzene	8.588	78	289813	18.639	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090825\
 Data File : VN087753.D
 Acq On : 08 Sep 2025 11:20
 Operator : JC\MD
 Sample : VN0908WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0908WBSD01

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/09/2025
 Supervised By :Mahesh Dadoda 09/09/2025

Quant Time: Sep 09 02:25:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:55:42 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.765	41	65320	18.204	ug/l	96
42) 1,2-Dichloroethane	8.647	62	110324	18.464	ug/l	100
43) Isopropyl Acetate	8.671	43	198675	18.370	ug/l	98
44) Trichloroethene	9.336	130	65759	18.525	ug/l	99
45) 1,2-Dichloropropane	9.600	63	75239	18.358	ug/l	98
46) Dibromomethane	9.688	93	56413	19.343	ug/l	97
47) Bromodichloromethane	9.871	83	114773	19.202	ug/l	94
48) Methyl methacrylate	9.659	41	90670	17.724	ug/l	99
49) 1,4-Dioxane	9.683	88	24445	368.674	ug/l #	91
51) 4-Methyl-2-Pentanone	10.424	43	618693	94.816	ug/l	100
52) Toluene	10.606	92	179456	18.617	ug/l	100
53) t-1,3-Dichloropropene	10.818	75	118072	19.084	ug/l	94
54) cis-1,3-Dichloropropene	10.288	75	121736	18.950	ug/l	98
55) 1,1,2-Trichloroethane	10.994	97	71727	19.236	ug/l	97
56) Ethyl methacrylate	10.853	69	117235	19.673	ug/l	98
57) 1,3-Dichloropropane	11.141	76	123780	18.757	ug/l	96
58) 2-Chloroethyl Vinyl ether	10.141	63	278264	86.272	ug/l	100
59) 2-Hexanone	11.177	43	401035	97.246	ug/l	100
60) Dibromochloromethane	11.335	129	81915	18.959	ug/l	99
61) 1,2-Dibromoethane	11.447	107	72863	18.531	ug/l	99
64) Tetrachloroethene	11.082	164	55013	19.055	ug/l	92
65) Chlorobenzene	11.871	112	194912	18.827	ug/l	96
66) 1,1,1,2-Tetrachloroethane	11.935	131	71101	20.687	ug/l	98
67) Ethyl Benzene	11.941	91	344565	19.223	ug/l	97
68) m/p-Xylenes	12.047	106	256482	39.015	ug/l	98
69) o-Xylene	12.377	106	121236	18.819	ug/l	96
70) Styrene	12.388	104	210348	19.489	ug/l	99
71) Bromoform	12.559	173	53879	20.440	ug/l	96
73) Isopropylbenzene	12.671	105	328324	19.167	ug/l	99
74) N-amyl acetate	12.518	43	114874m	17.696	ug/l	
75) 1,1,2,2-Tetrachloroethane	12.912	83	115548	19.593	ug/l	98
76) 1,2,3-Trichloropropane	12.971	75	101515m	20.421	ug/l	
77) Bromobenzene	12.959	156	76306	18.893	ug/l	98
78) n-propylbenzene	13.012	91	417970	19.810	ug/l	99
79) 2-Chlorotoluene	13.100	91	246367	19.403	ug/l	96
80) 1,3,5-Trimethylbenzene	13.147	105	275555	18.963	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.718	75	33950	18.080	ug/l	90
82) 4-Chlorotoluene	13.200	91	248629	18.624	ug/l	98
83) tert-Butylbenzene	13.412	119	246075	20.321	ug/l	97
84) 1,2,4-Trimethylbenzene	13.459	105	277689	19.089	ug/l	99
85) sec-Butylbenzene	13.594	105	365499	20.339	ug/l	99
86) p-Isopropyltoluene	13.706	119	300261	20.235	ug/l	99
87) 1,3-Dichlorobenzene	13.712	146	151214	19.012	ug/l	98
88) 1,4-Dichlorobenzene	13.788	146	154233	18.634	ug/l	99
89) n-Butylbenzene	14.029	91	303902	20.621	ug/l	100
90) Hexachloroethane	14.312	117	59984	21.218	ug/l	94
91) 1,2-Dichlorobenzene	14.082	146	143673	19.040	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.700	75	23538	16.801	ug/l	94
93) 1,2,4-Trichlorobenzene	15.365	180	83637	18.458	ug/l	99
94) Hexachlorobutadiene	15.470	225	34460	21.332	ug/l	98
95) Naphthalene	15.612	128	284818	17.746	ug/l	100
96) 1,2,3-Trichlorobenzene	15.812	180	81832	18.474	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090825\
Data File : VN087753.D
Acq On : 08 Sep 2025 11:20
Operator : JC\MD
Sample : VN0908WBSD01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0908WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 09/09/2025
Supervised By :Mahesh Dadoda 09/09/2025

Quant Time: Sep 09 02:25:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:55:42 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090825\
 Data File : VN087753.D
 Acq On : 08 Sep 2025 11:20
 Operator : JC\MD
 Sample : VN0908WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :

MSVOA_N

ClientSampleId :

VN0908WBSD01

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 09/09/2025

Supervised By :Mahesh Dadoda 09/09/2025


Quant Time: Sep 09 02:25:55 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M

Quant Title : SW846 8260

QLast Update : Fri Aug 22 02:55:42 2025

Response via : Initial Calibration



Manual Integration Report

Sequence:	VN082125	Instrument	MSVOA_n
-----------	----------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC001	VN087619.D	1,2,3-Trichloropropane	JOHN	8/22/2025 9:16:40 AM	MMDadoda	8/25/2025 8:11:19 AM	Peak Integrated by Software
VSTDICC001	VN087619.D	1,4-Dichlorobenzene	JOHN	8/22/2025 9:16:40 AM	MMDadoda	8/25/2025 8:11:19 AM	Peak Integrated by Software
VSTDICC001	VN087619.D	Diethyl Ether	JOHN	8/22/2025 9:16:40 AM	MMDadoda	8/25/2025 8:11:19 AM	Peak Integrated by Software
VSTDICC001	VN087619.D	Ethyl Acetate	JOHN	8/22/2025 9:16:40 AM	MMDadoda	8/25/2025 8:11:19 AM	Peak Integrated by Software
VSTDICC001	VN087619.D	Methyl Acetate	JOHN	8/22/2025 9:16:40 AM	MMDadoda	8/25/2025 8:11:19 AM	Peak Integrated by Software
VSTDICC001	VN087619.D	N-amyl acetate	JOHN	8/22/2025 9:16:40 AM	MMDadoda	8/25/2025 8:11:19 AM	Peak Integrated by Software
VSTDICC005	VN087620.D	1,2,3-Trichloropropane	JOHN	8/22/2025 9:16:49 AM	MMDadoda	8/25/2025 8:11:20 AM	Peak Integrated by Software
VSTDICC005	VN087620.D	Ethyl Acetate	JOHN	8/22/2025 9:16:49 AM	MMDadoda	8/25/2025 8:11:20 AM	Peak Integrated by Software
VSTDICC005	VN087620.D	N-amyl acetate	JOHN	8/22/2025 9:16:49 AM	MMDadoda	8/25/2025 8:11:20 AM	Peak Integrated by Software
VSTDICC020	VN087621.D	1,2,3-Trichloropropane	JOHN	8/22/2025 9:17:20 AM	MMDadoda	8/25/2025 8:11:22 AM	Peak Integrated by Software
VSTDICC020	VN087621.D	N-amyl acetate	JOHN	8/22/2025 9:17:20 AM	MMDadoda	8/25/2025 8:11:22 AM	Peak Integrated by Software
VSTDICCC050	VN087622.D	1,2,3-Trichloropropane	JOHN	8/22/2025 9:17:24 AM	MMDadoda	8/25/2025 8:11:23 AM	Peak Integrated by Software
VSTDICCC050	VN087622.D	N-amyl acetate	JOHN	8/22/2025 9:17:24 AM	MMDadoda	8/25/2025 8:11:23 AM	Peak Integrated by Software

Manual Integration Report

Sequence:

VN082125

Instrument

MSVOA_n

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC100	VN087623.D	1,2,3-Trichloropropane	JOHN	8/22/2025 9:17:28 AM	MMDadoda	8/25/2025 8:11:24 AM	Peak Integrated by Software
VSTDICC150	VN087624.D	1,2,3-Trichloropropane	JOHN	8/22/2025 9:17:33 AM	MMDadoda	8/25/2025 8:11:26 AM	Peak Integrated by Software
VSTDICV050	VN087626.D	1,2,3-Trichloropropane	JOHN	8/22/2025 9:17:37 AM	MMDadoda	8/25/2025 8:11:28 AM	Peak Integrated by Software
VSTDICV050	VN087626.D	N-amyl acetate	JOHN	8/22/2025 9:17:37 AM	MMDadoda	8/25/2025 8:11:28 AM	Peak Integrated by Software

Manual Integration Report

Sequence:	vn090825	Instrument	MSVOA_n
-----------	----------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VN087749.D	1,2,3-Trichloropropane	sam	9/9/2025 7:57:29 AM	MMDadoda	9/9/2025 3:13:20 PM	Peak Integrated by Software
VSTDCCC050	VN087749.D	N-amyl acetate	sam	9/9/2025 7:57:29 AM	MMDadoda	9/9/2025 3:13:20 PM	Peak Integrated by Software
VN0908WBS01	VN087752.D	1,2,3-Trichloropropane	sam	9/9/2025 7:57:35 AM	MMDadoda	9/9/2025 3:13:25 PM	Peak Integrated by Software
VN0908WBS01	VN087752.D	N-amyl acetate	sam	9/9/2025 7:57:35 AM	MMDadoda	9/9/2025 3:13:25 PM	Peak Integrated by Software
VN0908WBSD01	VN087753.D	1,2,3-Trichloropropane	sam	9/9/2025 7:57:39 AM	MMDadoda	9/9/2025 3:13:27 PM	Peak Integrated by Software
VN0908WBSD01	VN087753.D	N-amyl acetate	sam	9/9/2025 7:57:39 AM	MMDadoda	9/9/2025 3:13:27 PM	Peak Integrated by Software
Q3032-02	VN087759.D	Methylene Chloride	sam	9/9/2025 7:57:44 AM	MMDadoda	9/9/2025 3:13:29 PM	Peak Integrated by Software
Q3032-05	VN087760.D	Tert butyl alcohol	sam	9/9/2025 7:58:00 AM	MMDadoda	9/9/2025 3:13:32 PM	Peak Integrated by Software
VSTDCCC050	VN087762.D	1,2,3-Trichloropropane	sam	9/9/2025 7:57:54 AM	MMDadoda	9/9/2025 3:13:41 PM	Peak Integrated by Software
VSTDCCC050	VN087762.D	N-amyl acetate	sam	9/9/2025 7:57:54 AM	MMDadoda	9/9/2025 3:13:41 PM	Peak Integrated by Software

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # VN082125

Review By	John Carlone	Review On	8/22/2025 9:17:52 AM
Supervise By	Mahesh Dadoda	Supervise On	8/25/2025 8:11:38 AM
SubDirectory	VN082125	HP Acquire Method	HP Processing Method 82N082125W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP135214		
Initial Calibration Stds	VP135215,VP135216,VP135217,VP135218,VP135219,VP135220		
CCC			
Internal Standard/PEM			
ICV/I.BLK	VP135221		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN087614.D	21 Aug 2025 09:30	JC\MD	Ok
2	VSTDCCC020	VN087615.D	21 Aug 2025 10:04	JC\MD	Not Ok
3	VN0821WBS01	VN087616.D	21 Aug 2025 10:37	JC\MD	Not Ok
4	VN0821WBSD01	VN087617.D	21 Aug 2025 11:11	JC\MD	Not Ok
5	VN0821WBL01	VN087618.D	21 Aug 2025 11:36	JC\MD	Not Ok
6	VSTDICC001	VN087619.D	21 Aug 2025 12:09	JC\MD	Ok,M
7	VSTDICC005	VN087620.D	21 Aug 2025 13:08	JC\MD	Ok,M
8	VSTDICC020	VN087621.D	21 Aug 2025 13:41	JC\MD	Ok,M
9	VSTDICCC050	VN087622.D	21 Aug 2025 14:02	JC\MD	Ok,M
10	VSTDICC100	VN087623.D	21 Aug 2025 14:23	JC\MD	Ok,M
11	VSTDICC150	VN087624.D	21 Aug 2025 14:44	JC\MD	Ok,M
12	IBLK	VN087625.D	21 Aug 2025 15:05	JC\MD	Ok
13	VSTDICV050	VN087626.D	21 Aug 2025 15:47	JC\MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # VN090825

Review By	Semsettin Yesilyurt	Review On	9/9/2025 3:14:54 PM
Supervise By		Supervise On	
SubDirectory	VN090825	HP Acquire Method	MSVOA_N
		HP Processing Method	VN082125W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135400		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135401,VP135402 VP134956		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN087748.D	08 Sep 2025 07:42	JC\MD	Ok
2	VSTDCCC050	VN087749.D	08 Sep 2025 08:58	JC\MD	Ok,M
3	VN0908MBL01	VN087750.D	08 Sep 2025 09:39	JC\MD	Ok
4	VN0908WBL01	VN087751.D	08 Sep 2025 09:59	JC\MD	Ok
5	VN0908WBS01	VN087752.D	08 Sep 2025 10:59	JC\MD	Ok,M
6	VN0908WBSD01	VN087753.D	08 Sep 2025 11:20	JC\MD	Ok,M
7	Q3036-03	VN087754.D	08 Sep 2025 12:30	JC\MD	Ok
8	VIBLK	VN087755.D	08 Sep 2025 12:51	JC\MD	Ok
9	PB169581TB	VN087756.D	08 Sep 2025 13:19	JC\MD	Ok
10	Q3023-02	VN087757.D	08 Sep 2025 13:40	JC\MD	Ok
11	Q3023-05	VN087758.D	08 Sep 2025 14:01	JC\MD	Ok
12	Q3032-02	VN087759.D	08 Sep 2025 14:21	JC\MD	Ok,M
13	Q3032-05	VN087760.D	08 Sep 2025 14:42	JC\MD	Ok,M
14	VIBLK	VN087761.D	08 Sep 2025 15:46	JC\MD	Ok
15	VSTDCCC050	VN087762.D	08 Sep 2025 16:07	JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN082125

Review By	John Carlone	Review On	8/22/2025 9:17:52 AM
Supervise By	Maresh Dadoda	Supervise On	8/25/2025 8:11:38 AM
SubDirectory	VN082125	HP Acquire Method	HP Processing Method 82N082125W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135214 VP135215,VP135216,VP135217,VP135218,VP135219,VP135220 VP135221

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087614.D	21 Aug 2025 09:30		JC\MD	Ok
2	VSTDCCC020	VSTDCCC020	VN087615.D	21 Aug 2025 10:04	Need ICAL	JC\MD	Not Ok
3	VN0821WBS01	VN0821WBS01	VN087616.D	21 Aug 2025 10:37		JC\MD	Not Ok
4	VN0821WBSD01	VN0821WBSD01	VN087617.D	21 Aug 2025 11:11		JC\MD	Not Ok
5	VN0821WBL01	VN0821WBL01	VN087618.D	21 Aug 2025 11:36		JC\MD	Not Ok
6	VSTDICC001	VSTDICC001	VN087619.D	21 Aug 2025 12:09	LR-10,20	JC\MD	Ok,M
7	VSTDICC005	VSTDICC005	VN087620.D	21 Aug 2025 13:08		JC\MD	Ok,M
8	VSTDICC020	VSTDICC020	VN087621.D	21 Aug 2025 13:41	method not used for #N-amyle Acetate	JC\MD	Ok,M
9	VSTDICCC050	VSTDICCC050	VN087622.D	21 Aug 2025 14:02		JC\MD	Ok,M
10	VSTDICC100	VSTDICC100	VN087623.D	21 Aug 2025 14:23		JC\MD	Ok,M
11	VSTDICC150	VSTDICC150	VN087624.D	21 Aug 2025 14:44		JC\MD	Ok,M
12	IBLK	IBLK	VN087625.D	21 Aug 2025 15:05		JC\MD	Ok
13	VSTDICV050	ICVVN082125	VN087626.D	21 Aug 2025 15:47	Icv Failed for Com.# 64 for DOD	JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN090825

Review By	Semsettin Yesilyurt	Review On	9/9/2025 3:14:54 PM
Supervise By		Supervise On	
SubDirectory	VN090825	HP Acquire Method	MSVOA_N
		HP Processing Method	VN082125W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135400		
CCC	VP135401,VP135402		
Internal Standard/PEM	VP134956		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087748.D	08 Sep 2025 07:42		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN087749.D	08 Sep 2025 08:58		JC\MD	Ok,M
3	VN0908MBL01	VN0908MBL01	VN087750.D	08 Sep 2025 09:39		JC\MD	Ok
4	VN0908WBL01	VN0908WBL01	VN087751.D	08 Sep 2025 09:59		JC\MD	Ok
5	VN0908WBS01	VN0908WBS01	VN087752.D	08 Sep 2025 10:59		JC\MD	Ok,M
6	VN0908WBSD01	VN0908WBSD01	VN087753.D	08 Sep 2025 11:20		JC\MD	Ok,M
7	Q3036-03	EGR-2	VN087754.D	08 Sep 2025 12:30		JC\MD	Ok
8	VIBLK	VIBLK	VN087755.D	08 Sep 2025 12:51		JC\MD	Ok
9	PB169581TB	PB169581TB	VN087756.D	08 Sep 2025 13:19	pH#5.0 A	JC\MD	Ok
10	Q3023-02	TP-1	VN087757.D	08 Sep 2025 13:40	pH#5.0 A	JC\MD	Ok
11	Q3023-05	TP-2	VN087758.D	08 Sep 2025 14:01	pH#5.0 A	JC\MD	Ok
12	Q3032-02	SOIL-PILE-1	VN087759.D	08 Sep 2025 14:21	pH#5.0 A	JC\MD	Ok,M
13	Q3032-05	SOIL-PILE-2	VN087760.D	08 Sep 2025 14:42	pH#5.0 A	JC\MD	Ok,M
14	VIBLK	VIBLK	VN087761.D	08 Sep 2025 15:46		JC\MD	Ok
15	VSTDCCC050	VSTDCCC050EC	VN087762.D	08 Sep 2025 16:07		JC\MD	Ok,M

M : Manual Integration

SOP ID : M1311-TCLP-16
SDG No : N/A
Weigh By : JP
Balance ID : WC SC-7
pH Meter ID : WC PH METER-1
Extraction By : JP
Filter By : JP
Pipette ID : WC
Tumbler ID : ZHE-1
TCLP Filter ID : 60828725

Start Prep Date : 09/05/2025 **Time :** 16:00
End Prep Date : 09/06/2025 **Time :** 09:25
Combination Ratio : 20
ZHE Cleaning Batch : N/A
Initial Room Temperature: 23 °C
Final Room Temperature: 21 °C
TCLP Technician Signature : 
Supervisor By : 12

Standard Name	MLS USED	STD REF. # FROM LOG
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE U	Lot Number
TCLP-FLUID-1	N/A	WP114525
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
40ml VOA Vials	4127	N/A

Extraction Conformance/Non-Conformance Comments:

Matrix spikes are added after filtration and before preservation. TUMBLER ZHE-1 checked, 30 rpm. Leak checked after 10 minutes of tumbling.

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
09/08/25 08:45	JP TCLP Room	S-Y IVOC Lab
	Preparation Group	Analysis Group

Sample ID	ClientID	ZHE Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
PB169581TB	LEB581	05	N/A	500	N/A	N/A	N/A	4.93	N/A	ZHE-1
Q3023-02	TP-1	01	25.02	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
Q3023-05	TP-2	02	25.03	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
Q3032-02	SOIL-PILE-1	03	25.02	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
Q3032-05	SOIL-PILE-2	04	25.03	500	N/A	N/A	N/A	N/A	N/A	ZHE-1

SampleID	ClientID	Sample Weight (g)	Filter Weight (g)	Filtrate (mL)	Filter + Solid (After 100°C)	% solids	% Dry Solids
PB169581TB	LEB581	N/A	N/A	N/A	N/A	N/A	N/A
Q3023-02	TP-1	N/A	N/A	N/A	N/A	100	N/A
Q3023-05	TP-2	N/A	N/A	N/A	N/A	100	N/A
Q3032-02	SOIL-PILE-1	N/A	N/A	N/A	N/A	100	N/A
Q3032-05	SOIL-PILE-2	N/A	N/A	N/A	N/A	100	N/A

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tcip zhe q3023

WorkList ID : 191708

Department : TCLP Extraction

Date : 09-05-2025 12:02:07

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3023-02	TP-1	Solid	TCLP ZHE Extraction	Cool 4 deg C	PSEG03	D31	09/04/2025	1311 ZHE
Q3023-05	TP-2	Solid	TCLP ZHE Extraction	Cool 4 deg C	PSEG03	D41	09/04/2025	1311 ZHE
Q3032-02	SOIL-PILE-1	Solid	TCLP ZHE Extraction	Cool 4 deg C	ZUCC01	D31	09/05/2025	1311 ZHE
Q3032-05	SOIL-PILE-2	Solid	TCLP ZHE Extraction	Cool 4 deg C	ZUCC01	D31	09/05/2025	1311 ZHE

Date/Time 09/05/25 13:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 09/05/25 16:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]



SHIPPING DOCUMENTS

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: Zuccaro Inc.

ADDRESS: 3 Coal Ave

CITY: Morristown STATE: NJ ZIP:

ATTENTION: Sal Zuccaro

PHONE:

FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: 3 Coal Ave, Morristown, NJ

PROJECT NO.: LOCATION:

PROJECT MANAGER:

e-mail:

PHONE:

FAX:

CLIENT BILLING INFORMATION

BILL TO:

PO#:

ADDRESS:

CITY: Summit

STATE:

ZIP:

ATTENTION:

PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*

HARDCOPY (DATA PACKAGE): _____ DAYS*

EDD: _____ DAYS*

*TO BE APPROVED BY CHEMTECH

STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)

☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP

☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B

+ Raw Data ☐ Other _____

☐ EDD FORMAT _____

Metals TCL-TCL
SVOC-TCL BNA-20
PCB
TCL BNA
Mercury
TCL VOA
VOC-TCL VOA-10

PRESERVATIVES

COMMENTS

← Specify Preservatives

A-HCl D-NaOH
B-HNO3 E-ICE
C-H2SO4 F-OTHER

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES										
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	
1.	Soil pile #1	SOL	X		9.5.25	11:14	5	X	X	X	X	X					
2.	I	I		X	I	11:16	5						X	X			
3.	Soil pile #2	I	X		I	11:36	5	X	X	X	X	X					
4.	I	I		X	I	11:40	5						X	X			
5.																	
6.																	
7.																	
8.																	
9.																	
10.																	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER:	DATE/TIME: 12:06	RECEIVED BY:	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☒ NON COMPLIANT	COOLER TEMP: 4.2°C
1. [Signature]	9.5.2025	1. [Signature]	Comments: Equal volume of soil used, 5:1 composite	
RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	* Soil pile #1 - 24x16x7 = measurement *	
2. [Signature]		2. [Signature]	* Soil pile #2 - 24x22x7 = measurement *	
RELINQUISHED BY SAMPLER:	DATE/TIME: 12:36	RECEIVED BY:	Page 1 of 1	CLIENT: ☐ Hand Delivered ☐ Other
3. [Signature]	9.5.2025	3. [Signature]		Shipment Complete ☐ YES ☐ NO



Environmental Laboratory

www.chemtech.net | EMAIL: PM@chemtech.net

Project Name: 3001 AVE,

Morrisstown, N.J.

Service Order #:

Work Order #:

Labor WBS #:

Facility/Site:

Site Address: 3001 AVE

Morrisstown, NJ.

Chemtech Order ID:

Sampler Name: Katherine Carter

Client Project Coordinator & Phone:

Page #: 1 of 1

Date: 9.5.2025

Arrive Time: 10:38

Depart Time: 12:06

Waste Stream (circle one): drum / roll-off / soil pile / in-situ / linear construction / frac-tank

Sample Matrices (circle all that apply): Water / Solid / NAPL / Concrete / Wipe

Collection Depths:

Temp (range):

°C

PID Readings (range):

Dimensions/CY:

PPM

Odor:

Y

N

Color:

Y

N

Sample Description:


Field Observations:

Brown wet soil, rocks, grass.
Sampled, soil pile #1, soil pile #2


Grid/Area Composite Map:

QA Control # A3041134

soil pile #1



soil pile #2



Sampler Signature:

[Signature] 9.5.2025

Supervisor Review/Date:

Client Signature:

Date/Time Arrived at Lab:

Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312