

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3032</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Zuccaro Inc.</u>	Datafile :	<u>VN087752.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0908WBS01	Vinyl chloride	20	19.6	ug/L	98			70 (65)	130 (117)	
	1,1-Dichloroethene	20	18.7	ug/L	94			70 (74)	130 (110)	
	2-Butanone	100	84.1	ug/L	84			40 (65)	160 (122)	
	Carbon Tetrachloride	20	20.1	ug/L	101			70 (77)	130 (113)	
	Chloroform	20	18.0	ug/L	90			70 (79)	130 (113)	
	Benzene	20	18.0	ug/L	90			70 (82)	130 (109)	
	1,2-Dichloroethane	20	17.7	ug/L	89			70 (80)	130 (115)	
	Trichloroethene	20	18.4	ug/L	92			70 (77)	130 (113)	
	Tetrachloroethene	20	18.4	ug/L	92			70 (67)	130 (123)	
	Chlorobenzene	20	18.5	ug/L	93			70 (82)	130 (109)	