

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ZUCC01
 Lab Code: ACE SDG NO.: Q3032
 Lab File ID: VN087749.D Date Analyzed: 09/08/2025
 Instrument ID: MSVOA_N Time Analyzed: 08:58
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	232401	13.77				
UPPER LIMIT	464802	14.27				
LOWER LIMIT	116201	13.27				
EPA SAMPLE NO.						
SOIL-PILE-1	185969	13.77				
SOIL-PILE-2	180770	13.77				
VN0908WBL01	244144	13.77				
VN0908WBS01	222091	13.77				
VN0908WBSD01	211983	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.