

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ZUCC01
 Lab Code: ACE SDG NO.: Q3032
 Lab File ID: VY023290.D Date Analyzed: 09/05/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:44
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	340704	13.346				
UPPER LIMIT	681408	13.846				
LOWER LIMIT	170352	12.846				
EPA SAMPLE NO.						
SOIL-PILE-1	418128	13.35				
SOIL-PILE-2	401855	13.35				
VY0905SBL01	426836	13.35				
VY0905SBS01	326298	13.35				
VY0905SBSD01	311021	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.