

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>APEX04</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3037</u>
Instrument ID:	<u>MSVOA_L</u>	Calibration Date/Time:	<u>09/08/2025 11:05</u>
Lab File ID:	<u>VL042900.D</u>	Init. Calib. Date(s):	<u>08/13/2025 08/13/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>08:30 12:06</u>
GC Column:	<u>RTX-1</u>	ID:	<u>0.32</u> (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Vinyl Chloride	0.582	0.560		-3.78	30
1,2-Dichloroethane	0.500	0.635		27	30
Trichloroethene	0.437	0.455		4.12	30
Tetrachloroethene	0.346	0.331		-4.34	30
1-Bromo-4-Fluorobenzene	0.750	0.786		4.8	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.