

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: APEX04
 Lab Code: ACE SDG NO.: Q3037
 Lab File ID: VL042900.D Date Analyzed: 09/08/2025
 Instrument ID: MSVOA_L Time Analyzed: 11:05
 GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	126678	2.78	333606	3.94	282770	8.87
UPPER LIMIT	177349	3.11	467048	4.27	395878	9.20
LOWER LIMIT	76006.8	2.45	200164	3.61	169662	8.54
EPA SAMPLE NO.						
IA-2	138126	2.78	368558	3.95	302216	8.88
IA-2DUP	141221	2.78	372044	3.95	303028	8.88
VL0908ABL01	141293	2.78	355217	3.95	298537	8.88
VL0908ABS01	139945	2.78	365857	3.95	306385	8.87

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +40% of internal standard area
 AREA LOWER LIMIT = -40% of internal standard area
 RT UPPER LIMIT = +0.33 minutes of internal standard RT
 RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.