

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL090825\
 Data File : VL042900.D
 Acq On : 08 Sep 2025 11:05
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Sep 09 00:52:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL081325AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Sep 09 00:52:26 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	1.466	1.540	-5.0	109	0.00
3	Chlorodifluoromethane	1.620	1.767	-9.1	111	0.00
4	Chloromethane	0.591	0.591	0.0	105	0.00
5 T	Vinyl Chloride	0.582	0.560	3.8	106	0.00
6 T	Bromomethane	0.233	0.225	3.4	106	0.00
7	Chloroethane	0.227	0.225	0.9	106	0.00
8 T	Dichlorotetrafluoroethane	1.199	1.250	-4.3	106	0.00
9 T	Propene	0.700	0.700	0.0	102	0.00
10 T	Heptane	1.874	1.855	1.0	103	0.00
11 T	Trichlorofluoromethane	1.446	1.487	-2.8	111	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.152	1.163	-1.0	106	0.00
13	Ethanol	0.064	0.052	18.8	106	0.00
14 T	Bromoethene	0.417	0.424	-1.7	106	0.00
15 T	Acetone	1.323	1.147	13.3	115	0.00
16 T	1,3-Butadiene	0.600	0.558	7.0	106	0.00
17	tert-Butyl alcohol	1.569	1.345	14.3	93	0.00
18 T	1,1-Dichloroethene	0.507	0.502	1.0	102	0.00
19 T	Isopropyl Alcohol	0.790	0.703	11.0	103	0.00
20 T	Methylene Chloride	0.554	0.455	17.9	105	0.00
21 T	Allyl Chloride	0.935	0.903	3.4	101	0.00
22 T	trans-1,2-Dichloroethene	0.586	0.585	0.2	104	0.00
23 T	Vinyl Acetate	1.819	1.802	0.9	103	0.00
24 T	1,1-Dichloroethane	1.171	1.205	-2.9	110	0.00
25 T	Ethyl Acetate	3.202	3.279	-2.4	106	0.00
26 T	Hexane	1.478	1.482	-0.3	101	0.00
27 T	Carbon Disulfide	1.438	1.526	-6.1	110	0.00
28 T	Methyl tert-Butyl Ether	0.754	0.677	10.2	98	0.00
29 T	Chloroform	2.110	2.223	-5.4	110	0.00
30 T	Cyclohexane	1.296	1.208	6.8	95	0.00
31 T	cis-1,2-Dichloroethene	1.476	1.441	2.4	102	0.00
32 T	1,1,1-Trichloroethane	2.230	2.143	3.9	102	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	81	0.00
34 T	2-Butanone	0.675	0.818	-21.2	108	0.00
35 T	Carbon Tetrachloride	0.689	0.807	-17.1	102	0.00
36 T	Benzene	0.958	1.097	-14.5	97	0.00
37 T	1,2-Dichloroethane	0.500	0.635	-27.0	106	0.00
38 T	Trichloroethene	0.437	0.455	-4.1	92	0.00
39 T	1,2-Dichloropropane	0.350	0.419	-19.7	104	0.00
40 T	1,4-Dioxane	0.166	0.168	-1.2	91	0.00
41 T	Tetrahydrofuran	0.376	0.447	-18.9	101	0.00
42 T	Bromodichloromethane	0.745	0.905	-21.5	103	0.00
43	Methyl Methacrylate	0.394	0.425	-7.9	93	0.00
44 T	2,2,4-Trimethylpentane	1.625	1.892	-16.4	101	0.00
45 T	t-1,3-Dichloropropene	0.445	0.463	-4.0	87	0.00
46 T	cis-1,3-Dichloropropene	0.555	0.602	-8.5	90	0.00
47 T	1,1,2-Trichloroethane	0.371	0.432	-16.4	102	0.00
48 T	Dibromochloromethane	0.600	0.685	-14.2	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL090825\
 Data File : VL042900.D
 Acq On : 08 Sep 2025 11:05
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Sep 09 00:52:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL081325AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Sep 09 00:52:26 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.495	0.552	-11.5	91	0.00
50 T	4-Methyl-2-Pentanone	0.913	1.077	-18.0	99	0.00
51 T	2-Hexanone	0.718	0.853	-18.8	98	0.00
52 T	Tetrachloroethene	0.346	0.331	4.3	85	0.00
53 T	Toluene	1.138	1.205	-5.9	91	0.00
54 T	1,2-Dibromoethane	0.542	0.612	-12.9	98	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	76	0.00
56	1,1,1,2-Tetrachloroethane	0.517	0.602	-16.4	97	0.00
57 T	Chlorobenzene	0.951	1.101	-15.8	96	0.00
58 T	Ethyl Benzene	1.765	1.974	-11.8	92	0.00
59 T	m/p-Xylene	1.383	1.618	-17.0	96	0.00
60 T	o-Xylene	1.391	1.630	-17.2	98	0.00
61 T	Styrene	0.652	0.690	-5.8	86	0.00
62	Isopropylbenzene	2.033	2.317	-14.0	96	0.00
63 T	1,1,2,2-Tetrachloroethane	0.765	0.942	-23.1	107	0.00
64	n-propylbenzene	0.519	0.582	-12.1	94	0.00
65	tert-Butylbenzene	1.793	1.985	-10.7	94	0.00
66 T	Benzyl Chloride	0.228	0.180	21.1	60	0.00
67	sec-Butylbenzene	2.516	2.946	-17.1	100	0.00
68 S	1-Bromo-4-Fluorobenzene	0.750	0.786	-4.8	80	0.00
69	p-Isopropyltoluene	2.160	2.361	-9.3	95	0.00
70	n-Butylbenzene	2.142	2.514	-17.4	100	0.00
71	2-Chlorotoluene	1.579	1.761	-11.5	95	0.00
72 T	4-Ethyltoluene	1.687	1.961	-16.2	96	0.00
73 T	1,3,5-Trimethylbenzene	1.434	1.593	-11.1	94	0.00
74 T	1,2,4-Trimethylbenzene	1.574	1.799	-14.3	99	0.00
75 T	1,3-Dichlorobenzene	0.895	1.024	-14.4	94	0.00
76 T	1,4-Dichlorobenzene	0.905	1.016	-12.3	93	0.00
77 T	1,2-Dichlorobenzene	0.876	0.995	-13.6	97	0.00
78 T	Hexachloro-1,3-Butadiene	0.718	0.594	17.3	79	0.00
79 T	Naphthalene	1.427	1.515	-6.2	88	0.00
80 T	Naphthalene,2-methyl-	0.533	0.452	15.2	72	0.00
81 T	1,2,4-Trichlorobenzene	0.717	0.734	-2.4	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0