


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090925\
Data File : VN087763.D
Acq On : 09 Sep 2025 08:23
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090425W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Fri Sep 05 03:29:53 2025



AutoFind: Scans 1848, 1849, 1850; Background Corrected with Scan 1841

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	24.9	11676	PASS
75	95	30	60	57.4	26872	PASS
95	95	100	100	100.0	46840	PASS
96	95	5	9	6.4	2977	PASS
173	174	0.00	2	0.5	164	PASS
174	95	50	100	73.2	34283	PASS
175	174	5	9	7.1	2422	PASS
176	174	95	101	96.8	33179	PASS
177	176	5	9	7.6	2514	PASS