

Initial and Continuing Calibration Verification

Client: Tully Environmental, Inc

SDG No.: Q3041

Project: Transfer Station-SPDES

RunNo.: LB137121

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	0.96	1	96	90-110	09/09/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.97	1	97	90-110	09/09/2025
Sample ID: CCV2 Ammonia as N	mg/L	0.99	1	99	90-110	09/09/2025
Sample ID: CCV3 Ammonia as N	mg/L	0.99	1	99	90-110	09/09/2025
Sample ID: CCV4 Ammonia as N	mg/L	0.98	1	98	90-110	09/09/2025