

Duplicate Sample Summary

Client:	Tully Environmental, Inc	SDG No.:	Q3041
Project:	Transfer Station-SPDES	Sample ID:	Q3007-02
Client ID:	CompDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	228		226		1	0.88		09/08/2025

Duplicate Sample Summary

Client: Tully Environmental, Inc Project: Transfer Station-SPDES Client ID: 002 35th Ave(sep)DUP	SDG No.: Q3041 Sample ID: Q3041-04 Percent Solids for Spike Sample: 0
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	2.10	OR	2.00	OR	1	5		09/09/2025

Duplicate Sample Summary

Client: Tully Environmental, Inc	SDG No.: Q3041
Project: Transfer Station-SPDES	Sample ID: Q3041-04DL
Client ID: 002 35th Ave(sep)DUPDL	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	2.00	D	2.10	D	2	5		09/09/2025

Duplicate Sample Summary

Client: Tully Environmental, Inc	SDG No.: Q3041
Project: Transfer Station-SPDES	Sample ID: Q3041-04
Client ID: 002 35th Ave(sep)MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	3.00	OR	3.00	OR	1	0		09/09/2025

Duplicate Sample Summary

Client:	Tully Environmental, Inc	SDG No.:	Q3041
Project:	Transfer Station-SPDES	Sample ID:	Q3057-01
Client ID:	DRAIN-WATER-TANK-1MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	49.1		49.2		1	0.2		09/10/2025

Duplicate Sample Summary

Client:	Tully Environmental, Inc	SDG No.:	Q3041
Project:	Transfer Station-SPDES	Sample ID:	Q3067-02
Client ID:	COMPDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
BOD5	mg/L	+/-20	681		683		1	0.24		09/10/2025