

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q3041

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	1030	1000	103	95 - 105	P	09/09/2025	12:14	LB137122
	Iron	4010	4000	100	95 - 105	P	09/09/2025	12:14	LB137122
	Lead	3910	4000	98	95 - 105	P	09/09/2025	12:14	LB137122

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q3041

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	20.8	20.0	104	80 - 120	P	09/09/2025	12:26	LB137122
	Iron	103	100	104	80 - 120	P	09/09/2025	12:26	LB137122
	Lead	11.1	12.0	92	80 - 120	P	09/09/2025	12:26	LB137122

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q3041

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1230	1250	98	90 - 110	P	09/09/2025	12:59	LB137122
	Iron	4960	5000	99	90 - 110	P	09/09/2025	12:59	LB137122
	Lead	4900	5000	98	90 - 110	P	09/09/2025	12:59	LB137122
CCV02	Copper	1190	1250	95	90 - 110	P	09/09/2025	14:14	LB137122
	Iron	4890	5000	98	90 - 110	P	09/09/2025	14:14	LB137122
	Lead	4650	5000	93	90 - 110	P	09/09/2025	14:14	LB137122
CCV03	Copper	1190	1250	95	90 - 110	P	09/09/2025	15:11	LB137122
	Iron	4750	5000	95	90 - 110	P	09/09/2025	15:11	LB137122
	Lead	4630	5000	93	90 - 110	P	09/09/2025	15:11	LB137122
CCV04	Copper	1190	1250	95	90 - 110	P	09/09/2025	16:24	LB137122
	Iron	4750	5000	95	90 - 110	P	09/09/2025	16:24	LB137122
	Lead	4670	5000	93	90 - 110	P	09/09/2025	16:24	LB137122
CCV05	Copper	1220	1250	98	90 - 110	P	09/09/2025	17:13	LB137122
	Iron	4880	5000	98	90 - 110	P	09/09/2025	17:13	LB137122
	Lead	4780	5000	96	90 - 110	P	09/09/2025	17:13	LB137122
CCV06	Copper	1220	1250	97	90 - 110	P	09/09/2025	17:46	LB137122
	Iron	4880	5000	98	90 - 110	P	09/09/2025	17:46	LB137122
	Lead	4800	5000	96	90 - 110	P	09/09/2025	17:46	LB137122