

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tully Environmental, Inc

SDG No.: Q3041

Contract: TULL01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	-9.26	2.0	463	-18	22	09/09/2025	12:42	LB137122
	Iron	95100	101000	94	85600	116500	09/09/2025	12:42	LB137122
	Lead	-3.71			-12	12	09/09/2025	12:42	LB137122
ICSAB01	Copper	490	511	96	434	588	09/09/2025	12:46	LB137122
	Iron	104000	99300	105	84400	114500	09/09/2025	12:46	LB137122
	Lead	44.4	49.0	91	37	61	09/09/2025	12:46	LB137122
ICSA	Copper	-33.5	2.0	1675	-18	22	09/09/2025	12:50	LB137122
	Iron	93500	101000	93	85600	116500	09/09/2025	12:50	LB137122
	Lead	-143			-12	12	09/09/2025	12:50	LB137122
ICSAB	Copper	511	511	100	434	588	09/09/2025	12:55	LB137122
	Iron	99900	99300	101	84400	114500	09/09/2025	12:55	LB137122
	Lead	-94.9	49.0	194	37	61	09/09/2025	12:55	LB137122