

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Tully Environmental, Inc **level:** low **sdg no.:** Q3041
contract: TULL01 **lab code:** ACE
matrix: Water **sample id:** Q3041-04 **client id:** 002 35th Ave(sep)MS

Percent Solids for Sample: NA **Spiked ID:** Q3041-04MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	300		179		150	81		P
Iron	ug/L	75 - 125	17400		16700		1500	44		P
Lead	ug/L	75 - 125	521		80.7		500	88		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>Q3041</u>
contract: <u>TULL01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3041-04</u>	client id: <u>002 35th Ave(sep)MSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3041-04MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	311		179		150	89		P
Iron	ug/L	75 - 125	18000		16700		1500	85		P
Lead	ug/L	75 - 125	534		80.7		500	91		P