

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

|                 |                                        |                           |                          |
|-----------------|----------------------------------------|---------------------------|--------------------------|
| <b>SDG No.:</b> | <u><b>Q3041</b></u>                    | <b>Analytical Method:</b> | <u><b>SW624.1</b></u>    |
| <b>Client:</b>  | <u><b>Tully Environmental, Inc</b></u> | <b>Datafile :</b>         | <u><b>VN087766.D</b></u> |

| Lab Sample ID | Parameter     | Spike | Result | Unit | Rec | RPD | Qual | Low | Limits<br>High | RPD |
|---------------|---------------|-------|--------|------|-----|-----|------|-----|----------------|-----|
| VN0909WBSD01  | Benzene       | 20    | 18.0   | ug/L | 90  | 2   |      | 65  | 135            | 20  |
|               | Toluene       | 20    | 18.4   | ug/L | 92  | 1   |      | 70  | 130            | 20  |
|               | Ethyl Benzene | 20    | 18.4   | ug/L | 92  | 0   |      | 60  | 140            | 20  |
|               | m/p-Xylenes   | 40    | 37.4   | ug/L | 94  | 3   |      | 87  | 111            | 20  |
|               | o-Xylene      | 20    | 17.9   | ug/L | 90  | 1   |      | 87  | 111            | 20  |