

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TULL01
 Lab Code: ACE SDG NO.: Q3041
 Lab File ID: VN087764.D Date Analyzed: 09/09/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:16
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	51239	7.80	262160	9.08	242046	11.85
UPPER LIMIT	102478	8.3	524320	9.583	484092	12.347
LOWER LIMIT	25619.5	7.3	131080	8.583	121023	11.347
EPA SAMPLE NO.						
001 Willets Pt Blvd (Sep)	45351	7.79	240068	9.08	211456	11.84
002 35th Ave (sep)	46039	7.80	227372	9.08	210427	11.85
VN0909WBL01	46637	7.79	244324	9.08	216356	11.84
VN0909WBS01	49254	7.79	259436	9.08	243037	11.84
VN0909WBSD01	51766	7.79	268313	9.08	245687	11.84

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.