

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: PSEG

SDG No.: Q3046

Contract: PSEG03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	3750	4000	94	90 - 110	P	09/10/2025	11:31	LB137144

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: PSEG

SDG No.: Q3046

Contract: PSEG03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	11.9	12.0	99	80 - 120	P	09/10/2025	11:39	LB137144

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: PSEG

SDG No.: Q3046

Contract: PSEG03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	4880	5000	98	90 - 110	P	09/10/2025	12:09	LB137144
CCV02	Lead	4760	5000	95	90 - 110	P	09/10/2025	13:14	LB137144
CCV03	Lead	4680	5000	94	90 - 110	P	09/10/2025	14:04	LB137144
CCV04	Lead	4770	5000	95	90 - 110	P	09/10/2025	15:27	LB137144
CCV05	Lead	4820	5000	96	90 - 110	P	09/10/2025	16:16	LB137144
CCV06	Lead	4860	5000	97	90 - 110	P	09/10/2025	16:43	LB137144