



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: PSEG

SDG No.: Q3046

Contract: PSEG03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Lead	2.30	+/-6	U	12.0	P	09/10/2025	11:43	LB137144

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: PSEG

SDG No.: Q3046

Contract: PSEG03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Lead	2.30	+/-6	U	12.0	P	09/10/2025	12:13	LB137144
CCB02	Lead	2.30	+/-6	U	12.0	P	09/10/2025	13:18	LB137144
CCB03	Lead	2.30	+/-6	U	12.0	P	09/10/2025	14:25	LB137144
CCB04	Lead	2.30	+/-6	U	12.0	P	09/10/2025	15:31	LB137144
CCB05	Lead	2.30	+/-6	U	12.0	P	09/10/2025	16:20	LB137144
CCB06	Lead	2.30	+/-6	U	12.0	P	09/10/2025	16:48	LB137144

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: PSEG

SDG No.: Q3046

Contract: PSEG03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	------	---	------------------	------------------	---------------