

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: PSEG

SDG No.: Q3046

Contract: PSEG03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	3910	4000	98	95 - 105	P	09/09/2025	12:14	LB137122

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: PSEG

SDG No.: Q3046

Contract: PSEG03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	11.1	12.0	92	80 - 120	P	09/09/2025	12:26	LB137122

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: PSEG

SDG No.: Q3046

Contract: PSEG03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	4900	5000	98	90 - 110	P	09/09/2025	12:59	LB137122
CCV02	Lead	4650	5000	93	90 - 110	P	09/09/2025	14:14	LB137122
CCV03	Lead	4630	5000	93	90 - 110	P	09/09/2025	15:11	LB137122
CCV04	Lead	4670	5000	93	90 - 110	P	09/09/2025	16:24	LB137122
CCV05	Lead	4780	5000	96	90 - 110	P	09/09/2025	17:13	LB137122
CCV06	Lead	4800	5000	96	90 - 110	P	09/09/2025	17:46	LB137122