

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: PSEG

SDG No.: Q3046

Contract: PSEG03

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Lead	-3.71			-12	12	09/09/2025	12:42	LB137122
ICSAB01	Lead	44.4	49.0	91	37	61	09/09/2025	12:46	LB137122
ICSA	Lead	-143			-12	12	09/09/2025	12:50	LB137122
ICSAB	Lead	-94.9	49.0	194	37	61	09/09/2025	12:55	LB137122