

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: PSEG

SDG No.: Q3046

Contract: PSEG03

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169617BS Lead	ug/L	5000	4690		94	80 - 120	P