

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>GFEL01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3052</u>
Instrument ID:	<u>MSVOA_L</u>	Calibration Date/Time:	<u>09/10/2025</u> <u>09:36</u>
Lab File ID:	<u>VL042921.D</u>	Init. Calib. Date(s):	<u>08/13/2025</u> <u>08/13/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>08:30</u> <u>12:06</u>
GC Column:	<u>RTX-1</u>	ID:	<u>0.32</u> (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Vinyl Chloride	0.582	0.557		-4.3	30
trans-1,2-Dichloroethene	0.586	0.590		0.68	30
1,1-Dichloroethane	1.171	1.178		0.6	30
cis-1,2-Dichloroethene	1.476	1.360		-7.86	30
1,1,1-Trichloroethane	2.230	1.999		-10.36	30
1,2-Dichloroethane	0.500	0.609		21.8	30
Trichloroethene	0.437	0.443		1.37	30
Tetrachloroethene	0.346	0.322		-6.94	30
1-Bromo-4-Fluorobenzene	0.750	0.787		4.93	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.