

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3052</u>	Analytical Method:	<u>SWTO-15</u>
Client:	<u>GFE LLC</u>	Datafile :	<u>VL042902.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL0908ABS01	Vinyl Chloride	10	9.60	ppbv	96			70	130	
	trans-1,2-Dichloroethene	10	10.0	ppbv	100			70	130	
	1,1-Dichloroethane	10	10.1	ppbv	101			70	130	
	cis-1,2-Dichloroethene	10	9.80	ppbv	98			70	130	
	1,1,1-Trichloroethane	10	9.40	ppbv	94			70	130	
	1,2-Dichloroethane	10	12.7	ppbv	127			70	130	
	Trichloroethene	10	10.4	ppbv	104			70	130	
	Tetrachloroethene	10	9.60	ppbv	96			70	130	