

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3052</u>	Analytical Method:	<u>SWTO-15</u>
Client:	<u>GFE LLC</u>	Datafile :	<u>VL042923.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL0910ABS01	Vinyl Chloride	10	9.10	ppbv	91			70	130	
	trans-1,2-Dichloroethene	10	9.60	ppbv	96			70	130	
	1,1-Dichloroethane	10	9.60	ppbv	96			70	130	
	cis-1,2-Dichloroethene	10	8.80	ppbv	88			70	130	
	1,1,1-Trichloroethane	10	8.60	ppbv	86			70	130	
	1,2-Dichloroethane	10	12.5	ppbv	125			70	130	
	Trichloroethene	10	10.3	ppbv	103			70	130	
	Tetrachloroethene	10	9.20	ppbv	92			70	130	