

Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB137149

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/11/2025	08:30
2	CAL2	1	Water	NA	NA	20.2	7.00	09/11/2025	08:31
3	CAL3	1	Water	NA	NA	20.2	10.02	09/11/2025	08:33
4	ICV	1	Water	NA	NA	20.3	7.02	09/11/2025	08:35
5	CCV1	1	Water	NA	NA	20.1	2.02	09/11/2025	09:00
6	Q3057-01	1	Water	NA	NA	20.5	6.99	09/11/2025	09:10
7	Q3070-01	1	Water	NA	NA	20.8	7.38	09/11/2025	09:20
8	Q3071-02	1	Water	NA	NA	20.6	6.50	09/11/2025	09:30
9	Q3074-02	1	Water	NA	NA	20.8	6.30	09/11/2025	09:38
10	Q3074-02DUP	1	Water	NA	NA	20.9	6.32	09/11/2025	09:40
11	CCV2	1	Water	NA	NA	20.2	12.02	09/11/2025	09:41

WORKLIST(Hardcopy Internal Chain)

137149

WorkList Name : ph q3074 w

WorkList ID : 191795

Department : Wet-Chemistry

Date : 09-11-2025 08:11:14

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3057-01	DRAIN-WATER-TANK-1	Water	pH	Cool 4 deg C	MAJO01	J12	09/09/2025	9040C
Q3070-01	DRAIN-WATER-TANK-1	Water	pH	Cool 4 deg C	MAJO01	D21	09/10/2025	9040C
Q3071-02	4089	Water	pH	Cool 4 deg C	PSEG03	D21	09/10/2025	9040C
Q3074-02	MOO-25-0140	Water	pH	Cool 4 deg C	PSEG03	D31	09/10/2025	9040C

Date/Time 09-11-25 08:15

Raw Sample Received by: JB WDC

Raw Sample Relinquished by: JB WDC

Date/Time 09-11-25

Raw Sample Received by: JB WDC

Raw Sample Relinquished by: JB WDC