

Initial and Continuing Calibration Verification

Client: Major Products

SDG No.: Q3057

Project: Floor Drain Water 2025

RunNo.: LB137149

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	7.02	7	100	90-110	09/11/2025
Sample ID: CCV1 pH	pH	2.02	2.00	101	90-110	09/11/2025
Sample ID: CCV2 pH	pH	12.02	12.00	100	90-110	09/11/2025