

Matrix Spike Summary

Client:	Major Products	SDG No.:	Q3057
Project:	Floor Drain Water 2025	Sample ID:	Q3057-01
Client ID:	DRAIN-WATER-TANK-1MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	49.1		29.0		20.0	1	101		09/10/2025

Matrix Spike Summary

Client:	Major Products	SDG No.:	Q3057
Project:	Floor Drain Water 2025	Sample ID:	Q3057-01
Client ID:	DRAIN-WATER-TANK-1MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	49.2		29.0		20.0	1	101		09/10/2025