

## **DATA PACKAGE**

METALS  
VOLATILE ORGANICS

**PROJECT NAME : BUFFINGTON**

**G ENVIRONMENTAL**

**8 Carriage Ln**

**Succasunna, NJ - 07876**

**Phone No: 973-294-1771**

**ORDER ID : Q3058**

**ATTENTION : Gary Landis**



**Laboratory Certification ID # 20012**



|                                                 |     |   |
|-------------------------------------------------|-----|---|
| 1) Signature Page                               | 3   | 1 |
| 2) Case Narrative                               | 5   | 2 |
| 2.1) VOCMS Group1_VOC-TCLVOA-10- Case Narrative | 5   | 3 |
| 2.2) Metals-AES- Case Narrative                 | 7   | 4 |
| 3) Qualifier Page                               | 8   | 5 |
| 4) QA Checklist                                 | 10  | 6 |
| 5) VOC-TCLVOA-10 Data                           | 11  | 7 |
| 6) VOCMS Group1 Data                            | 149 | 8 |
| 7) Metals-AES Data                              | 226 |   |
| 8) Shipping Document                            | 272 |   |
| 8.1) CHAIN OF CUSTODY                           | 273 |   |
| 8.2) Lab Certificate                            | 274 |   |
| 8.3) Internal COC                               | 275 |   |

## DATA OF KNOWN QUALITY CONFORMANCE/NON-CONFORMANCE SUMMARY QUESTIONNAIRE

Laboratory Name : Alliance Technical Group LLC Client : G Environmental  
 Project Location : \_\_\_\_\_ Project Number : \_\_\_\_\_  
 Laboratory Sample ID(s) : Q3058 Sampling Date(s) : 9/09/2025  
 List DKQP Methods Used (e.g., 8260,8270, et Cetra) **6010D,8260D,8260D,SOP**

|    |                                                                                                                                                                                                                                                                                                     |                                                                                                                                                                             |
|----|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| 1  | For each analytical method referenced in this laboratory report package, were all specified QA/QC performance criteria followed, including the requirement to explain any criteria falling outside of acceptable guidelines, as specified in the NJDEP Data of Known Quality performance standards? | <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No                                                                                                         |
| 1A | Were the method specified handling, preservation, and holding time requirements met?                                                                                                                                                                                                                | <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No                                                                                                         |
| 1B | EPH Method: Was the EPH method conducted without significant modifications (see Section 11.3 of respective DKQ methods)                                                                                                                                                                             | <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A                                                                            |
| 2  | Were all samples received by the laboratory in a condition consistent with that described on the associated chain-of-custody document(s)?                                                                                                                                                           | <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No                                                                                                         |
| 3  | Were samples received at an appropriate temperature (4±2° C)?                                                                                                                                                                                                                                       | <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A                                                                            |
| 4  | Were all QA/QC performance criteria specified in the NJDEP DKQP standards achieved?                                                                                                                                                                                                                 | <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No                                                                                                         |
| 5  | a)Were reporting limits specified or referenced on the chain-of-custody or communicated to the laboratory prior to sample receipt?<br><br>b)Were these reporting limits met?                                                                                                                        | <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No<br><br><input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A |
| 6  | For each analytical method referenced in this laboratory report package, were results reported for all constituents identified in the method-specific analyte lists presented in the DKQP documents and/or site-specific QAPP?                                                                      | <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No                                                                                                         |
| 7  | Are project-specific matrix spikes and/or laboratory duplicates included in this data set?                                                                                                                                                                                                          | <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No                                                                                                         |

Notes: For all questions to which the response was "No" (with the exception of question #7), additional information should be provided in an attached narrative. If the answer to question #1, #1A, or #1B is "No", the data package does not meet the requirements for "Data of Known Quality."

## Cover Page

**Order ID :** Q3058

**Project ID :** Buffington

**Client :** G Environmental

**Lab Sample Number**

Q3058-01  
Q3058-02  
Q3058-03

**Client Sample Number**

RPXY5  
RPXY4  
MW4

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : \_\_\_\_\_

Date: 9/19/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

## **CASE NARRATIVE**

### **G Environmental**

**Project Name: Buffington**

**Project # N/A**

**Order ID # Q3058**

**Test Name: VOCMS Group1,VOC-TCLVOA-10**

#### **A. Number of Samples and Date of Receipt:**

2 Solid samples were received on 09/09/2025.

1 Water sample was received on 09/09/2025.

#### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested:

. This data package contains results for VOCMS Group1(8260-Low), VOC-TCLVOA-10(8260D).

#### **C. Analytical Techniques:**

VOCMS Group1 : The analysis performed on instrument MSVOA\_N were done using GC column Rxi-624SIL MS 30m, 0.25mm, 1.4 um, Cat. #13868.The analysis of VOCMS Group1 was based on method 8260-Low.

VOC-TCLVOA-10 : The analysis performed on instrument MSVOA\_N were done using GC column Rxi-624SIL MS 30m, 0.25mm, 1.4 um, Cat. #13868.The analysis performed on instrument MSVOA\_Y were done using GC column Rxi-624SIL MS 30m, 0.25mm, 1.4 um, Cat. #13868.The analysis of VOC-TCLVOA-10 was based on method 8260D.

#### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis except following  
VOC-TCLVOA-10 : RPXY4 [1,2-Dichloroethane-d4 - 136%] this compound did not meet the NJDKQP criteria but met the in-house criteria.

The Internal Standards Areas were met for all analysis.

The Retention Times were met for all analysis.

The RPD were met for all analysis.

The Blank Spike met requirements for all compounds.

The Blank Spike Duplicate met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements except following

VOCMS Group1 : The %RSD is greater than 20% in the Initial Calibration method (82N082125W.M) for Methylene Chloride passing on Linear regression.

VOC-TCLVOA-10 : The %RSD is greater than 20% in the Initial Calibration method (82N082125W.M) for Methylene Chloride passing on Linear regression.

The Continuous Calibration met the requirements.  
The Tuning criteria met requirements.

VOC-TCLVOA-10 : Sample RPXY5 was diluted due to high concentration.

**E. Additional Comments:**

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

Trip Blank was not provided with this set of samples.

The soil samples results are based on a dry weight basis.

The Sample # RPXY4 have the concentration of target compound below Method detection limits, therefore it is not reported as Hit in Form1.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

**F. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

### **G Environmental**

**Project Name: Buffington**

**Project # N/A**

**Order ID # Q3058**

**Test Name: Metals Group4**

### **A. Number of Samples and Date of Receipt:**

2 Solid samples were received on 09/09/2025.

1 Water sample was received on 09/09/2025.

### **B. Parameters:**

According to the Chain of Custody document, the following analyses were requested: Metals Group4. This data package contains results for Metals Group4.

### **C. Analytical Techniques:**

The analysis of Metals Group4 was based on method 6010D and digestion based on method 3010 (waters).

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Blank Spike met requirements for all compounds.

The Duplicate (4089DUP) analysis met criteria for all compounds except for Iron, Manganese and Sodium due to sample matrix interference.

The Matrix Spike (4089MS) analysis met criteria for all compounds except for Iron due to Chemical interference during Digestion Process.

The Matrix Spike Duplicate (4089MSD) analysis met criteria for all compounds except for Iron due to Chemical interference during Digestion Process.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.

### **E. Additional Comments:**

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_

## DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following “ Results Qualifiers” are used:

|           |                                                                                                                                                                                                                                                                                                                                                                               |
|-----------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <b>J</b>  | Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL).                                                                                                                                                                                  |
| <b>U</b>  | Indicates the analyte was analyzed for, but not detected.                                                                                                                                                                                                                                                                                                                     |
| <b>ND</b> | Indicates the analyte was analyzed for, but not detected                                                                                                                                                                                                                                                                                                                      |
| <b>E</b>  | Indicates the reported value is estimated because of the presence of interference                                                                                                                                                                                                                                                                                             |
| <b>M</b>  | Indicates Duplicate injection precision not met.                                                                                                                                                                                                                                                                                                                              |
| <b>N</b>  | Indicates the spiked sample recovery is not within control limits.                                                                                                                                                                                                                                                                                                            |
| <b>S</b>  | Indicates the reported value was determined by the Method of Standard Addition (MSA).                                                                                                                                                                                                                                                                                         |
| <b>*</b>  | Indicates that the duplicate analysis is not within control limits.                                                                                                                                                                                                                                                                                                           |
| <b>+</b>  | Indicates the correlation coefficient for the MSA is less than 0.995.                                                                                                                                                                                                                                                                                                         |
| <b>D</b>  | Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.                                                                                                                                                                                                                                       |
| <b>M</b>  | Method qualifiers<br>“P” for ICP instrument<br>“PM” for ICP when Microwave Digestion is used<br>“CV” for Manual Cold Vapor AA<br>“AV” for automated Cold Vapor AA<br>“CA” for MIDI-Distillation Spectrophotometric<br>“AS” for Semi -Automated Spectrophotometric<br>“C” for Manual Spectrophotometric<br>“T” for Titrimetric<br>“NR” for analyte not required to be analyzed |
| <b>OR</b> | Indicates the analyte’s concentration exceeds the calibrated range of the instrument for that specific analysis.                                                                                                                                                                                                                                                              |
| <b>Q</b>  | Indicates the LCS did not meet the control limits requirements                                                                                                                                                                                                                                                                                                                |
| <b>H</b>  | Sample Analysis Out Of Hold Time                                                                                                                                                                                                                                                                                                                                              |

## DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “Results Qualifiers” are used:

|       |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                |
|-------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Value | If the result is a value greater than or equal to the detection limit, report the value                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                        |
| U     | Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.                                                                                                                                                                                                                                                                                      |
| ND    | Indicates the analyte was analyzed for, but not detected                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                       |
| J     | Indicates an estimated value. This flag is used: <ul style="list-style-type: none"> <li>(1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.)</li> <li>(2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.</li> </ul> |
| B     | Indicates the analyte was found in the blank as well as the sample report as “12 B”.                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                           |
| E     | Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.                                                                                                                                                                                                                                                                                                                                                                                                                                                                              |
| D     | This flag identifies all compounds identified in an analysis at a secondary dilution factor.                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                   |
| P     | This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.                                                                                                                                                                                                                                                                                                                                                                             |
| N     | This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.                                                                                                                                                                                                                                                                       |
| A     | This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.                                                                                                                                                                                                                                                                                                                                                                                                                                                                                          |
| Q     | Indicates the LCS did not meet the control limits requirements                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                 |

## APPENDIX A

### QA REVIEW GENERAL DOCUMENTATION

Project #: Q3058

Completed

For thorough review, the report must have the following:

#### GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

#### COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

#### CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

#### ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 09/19/2025

**Hit Summary Sheet**  
**SW-846**

**SDG No.:** Q3058  
**Client:** G Environmental

| Sample ID         | Client ID      | Matrix | Parameter                   | Concentration | C  | MDL  | RDL  | Units |
|-------------------|----------------|--------|-----------------------------|---------------|----|------|------|-------|
| <b>Client ID:</b> | <b>RPXY5</b>   |        |                             |               |    |      |      |       |
| Q3058-01          | RPXY5          | SOIL   | Acetone                     | 10.2          | J  | 4.30 | 22.5 | ug/Kg |
| Q3058-01          | RPXY5          | SOIL   | Methylene Chloride          | 3.50          | J  | 3.20 | 9.00 | ug/Kg |
| Q3058-01          | RPXY5          | SOIL   | Toluene                     | 2.30          | J  | 0.70 | 4.50 | ug/Kg |
| Q3058-01          | RPXY5          | SOIL   | Ethyl Benzene               | 95.9          |    | 0.60 | 4.50 | ug/Kg |
| Q3058-01          | RPXY5          | SOIL   | m/p-Xylenes                 | 340           | E  | 1.10 | 9.00 | ug/Kg |
| Q3058-01          | RPXY5          | SOIL   | o-Xylene                    | 87.3          |    | 0.74 | 4.50 | ug/Kg |
|                   |                |        | <b>Total Voc :</b>          |               |    | 539  |      |       |
|                   |                |        | <b>Total Concentration:</b> |               |    | 539  |      |       |
| <b>Client ID:</b> | <b>RPXY5ME</b> |        |                             |               |    |      |      |       |
| Q3058-01ME        | RPXY5ME        | SOIL   | Ethyl Benzene               | 72.9          | JD | 31.5 | 240  | ug/Kg |
| Q3058-01ME        | RPXY5ME        | SOIL   | m/p-Xylenes                 | 220           | JD | 58.3 | 470  | ug/Kg |
| Q3058-01ME        | RPXY5ME        | SOIL   | o-Xylene                    | 60.7          | JD | 38.5 | 240  | ug/Kg |
|                   |                |        | <b>Total Voc :</b>          |               |    | 354  |      |       |
|                   |                |        | <b>Total Concentration:</b> |               |    | 354  |      |       |
| <b>Client ID:</b> | <b>RPXY4</b>   |        |                             |               |    |      |      |       |
| Q3058-02          | RPXY4          | SOIL   | Chloromethane               | 120           |    | 0.94 | 4.10 | ug/Kg |
| Q3058-02          | RPXY4          | SOIL   | Bromomethane                | 13.5          |    | 0.88 | 4.10 | ug/Kg |
| Q3058-02          | RPXY4          | SOIL   | Acetone                     | 140           |    | 3.90 | 20.6 | ug/Kg |
| Q3058-02          | RPXY4          | SOIL   | Methylene Chloride          | 4.30          | J  | 2.90 | 8.30 | ug/Kg |
|                   |                |        | <b>Total Voc :</b>          |               |    | 278  |      |       |
| Q3058-02          | RPXY4          | SOIL   | Methyl Iodide               | * 3.00        | J  | 0.75 | 4.10 | ug/Kg |
|                   |                |        | <b>Total Tics :</b>         |               |    | 3.00 |      |       |
|                   |                |        | <b>Total Concentration:</b> |               |    | 281  |      |       |



# SAMPLE DATA

A

B

C

D

E

F

G

H

I

J

## Report of Analysis

|                    |                 |                 |                     |
|--------------------|-----------------|-----------------|---------------------|
| Client:            | G Environmental | Date Collected: | 09/09/25            |
| Project:           | Buffington      | Date Received:  | 09/09/25            |
| Client Sample ID:  | RPXY5           | SDG No.:        | Q3058               |
| Lab Sample ID:     | Q3058-01        | Matrix:         | SOIL                |
| Analytical Method: | 8260D           | % Solid:        | 84.7                |
| Sample Wt/Vol:     | 6.55            | Units: g        | Final Vol: 5000 uL  |
| Soil Aliquot Vol:  |                 | uL              | Test: VOC-TCLVOA-10 |
| GC Column:         | RXI-624         | ID : 0.25       | Level : LOW         |
| Prep Method :      |                 |                 |                     |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VY023328.D        | 1         | 09/10/25 12:49 | VY091025      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units(Dry Weight) |
|----------------|--------------------------------|-------|-----------|------|------------|-------------------|
| <b>TARGETS</b> |                                |       |           |      |            |                   |
| 75-71-8        | Dichlorodifluoromethane        | 1.00  | U         | 1.00 | 4.50       | ug/Kg             |
| 74-87-3        | Chloromethane                  | 1.00  | U         | 1.00 | 4.50       | ug/Kg             |
| 75-01-4        | Vinyl Chloride                 | 0.71  | U         | 0.71 | 4.50       | ug/Kg             |
| 74-83-9        | Bromomethane                   | 0.96  | U         | 0.96 | 4.50       | ug/Kg             |
| 75-00-3        | Chloroethane                   | 1.10  | U         | 1.10 | 4.50       | ug/Kg             |
| 75-69-4        | Trichlorofluoromethane         | 1.10  | U         | 1.10 | 4.50       | ug/Kg             |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 0.96  | U         | 0.96 | 4.50       | ug/Kg             |
| 75-35-4        | 1,1-Dichloroethene             | 0.90  | U         | 0.90 | 4.50       | ug/Kg             |
| 67-64-1        | Acetone                        | 10.2  | J         | 4.30 | 22.5       | ug/Kg             |
| 75-15-0        | Carbon Disulfide               | 0.96  | U         | 0.96 | 4.50       | ug/Kg             |
| 1634-04-4      | Methyl tert-butyl Ether        | 0.66  | U         | 0.66 | 4.50       | ug/Kg             |
| 79-20-9        | Methyl Acetate                 | 1.40  | U         | 1.40 | 4.50       | ug/Kg             |
| 75-09-2        | Methylene Chloride             | 3.50  | J         | 3.20 | 9.00       | ug/Kg             |
| 156-60-5       | trans-1,2-Dichloroethene       | 0.78  | U         | 0.78 | 4.50       | ug/Kg             |
| 75-34-3        | 1,1-Dichloroethane             | 0.72  | U         | 0.72 | 4.50       | ug/Kg             |
| 110-82-7       | Cyclohexane                    | 0.71  | U         | 0.71 | 4.50       | ug/Kg             |
| 78-93-3        | 2-Butanone                     | 5.90  | U         | 5.90 | 22.5       | ug/Kg             |
| 56-23-5        | Carbon Tetrachloride           | 0.87  | U         | 0.87 | 4.50       | ug/Kg             |
| 156-59-2       | cis-1,2-Dichloroethene         | 0.68  | U         | 0.68 | 4.50       | ug/Kg             |
| 74-97-5        | Bromochloromethane             | 1.00  | U         | 1.00 | 4.50       | ug/Kg             |
| 67-66-3        | Chloroform                     | 0.76  | U         | 0.76 | 4.50       | ug/Kg             |
| 71-55-6        | 1,1,1-Trichloroethane          | 0.84  | U         | 0.84 | 4.50       | ug/Kg             |
| 108-87-2       | Methylcyclohexane              | 0.82  | U         | 0.82 | 4.50       | ug/Kg             |
| 71-43-2        | Benzene                        | 0.71  | U         | 0.71 | 4.50       | ug/Kg             |
| 107-06-2       | 1,2-Dichloroethane             | 0.71  | U         | 0.71 | 4.50       | ug/Kg             |
| 79-01-6        | Trichloroethene                | 0.73  | U         | 0.73 | 4.50       | ug/Kg             |
| 78-87-5        | 1,2-Dichloropropane            | 0.82  | U         | 0.82 | 4.50       | ug/Kg             |
| 75-27-4        | Bromodichloromethane           | 0.70  | U         | 0.70 | 4.50       | ug/Kg             |
| 108-10-1       | 4-Methyl-2-Pentanone           | 3.20  | U         | 3.20 | 22.5       | ug/Kg             |
| 108-88-3       | Toluene                        | 2.30  | J         | 0.70 | 4.50       | ug/Kg             |

## Report of Analysis

|                    |                                   |                 |                      |
|--------------------|-----------------------------------|-----------------|----------------------|
| Client:            | G Environmental                   | Date Collected: | 09/09/25             |
| Project:           | Buffington                        | Date Received:  | 09/09/25             |
| Client Sample ID:  | RPXY5                             | SDG No.:        | Q3058                |
| Lab Sample ID:     | Q3058-01                          | Matrix:         | SOIL                 |
| Analytical Method: | 8260D                             | % Solid:        | 84.7                 |
| Sample Wt/Vol:     | 6.55      Units:    g             | Final Vol:      | 5000              uL |
| Soil Aliquot Vol:  | uL                                | Test:           | VOC-TCLVOA-10        |
| GC Column:         | RXI-624              ID :    0.25 | Level :         | LOW                  |
| Prep Method :      |                                   |                 |                      |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VY023328.D        | 1         | 09/10/25 12:49 | VY091025      |

| CAS Number                | Parameter                   | Conc.   | Qualifier | MDL                 | LOQ / CRQL | Units(Dry Weight) |
|---------------------------|-----------------------------|---------|-----------|---------------------|------------|-------------------|
| 10061-02-6                | t-1,3-Dichloropropene       | 0.59    | U         | 0.59                | 4.50       | ug/Kg             |
| 10061-01-5                | cis-1,3-Dichloropropene     | 0.56    | U         | 0.56                | 4.50       | ug/Kg             |
| 79-00-5                   | 1,1,2-Trichloroethane       | 0.83    | U         | 0.83                | 4.50       | ug/Kg             |
| 591-78-6                  | 2-Hexanone                  | 3.30    | U         | 3.30                | 22.5       | ug/Kg             |
| 124-48-1                  | Dibromochloromethane        | 0.78    | U         | 0.78                | 4.50       | ug/Kg             |
| 106-93-4                  | 1,2-Dibromoethane           | 0.79    | U         | 0.79                | 4.50       | ug/Kg             |
| 127-18-4                  | Tetrachloroethene           | 0.95    | U         | 0.95                | 4.50       | ug/Kg             |
| 108-90-7                  | Chlorobenzene               | 0.82    | U         | 0.82                | 4.50       | ug/Kg             |
| 100-41-4                  | Ethyl Benzene               | 95.9    |           | 0.60                | 4.50       | ug/Kg             |
| 179601-23-1               | m/p-Xylenes                 | 340     | E         | 1.10                | 9.00       | ug/Kg             |
| 95-47-6                   | o-Xylene                    | 87.3    |           | 0.74                | 4.50       | ug/Kg             |
| 100-42-5                  | Styrene                     | 0.64    | U         | 0.64                | 4.50       | ug/Kg             |
| 75-25-2                   | Bromoform                   | 0.78    | U         | 0.78                | 4.50       | ug/Kg             |
| 98-82-8                   | Isopropylbenzene            | 0.70    | U         | 0.70                | 4.50       | ug/Kg             |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 1.10    | U         | 1.10                | 4.50       | ug/Kg             |
| 541-73-1                  | 1,3-Dichlorobenzene         | 1.50    | U         | 1.50                | 4.50       | ug/Kg             |
| 106-46-7                  | 1,4-Dichlorobenzene         | 1.40    | U         | 1.40                | 4.50       | ug/Kg             |
| 95-50-1                   | 1,2-Dichlorobenzene         | 1.30    | U         | 1.30                | 4.50       | ug/Kg             |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 1.70    | U         | 1.70                | 4.50       | ug/Kg             |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 2.70    | U         | 2.70                | 4.50       | ug/Kg             |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 2.90    | U         | 2.90                | 4.50       | ug/Kg             |
| <b>SURROGATES</b>         |                             |         |           |                     |            |                   |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 62.0    |           | 70 (63) - 130 (155) | 124%       | SPK: 50           |
| 1868-53-7                 | Dibromofluoromethane        | 54.2    |           | 70 (70) - 130 (134) | 108%       | SPK: 50           |
| 2037-26-5                 | Toluene-d8                  | 52.1    |           | 70 (74) - 130 (123) | 104%       | SPK: 50           |
| 460-00-4                  | 4-Bromofluorobenzene        | 54.6    |           | 70 (17) - 130 (146) | 109%       | SPK: 50           |
| <b>INTERNAL STANDARDS</b> |                             |         |           |                     |            |                   |
| 363-72-4                  | Pentafluorobenzene          | 542000  | 7.713     |                     |            |                   |
| 540-36-3                  | 1,4-Difluorobenzene         | 1030000 | 8.615     |                     |            |                   |
| 3114-55-4                 | Chlorobenzene-d5            | 977000  | 11.414    |                     |            |                   |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 419000  | 13.346    |                     |            |                   |

## Report of Analysis

|                    |                 |           |                 |               |    |
|--------------------|-----------------|-----------|-----------------|---------------|----|
| Client:            | G Environmental |           | Date Collected: | 09/09/25      |    |
| Project:           | Buffington      |           | Date Received:  | 09/09/25      |    |
| Client Sample ID:  | RPXY5           |           | SDG No.:        | Q3058         |    |
| Lab Sample ID:     | Q3058-01        |           | Matrix:         | SOIL          |    |
| Analytical Method: | 8260D           |           | % Solid:        | 84.7          |    |
| Sample Wt/Vol:     | 6.55            | Units: g  | Final Vol:      | 5000          | uL |
| Soil Aliquot Vol:  |                 | uL        | Test:           | VOC-TCLVOA-10 |    |
| GC Column:         | RXI-624         | ID : 0.25 | Level :         | LOW           |    |
| Prep Method :      |                 |           |                 |               |    |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VY023328.D        | 1         | 09/10/25 12:49 | VY091025      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

### Report of Analysis

|                    |                 |           |                 |               |    |
|--------------------|-----------------|-----------|-----------------|---------------|----|
| Client:            | G Environmental |           | Date Collected: | 09/09/25      |    |
| Project:           | Buffington      |           | Date Received:  | 09/09/25      |    |
| Client Sample ID:  | RPXY5ME         |           | SDG No.:        | Q3058         |    |
| Lab Sample ID:     | Q3058-01ME      |           | Matrix:         | SOIL          |    |
| Analytical Method: | 8260D           |           | % Solid:        | 84.7          |    |
| Sample Wt/Vol:     | 6.28            | Units: g  | Final Vol:      | 5000          | uL |
| Soil Aliquot Vol:  | 100             | uL        | Test:           | VOC-TCLVOA-10 |    |
| GC Column:         | RXI-624         | ID : 0.25 | Level :         | MED           |    |
| Prep Method :      |                 |           |                 |               |    |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VN087839.D        | 1         | 09/15/25 15:43 | VN091525      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units(Dry Weight) |
|----------------|--------------------------------|-------|-----------|------|------------|-------------------|
| <b>TARGETS</b> |                                |       |           |      |            |                   |
| 75-71-8        | Dichlorodifluoromethane        | 53.6  | UD        | 53.6 | 240        | ug/Kg             |
| 74-87-3        | Chloromethane                  | 53.6  | UD        | 53.6 | 240        | ug/Kg             |
| 75-01-4        | Vinyl Chloride                 | 37.1  | UD        | 37.1 | 240        | ug/Kg             |
| 74-83-9        | Bromomethane                   | 50.3  | UD        | 50.3 | 240        | ug/Kg             |
| 75-00-3        | Chloroethane                   | 59.2  | UD        | 59.2 | 240        | ug/Kg             |
| 75-69-4        | Trichlorofluoromethane         | 56.9  | UD        | 56.9 | 240        | ug/Kg             |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 49.8  | UD        | 49.8 | 240        | ug/Kg             |
| 75-35-4        | 1,1-Dichloroethene             | 47.0  | UD        | 47.0 | 240        | ug/Kg             |
| 67-64-1        | Acetone                        | 220   | UD        | 220  | 1200       | ug/Kg             |
| 75-15-0        | Carbon Disulfide               | 49.8  | UD        | 49.8 | 240        | ug/Kg             |
| 1634-04-4      | Methyl tert-butyl Ether        | 34.3  | UD        | 34.3 | 240        | ug/Kg             |
| 79-20-9        | Methyl Acetate                 | 72.4  | UD        | 72.4 | 240        | ug/Kg             |
| 75-09-2        | Methylene Chloride             | 170   | UD        | 170  | 470        | ug/Kg             |
| 156-60-5       | trans-1,2-Dichloroethene       | 40.4  | UD        | 40.4 | 240        | ug/Kg             |
| 75-34-3        | 1,1-Dichloroethane             | 37.6  | UD        | 37.6 | 240        | ug/Kg             |
| 110-82-7       | Cyclohexane                    | 37.1  | UD        | 37.1 | 240        | ug/Kg             |
| 78-93-3        | 2-Butanone                     | 310   | UD        | 310  | 1200       | ug/Kg             |
| 56-23-5        | Carbon Tetrachloride           | 45.6  | UD        | 45.6 | 240        | ug/Kg             |
| 156-59-2       | cis-1,2-Dichloroethene         | 35.2  | UD        | 35.2 | 240        | ug/Kg             |
| 74-97-5        | Bromochloromethane             | 54.0  | UD        | 54.0 | 240        | ug/Kg             |
| 67-66-3        | Chloroform                     | 39.5  | UD        | 39.5 | 240        | ug/Kg             |
| 71-55-6        | 1,1,1-Trichloroethane          | 43.7  | UD        | 43.7 | 240        | ug/Kg             |
| 108-87-2       | Methylcyclohexane              | 42.8  | UD        | 42.8 | 240        | ug/Kg             |
| 71-43-2        | Benzene                        | 37.1  | UD        | 37.1 | 240        | ug/Kg             |
| 107-06-2       | 1,2-Dichloroethane             | 37.1  | UD        | 37.1 | 240        | ug/Kg             |
| 79-01-6        | Trichloroethene                | 38.1  | UD        | 38.1 | 240        | ug/Kg             |
| 78-87-5        | 1,2-Dichloropropane            | 42.8  | UD        | 42.8 | 240        | ug/Kg             |
| 75-27-4        | Bromodichloromethane           | 36.7  | UD        | 36.7 | 240        | ug/Kg             |
| 108-10-1       | 4-Methyl-2-Pentanone           | 170   | UD        | 170  | 1200       | ug/Kg             |
| 108-88-3       | Toluene                        | 36.7  | UD        | 36.7 | 240        | ug/Kg             |

## Report of Analysis

|                    |                                   |                 |                      |
|--------------------|-----------------------------------|-----------------|----------------------|
| Client:            | G Environmental                   | Date Collected: | 09/09/25             |
| Project:           | Buffington                        | Date Received:  | 09/09/25             |
| Client Sample ID:  | RPXY5ME                           | SDG No.:        | Q3058                |
| Lab Sample ID:     | Q3058-01ME                        | Matrix:         | SOIL                 |
| Analytical Method: | 8260D                             | % Solid:        | 84.7                 |
| Sample Wt/Vol:     | 6.28      Units:    g             | Final Vol:      | 5000              uL |
| Soil Aliquot Vol:  | 100                      uL       | Test:           | VOC-TCLVOA-10        |
| GC Column:         | RXI-624              ID :    0.25 | Level :         | MED                  |
| Prep Method :      |                                   |                 |                      |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VN087839.D        | 1         | 09/15/25 15:43 | VN091525      |

| CAS Number                | Parameter                   | Conc.  | Qualifier | MDL                 | LOQ / CRQL | Units(Dry Weight) |
|---------------------------|-----------------------------|--------|-----------|---------------------|------------|-------------------|
| 10061-02-6                | t-1,3-Dichloropropene       | 30.5   | UD        | 30.5                | 240        | ug/Kg             |
| 10061-01-5                | cis-1,3-Dichloropropene     | 29.1   | UD        | 29.1                | 240        | ug/Kg             |
| 79-00-5                   | 1,1,2-Trichloroethane       | 43.2   | UD        | 43.2                | 240        | ug/Kg             |
| 591-78-6                  | 2-Hexanone                  | 170    | UD        | 170                 | 1200       | ug/Kg             |
| 124-48-1                  | Dibromochloromethane        | 40.9   | UD        | 40.9                | 240        | ug/Kg             |
| 106-93-4                  | 1,2-Dibromoethane           | 41.4   | UD        | 41.4                | 240        | ug/Kg             |
| 127-18-4                  | Tetrachloroethene           | 49.3   | UD        | 49.3                | 240        | ug/Kg             |
| 108-90-7                  | Chlorobenzene               | 42.8   | UD        | 42.8                | 240        | ug/Kg             |
| 100-41-4                  | Ethyl Benzene               | 72.9   | JD        | 31.5                | 240        | ug/Kg             |
| 179601-23-1               | m/p-Xylenes                 | 220    | JD        | 58.3                | 470        | ug/Kg             |
| 95-47-6                   | o-Xylene                    | 60.7   | JD        | 38.5                | 240        | ug/Kg             |
| 100-42-5                  | Styrene                     | 33.4   | UD        | 33.4                | 240        | ug/Kg             |
| 75-25-2                   | Bromoform                   | 40.4   | UD        | 40.4                | 240        | ug/Kg             |
| 98-82-8                   | Isopropylbenzene            | 36.7   | UD        | 36.7                | 240        | ug/Kg             |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 56.9   | UD        | 56.9                | 240        | ug/Kg             |
| 541-73-1                  | 1,3-Dichlorobenzene         | 80.4   | UD        | 80.4                | 240        | ug/Kg             |
| 106-46-7                  | 1,4-Dichlorobenzene         | 73.3   | UD        | 73.3                | 240        | ug/Kg             |
| 95-50-1                   | 1,2-Dichlorobenzene         | 68.1   | UD        | 68.1                | 240        | ug/Kg             |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 86.5   | UD        | 86.5                | 240        | ug/Kg             |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 140    | UD        | 140                 | 240        | ug/Kg             |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 150    | UD        | 150                 | 240        | ug/Kg             |
| <b>SURROGATES</b>         |                             |        |           |                     |            |                   |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 50.6   |           | 70 (63) - 130 (155) | 101%       | SPK: 50           |
| 1868-53-7                 | Dibromofluoromethane        | 50.6   |           | 70 (70) - 130 (134) | 101%       | SPK: 50           |
| 2037-26-5                 | Toluene-d8                  | 46.8   |           | 70 (74) - 130 (123) | 94%        | SPK: 50           |
| 460-00-4                  | 4-Bromofluorobenzene        | 39.8   |           | 70 (17) - 130 (146) | 80%        | SPK: 50           |
| <b>INTERNAL STANDARDS</b> |                             |        |           |                     |            |                   |
| 363-72-4                  | Pentafluorobenzene          | 191000 | 8.2       |                     |            |                   |
| 540-36-3                  | 1,4-Difluorobenzene         | 435000 | 9.082     |                     |            |                   |
| 3114-55-4                 | Chlorobenzene-d5            | 387000 | 11.847    |                     |            |                   |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 170000 | 13.77     |                     |            |                   |

## Report of Analysis

|                    |                 |           |                 |               |    |
|--------------------|-----------------|-----------|-----------------|---------------|----|
| Client:            | G Environmental |           | Date Collected: | 09/09/25      |    |
| Project:           | Buffington      |           | Date Received:  | 09/09/25      |    |
| Client Sample ID:  | RPXY5ME         |           | SDG No.:        | Q3058         |    |
| Lab Sample ID:     | Q3058-01ME      |           | Matrix:         | SOIL          |    |
| Analytical Method: | 8260D           |           | % Solid:        | 84.7          |    |
| Sample Wt/Vol:     | 6.28            | Units: g  | Final Vol:      | 5000          | uL |
| Soil Aliquot Vol:  | 100             | uL        | Test:           | VOC-TCLVOA-10 |    |
| GC Column:         | RXI-624         | ID : 0.25 | Level :         | MED           |    |
| Prep Method :      |                 |           |                 |               |    |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VN087839.D        | 1         | 09/15/25 15:43 | VN091525      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected  
 LOQ = Limit of Quantitation  
 MDL = Method Detection Limit  
 LOD = Limit of Detection  
 E = Value Exceeds Calibration Range  
 Q = indicates LCS control criteria did not meet requirements  
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value  
 B = Analyte Found in Associated Method Blank  
 N = Presumptive Evidence of a Compound  
 \* = Values outside of QC limits  
 D = Dilution  
 () = Laboratory InHouse Limit  
 A = Aldol-Condensation Reaction Products

### Report of Analysis

|                    |                 |           |                 |               |    |
|--------------------|-----------------|-----------|-----------------|---------------|----|
| Client:            | G Environmental |           | Date Collected: | 09/09/25      |    |
| Project:           | Buffington      |           | Date Received:  | 09/09/25      |    |
| Client Sample ID:  | RPXY4           |           | SDG No.:        | Q3058         |    |
| Lab Sample ID:     | Q3058-02        |           | Matrix:         | SOIL          |    |
| Analytical Method: | 8260D           |           | % Solid:        | 85.9          |    |
| Sample Wt/Vol:     | 7.05            | Units: g  | Final Vol:      | 5000          | uL |
| Soil Aliquot Vol:  |                 | uL        | Test:           | VOC-TCLVOA-10 |    |
| GC Column:         | RXI-624         | ID : 0.25 | Level :         | LOW           |    |
| Prep Method :      |                 |           |                 |               |    |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VY023329.D        | 1         | 09/10/25 13:12 | VY091025      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units(Dry Weight) |
|----------------|--------------------------------|-------|-----------|------|------------|-------------------|
| <b>TARGETS</b> |                                |       |           |      |            |                   |
| 75-71-8        | Dichlorodifluoromethane        | 0.94  | U         | 0.94 | 4.10       | ug/Kg             |
| 74-87-3        | Chloromethane                  | 120   |           | 0.94 | 4.10       | ug/Kg             |
| 75-01-4        | Vinyl Chloride                 | 0.65  | U         | 0.65 | 4.10       | ug/Kg             |
| 74-83-9        | Bromomethane                   | 13.5  |           | 0.88 | 4.10       | ug/Kg             |
| 75-00-3        | Chloroethane                   | 1.00  | U         | 1.00 | 4.10       | ug/Kg             |
| 75-69-4        | Trichlorofluoromethane         | 1.00  | U         | 1.00 | 4.10       | ug/Kg             |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 0.88  | U         | 0.88 | 4.10       | ug/Kg             |
| 75-35-4        | 1,1-Dichloroethene             | 0.83  | U         | 0.83 | 4.10       | ug/Kg             |
| 67-64-1        | Acetone                        | 140   |           | 3.90 | 20.6       | ug/Kg             |
| 75-15-0        | Carbon Disulfide               | 0.88  | U         | 0.88 | 4.10       | ug/Kg             |
| 1634-04-4      | Methyl tert-butyl Ether        | 0.60  | U         | 0.60 | 4.10       | ug/Kg             |
| 79-20-9        | Methyl Acetate                 | 1.30  | U         | 1.30 | 4.10       | ug/Kg             |
| 75-09-2        | Methylene Chloride             | 4.30  | J         | 2.90 | 8.30       | ug/Kg             |
| 156-60-5       | trans-1,2-Dichloroethene       | 0.71  | U         | 0.71 | 4.10       | ug/Kg             |
| 75-34-3        | 1,1-Dichloroethane             | 0.66  | U         | 0.66 | 4.10       | ug/Kg             |
| 110-82-7       | Cyclohexane                    | 0.65  | U         | 0.65 | 4.10       | ug/Kg             |
| 78-93-3        | 2-Butanone                     | 5.40  | U         | 5.40 | 20.6       | ug/Kg             |
| 56-23-5        | Carbon Tetrachloride           | 0.80  | U         | 0.80 | 4.10       | ug/Kg             |
| 156-59-2       | cis-1,2-Dichloroethene         | 0.62  | U         | 0.62 | 4.10       | ug/Kg             |
| 74-97-5        | Bromochloromethane             | 0.95  | U         | 0.95 | 4.10       | ug/Kg             |
| 67-66-3        | Chloroform                     | 0.69  | U         | 0.69 | 4.10       | ug/Kg             |
| 71-55-6        | 1,1,1-Trichloroethane          | 0.77  | U         | 0.77 | 4.10       | ug/Kg             |
| 108-87-2       | Methylcyclohexane              | 0.75  | U         | 0.75 | 4.10       | ug/Kg             |
| 71-43-2        | Benzene                        | 0.65  | U         | 0.65 | 4.10       | ug/Kg             |
| 107-06-2       | 1,2-Dichloroethane             | 0.65  | U         | 0.65 | 4.10       | ug/Kg             |
| 79-01-6        | Trichloroethene                | 0.67  | U         | 0.67 | 4.10       | ug/Kg             |
| 78-87-5        | 1,2-Dichloropropane            | 0.75  | U         | 0.75 | 4.10       | ug/Kg             |
| 75-27-4        | Bromodichloromethane           | 0.64  | U         | 0.64 | 4.10       | ug/Kg             |
| 108-10-1       | 4-Methyl-2-Pentanone           | 3.00  | U         | 3.00 | 20.6       | ug/Kg             |
| 108-88-3       | Toluene                        | 0.64  | U         | 0.64 | 4.10       | ug/Kg             |

### Report of Analysis

|                    |                 |           |                 |               |    |
|--------------------|-----------------|-----------|-----------------|---------------|----|
| Client:            | G Environmental |           | Date Collected: | 09/09/25      |    |
| Project:           | Buffington      |           | Date Received:  | 09/09/25      |    |
| Client Sample ID:  | RPXY4           |           | SDG No.:        | Q3058         |    |
| Lab Sample ID:     | Q3058-02        |           | Matrix:         | SOIL          |    |
| Analytical Method: | 8260D           |           | % Solid:        | 85.9          |    |
| Sample Wt/Vol:     | 7.05            | Units: g  | Final Vol:      | 5000          | uL |
| Soil Aliquot Vol:  |                 | uL        | Test:           | VOC-TCLVOA-10 |    |
| GC Column:         | RXI-624         | ID : 0.25 | Level :         | LOW           |    |
| Prep Method :      |                 |           |                 |               |    |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VY023329.D        | 1         | 09/10/25 13:12 | VY091025      |

| CAS Number                            | Parameter                   | Conc.   | Qualifier | MDL                 | LOQ / CRQL | Units(Dry Weight) |
|---------------------------------------|-----------------------------|---------|-----------|---------------------|------------|-------------------|
| 10061-02-6                            | t-1,3-Dichloropropene       | 0.54    | U         | 0.54                | 4.10       | ug/Kg             |
| 10061-01-5                            | cis-1,3-Dichloropropene     | 0.51    | U         | 0.51                | 4.10       | ug/Kg             |
| 79-00-5                               | 1,1,2-Trichloroethane       | 0.76    | U         | 0.76                | 4.10       | ug/Kg             |
| 591-78-6                              | 2-Hexanone                  | 3.00    | U         | 3.00                | 20.6       | ug/Kg             |
| 124-48-1                              | Dibromochloromethane        | 0.72    | U         | 0.72                | 4.10       | ug/Kg             |
| 106-93-4                              | 1,2-Dibromoethane           | 0.73    | U         | 0.73                | 4.10       | ug/Kg             |
| 127-18-4                              | Tetrachloroethene           | 0.87    | U         | 0.87                | 4.10       | ug/Kg             |
| 108-90-7                              | Chlorobenzene               | 0.75    | U         | 0.75                | 4.10       | ug/Kg             |
| 100-41-4                              | Ethyl Benzene               | 0.55    | U         | 0.55                | 4.10       | ug/Kg             |
| 179601-23-1                           | m/p-Xylenes                 | 1.00    | U         | 1.00                | 8.30       | ug/Kg             |
| 95-47-6                               | o-Xylene                    | 0.68    | U         | 0.68                | 4.10       | ug/Kg             |
| 100-42-5                              | Styrene                     | 0.59    | U         | 0.59                | 4.10       | ug/Kg             |
| 75-25-2                               | Bromoform                   | 0.71    | U         | 0.71                | 4.10       | ug/Kg             |
| 98-82-8                               | Isopropylbenzene            | 0.64    | U         | 0.64                | 4.10       | ug/Kg             |
| 79-34-5                               | 1,1,2,2-Tetrachloroethane   | 1.00    | U         | 1.00                | 4.10       | ug/Kg             |
| 541-73-1                              | 1,3-Dichlorobenzene         | 1.40    | U         | 1.40                | 4.10       | ug/Kg             |
| 106-46-7                              | 1,4-Dichlorobenzene         | 1.30    | U         | 1.30                | 4.10       | ug/Kg             |
| 95-50-1                               | 1,2-Dichlorobenzene         | 1.20    | U         | 1.20                | 4.10       | ug/Kg             |
| 96-12-8                               | 1,2-Dibromo-3-Chloropropane | 1.50    | U         | 1.50                | 4.10       | ug/Kg             |
| 120-82-1                              | 1,2,4-Trichlorobenzene      | 2.50    | U         | 2.50                | 4.10       | ug/Kg             |
| 87-61-6                               | 1,2,3-Trichlorobenzene      | 2.60    | U         | 2.60                | 4.10       | ug/Kg             |
| <b>SURROGATES</b>                     |                             |         |           |                     |            |                   |
| 17060-07-0                            | 1,2-Dichloroethane-d4       | 67.9    | *         | 70 (63) - 130 (155) | 136%       | SPK: 50           |
| 1868-53-7                             | Dibromofluoromethane        | 57.9    |           | 70 (70) - 130 (134) | 116%       | SPK: 50           |
| 2037-26-5                             | Toluene-d8                  | 50.9    |           | 70 (74) - 130 (123) | 102%       | SPK: 50           |
| 460-00-4                              | 4-Bromofluorobenzene        | 55.5    |           | 70 (17) - 130 (146) | 111%       | SPK: 50           |
| <b>INTERNAL STANDARDS</b>             |                             |         |           |                     |            |                   |
| 363-72-4                              | Pentafluorobenzene          | 539000  | 7.713     |                     |            |                   |
| 540-36-3                              | 1,4-Difluorobenzene         | 1020000 | 8.616     |                     |            |                   |
| 3114-55-4                             | Chlorobenzene-d5            | 1000000 | 11.414    |                     |            |                   |
| 3855-82-1                             | 1,4-Dichlorobenzene-d4      | 442000  | 13.346    |                     |            |                   |
| <b>TENTATIVE IDENTIFIED COMPOUNDS</b> |                             |         |           |                     |            |                   |

## Report of Analysis

|                    |                 |                 |               |
|--------------------|-----------------|-----------------|---------------|
| Client:            | G Environmental | Date Collected: | 09/09/25      |
| Project:           | Buffington      | Date Received:  | 09/09/25      |
| Client Sample ID:  | RPXY4           | SDG No.:        | Q3058         |
| Lab Sample ID:     | Q3058-02        | Matrix:         | SOIL          |
| Analytical Method: | 8260D           | % Solid:        | 85.9          |
| Sample Wt/Vol:     | 7.05            | Units:          | g             |
| Soil Aliquot Vol:  |                 | Final Vol:      | 5000 uL       |
|                    |                 | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624         | Level :         | LOW           |
| Prep Method :      | ID : 0.25       |                 |               |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VY023329.D        | 1         | 09/10/25 13:12 | VY091025      |

| CAS Number | Parameter     | Conc. | Qualifier | MDL | LOQ / CRQL | Units(Dry Weight) |
|------------|---------------|-------|-----------|-----|------------|-------------------|
| 74-88-4    | Methyl Iodide | 3.00  | J         |     | 4.01       | ug/Kg             |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



# QC SUMMARY

A

B

C

D

E

F

G

H

I

J

### Surrogate Summary

SDG No.: **Q3058**

Client: **G Environmental**

Analytical Method: **SW8260D**

| Lab Sample ID | Client ID    | Parameter             | Spike | Result | Recovery (%) | Qual | Limits (%) |           |
|---------------|--------------|-----------------------|-------|--------|--------------|------|------------|-----------|
|               |              |                       |       |        |              |      | Low        | High      |
| Q3058-01      | RPXY5        | 1,2-Dichloroethane-d4 | 50    | 62.0   | 124          |      | 70 (63)    | 130 (155) |
|               |              | Dibromofluoromethane  | 50    | 54.2   | 108          |      | 70 (70)    | 130 (134) |
|               |              | Toluene-d8            | 50    | 52.0   | 104          |      | 70 (74)    | 130 (123) |
|               |              | 4-Bromofluorobenzene  | 50    | 54.6   | 109          |      | 70 (17)    | 130 (146) |
| Q3058-01ME    | RPXY5ME      | 1,2-Dichloroethane-d4 | 50    | 50.6   | 101          |      | 70 (63)    | 130 (155) |
|               |              | Dibromofluoromethane  | 50    | 50.6   | 101          |      | 70 (70)    | 130 (134) |
|               |              | Toluene-d8            | 50    | 46.8   | 94           |      | 70 (74)    | 130 (123) |
|               |              | 4-Bromofluorobenzene  | 50    | 39.8   | 80           |      | 70 (17)    | 130 (146) |
| Q3058-02      | RPXY4        | 1,2-Dichloroethane-d4 | 50    | 67.9   | 136          | *    | 70 (63)    | 130 (155) |
|               |              | Dibromofluoromethane  | 50    | 57.9   | 116          |      | 70 (70)    | 130 (134) |
|               |              | Toluene-d8            | 50    | 50.9   | 102          |      | 70 (74)    | 130 (123) |
|               |              | 4-Bromofluorobenzene  | 50    | 55.5   | 111          |      | 70 (17)    | 130 (146) |
| VN0915MBL01   | VN0915MBL01  | 1,2-Dichloroethane-d4 | 50    | 49.9   | 100          |      | 70 (63)    | 130 (155) |
|               |              | Dibromofluoromethane  | 50    | 52.3   | 105          |      | 70 (70)    | 130 (134) |
|               |              | Toluene-d8            | 50    | 48.3   | 97           |      | 70 (74)    | 130 (123) |
|               |              | 4-Bromofluorobenzene  | 50    | 45.6   | 91           |      | 70 (17)    | 130 (146) |
| VN0915MBS01   | VN0915MBS01  | 1,2-Dichloroethane-d4 | 50    | 48.5   | 97           |      | 70 (63)    | 130 (155) |
|               |              | Dibromofluoromethane  | 50    | 49.4   | 99           |      | 70 (70)    | 130 (134) |
|               |              | Toluene-d8            | 50    | 45.6   | 91           |      | 70 (74)    | 130 (123) |
|               |              | 4-Bromofluorobenzene  | 50    | 46.4   | 93           |      | 70 (17)    | 130 (146) |
| VY0910SBL01   | VY0910SBL01  | 1,2-Dichloroethane-d4 | 50    | 61.2   | 122          |      | 70 (63)    | 130 (155) |
|               |              | Dibromofluoromethane  | 50    | 53.7   | 107          |      | 70 (70)    | 130 (134) |
|               |              | Toluene-d8            | 50    | 51.7   | 103          |      | 70 (74)    | 130 (123) |
|               |              | 4-Bromofluorobenzene  | 50    | 49.4   | 99           |      | 70 (17)    | 130 (146) |
| VY0910SBS01   | VY0910SBS01  | 1,2-Dichloroethane-d4 | 50    | 49.6   | 99           |      | 70 (63)    | 130 (155) |
|               |              | Dibromofluoromethane  | 50    | 49.3   | 99           |      | 70 (70)    | 130 (134) |
|               |              | Toluene-d8            | 50    | 50.5   | 101          |      | 70 (74)    | 130 (123) |
|               |              | 4-Bromofluorobenzene  | 50    | 50.6   | 101          |      | 70 (17)    | 130 (146) |
| VY0910SBSD01  | VY0910SBSD01 | 1,2-Dichloroethane-d4 | 50    | 48.3   | 97           |      | 70 (63)    | 130 (155) |
|               |              | Dibromofluoromethane  | 50    | 48.5   | 97           |      | 70 (70)    | 130 (134) |
|               |              | Toluene-d8            | 50    | 49.0   | 98           |      | 70 (74)    | 130 (123) |
|               |              | 4-Bromofluorobenzene  | 50    | 48.7   | 97           |      | 70 (17)    | 130 (146) |

( ) = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

|          |                        |                    |                   |
|----------|------------------------|--------------------|-------------------|
| SDG No.: | <u>Q3058</u>           | Analytical Method: | <u>SW8260D</u>    |
| Client:  | <u>G Environmental</u> | Datafile :         | <u>VN087837.D</u> |

| Lab Sample ID | Parameter                      | Spike | Result | Unit  | Rec | RPD | Qual | Low     | Limits<br>High | RPD |
|---------------|--------------------------------|-------|--------|-------|-----|-----|------|---------|----------------|-----|
| VN0915MBS01   | Dichlorodifluoromethane        | 2000  | 1800   | ug/Kg | 90  |     |      | 40 (64) | 160 (136)      |     |
|               | Chloromethane                  | 2000  | 1900   | ug/Kg | 95  |     |      | 40 (52) | 160 (151)      |     |
|               | Vinyl chloride                 | 2000  | 1900   | ug/Kg | 95  |     |      | 70 (56) | 130 (148)      |     |
|               | Bromomethane                   | 2000  | 1600   | ug/Kg | 80  |     |      | 40 (58) | 160 (141)      |     |
|               | Chloroethane                   | 2000  | 1800   | ug/Kg | 90  |     |      | 40 (69) | 160 (130)      |     |
|               | Trichlorofluoromethane         | 2000  | 2000   | ug/Kg | 100 |     |      | 40 (69) | 160 (134)      |     |
|               | 1,1,2-Trichlorotrifluoroethane | 2000  | 1800   | ug/Kg | 90  |     |      | 70 (81) | 130 (123)      |     |
|               | 1,1-Dichloroethene             | 2000  | 1800   | ug/Kg | 90  |     |      | 70 (79) | 130 (121)      |     |
|               | Acetone                        | 10000 | 7900   | ug/Kg | 79  |     |      | 40 (40) | 160 (171)      |     |
|               | Carbon disulfide               | 2000  | 2000   | ug/Kg | 100 |     |      | 40 (59) | 160 (130)      |     |
|               | Methyl tert-butyl Ether        | 2000  | 1800   | ug/Kg | 90  |     |      | 70 (77) | 130 (129)      |     |
|               | Methyl Acetate                 | 2000  | 2000   | ug/Kg | 100 |     |      | 70 (69) | 130 (149)      |     |
|               | Methylene Chloride             | 2000  | 1900   | ug/Kg | 95  |     |      | 70 (72) | 130 (131)      |     |
|               | trans-1,2-Dichloroethene       | 2000  | 1800   | ug/Kg | 90  |     |      | 70 (80) | 130 (123)      |     |
|               | 1,1-Dichloroethane             | 2000  | 1800   | ug/Kg | 90  |     |      | 70 (82) | 130 (123)      |     |
|               | Cyclohexane                    | 2000  | 1700   | ug/Kg | 85  |     |      | 70 (76) | 130 (122)      |     |
|               | 2-Butanone                     | 10000 | 8600   | ug/Kg | 86  |     |      | 40 (69) | 160 (131)      |     |
|               | Carbon Tetrachloride           | 2000  | 2000   | ug/Kg | 100 |     |      | 70 (76) | 130 (129)      |     |
|               | cis-1,2-Dichloroethene         | 2000  | 1800   | ug/Kg | 90  |     |      | 70 (82) | 130 (123)      |     |
|               | Bromochloromethane             | 2000  | 2200   | ug/Kg | 110 |     |      | 70 (80) | 130 (127)      |     |
|               | Chloroform                     | 2000  | 1900   | ug/Kg | 95  |     |      | 70 (82) | 130 (125)      |     |
|               | 1,1,1-Trichloroethane          | 2000  | 1800   | ug/Kg | 90  |     |      | 70 (80) | 130 (126)      |     |
|               | Methylcyclohexane              | 2000  | 1600   | ug/Kg | 80  |     |      | 70 (77) | 130 (123)      |     |
|               | Benzene                        | 2000  | 1900   | ug/Kg | 95  |     |      | 70 (84) | 130 (121)      |     |
|               | 1,2-Dichloroethane             | 2000  | 1800   | ug/Kg | 90  |     |      | 70 (81) | 130 (126)      |     |
|               | Trichloroethene                | 2000  | 1900   | ug/Kg | 95  |     |      | 70 (83) | 130 (122)      |     |
|               | 1,2-Dichloropropane            | 2000  | 1900   | ug/Kg | 95  |     |      | 70 (83) | 130 (122)      |     |
|               | Bromodichloromethane           | 2000  | 1900   | ug/Kg | 95  |     |      | 70 (82) | 130 (123)      |     |
|               | 4-Methyl-2-Pentanone           | 10000 | 9300   | ug/Kg | 93  |     |      | 40 (70) | 160 (135)      |     |
|               | Toluene                        | 2000  | 1900   | ug/Kg | 95  |     |      | 70 (83) | 130 (122)      |     |
|               | t-1,3-Dichloropropene          | 2000  | 1800   | ug/Kg | 90  |     |      | 70 (78) | 130 (124)      |     |
|               | cis-1,3-Dichloropropene        | 2000  | 1800   | ug/Kg | 90  |     |      | 70 (81) | 130 (122)      |     |
|               | 1,1,2-Trichloroethane          | 2000  | 1900   | ug/Kg | 95  |     |      | 70 (82) | 130 (125)      |     |
|               | 2-Hexanone                     | 10000 | 9400   | ug/Kg | 94  |     |      | 40 (66) | 160 (138)      |     |
|               | Dibromochloromethane           | 2000  | 1900   | ug/Kg | 95  |     |      | 70 (79) | 130 (125)      |     |
|               | 1,2-Dibromoethane              | 2000  | 1900   | ug/Kg | 95  |     |      | 70 (80) | 130 (125)      |     |
|               | Tetrachloroethene              | 2000  | 1800   | ug/Kg | 90  |     |      | 70 (83) | 130 (125)      |     |
|               | Chlorobenzene                  | 2000  | 1700   | ug/Kg | 85  |     |      | 70 (84) | 130 (122)      |     |
|               | Ethyl Benzene                  | 2000  | 1700   | ug/Kg | 85  |     |      | 70 (82) | 130 (124)      |     |
|               | m/p-Xylenes                    | 4000  | 3500   | ug/Kg | 88  |     |      | 70 (83) | 130 (124)      |     |
|               | o-Xylene                       | 2000  | 1700   | ug/Kg | 85  |     |      | 70 (83) | 130 (123)      |     |
|               | Styrene                        | 2000  | 1800   | ug/Kg | 90  |     |      | 70 (82) | 130 (124)      |     |
|               | Bromoform                      | 2000  | 1900   | ug/Kg | 95  |     |      | 70 (75) | 130 (127)      |     |
|               | Isopropylbenzene               | 2000  | 1600   | ug/Kg | 80  |     |      | 70 (82) | 130 (124)      |     |
|               | 1,1,2,2-Tetrachloroethane      | 2000  | 1700   | ug/Kg | 85  |     |      | 70 (77) | 130 (127)      |     |
|               | 1,3-Dichlorobenzene            | 2000  | 1700   | ug/Kg | 85  |     |      | 70 (83) | 130 (122)      |     |
|               | 1,4-Dichlorobenzene            | 2000  | 1700   | ug/Kg | 85  |     |      | 70 (84) | 130 (121)      |     |
|               | 1,2-Dichlorobenzene            | 2000  | 1700   | ug/Kg | 85  |     |      | 70 (83) | 130 (124)      |     |

() = LABORATORY INHOUSE LIMIT

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

|                 |                               |                           |                          |
|-----------------|-------------------------------|---------------------------|--------------------------|
| <b>SDG No.:</b> | <u><b>Q3058</b></u>           | <b>Analytical Method:</b> | <u><b>SW8260D</b></u>    |
| <b>Client:</b>  | <u><b>G Environmental</b></u> | <b>Datafile :</b>         | <u><b>VN087837.D</b></u> |

| Lab Sample ID | Parameter                   | Spike | Result | Unit  | Rec | RPD | Qual | Low     | Limits<br>High | RPD |
|---------------|-----------------------------|-------|--------|-------|-----|-----|------|---------|----------------|-----|
| VN0915MBS01   | 1,2-Dibromo-3-Chloropropane | 2000  | 1600   | ug/Kg | 80  |     |      | 40 (66) | 160 (134)      |     |
|               | 1,2,4-Trichlorobenzene      | 2000  | 1600   | ug/Kg | 80  |     |      | 70 (78) | 130 (127)      |     |
|               | 1,2,3-Trichlorobenzene      | 2000  | 1600   | ug/Kg | 80  |     |      | 70 (70) | 130 (137)      |     |

() = LABORATORY INHOUSE LIMIT

## Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

|          |                 |                    |            |
|----------|-----------------|--------------------|------------|
| SDG No.: | Q3058           | Analytical Method: | SW8260D    |
| Client:  | G Environmental | Datafile :         | VY023323.D |

| Lab Sample ID | Parameter                      | Spike | Result | Unit  | Rec | RPD | Qual | Low     | Limits High | RPD |
|---------------|--------------------------------|-------|--------|-------|-----|-----|------|---------|-------------|-----|
| VY0910SBS01   | Dichlorodifluoromethane        | 20    | 20.8   | ug/Kg | 104 |     |      | 40 (64) | 160 (136)   |     |
|               | Chloromethane                  | 20    | 21.7   | ug/Kg | 109 |     |      | 40 (52) | 160 (151)   |     |
|               | Vinyl chloride                 | 20    | 21.3   | ug/Kg | 106 |     |      | 70 (56) | 130 (148)   |     |
|               | Bromomethane                   | 20    | 22.1   | ug/Kg | 111 |     |      | 40 (58) | 160 (141)   |     |
|               | Chloroethane                   | 20    | 21.3   | ug/Kg | 106 |     |      | 40 (69) | 160 (130)   |     |
|               | Trichlorofluoromethane         | 20    | 20.9   | ug/Kg | 104 |     |      | 40 (69) | 160 (134)   |     |
|               | 1,1,2-Trichlorotrifluoroethane | 20    | 21.2   | ug/Kg | 106 |     |      | 70 (81) | 130 (123)   |     |
|               | 1,1-Dichloroethene             | 20    | 20.1   | ug/Kg | 101 |     |      | 70 (79) | 130 (121)   |     |
|               | Acetone                        | 100   | 110    | ug/Kg | 110 |     |      | 40 (40) | 160 (171)   |     |
|               | Carbon disulfide               | 20    | 20.5   | ug/Kg | 103 |     |      | 40 (59) | 160 (130)   |     |
|               | Methyl tert-butyl Ether        | 20    | 20.1   | ug/Kg | 101 |     |      | 70 (77) | 130 (129)   |     |
|               | Methyl Acetate                 | 20    | 19.1   | ug/Kg | 96  |     |      | 70 (69) | 130 (149)   |     |
|               | Methylene Chloride             | 20    | 23.2   | ug/Kg | 116 |     |      | 70 (72) | 130 (131)   |     |
|               | trans-1,2-Dichloroethene       | 20    | 20.7   | ug/Kg | 104 |     |      | 70 (80) | 130 (123)   |     |
|               | 1,1-Dichloroethane             | 20    | 20.7   | ug/Kg | 104 |     |      | 70 (82) | 130 (123)   |     |
|               | Cyclohexane                    | 20    | 20.0   | ug/Kg | 100 |     |      | 70 (76) | 130 (122)   |     |
|               | 2-Butanone                     | 100   | 100    | ug/Kg | 100 |     |      | 40 (69) | 160 (131)   |     |
|               | Carbon Tetrachloride           | 20    | 20.5   | ug/Kg | 103 |     |      | 70 (76) | 130 (129)   |     |
|               | cis-1,2-Dichloroethene         | 20    | 20.6   | ug/Kg | 103 |     |      | 70 (82) | 130 (123)   |     |
|               | Bromochloromethane             | 20    | 19.9   | ug/Kg | 100 |     |      | 70 (80) | 130 (127)   |     |
|               | Chloroform                     | 20    | 20.7   | ug/Kg | 104 |     |      | 70 (82) | 130 (125)   |     |
|               | 1,1,1-Trichloroethane          | 20    | 20.9   | ug/Kg | 104 |     |      | 70 (80) | 130 (126)   |     |
|               | Methylcyclohexane              | 20    | 20.2   | ug/Kg | 101 |     |      | 70 (77) | 130 (123)   |     |
|               | Benzene                        | 20    | 20.7   | ug/Kg | 104 |     |      | 70 (84) | 130 (121)   |     |
|               | 1,2-Dichloroethane             | 20    | 20.5   | ug/Kg | 103 |     |      | 70 (81) | 130 (126)   |     |
|               | Trichloroethene                | 20    | 20.9   | ug/Kg | 104 |     |      | 70 (83) | 130 (122)   |     |
|               | 1,2-Dichloropropane            | 20    | 20.6   | ug/Kg | 103 |     |      | 70 (83) | 130 (122)   |     |
|               | Bromodichloromethane           | 20    | 20.3   | ug/Kg | 102 |     |      | 70 (82) | 130 (123)   |     |
|               | 4-Methyl-2-Pentanone           | 100   | 98.3   | ug/Kg | 98  |     |      | 40 (70) | 160 (135)   |     |
|               | Toluene                        | 20    | 21.1   | ug/Kg | 106 |     |      | 70 (83) | 130 (122)   |     |
|               | t-1,3-Dichloropropene          | 20    | 20.0   | ug/Kg | 100 |     |      | 70 (78) | 130 (124)   |     |
|               | cis-1,3-Dichloropropene        | 20    | 20.1   | ug/Kg | 101 |     |      | 70 (81) | 130 (122)   |     |
|               | 1,1,2-Trichloroethane          | 20    | 20.6   | ug/Kg | 103 |     |      | 70 (82) | 130 (125)   |     |
|               | 2-Hexanone                     | 100   | 100    | ug/Kg | 100 |     |      | 40 (66) | 160 (138)   |     |
|               | Dibromochloromethane           | 20    | 20.5   | ug/Kg | 103 |     |      | 70 (79) | 130 (125)   |     |
|               | 1,2-Dibromoethane              | 20    | 20.2   | ug/Kg | 101 |     |      | 70 (80) | 130 (125)   |     |
|               | Tetrachloroethene              | 20    | 21.5   | ug/Kg | 108 |     |      | 70 (83) | 130 (125)   |     |
|               | Chlorobenzene                  | 20    | 20.2   | ug/Kg | 101 |     |      | 70 (84) | 130 (122)   |     |
|               | Ethyl Benzene                  | 20    | 20.2   | ug/Kg | 101 |     |      | 70 (82) | 130 (124)   |     |
|               | m/p-Xylenes                    | 40    | 41.2   | ug/Kg | 103 |     |      | 70 (83) | 130 (124)   |     |
|               | o-Xylene                       | 20    | 20.1   | ug/Kg | 101 |     |      | 70 (83) | 130 (123)   |     |
|               | Styrene                        | 20    | 20.5   | ug/Kg | 103 |     |      | 70 (82) | 130 (124)   |     |
|               | Bromoform                      | 20    | 20.1   | ug/Kg | 101 |     |      | 70 (75) | 130 (127)   |     |
|               | Isopropylbenzene               | 20    | 20.1   | ug/Kg | 101 |     |      | 70 (82) | 130 (124)   |     |
|               | 1,1,2,2-Tetrachloroethane      | 20    | 19.2   | ug/Kg | 96  |     |      | 70 (77) | 130 (127)   |     |
|               | 1,3-Dichlorobenzene            | 20    | 20.0   | ug/Kg | 100 |     |      | 70 (83) | 130 (122)   |     |
|               | 1,4-Dichlorobenzene            | 20    | 20.3   | ug/Kg | 102 |     |      | 70 (84) | 130 (121)   |     |
|               | 1,2-Dichlorobenzene            | 20    | 20.2   | ug/Kg | 101 |     |      | 70 (83) | 130 (124)   |     |

() = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3058 Analytical Method: SW8260D  
Client: G Environmental Datafile : VY023323.D

| Lab Sample ID | Parameter                   | Spike | Result | Unit  | Rec | RPD | Qual | Low     | Limits<br>High | RPD |
|---------------|-----------------------------|-------|--------|-------|-----|-----|------|---------|----------------|-----|
| VY0910SBS01   | 1,2-Dibromo-3-Chloropropane | 20    | 19.3   | ug/Kg | 97  |     |      | 40 (66) | 160 (134)      |     |
|               | 1,2,4-Trichlorobenzene      | 20    | 19.9   | ug/Kg | 100 |     |      | 70 (78) | 130 (127)      |     |
|               | 1,2,3-Trichlorobenzene      | 20    | 20.0   | ug/Kg | 100 |     |      | 70 (70) | 130 (137)      |     |

() = LABORATORY INHOUSE LIMIT

## Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

|          |                        |                    |                   |
|----------|------------------------|--------------------|-------------------|
| SDG No.: | <u>Q3058</u>           | Analytical Method: | <u>SW8260D</u>    |
| Client:  | <u>G Environmental</u> | Datafile :         | <u>VY023324.D</u> |

| Lab Sample ID | Parameter                      | Spike | Result | Unit  | Rec | RPD | Qual | Low     | Limits<br>High | RPD     |
|---------------|--------------------------------|-------|--------|-------|-----|-----|------|---------|----------------|---------|
| VY0910SBSD01  | Dichlorodifluoromethane        | 20    | 20.9   | ug/Kg | 104 | 0   |      | 40 (64) | 160 (136)      | 30 (20) |
|               | Chloromethane                  | 20    | 21.0   | ug/Kg | 105 | 4   |      | 40 (52) | 160 (151)      | 30 (20) |
|               | Vinyl chloride                 | 20    | 20.8   | ug/Kg | 104 | 2   |      | 70 (56) | 130 (148)      | 30 (20) |
|               | Bromomethane                   | 20    | 21.8   | ug/Kg | 109 | 2   |      | 40 (58) | 160 (141)      | 30 (20) |
|               | Chloroethane                   | 20    | 20.7   | ug/Kg | 104 | 2   |      | 40 (69) | 160 (130)      | 30 (20) |
|               | Trichlorofluoromethane         | 20    | 20.5   | ug/Kg | 103 | 1   |      | 40 (69) | 160 (134)      | 30 (20) |
|               | 1,1,2-Trichlorotrifluoroethane | 20    | 20.2   | ug/Kg | 101 | 5   |      | 70 (81) | 130 (123)      | 30 (20) |
|               | 1,1-Dichloroethene             | 20    | 19.5   | ug/Kg | 98  | 3   |      | 70 (79) | 130 (121)      | 30 (20) |
|               | Acetone                        | 100   | 110    | ug/Kg | 110 | 0   |      | 40 (40) | 160 (171)      | 30 (20) |
|               | Carbon disulfide               | 20    | 19.8   | ug/Kg | 99  | 4   |      | 40 (59) | 160 (130)      | 30 (20) |
|               | Methyl tert-butyl Ether        | 20    | 19.8   | ug/Kg | 99  | 2   |      | 70 (77) | 130 (129)      | 30 (20) |
|               | Methyl Acetate                 | 20    | 20.4   | ug/Kg | 102 | 6   |      | 70 (69) | 130 (149)      | 30 (20) |
|               | Methylene Chloride             | 20    | 22.3   | ug/Kg | 112 | 4   |      | 70 (72) | 130 (131)      | 30 (20) |
|               | trans-1,2-Dichloroethene       | 20    | 20.1   | ug/Kg | 101 | 3   |      | 70 (80) | 130 (123)      | 30 (20) |
|               | 1,1-Dichloroethane             | 20    | 20.1   | ug/Kg | 101 | 3   |      | 70 (82) | 130 (123)      | 30 (20) |
|               | Cyclohexane                    | 20    | 19.7   | ug/Kg | 99  | 1   |      | 70 (76) | 130 (122)      | 30 (20) |
|               | 2-Butanone                     | 100   | 100    | ug/Kg | 100 | 0   |      | 40 (69) | 160 (131)      | 30 (20) |
|               | Carbon Tetrachloride           | 20    | 19.7   | ug/Kg | 99  | 4   |      | 70 (76) | 130 (129)      | 30 (20) |
|               | cis-1,2-Dichloroethene         | 20    | 20.0   | ug/Kg | 100 | 3   |      | 70 (82) | 130 (123)      | 30 (20) |
|               | Bromochloromethane             | 20    | 19.8   | ug/Kg | 99  | 1   |      | 70 (80) | 130 (127)      | 30 (20) |
|               | Chloroform                     | 20    | 20.1   | ug/Kg | 101 | 3   |      | 70 (82) | 130 (125)      | 30 (20) |
|               | 1,1,1-Trichloroethane          | 20    | 20.0   | ug/Kg | 100 | 4   |      | 70 (80) | 130 (126)      | 30 (20) |
|               | Methylcyclohexane              | 20    | 19.6   | ug/Kg | 98  | 3   |      | 70 (77) | 130 (123)      | 30 (20) |
|               | Benzene                        | 20    | 20.3   | ug/Kg | 102 | 2   |      | 70 (84) | 130 (121)      | 30 (20) |
|               | 1,2-Dichloroethane             | 20    | 20.3   | ug/Kg | 102 | 1   |      | 70 (81) | 130 (126)      | 30 (20) |
|               | Trichloroethene                | 20    | 20.1   | ug/Kg | 101 | 3   |      | 70 (83) | 130 (122)      | 30 (20) |
|               | 1,2-Dichloropropane            | 20    | 20.0   | ug/Kg | 100 | 3   |      | 70 (83) | 130 (122)      | 30 (20) |
|               | Bromodichloromethane           | 20    | 20.1   | ug/Kg | 101 | 1   |      | 70 (82) | 130 (123)      | 30 (20) |
|               | 4-Methyl-2-Pentanone           | 100   | 99.0   | ug/Kg | 99  | 1   |      | 40 (70) | 160 (135)      | 30 (20) |
|               | Toluene                        | 20    | 20.1   | ug/Kg | 101 | 5   |      | 70 (83) | 130 (122)      | 30 (20) |
|               | t-1,3-Dichloropropene          | 20    | 19.7   | ug/Kg | 99  | 1   |      | 70 (78) | 130 (124)      | 30 (20) |
|               | cis-1,3-Dichloropropene        | 20    | 19.7   | ug/Kg | 99  | 2   |      | 70 (81) | 130 (122)      | 30 (20) |
|               | 1,1,2-Trichloroethane          | 20    | 20.1   | ug/Kg | 101 | 2   |      | 70 (82) | 130 (125)      | 30 (20) |
|               | 2-Hexanone                     | 100   | 100    | ug/Kg | 100 | 0   |      | 40 (66) | 160 (138)      | 30 (20) |
|               | Dibromochloromethane           | 20    | 20.1   | ug/Kg | 101 | 2   |      | 70 (79) | 130 (125)      | 30 (20) |
|               | 1,2-Dibromoethane              | 20    | 20.0   | ug/Kg | 100 | 1   |      | 70 (80) | 130 (125)      | 30 (20) |
|               | Tetrachloroethene              | 20    | 21.2   | ug/Kg | 106 | 2   |      | 70 (83) | 130 (125)      | 30 (20) |
|               | Chlorobenzene                  | 20    | 20.2   | ug/Kg | 101 | 0   |      | 70 (84) | 130 (122)      | 30 (20) |
|               | Ethyl Benzene                  | 20    | 19.9   | ug/Kg | 100 | 1   |      | 70 (82) | 130 (124)      | 30 (20) |
|               | m/p-Xylenes                    | 40    | 40.5   | ug/Kg | 101 | 2   |      | 70 (83) | 130 (124)      | 30 (20) |
|               | o-Xylene                       | 20    | 19.9   | ug/Kg | 100 | 1   |      | 70 (83) | 130 (123)      | 30 (20) |
|               | Styrene                        | 20    | 20.3   | ug/Kg | 102 | 1   |      | 70 (82) | 130 (124)      | 30 (20) |
|               | Bromoform                      | 20    | 20.4   | ug/Kg | 102 | 1   |      | 70 (75) | 130 (127)      | 30 (20) |
|               | Isopropylbenzene               | 20    | 19.8   | ug/Kg | 99  | 2   |      | 70 (82) | 130 (124)      | 30 (20) |
|               | 1,1,2,2-Tetrachloroethane      | 20    | 19.3   | ug/Kg | 97  | 1   |      | 70 (77) | 130 (127)      | 30 (20) |
|               | 1,3-Dichlorobenzene            | 20    | 20.3   | ug/Kg | 102 | 2   |      | 70 (83) | 130 (122)      | 30 (20) |
|               | 1,4-Dichlorobenzene            | 20    | 20.0   | ug/Kg | 100 | 2   |      | 70 (84) | 130 (121)      | 30 (20) |
|               | 1,2-Dichlorobenzene            | 20    | 20.2   | ug/Kg | 101 | 0   |      | 70 (83) | 130 (124)      | 30 (20) |

() = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3058 Analytical Method: SW8260D  
Client: G Environmental Datafile : VY023324.D

| Lab Sample ID | Parameter                   | Spike | Result | Unit  | Rec | RPD | Qual | Low     | Limits<br>High | RPD     |
|---------------|-----------------------------|-------|--------|-------|-----|-----|------|---------|----------------|---------|
| VY0910SBSD01  | 1,2-Dibromo-3-Chloropropane | 20    | 20.0   | ug/Kg | 100 | 3   |      | 40 (66) | 160 (134)      | 30 (20) |
|               | 1,2,4-Trichlorobenzene      | 20    | 20.3   | ug/Kg | 102 | 2   |      | 70 (78) | 130 (127)      | 30 (20) |
|               | 1,2,3-Trichlorobenzene      | 20    | 20.2   | ug/Kg | 101 | 1   |      | 70 (70) | 130 (137)      | 30 (20) |

() = LABORATORY INHOUSE LIMIT

VOLATILE METHOD BLANK SUMMARY

Client ID

VN0915MBL01

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG NO.: Q3058

Lab File ID: VN087821.D

Lab Sample ID: VN0915MBL01

Date Analyzed: 09/15/2025

Time Analyzed: 09:14

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA\_N

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

| EPA<br>SAMPLE NO. | LAB<br>SAMPLE ID | LAB<br>FILE ID | DATE<br>ANALYZED |
|-------------------|------------------|----------------|------------------|
| VN0915MBS01       | VN0915MBS01      | VN087837.D     | 09/15/2025       |
| RPHY5ME           | Q3058-01ME       | VN087839.D     | 09/15/2025       |

COMMENTS: \_\_\_\_\_

VOLATILE METHOD BLANK SUMMARY

Client ID

VY0910SBL01

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG NO.: Q3058

Lab File ID: VY023322.D

Lab Sample ID: VY0910SBL01

Date Analyzed: 09/10/2025

Time Analyzed: 09:39

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) Y

Instrument ID: MSVOA\_Y

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

| EPA<br>SAMPLE NO. | LAB<br>SAMPLE ID | LAB<br>FILE ID | DATE<br>ANALYZED |
|-------------------|------------------|----------------|------------------|
| VY0910SBS01       | VY0910SBS01      | VY023323.D     | 09/10/2025       |
| VY0910SBSD01      | VY0910SBSD01     | VY023324.D     | 09/10/2025       |
| RPXY5             | Q3058-01         | VY023328.D     | 09/10/2025       |
| RPXY4             | Q3058-02         | VY023329.D     | 09/10/2025       |

COMMENTS: \_\_\_\_\_

\_\_\_\_\_

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance  
Lab Code: ACE  
Lab File ID: VN087614.D  
Instrument ID: MSVOA\_N  
GC Column: RXI-624 ID: 0.25 (mm)

Contract: GENV01  
SDG NO.: Q3058  
BFB Injection Date: 08/21/2025  
BFB Injection Time: 09:30  
Heated Purge: Y/N N

| m/e | ION ABUNDANCE CRITERIA             | % RELATIVE ABUNDANCE |
|-----|------------------------------------|----------------------|
| 50  | 15.0 - 40.0% of mass 95            | 25.8                 |
| 75  | 30.0 - 60.0% of mass 95            | 58.1                 |
| 95  | Base Peak, 100% relative abundance | 100                  |
| 96  | 5.0 - 9.0% of mass 95              | 7.3                  |
| 173 | Less than 2.0% of mass 174         | 1.1 ( 1.6 ) 1        |
| 174 | 50.0 - 100.0% of mass 95           | 69                   |
| 175 | 5.0 - 9.0% of mass 174             | 5.6 ( 8.1 ) 1        |
| 176 | 95.0 - 101.0% of mass 174          | 67.7 ( 98.1 ) 1      |
| 177 | 5.0 - 9.0% of mass 176             | 4.8 ( 7.1 ) 2        |

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| CLIENT ID | LAB SAMPLE ID | LAB FILE ID | DATE ANALYZED | TIME ANALYZED |
|-----------|---------------|-------------|---------------|---------------|
| VSTDIC001 | VSTDIC001     | VN087619.D  | 08/21/2025    | 12:09         |
| VSTDIC005 | VSTDIC005     | VN087620.D  | 08/21/2025    | 13:08         |
| VSTDIC020 | VSTDIC020     | VN087621.D  | 08/21/2025    | 13:41         |
| VSTDIC050 | VSTDIC050     | VN087622.D  | 08/21/2025    | 14:02         |
| VSTDIC100 | VSTDIC100     | VN087623.D  | 08/21/2025    | 14:23         |
| VSTDIC150 | VSTDIC150     | VN087624.D  | 08/21/2025    | 14:44         |

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG NO.: Q3058

Lab File ID: VN087819.D

BFB Injection Date: 09/15/2025

Instrument ID: MSVOA\_N

BFB Injection Time: 08:18

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N N

| m/e | ION ABUNDANCE CRITERIA             | % RELATIVE ABUNDANCE |
|-----|------------------------------------|----------------------|
| 50  | 15.0 - 40.0% of mass 95            | 23.1                 |
| 75  | 30.0 - 60.0% of mass 95            | 51.9                 |
| 95  | Base Peak, 100% relative abundance | 100                  |
| 96  | 5.0 - 9.0% of mass 95              | 7.1                  |
| 173 | Less than 2.0% of mass 174         | 0.9 ( 1.2 ) 1        |
| 174 | 50.0 - 100.0% of mass 95           | 74.1                 |
| 175 | 5.0 - 9.0% of mass 174             | 6.3 ( 8.5 ) 1        |
| 176 | 95.0 - 101.0% of mass 174          | 70.8 ( 95.4 ) 1      |
| 177 | 5.0 - 9.0% of mass 176             | 4.8 ( 6.8 ) 2        |

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| CLIENT ID   | LAB SAMPLE ID | LAB FILE ID | DATE ANALYZED | TIME ANALYZED |
|-------------|---------------|-------------|---------------|---------------|
| VSTDCCC050  | VSTDCCC050    | VN087820.D  | 09/15/2025    | 08:39         |
| VN0915MBL01 | VN0915MBL01   | VN087821.D  | 09/15/2025    | 09:14         |
| VN0915MBS01 | VN0915MBS01   | VN087837.D  | 09/15/2025    | 15:02         |
| RPXY5ME     | Q3058-01ME    | VN087839.D  | 09/15/2025    | 15:43         |

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG NO.: Q3058

Lab File ID: VY023302.D

BFB Injection Date: 09/08/2025

Instrument ID: MSVOA\_Y

BFB Injection Time: 09:07

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N Y

| m/e | ION ABUNDANCE CRITERIA             | % RELATIVE ABUNDANCE |
|-----|------------------------------------|----------------------|
| 50  | 15.0 - 40.0% of mass 95            | 18.3                 |
| 75  | 30.0 - 60.0% of mass 95            | 49.4                 |
| 95  | Base Peak, 100% relative abundance | 100                  |
| 96  | 5.0 - 9.0% of mass 95              | 6.2                  |
| 173 | Less than 2.0% of mass 174         | 1.1 ( 1.2 ) 1        |
| 174 | 50.0 - 100.0% of mass 95           | 90.6                 |
| 175 | 5.0 - 9.0% of mass 174             | 6.9 ( 7.6 ) 1        |
| 176 | 95.0 - 101.0% of mass 174          | 89.4 ( 98.7 ) 1      |
| 177 | 5.0 - 9.0% of mass 176             | 5.9 ( 6.6 ) 2        |

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| CLIENT ID | LAB SAMPLE ID | LAB FILE ID | DATE ANALYZED | TIME ANALYZED |
|-----------|---------------|-------------|---------------|---------------|
| VSTDIC005 | VSTDIC005     | VY023303.D  | 09/08/2025    | 10:00         |
| VSTDIC010 | VSTDIC010     | VY023304.D  | 09/08/2025    | 10:23         |
| VSTDIC020 | VSTDIC020     | VY023305.D  | 09/08/2025    | 11:01         |
| VSTDIC050 | VSTDIC050     | VY023306.D  | 09/08/2025    | 11:23         |
| VSTDIC100 | VSTDIC100     | VY023307.D  | 09/08/2025    | 11:46         |
| VSTDIC150 | VSTDIC150     | VY023308.D  | 09/08/2025    | 12:08         |

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance  
Lab Code: ACE  
Lab File ID: VY023320.D  
Instrument ID: MSVOA\_Y  
GC Column: RXI-624 ID: 0.25 (mm)

Contract: GENV01  
SDG NO.: Q3058  
BFB Injection Date: 09/10/2025  
BFB Injection Time: 08:38  
Heated Purge: Y/N Y

| m/e | ION ABUNDANCE CRITERIA             | % RELATIVE ABUNDANCE |
|-----|------------------------------------|----------------------|
| 50  | 15.0 - 40.0% of mass 95            | 17.4                 |
| 75  | 30.0 - 60.0% of mass 95            | 49.8                 |
| 95  | Base Peak, 100% relative abundance | 100                  |
| 96  | 5.0 - 9.0% of mass 95              | 6.7                  |
| 173 | Less than 2.0% of mass 174         | 1 ( 1.1 ) 1          |
| 174 | 50.0 - 100.0% of mass 95           | 93.1                 |
| 175 | 5.0 - 9.0% of mass 174             | 6.9 ( 7.4 ) 1        |
| 176 | 95.0 - 101.0% of mass 174          | 88.6 ( 95.1 ) 1      |
| 177 | 5.0 - 9.0% of mass 176             | 5.8 ( 6.6 ) 2        |

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| CLIENT ID    | LAB SAMPLE ID | LAB FILE ID | DATE ANALYZED | TIME ANALYZED |
|--------------|---------------|-------------|---------------|---------------|
| VSTDCCC050   | VSTDCCC050    | VY023321.D  | 09/10/2025    | 09:09         |
| VY0910SBL01  | VY0910SBL01   | VY023322.D  | 09/10/2025    | 09:39         |
| VY0910SBS01  | VY0910SBS01   | VY023323.D  | 09/10/2025    | 10:11         |
| VY0910SBSD01 | VY0910SBSD01  | VY023324.D  | 09/10/2025    | 10:33         |
| RPXY5        | Q3058-01      | VY023328.D  | 09/10/2025    | 12:49         |
| RPXY4        | Q3058-02      | VY023329.D  | 09/10/2025    | 13:12         |

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01  
Lab Code: ACE SDG NO.: Q3058  
Lab File ID: VN087820.D Date Analyzed: 09/15/2025  
Instrument ID: MSVOA\_N Time Analyzed: 08:39  
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #   |
|----------------|---------------|-------|---------------|-------|---------------|--------|
| 12 HOUR STD    | 257288        | 8.21  | 490599        | 9.08  | 467132        | 11.84  |
| UPPER LIMIT    | 514576        | 8.706 | 981198        | 9.582 | 934264        | 12.341 |
| LOWER LIMIT    | 128644        | 7.706 | 245300        | 8.582 | 233566        | 11.341 |
| EPA SAMPLE NO. |               |       |               |       |               |        |
| RPXY5ME        | 191442        | 8.20  | 434787        | 9.08  | 386645        | 11.85  |
| VN0915MBL01    | 198222        | 8.21  | 425993        | 9.08  | 408687        | 11.85  |
| VN0915MBS01    | 204697        | 8.21  | 405190        | 9.08  | 394337        | 11.85  |

IS1 = Pentafluorobenzene  
IS2 = 1,4-Difluorobenzene  
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
AREA LOWER LIMIT = -50% of internal standard area  
RT UPPER LIMIT = +0.50 minutes of internal standard RT  
RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
\* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01  
 Lab Code: ACE SDG NO.: Q3058  
 Lab File ID: VN087820.D Date Analyzed: 09/15/2025  
 Instrument ID: MSVOA\_N Time Analyzed: 08:39  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 259703        | 13.765 |  |  |  |  |
| UPPER LIMIT    | 519406        | 14.265 |  |  |  |  |
| LOWER LIMIT    | 129852        | 13.265 |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| RPXY5ME        | 170499        | 13.77  |  |  |  |  |
| VN0915MBL01    | 193776        | 13.77  |  |  |  |  |
| VN0915MBS01    | 213436        | 13.77  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01  
Lab Code: ACE SDG NO.: Q3058  
Lab File ID: VY023321.D Date Analyzed: 09/10/2025  
Instrument ID: MSVOA\_Y Time Analyzed: 09:09  
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

|                | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #  |
|----------------|---------------|-------|---------------|-------|---------------|-------|
| 12 HOUR STD    | 519574        | 7.71  | 748186        | 8.62  | 662487        | 11.42 |
| UPPER LIMIT    | 1039150       | 8.213 | 1496370       | 9.115 | 1324970       | 11.92 |
| LOWER LIMIT    | 259787        | 7.213 | 374093        | 8.115 | 331244        | 10.92 |
| EPA SAMPLE NO. |               |       |               |       |               |       |
| RPXY5          | 542076        | 7.71  | 1032321       | 8.62  | 977441        | 11.41 |
| RPXY4          | 539229        | 7.71  | 1018906       | 8.62  | 1004230       | 11.41 |
| VY0910SBL01    | 583431        | 7.71  | 1118169       | 8.62  | 1034716       | 11.42 |
| VY0910SBS01    | 502093        | 7.71  | 763457        | 8.62  | 676997        | 11.42 |
| VY0910SBSD01   | 502400        | 7.71  | 765892        | 8.62  | 669720        | 11.41 |

IS1 = Pentafluorobenzene  
IS2 = 1,4-Difluorobenzene  
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
AREA LOWER LIMIT = -50% of internal standard area  
RT UPPER LIMIT = +0.50 minutes of internal standard RT  
RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
\* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01  
 Lab Code: ACE SDG NO.: Q3058  
 Lab File ID: VY023321.D Date Analyzed: 09/10/2025  
 Instrument ID: MSVOA\_Y Time Analyzed: 09:09  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 341214        | 13.346 |  |  |  |  |
| UPPER LIMIT    | 682428        | 13.846 |  |  |  |  |
| LOWER LIMIT    | 170607        | 12.846 |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| RPXY5          | 418953        | 13.35  |  |  |  |  |
| RPXY4          | 442394        | 13.35  |  |  |  |  |
| VY0910SBL01    | 415042        | 13.35  |  |  |  |  |
| VY0910SBS01    | 350574        | 13.35  |  |  |  |  |
| VY0910SBSD01   | 341290        | 13.35  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.



# QC SAMPLE DATA

### Report of Analysis

|                    |                 |           |                 |               |
|--------------------|-----------------|-----------|-----------------|---------------|
| Client:            | G Environmental |           | Date Collected: |               |
| Project:           | Buffington      |           | Date Received:  |               |
| Client Sample ID:  | VN0915MBL01     |           | SDG No.:        | Q3058         |
| Lab Sample ID:     | VN0915MBL01     |           | Matrix:         | SOIL          |
| Analytical Method: | 8260D           |           | % Solid:        | 100           |
| Sample Wt/Vol:     | 5               | Units: g  | Final Vol:      | 10000 uL      |
| Soil Aliquot Vol:  | 100             | uL        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624         | ID : 0.25 | Level :         | MED           |
| Prep Method :      |                 |           |                 |               |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VN087821.D        | 1         | 09/15/25 09:14 | VN091525      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units(Dry Weight) |
|----------------|--------------------------------|-------|-----------|------|------------|-------------------|
| <b>TARGETS</b> |                                |       |           |      |            |                   |
| 75-71-8        | Dichlorodifluoromethane        | 110   | U         | 110  | 500        | ug/Kg             |
| 74-87-3        | Chloromethane                  | 110   | U         | 110  | 500        | ug/Kg             |
| 75-01-4        | Vinyl Chloride                 | 79.0  | U         | 79.0 | 500        | ug/Kg             |
| 74-83-9        | Bromomethane                   | 110   | U         | 110  | 500        | ug/Kg             |
| 75-00-3        | Chloroethane                   | 130   | U         | 130  | 500        | ug/Kg             |
| 75-69-4        | Trichlorofluoromethane         | 120   | U         | 120  | 500        | ug/Kg             |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 110   | U         | 110  | 500        | ug/Kg             |
| 75-35-4        | 1,1-Dichloroethene             | 100   | U         | 100  | 500        | ug/Kg             |
| 67-64-1        | Acetone                        | 470   | U         | 470  | 2500       | ug/Kg             |
| 75-15-0        | Carbon Disulfide               | 110   | U         | 110  | 500        | ug/Kg             |
| 1634-04-4      | Methyl tert-butyl Ether        | 73.0  | U         | 73.0 | 500        | ug/Kg             |
| 79-20-9        | Methyl Acetate                 | 150   | U         | 150  | 500        | ug/Kg             |
| 75-09-2        | Methylene Chloride             | 350   | U         | 350  | 1000       | ug/Kg             |
| 156-60-5       | trans-1,2-Dichloroethene       | 86.0  | U         | 86.0 | 500        | ug/Kg             |
| 75-34-3        | 1,1-Dichloroethane             | 80.0  | U         | 80.0 | 500        | ug/Kg             |
| 110-82-7       | Cyclohexane                    | 79.0  | U         | 79.0 | 500        | ug/Kg             |
| 78-93-3        | 2-Butanone                     | 650   | U         | 650  | 2500       | ug/Kg             |
| 56-23-5        | Carbon Tetrachloride           | 97.0  | U         | 97.0 | 500        | ug/Kg             |
| 156-59-2       | cis-1,2-Dichloroethene         | 75.0  | U         | 75.0 | 500        | ug/Kg             |
| 74-97-5        | Bromochloromethane             | 120   | U         | 120  | 500        | ug/Kg             |
| 67-66-3        | Chloroform                     | 84.0  | U         | 84.0 | 500        | ug/Kg             |
| 71-55-6        | 1,1,1-Trichloroethane          | 93.0  | U         | 93.0 | 500        | ug/Kg             |
| 108-87-2       | Methylcyclohexane              | 91.0  | U         | 91.0 | 500        | ug/Kg             |
| 71-43-2        | Benzene                        | 79.0  | U         | 79.0 | 500        | ug/Kg             |
| 107-06-2       | 1,2-Dichloroethane             | 79.0  | U         | 79.0 | 500        | ug/Kg             |
| 79-01-6        | Trichloroethene                | 81.0  | U         | 81.0 | 500        | ug/Kg             |
| 78-87-5        | 1,2-Dichloropropane            | 91.0  | U         | 91.0 | 500        | ug/Kg             |
| 75-27-4        | Bromodichloromethane           | 78.0  | U         | 78.0 | 500        | ug/Kg             |
| 108-10-1       | 4-Methyl-2-Pentanone           | 360   | U         | 360  | 2500       | ug/Kg             |
| 108-88-3       | Toluene                        | 78.0  | U         | 78.0 | 500        | ug/Kg             |

### Report of Analysis

|                    |                 |           |                 |               |
|--------------------|-----------------|-----------|-----------------|---------------|
| Client:            | G Environmental |           | Date Collected: |               |
| Project:           | Buffington      |           | Date Received:  |               |
| Client Sample ID:  | VN0915MBL01     |           | SDG No.:        | Q3058         |
| Lab Sample ID:     | VN0915MBL01     |           | Matrix:         | SOIL          |
| Analytical Method: | 8260D           |           | % Solid:        | 100           |
| Sample Wt/Vol:     | 5               | Units: g  | Final Vol:      | 10000 uL      |
| Soil Aliquot Vol:  | 100             | uL        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624         | ID : 0.25 | Level :         | MED           |
| Prep Method :      |                 |           |                 |               |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VN087821.D        | 1         | 09/15/25 09:14 | VN091525      |

| CAS Number                | Parameter                   | Conc.  | Qualifier | MDL                 | LOQ / CRQL | Units(Dry Weight) |
|---------------------------|-----------------------------|--------|-----------|---------------------|------------|-------------------|
| 10061-02-6                | t-1,3-Dichloropropene       | 65.0   | U         | 65.0                | 500        | ug/Kg             |
| 10061-01-5                | cis-1,3-Dichloropropene     | 62.0   | U         | 62.0                | 500        | ug/Kg             |
| 79-00-5                   | 1,1,2-Trichloroethane       | 92.0   | U         | 92.0                | 500        | ug/Kg             |
| 591-78-6                  | 2-Hexanone                  | 370    | U         | 370                 | 2500       | ug/Kg             |
| 124-48-1                  | Dibromochloromethane        | 87.0   | U         | 87.0                | 500        | ug/Kg             |
| 106-93-4                  | 1,2-Dibromoethane           | 88.0   | U         | 88.0                | 500        | ug/Kg             |
| 127-18-4                  | Tetrachloroethene           | 110    | U         | 110                 | 500        | ug/Kg             |
| 108-90-7                  | Chlorobenzene               | 91.0   | U         | 91.0                | 500        | ug/Kg             |
| 100-41-4                  | Ethyl Benzene               | 67.0   | U         | 67.0                | 500        | ug/Kg             |
| 179601-23-1               | m/p-Xylenes                 | 120    | U         | 120                 | 1000       | ug/Kg             |
| 95-47-6                   | o-Xylene                    | 82.0   | U         | 82.0                | 500        | ug/Kg             |
| 100-42-5                  | Styrene                     | 71.0   | U         | 71.0                | 500        | ug/Kg             |
| 75-25-2                   | Bromoform                   | 86.0   | U         | 86.0                | 500        | ug/Kg             |
| 98-82-8                   | Isopropylbenzene            | 78.0   | U         | 78.0                | 500        | ug/Kg             |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 120    | U         | 120                 | 500        | ug/Kg             |
| 541-73-1                  | 1,3-Dichlorobenzene         | 170    | U         | 170                 | 500        | ug/Kg             |
| 106-46-7                  | 1,4-Dichlorobenzene         | 160    | U         | 160                 | 500        | ug/Kg             |
| 95-50-1                   | 1,2-Dichlorobenzene         | 150    | U         | 150                 | 500        | ug/Kg             |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 180    | U         | 180                 | 500        | ug/Kg             |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 300    | U         | 300                 | 500        | ug/Kg             |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 320    | U         | 320                 | 500        | ug/Kg             |
| <b>SURROGATES</b>         |                             |        |           |                     |            |                   |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 49.9   |           | 70 (63) - 130 (155) | 100%       | SPK: 50           |
| 1868-53-7                 | Dibromofluoromethane        | 52.3   |           | 70 (70) - 130 (134) | 105%       | SPK: 50           |
| 2037-26-5                 | Toluene-d8                  | 48.3   |           | 70 (74) - 130 (123) | 97%        | SPK: 50           |
| 460-00-4                  | 4-Bromofluorobenzene        | 45.6   |           | 70 (17) - 130 (146) | 91%        | SPK: 50           |
| <b>INTERNAL STANDARDS</b> |                             |        |           |                     |            |                   |
| 363-72-4                  | Pentafluorobenzene          | 198000 | 8.206     |                     |            |                   |
| 540-36-3                  | 1,4-Difluorobenzene         | 426000 | 9.083     |                     |            |                   |
| 3114-55-4                 | Chlorobenzene-d5            | 409000 | 11.847    |                     |            |                   |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 194000 | 13.771    |                     |            |                   |

## Report of Analysis

|                    |                 |           |                 |               |
|--------------------|-----------------|-----------|-----------------|---------------|
| Client:            | G Environmental |           | Date Collected: |               |
| Project:           | Buffington      |           | Date Received:  |               |
| Client Sample ID:  | VN0915MBL01     |           | SDG No.:        | Q3058         |
| Lab Sample ID:     | VN0915MBL01     |           | Matrix:         | SOIL          |
| Analytical Method: | 8260D           |           | % Solid:        | 100           |
| Sample Wt/Vol:     | 5               | Units: g  | Final Vol:      | 10000 uL      |
| Soil Aliquot Vol:  | 100             | uL        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624         | ID : 0.25 | Level :         | MED           |
| Prep Method :      |                 |           |                 |               |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VN087821.D        | 1         | 09/15/25 09:14 | VN091525      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## Report of Analysis

|                    |                           |                 |               |
|--------------------|---------------------------|-----------------|---------------|
| Client:            | G Environmental           | Date Collected: |               |
| Project:           | Buffington                | Date Received:  |               |
| Client Sample ID:  | VY0910SBL01               | SDG No.:        | Q3058         |
| Lab Sample ID:     | VY0910SBL01               | Matrix:         | SOIL          |
| Analytical Method: | 8260D                     | % Solid:        | 100           |
| Sample Wt/Vol:     | 5      Units:    g        | Final Vol:      | 5000      uL  |
| Soil Aliquot Vol:  | uL                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624      ID :    0.25 | Level :         | LOW           |
| Prep Method :      |                           |                 |               |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VY023322.D        | 1         | 09/10/25 09:39 | VY091025      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units(Dry Weight) |
|----------------|--------------------------------|-------|-----------|------|------------|-------------------|
| <b>TARGETS</b> |                                |       |           |      |            |                   |
| 75-71-8        | Dichlorodifluoromethane        | 1.10  | U         | 1.10 | 5.00       | ug/Kg             |
| 74-87-3        | Chloromethane                  | 1.10  | U         | 1.10 | 5.00       | ug/Kg             |
| 75-01-4        | Vinyl Chloride                 | 0.79  | U         | 0.79 | 5.00       | ug/Kg             |
| 74-83-9        | Bromomethane                   | 1.10  | U         | 1.10 | 5.00       | ug/Kg             |
| 75-00-3        | Chloroethane                   | 1.30  | U         | 1.30 | 5.00       | ug/Kg             |
| 75-69-4        | Trichlorofluoromethane         | 1.20  | U         | 1.20 | 5.00       | ug/Kg             |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 1.10  | U         | 1.10 | 5.00       | ug/Kg             |
| 75-35-4        | 1,1-Dichloroethene             | 1.00  | U         | 1.00 | 5.00       | ug/Kg             |
| 67-64-1        | Acetone                        | 4.70  | U         | 4.70 | 25.0       | ug/Kg             |
| 75-15-0        | Carbon Disulfide               | 1.10  | U         | 1.10 | 5.00       | ug/Kg             |
| 1634-04-4      | Methyl tert-butyl Ether        | 0.73  | U         | 0.73 | 5.00       | ug/Kg             |
| 79-20-9        | Methyl Acetate                 | 1.50  | U         | 1.50 | 5.00       | ug/Kg             |
| 75-09-2        | Methylene Chloride             | 3.50  | U         | 3.50 | 10.0       | ug/Kg             |
| 156-60-5       | trans-1,2-Dichloroethene       | 0.86  | U         | 0.86 | 5.00       | ug/Kg             |
| 75-34-3        | 1,1-Dichloroethane             | 0.80  | U         | 0.80 | 5.00       | ug/Kg             |
| 110-82-7       | Cyclohexane                    | 0.79  | U         | 0.79 | 5.00       | ug/Kg             |
| 78-93-3        | 2-Butanone                     | 6.50  | U         | 6.50 | 25.0       | ug/Kg             |
| 56-23-5        | Carbon Tetrachloride           | 0.97  | U         | 0.97 | 5.00       | ug/Kg             |
| 156-59-2       | cis-1,2-Dichloroethene         | 0.75  | U         | 0.75 | 5.00       | ug/Kg             |
| 74-97-5        | Bromochloromethane             | 1.20  | U         | 1.20 | 5.00       | ug/Kg             |
| 67-66-3        | Chloroform                     | 0.84  | U         | 0.84 | 5.00       | ug/Kg             |
| 71-55-6        | 1,1,1-Trichloroethane          | 0.93  | U         | 0.93 | 5.00       | ug/Kg             |
| 108-87-2       | Methylcyclohexane              | 0.91  | U         | 0.91 | 5.00       | ug/Kg             |
| 71-43-2        | Benzene                        | 0.79  | U         | 0.79 | 5.00       | ug/Kg             |
| 107-06-2       | 1,2-Dichloroethane             | 0.79  | U         | 0.79 | 5.00       | ug/Kg             |
| 79-01-6        | Trichloroethene                | 0.81  | U         | 0.81 | 5.00       | ug/Kg             |
| 78-87-5        | 1,2-Dichloropropane            | 0.91  | U         | 0.91 | 5.00       | ug/Kg             |
| 75-27-4        | Bromodichloromethane           | 0.78  | U         | 0.78 | 5.00       | ug/Kg             |
| 108-10-1       | 4-Methyl-2-Pentanone           | 3.60  | U         | 3.60 | 25.0       | ug/Kg             |
| 108-88-3       | Toluene                        | 0.78  | U         | 0.78 | 5.00       | ug/Kg             |



## Report of Analysis

|                    |                 |           |                 |               |
|--------------------|-----------------|-----------|-----------------|---------------|
| Client:            | G Environmental |           | Date Collected: |               |
| Project:           | Buffington      |           | Date Received:  |               |
| Client Sample ID:  | VY0910SBL01     |           | SDG No.:        | Q3058         |
| Lab Sample ID:     | VY0910SBL01     |           | Matrix:         | SOIL          |
| Analytical Method: | 8260D           |           | % Solid:        | 100           |
| Sample Wt/Vol:     | 5               | Units: g  | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  |                 | uL        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624         | ID : 0.25 | Level :         | LOW           |
| Prep Method :      |                 |           |                 |               |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VY023322.D        | 1         | 09/10/25 09:39 | VY091025      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## Report of Analysis

|                    |                                           |                 |                               |
|--------------------|-------------------------------------------|-----------------|-------------------------------|
| Client:            | G Environmental                           | Date Collected: |                               |
| Project:           | Buffington                                | Date Received:  |                               |
| Client Sample ID:  | VN0915MBS01                               | SDG No.:        | Q3058                         |
| Lab Sample ID:     | VN0915MBS01                               | Matrix:         | SOIL                          |
| Analytical Method: | 8260D                                     | % Solid:        | 100                           |
| Sample Wt/Vol:     | 5                      Units:    g        | Final Vol:      | 10000                      uL |
| Soil Aliquot Vol:  | 100                      uL               | Test:           | VOC-TCLVOA-10                 |
| GC Column:         | RXI-624                      ID :    0.25 | Level :         | MED                           |
| Prep Method :      |                                           |                 |                               |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VN087837.D        | 1         | 09/15/25 15:02 | VN091525      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units(Dry Weight) |
|----------------|--------------------------------|-------|-----------|------|------------|-------------------|
| <b>TARGETS</b> |                                |       |           |      |            |                   |
| 75-71-8        | Dichlorodifluoromethane        | 1800  |           | 110  | 500        | ug/Kg             |
| 74-87-3        | Chloromethane                  | 1900  |           | 110  | 500        | ug/Kg             |
| 75-01-4        | Vinyl Chloride                 | 1900  |           | 79.0 | 500        | ug/Kg             |
| 74-83-9        | Bromomethane                   | 1600  |           | 110  | 500        | ug/Kg             |
| 75-00-3        | Chloroethane                   | 1800  |           | 130  | 500        | ug/Kg             |
| 75-69-4        | Trichlorofluoromethane         | 2000  |           | 120  | 500        | ug/Kg             |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 1800  |           | 110  | 500        | ug/Kg             |
| 75-35-4        | 1,1-Dichloroethene             | 1800  |           | 100  | 500        | ug/Kg             |
| 67-64-1        | Acetone                        | 7900  |           | 470  | 2500       | ug/Kg             |
| 75-15-0        | Carbon Disulfide               | 2000  |           | 110  | 500        | ug/Kg             |
| 1634-04-4      | Methyl tert-butyl Ether        | 1800  |           | 73.0 | 500        | ug/Kg             |
| 79-20-9        | Methyl Acetate                 | 2000  |           | 150  | 500        | ug/Kg             |
| 75-09-2        | Methylene Chloride             | 1900  |           | 350  | 1000       | ug/Kg             |
| 156-60-5       | trans-1,2-Dichloroethene       | 1800  |           | 86.0 | 500        | ug/Kg             |
| 75-34-3        | 1,1-Dichloroethane             | 1800  |           | 80.0 | 500        | ug/Kg             |
| 110-82-7       | Cyclohexane                    | 1700  |           | 79.0 | 500        | ug/Kg             |
| 78-93-3        | 2-Butanone                     | 8600  |           | 650  | 2500       | ug/Kg             |
| 56-23-5        | Carbon Tetrachloride           | 2000  |           | 97.0 | 500        | ug/Kg             |
| 156-59-2       | cis-1,2-Dichloroethene         | 1800  |           | 75.0 | 500        | ug/Kg             |
| 74-97-5        | Bromochloromethane             | 2200  |           | 120  | 500        | ug/Kg             |
| 67-66-3        | Chloroform                     | 1900  |           | 84.0 | 500        | ug/Kg             |
| 71-55-6        | 1,1,1-Trichloroethane          | 1800  |           | 93.0 | 500        | ug/Kg             |
| 108-87-2       | Methylcyclohexane              | 1600  |           | 91.0 | 500        | ug/Kg             |
| 71-43-2        | Benzene                        | 1900  |           | 79.0 | 500        | ug/Kg             |
| 107-06-2       | 1,2-Dichloroethane             | 1800  |           | 79.0 | 500        | ug/Kg             |
| 79-01-6        | Trichloroethene                | 1900  |           | 81.0 | 500        | ug/Kg             |
| 78-87-5        | 1,2-Dichloropropane            | 1900  |           | 91.0 | 500        | ug/Kg             |
| 75-27-4        | Bromodichloromethane           | 1900  |           | 78.0 | 500        | ug/Kg             |
| 108-10-1       | 4-Methyl-2-Pentanone           | 9300  |           | 360  | 2500       | ug/Kg             |
| 108-88-3       | Toluene                        | 1900  |           | 78.0 | 500        | ug/Kg             |

### Report of Analysis

|                    |                 |           |                 |               |
|--------------------|-----------------|-----------|-----------------|---------------|
| Client:            | G Environmental |           | Date Collected: |               |
| Project:           | Buffington      |           | Date Received:  |               |
| Client Sample ID:  | VN0915MBS01     |           | SDG No.:        | Q3058         |
| Lab Sample ID:     | VN0915MBS01     |           | Matrix:         | SOIL          |
| Analytical Method: | 8260D           |           | % Solid:        | 100           |
| Sample Wt/Vol:     | 5               | Units: g  | Final Vol:      | 10000 uL      |
| Soil Aliquot Vol:  | 100             | uL        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624         | ID : 0.25 | Level :         | MED           |
| Prep Method :      |                 |           |                 |               |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VN087837.D        | 1         | 09/15/25 15:02 | VN091525      |

| CAS Number                | Parameter                   | Conc.  | Qualifier | MDL                 | LOQ / CRQL | Units(Dry Weight) |
|---------------------------|-----------------------------|--------|-----------|---------------------|------------|-------------------|
| 10061-02-6                | t-1,3-Dichloropropene       | 1800   |           | 65.0                | 500        | ug/Kg             |
| 10061-01-5                | cis-1,3-Dichloropropene     | 1800   |           | 62.0                | 500        | ug/Kg             |
| 79-00-5                   | 1,1,2-Trichloroethane       | 1900   |           | 92.0                | 500        | ug/Kg             |
| 591-78-6                  | 2-Hexanone                  | 9400   |           | 370                 | 2500       | ug/Kg             |
| 124-48-1                  | Dibromochloromethane        | 1900   |           | 87.0                | 500        | ug/Kg             |
| 106-93-4                  | 1,2-Dibromoethane           | 1900   |           | 88.0                | 500        | ug/Kg             |
| 127-18-4                  | Tetrachloroethene           | 1800   |           | 110                 | 500        | ug/Kg             |
| 108-90-7                  | Chlorobenzene               | 1700   |           | 91.0                | 500        | ug/Kg             |
| 100-41-4                  | Ethyl Benzene               | 1700   |           | 67.0                | 500        | ug/Kg             |
| 179601-23-1               | m/p-Xylenes                 | 3500   |           | 120                 | 1000       | ug/Kg             |
| 95-47-6                   | o-Xylene                    | 1700   |           | 82.0                | 500        | ug/Kg             |
| 100-42-5                  | Styrene                     | 1800   |           | 71.0                | 500        | ug/Kg             |
| 75-25-2                   | Bromoform                   | 1900   |           | 86.0                | 500        | ug/Kg             |
| 98-82-8                   | Isopropylbenzene            | 1600   |           | 78.0                | 500        | ug/Kg             |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 1700   |           | 120                 | 500        | ug/Kg             |
| 541-73-1                  | 1,3-Dichlorobenzene         | 1700   |           | 170                 | 500        | ug/Kg             |
| 106-46-7                  | 1,4-Dichlorobenzene         | 1700   |           | 160                 | 500        | ug/Kg             |
| 95-50-1                   | 1,2-Dichlorobenzene         | 1700   |           | 150                 | 500        | ug/Kg             |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 1600   |           | 180                 | 500        | ug/Kg             |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 1600   |           | 300                 | 500        | ug/Kg             |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 1600   |           | 320                 | 500        | ug/Kg             |
| <b>SURROGATES</b>         |                             |        |           |                     |            |                   |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 48.5   |           | 70 (63) - 130 (155) | 97%        | SPK: 50           |
| 1868-53-7                 | Dibromofluoromethane        | 49.4   |           | 70 (70) - 130 (134) | 99%        | SPK: 50           |
| 2037-26-5                 | Toluene-d8                  | 45.6   |           | 70 (74) - 130 (123) | 91%        | SPK: 50           |
| 460-00-4                  | 4-Bromofluorobenzene        | 46.4   |           | 70 (17) - 130 (146) | 93%        | SPK: 50           |
| <b>INTERNAL STANDARDS</b> |                             |        |           |                     |            |                   |
| 363-72-4                  | Pentafluorobenzene          | 205000 | 8.212     |                     |            |                   |
| 540-36-3                  | 1,4-Difluorobenzene         | 405000 | 9.082     |                     |            |                   |
| 3114-55-4                 | Chlorobenzene-d5            | 394000 | 11.847    |                     |            |                   |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 213000 | 13.77     |                     |            |                   |

## Report of Analysis

|                    |                 |           |                 |               |
|--------------------|-----------------|-----------|-----------------|---------------|
| Client:            | G Environmental |           | Date Collected: |               |
| Project:           | Buffington      |           | Date Received:  |               |
| Client Sample ID:  | VN0915MBS01     |           | SDG No.:        | Q3058         |
| Lab Sample ID:     | VN0915MBS01     |           | Matrix:         | SOIL          |
| Analytical Method: | 8260D           |           | % Solid:        | 100           |
| Sample Wt/Vol:     | 5               | Units: g  | Final Vol:      | 10000 uL      |
| Soil Aliquot Vol:  | 100             | uL        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624         | ID : 0.25 | Level :         | MED           |
| Prep Method :      |                 |           |                 |               |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VN087837.D        | 1         | 09/15/25 15:02 | VN091525      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected  
 LOQ = Limit of Quantitation  
 MDL = Method Detection Limit  
 LOD = Limit of Detection  
 E = Value Exceeds Calibration Range  
 Q = indicates LCS control criteria did not meet requirements  
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value  
 B = Analyte Found in Associated Method Blank  
 N = Presumptive Evidence of a Compound  
 \* = Values outside of QC limits  
 D = Dilution  
 () = Laboratory InHouse Limit  
 A = Aldol-Condensation Reaction Products



### Report of Analysis

|                    |                 |           |                 |               |    |
|--------------------|-----------------|-----------|-----------------|---------------|----|
| Client:            | G Environmental |           | Date Collected: |               |    |
| Project:           | Buffington      |           | Date Received:  |               |    |
| Client Sample ID:  | VY0910SBS01     |           | SDG No.:        | Q3058         |    |
| Lab Sample ID:     | VY0910SBS01     |           | Matrix:         | SOIL          |    |
| Analytical Method: | 8260D           |           | % Solid:        | 100           |    |
| Sample Wt/Vol:     | 5               | Units: g  | Final Vol:      | 5000          | uL |
| Soil Aliquot Vol:  |                 | uL        | Test:           | VOC-TCLVOA-10 |    |
| GC Column:         | RXI-624         | ID : 0.25 | Level :         | LOW           |    |
| Prep Method :      |                 |           |                 |               |    |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VY023323.D        | 1         | 09/10/25 10:11 | VY091025      |

| CAS Number                | Parameter                   | Conc.  | Qualifier | MDL                 | LOQ / CRQL | Units(Dry Weight) |
|---------------------------|-----------------------------|--------|-----------|---------------------|------------|-------------------|
| 10061-02-6                | t-1,3-Dichloropropene       | 20.0   |           | 0.65                | 5.00       | ug/Kg             |
| 10061-01-5                | cis-1,3-Dichloropropene     | 20.1   |           | 0.62                | 5.00       | ug/Kg             |
| 79-00-5                   | 1,1,2-Trichloroethane       | 20.6   |           | 0.92                | 5.00       | ug/Kg             |
| 591-78-6                  | 2-Hexanone                  | 100    |           | 3.70                | 25.0       | ug/Kg             |
| 124-48-1                  | Dibromochloromethane        | 20.5   |           | 0.87                | 5.00       | ug/Kg             |
| 106-93-4                  | 1,2-Dibromoethane           | 20.2   |           | 0.88                | 5.00       | ug/Kg             |
| 127-18-4                  | Tetrachloroethene           | 21.5   |           | 1.10                | 5.00       | ug/Kg             |
| 108-90-7                  | Chlorobenzene               | 20.2   |           | 0.91                | 5.00       | ug/Kg             |
| 100-41-4                  | Ethyl Benzene               | 20.2   |           | 0.67                | 5.00       | ug/Kg             |
| 179601-23-1               | m/p-Xylenes                 | 41.2   |           | 1.20                | 10.0       | ug/Kg             |
| 95-47-6                   | o-Xylene                    | 20.1   |           | 0.82                | 5.00       | ug/Kg             |
| 100-42-5                  | Styrene                     | 20.5   |           | 0.71                | 5.00       | ug/Kg             |
| 75-25-2                   | Bromoform                   | 20.1   |           | 0.86                | 5.00       | ug/Kg             |
| 98-82-8                   | Isopropylbenzene            | 20.1   |           | 0.78                | 5.00       | ug/Kg             |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 19.2   |           | 1.20                | 5.00       | ug/Kg             |
| 541-73-1                  | 1,3-Dichlorobenzene         | 20.0   |           | 1.70                | 5.00       | ug/Kg             |
| 106-46-7                  | 1,4-Dichlorobenzene         | 20.3   |           | 1.60                | 5.00       | ug/Kg             |
| 95-50-1                   | 1,2-Dichlorobenzene         | 20.2   |           | 1.50                | 5.00       | ug/Kg             |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 19.3   |           | 1.80                | 5.00       | ug/Kg             |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 19.9   |           | 3.00                | 5.00       | ug/Kg             |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 20.0   |           | 3.20                | 5.00       | ug/Kg             |
| <b>SURROGATES</b>         |                             |        |           |                     |            |                   |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 49.6   |           | 70 (63) - 130 (155) | 99%        | SPK: 50           |
| 1868-53-7                 | Dibromofluoromethane        | 49.3   |           | 70 (70) - 130 (134) | 99%        | SPK: 50           |
| 2037-26-5                 | Toluene-d8                  | 50.5   |           | 70 (74) - 130 (123) | 101%       | SPK: 50           |
| 460-00-4                  | 4-Bromofluorobenzene        | 50.6   |           | 70 (17) - 130 (146) | 101%       | SPK: 50           |
| <b>INTERNAL STANDARDS</b> |                             |        |           |                     |            |                   |
| 363-72-4                  | Pentafluorobenzene          | 502000 | 7.713     |                     |            |                   |
| 540-36-3                  | 1,4-Difluorobenzene         | 763000 | 8.616     |                     |            |                   |
| 3114-55-4                 | Chlorobenzene-d5            | 677000 | 11.42     |                     |            |                   |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 351000 | 13.346    |                     |            |                   |

## Report of Analysis

|                    |                 |                 |               |
|--------------------|-----------------|-----------------|---------------|
| Client:            | G Environmental | Date Collected: |               |
| Project:           | Buffington      | Date Received:  |               |
| Client Sample ID:  | VY0910SBS01     | SDG No.:        | Q3058         |
| Lab Sample ID:     | VY0910SBS01     | Matrix:         | SOIL          |
| Analytical Method: | 8260D           | % Solid:        | 100           |
| Sample Wt/Vol:     | 5               | Units:          | g             |
| Soil Aliquot Vol:  |                 | Final Vol:      | 5000 uL       |
|                    |                 | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624         | Level :         | LOW           |
| Prep Method :      | ID : 0.25       |                 |               |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VY023323.D        | 1         | 09/10/25 10:11 | VY091025      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



### Report of Analysis

|                    |                 |           |                 |               |    |
|--------------------|-----------------|-----------|-----------------|---------------|----|
| Client:            | G Environmental |           | Date Collected: |               |    |
| Project:           | Buffington      |           | Date Received:  |               |    |
| Client Sample ID:  | VY0910SBSD01    |           | SDG No.:        | Q3058         |    |
| Lab Sample ID:     | VY0910SBSD01    |           | Matrix:         | SOIL          |    |
| Analytical Method: | 8260D           |           | % Solid:        | 100           |    |
| Sample Wt/Vol:     | 5               | Units: g  | Final Vol:      | 5000          | uL |
| Soil Aliquot Vol:  |                 | uL        | Test:           | VOC-TCLVOA-10 |    |
| GC Column:         | RXI-624         | ID : 0.25 | Level :         | LOW           |    |
| Prep Method :      |                 |           |                 |               |    |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VY023324.D        | 1         | 09/10/25 10:33 | VY091025      |

| CAS Number                | Parameter                   | Conc.  | Qualifier | MDL                 | LOQ / CRQL | Units(Dry Weight) |
|---------------------------|-----------------------------|--------|-----------|---------------------|------------|-------------------|
| 10061-02-6                | t-1,3-Dichloropropene       | 19.7   |           | 0.65                | 5.00       | ug/Kg             |
| 10061-01-5                | cis-1,3-Dichloropropene     | 19.7   |           | 0.62                | 5.00       | ug/Kg             |
| 79-00-5                   | 1,1,2-Trichloroethane       | 20.1   |           | 0.92                | 5.00       | ug/Kg             |
| 591-78-6                  | 2-Hexanone                  | 100    |           | 3.70                | 25.0       | ug/Kg             |
| 124-48-1                  | Dibromochloromethane        | 20.1   |           | 0.87                | 5.00       | ug/Kg             |
| 106-93-4                  | 1,2-Dibromoethane           | 20.0   |           | 0.88                | 5.00       | ug/Kg             |
| 127-18-4                  | Tetrachloroethene           | 21.2   |           | 1.10                | 5.00       | ug/Kg             |
| 108-90-7                  | Chlorobenzene               | 20.2   |           | 0.91                | 5.00       | ug/Kg             |
| 100-41-4                  | Ethyl Benzene               | 19.9   |           | 0.67                | 5.00       | ug/Kg             |
| 179601-23-1               | m/p-Xylenes                 | 40.5   |           | 1.20                | 10.0       | ug/Kg             |
| 95-47-6                   | o-Xylene                    | 19.9   |           | 0.82                | 5.00       | ug/Kg             |
| 100-42-5                  | Styrene                     | 20.3   |           | 0.71                | 5.00       | ug/Kg             |
| 75-25-2                   | Bromoform                   | 20.4   |           | 0.86                | 5.00       | ug/Kg             |
| 98-82-8                   | Isopropylbenzene            | 19.8   |           | 0.78                | 5.00       | ug/Kg             |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 19.3   |           | 1.20                | 5.00       | ug/Kg             |
| 541-73-1                  | 1,3-Dichlorobenzene         | 20.3   |           | 1.70                | 5.00       | ug/Kg             |
| 106-46-7                  | 1,4-Dichlorobenzene         | 20.0   |           | 1.60                | 5.00       | ug/Kg             |
| 95-50-1                   | 1,2-Dichlorobenzene         | 20.2   |           | 1.50                | 5.00       | ug/Kg             |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 20.0   |           | 1.80                | 5.00       | ug/Kg             |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 20.3   |           | 3.00                | 5.00       | ug/Kg             |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 20.2   |           | 3.20                | 5.00       | ug/Kg             |
| <b>SURROGATES</b>         |                             |        |           |                     |            |                   |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 48.3   |           | 70 (63) - 130 (155) | 97%        | SPK: 50           |
| 1868-53-7                 | Dibromofluoromethane        | 48.5   |           | 70 (70) - 130 (134) | 97%        | SPK: 50           |
| 2037-26-5                 | Toluene-d8                  | 49.0   |           | 70 (74) - 130 (123) | 98%        | SPK: 50           |
| 460-00-4                  | 4-Bromofluorobenzene        | 48.7   |           | 70 (17) - 130 (146) | 97%        | SPK: 50           |
| <b>INTERNAL STANDARDS</b> |                             |        |           |                     |            |                   |
| 363-72-4                  | Pentafluorobenzene          | 502000 | 7.713     |                     |            |                   |
| 540-36-3                  | 1,4-Difluorobenzene         | 766000 | 8.616     |                     |            |                   |
| 3114-55-4                 | Chlorobenzene-d5            | 670000 | 11.414    |                     |            |                   |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 341000 | 13.346    |                     |            |                   |

## Report of Analysis

|                    |                 |           |                 |               |
|--------------------|-----------------|-----------|-----------------|---------------|
| Client:            | G Environmental |           | Date Collected: |               |
| Project:           | Buffington      |           | Date Received:  |               |
| Client Sample ID:  | VY0910SBSD01    |           | SDG No.:        | Q3058         |
| Lab Sample ID:     | VY0910SBSD01    |           | Matrix:         | SOIL          |
| Analytical Method: | 8260D           |           | % Solid:        | 100           |
| Sample Wt/Vol:     | 5               | Units: g  | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  |                 | uL        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624         | ID : 0.25 | Level :         | LOW           |
| Prep Method :      |                 |           |                 |               |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VY023324.D        | 1         | 09/10/25 10:33 | VY091025      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



# CALIBRATION SUMMARY

# VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance Contract: GENV01  
 Lab Code: ACE SDG No.: Q3058  
 Instrument ID: MSVOA\_N Calibration Date(s): 08/21/2025 08/21/2025  
 Heated Purge: (Y/N) N Calibration Time(s): 12:09 14:44  
 GC Column: RXI-624 ID: 0.25 (mm)

| LAB FILE ID:                   |        |        |                     |        |        |                     |       |       |
|--------------------------------|--------|--------|---------------------|--------|--------|---------------------|-------|-------|
| RRF001 = VN087619.D            |        |        | RRF005 = VN087620.D |        |        | RRF020 = VN087621.D |       |       |
| RRF050 = VN087622.D            |        |        | RRF100 = VN087623.D |        |        | RRF150 = VN087624.D |       |       |
| COMPOUND                       | RRF001 | RRF005 | RRF020              | RRF050 | RRF100 | RRF150              | RRF   | % RSD |
| Dichlorodifluoromethane        | 0.707  | 0.579  | 0.606               | 0.810  | 0.777  | 0.782               | 0.710 | 13.7  |
| Chloromethane                  | 0.739  | 0.637  | 0.710               | 0.795  | 0.752  | 0.747               | 0.730 | 7.3   |
| Vinyl Chloride                 | 0.909  | 0.739  | 0.832               | 0.903  | 0.847  | 0.846               | 0.846 | 7.3   |
| Bromomethane                   |        | 0.388  | 0.393               | 0.448  | 0.466  | 0.488               | 0.437 | 10.2  |
| Chloroethane                   | 0.727  | 0.528  | 0.599               | 0.603  | 0.558  | 0.554               | 0.595 | 11.9  |
| Trichlorofluoromethane         | 1.323  | 1.139  | 1.242               | 1.274  | 1.227  | 1.233               | 1.240 | 4.9   |
| 1,1,2-Trichlorotrifluoroethane | 0.879  | 0.648  | 0.677               | 0.695  | 0.654  | 0.656               | 0.701 | 12.6  |
| 1,1-Dichloroethene             | 0.892  | 0.649  | 0.730               | 0.700  | 0.667  | 0.687               | 0.721 | 12.3  |
| Acetone                        | 0.633  | 0.634  | 0.584               | 0.522  | 0.484  | 0.489               | 0.558 | 12.3  |
| Carbon Disulfide               | 2.272  | 1.746  | 1.970               | 2.003  | 1.937  | 1.979               | 1.985 | 8.5   |
| Methyl tert-butyl Ether        | 2.859  | 2.508  | 2.740               | 2.733  | 2.643  | 2.742               | 2.704 | 4.4   |
| Methyl Acetate                 | 1.450  | 1.043  | 1.097               | 1.157  | 1.117  | 1.141               | 1.168 | 12.3  |
| Methylene Chloride             | 1.494  | 0.916  | 0.875               | 0.823  | 0.779  | 0.799               | 0.948 | 28.8  |
| trans-1,2-Dichloroethene       | 0.952  | 0.734  | 0.768               | 0.750  | 0.737  | 0.745               | 0.781 | 10.8  |
| 1,1-Dichloroethane             | 1.798  | 1.500  | 1.549               | 1.501  | 1.454  | 1.471               | 1.546 | 8.3   |
| Cyclohexane                    |        | 1.490  | 1.310               | 1.234  | 1.201  | 1.207               | 1.289 | 9.4   |
| 2-Butanone                     | 0.783  | 0.722  | 0.757               | 0.732  | 0.700  | 0.707               | 0.734 | 4.3   |
| Carbon Tetrachloride           | 0.571  | 0.509  | 0.568               | 0.542  | 0.549  | 0.542               | 0.547 | 4.1   |
| cis-1,2-Dichloroethene         | 0.997  | 0.916  | 0.937               | 0.941  | 0.908  | 0.905               | 0.934 | 3.7   |
| Bromochloromethane             | 0.800  | 0.781  | 0.722               | 0.701  | 0.748  | 0.731               | 0.747 | 5     |
| Chloroform                     | 1.677  | 1.535  | 1.629               | 1.579  | 1.519  | 1.528               | 1.578 | 4     |
| 1,1,1-Trichloroethane          | 1.510  | 1.297  | 1.359               | 1.303  | 1.265  | 1.288               | 1.337 | 6.8   |
| Methylcyclohexane              | 0.686  | 0.580  | 0.610               | 0.590  | 0.597  | 0.596               | 0.610 | 6.3   |
| Benzene                        | 1.853  | 1.637  | 1.727               | 1.684  | 1.639  | 1.652               | 1.699 | 4.9   |
| 1,2-Dichloroethane             | 0.714  | 0.627  | 0.675               | 0.650  | 0.624  | 0.626               | 0.653 | 5.5   |
| Trichloroethene                | 0.435  | 0.387  | 0.390               | 0.377  | 0.367  | 0.370               | 0.388 | 6.5   |
| 1,2-Dichloropropane            | 0.544  | 0.435  | 0.447               | 0.433  | 0.415  | 0.412               | 0.448 | 11    |
| Bromodichloromethane           | 0.686  | 0.649  | 0.667               | 0.638  | 0.638  | 0.639               | 0.653 | 3     |
| 4-Methyl-2-Pentanone           | 0.701  | 0.675  | 0.744               | 0.736  | 0.718  | 0.704               | 0.713 | 3.5   |
| Toluene                        | 1.148  | 0.987  | 1.062               | 1.050  | 1.037  | 1.033               | 1.053 | 5     |

\* Compounds with required minimum RRF and maximum %RSD values.  
 All other compounds must meet a minimum RRF of 0.010.  
 RRF of 1,4-Dioxane = Value should be divide by 1000.

# VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG No.: Q3058

Instrument ID: MSVOA\_N

Calibration Date(s): 08/21/2025 08/21/2025

Heated Purge: (Y/N) N

Calibration Time(s): 12:09 14:44

GC Column: RXI-624 ID: 0.25 (mm)

| LAB FILE ID:                |        | RRF001 = VN087619.D |        | RRF005 = VN087620.D |        | RRF020 = VN087621.D |       |       |  |
|-----------------------------|--------|---------------------|--------|---------------------|--------|---------------------|-------|-------|--|
|                             |        | RRF050 = VN087622.D |        | RRF100 = VN087623.D |        | RRF150 = VN087624.D |       |       |  |
| COMPOUND                    | RRF001 | RRF005              | RRF020 | RRF050              | RRF100 | RRF150              | RRF   | % RSD |  |
| t-1,3-Dichloropropene       | 0.620  | 0.628               | 0.691  | 0.703               | 0.699  | 0.715               | 0.676 | 6.1   |  |
| cis-1,3-Dichloropropene     | 0.717  | 0.640               | 0.712  | 0.718               | 0.714  | 0.710               | 0.702 | 4.3   |  |
| 1,1,2-Trichloroethane       | 0.439  | 0.383               | 0.417  | 0.408               | 0.400  | 0.397               | 0.407 | 4.7   |  |
| 2-Hexanone                  | 0.295  | 0.390               | 0.493  | 0.510               | 0.508  | 0.507               | 0.451 | 19.8  |  |
| Dibromochloromethane        | 0.535  | 0.440               | 0.458  | 0.466               | 0.465  | 0.468               | 0.472 | 6.8   |  |
| 1,2-Dibromoethane           | 0.412  | 0.433               | 0.444  | 0.432               | 0.429  | 0.427               | 0.430 | 2.4   |  |
| Tetrachloroethene           | 0.447  | 0.353               | 0.340  | 0.313               | 0.315  | 0.313               | 0.347 | 14.9  |  |
| Chlorobenzene               | 1.408  | 1.164               | 1.253  | 1.210               | 1.216  | 1.212               | 1.244 | 6.9   |  |
| Ethyl Benzene               | 2.232  | 1.993               | 2.224  | 2.144               | 2.152  | 2.178               | 2.154 | 4     |  |
| m/p-Xylenes                 | 0.759  | 0.735               | 0.811  | 0.812               | 0.808  | 0.814               | 0.790 | 4.3   |  |
| o-Xylene                    | 0.777  | 0.707               | 0.794  | 0.796               | 0.788  | 0.784               | 0.774 | 4.3   |  |
| Styrene                     | 1.142  | 1.086               | 1.391  | 1.382               | 1.386  | 1.393               | 1.297 | 11    |  |
| Bromoform                   | 0.303  | 0.285               | 0.320  | 0.320               | 0.336  | 0.336               | 0.317 | 6.3   |  |
| Isopropylbenzene            | 4.328  | 3.674               | 4.096  | 4.026               | 3.998  | 4.120               | 4.040 | 5.3   |  |
| 1,1,2,2-Tetrachloroethane   | 1.488  | 1.381               | 1.421  | 1.363               | 1.322  | 1.371               | 1.391 | 4.1   |  |
| 1,3-Dichlorobenzene         | 2.087  | 1.770               | 1.940  | 1.827               | 1.806  | 1.826               | 1.876 | 6.3   |  |
| 1,4-Dichlorobenzene         | 2.348  | 1.884               | 1.956  | 1.867               | 1.818  | 1.841               | 1.952 | 10.2  |  |
| 1,2-Dichlorobenzene         | 1.942  | 1.662               | 1.853  | 1.755               | 1.723  | 1.743               | 1.780 | 5.7   |  |
| 1,2-Dibromo-3-Chloropropane | 0.370  | 0.323               | 0.329  | 0.317               | 0.313  | 0.330               | 0.330 | 6.2   |  |
| 1,2,4-Trichlorobenzene      | 1.192  | 0.995               | 1.081  | 1.026               | 1.040  | 1.079               | 1.069 | 6.4   |  |
| 1,2,3-Trichlorobenzene      | 1.140  | 0.972               | 1.059  | 1.006               | 1.019  | 1.072               | 1.045 | 5.7   |  |
| 1,2-Dichloroethane-d4       |        | 1.025               | 1.016  | 0.959               | 1.035  | 1.089               | 1.024 | 4.5   |  |
| Dibromofluoromethane        |        | 0.369               | 0.344  | 0.334               | 0.373  | 0.386               | 0.361 | 6     |  |
| Toluene-d8                  |        | 1.384               | 1.266  | 1.223               | 1.408  | 1.475               | 1.351 | 7.7   |  |
| 4-Bromofluorobenzene        |        | 0.446               | 0.454  | 0.450               | 0.514  | 0.540               | 0.481 | 9     |  |

\* Compounds with required minimum RRF and maximum %RSD values.  
All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

# VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG No.: Q3058

Instrument ID: MSVOA\_Y

Calibration Date(s): 09/08/2025 09/08/2025

Heated Purge: (Y/N) Y

Calibration Time(s): 10:00 12:08

GC Column: RXI-624 ID: 0.25 (mm)

| LAB FILE ID:                   |  | RRF005 = VY023303.D |        | RRF010 = VY023304.D |        | RRF020 = VY023305.D |        | RRF050 = VY023306.D |       | RRF100 = VY023307.D |  | RRF150 = VY023308.D |  |
|--------------------------------|--|---------------------|--------|---------------------|--------|---------------------|--------|---------------------|-------|---------------------|--|---------------------|--|
| COMPOUND                       |  | RRF005              | RRF010 | RRF020              | RRF050 | RRF100              | RRF150 | RRF                 | % RSD |                     |  |                     |  |
| Dichlorodifluoromethane        |  | 0.471               | 0.447  | 0.434               | 0.365  | 0.345               | 0.345  | 0.401               | 13.9  |                     |  |                     |  |
| Chloromethane                  |  | 0.661               | 0.637  | 0.604               | 0.534  | 0.507               | 0.502  | 0.574               | 12    |                     |  |                     |  |
| Vinyl Chloride                 |  | 0.800               | 0.786  | 0.757               | 0.695  | 0.653               | 0.647  | 0.723               | 9.3   |                     |  |                     |  |
| Bromomethane                   |  | 0.709               | 0.630  | 0.611               | 0.533  | 0.524               | 0.537  | 0.590               | 12.4  |                     |  |                     |  |
| Chloroethane                   |  | 0.544               | 0.529  | 0.504               | 0.471  | 0.447               | 0.441  | 0.489               | 8.8   |                     |  |                     |  |
| Trichlorofluoromethane         |  | 0.990               | 0.998  | 0.953               | 0.870  | 0.843               | 0.822  | 0.913               | 8.4   |                     |  |                     |  |
| 1,1,2-Trichlorotrifluoroethane |  | 0.556               | 0.524  | 0.511               | 0.470  | 0.446               | 0.437  | 0.491               | 9.6   |                     |  |                     |  |
| 1,1-Dichloroethene             |  | 0.501               | 0.504  | 0.481               | 0.458  | 0.446               | 0.443  | 0.472               | 5.7   |                     |  |                     |  |
| Acetone                        |  | 0.112               | 0.110  | 0.096               | 0.100  | 0.095               | 0.087  | 0.100               | 9.3   |                     |  |                     |  |
| Carbon Disulfide               |  | 1.643               | 1.589  | 1.553               | 1.452  | 1.383               | 1.351  | 1.495               | 7.9   |                     |  |                     |  |
| Methyl tert-butyl Ether        |  | 1.070               | 1.130  | 1.123               | 1.121  | 1.121               | 1.096  | 1.110               | 2     |                     |  |                     |  |
| Methyl Acetate                 |  | 0.255               | 0.275  | 0.254               | 0.233  | 0.235               | 0.220  | 0.245               | 8.1   |                     |  |                     |  |
| Methylene Chloride             |  | 0.654               | 0.569  | 0.542               | 0.503  | 0.478               | 0.467  | 0.535               | 13    |                     |  |                     |  |
| trans-1,2-Dichloroethene       |  | 0.555               | 0.544  | 0.542               | 0.519  | 0.506               | 0.498  | 0.528               | 4.4   |                     |  |                     |  |
| 1,1-Dichloroethane             |  | 0.918               | 0.913  | 0.899               | 0.844  | 0.823               | 0.816  | 0.869               | 5.3   |                     |  |                     |  |
| Cyclohexane                    |  | 0.882               | 0.799  | 0.790               | 0.728  | 0.711               | 0.704  | 0.769               | 8.9   |                     |  |                     |  |
| 2-Butanone                     |  | 0.123               | 0.130  | 0.126               | 0.122  | 0.123               | 0.116  | 0.123               | 3.7   |                     |  |                     |  |
| Carbon Tetrachloride           |  | 0.513               | 0.525  | 0.506               | 0.503  | 0.483               | 0.475  | 0.501               | 3.7   |                     |  |                     |  |
| cis-1,2-Dichloroethene         |  | 0.595               | 0.617  | 0.606               | 0.592  | 0.590               | 0.586  | 0.597               | 2     |                     |  |                     |  |
| Bromochloromethane             |  | 0.376               | 0.357  | 0.355               | 0.329  | 0.317               | 0.317  | 0.342               | 7.1   |                     |  |                     |  |
| Chloroform                     |  | 0.994               | 0.996  | 0.961               | 0.907  | 0.878               | 0.865  | 0.934               | 6.2   |                     |  |                     |  |
| 1,1,1-Trichloroethane          |  | 0.885               | 0.889  | 0.860               | 0.807  | 0.785               | 0.778  | 0.834               | 6     |                     |  |                     |  |
| Methylcyclohexane              |  | 0.529               | 0.531  | 0.559               | 0.580  | 0.571               | 0.571  | 0.557               | 3.9   |                     |  |                     |  |
| Benzene                        |  | 1.380               | 1.361  | 1.391               | 1.369  | 1.318               | 1.299  | 1.353               | 2.7   |                     |  |                     |  |
| 1,2-Dichloroethane             |  | 0.357               | 0.367  | 0.352               | 0.344  | 0.330               | 0.318  | 0.344               | 5.2   |                     |  |                     |  |
| Trichloroethene                |  | 0.387               | 0.378  | 0.388               | 0.381  | 0.369               | 0.359  | 0.377               | 3     |                     |  |                     |  |
| 1,2-Dichloropropane            |  | 0.313               | 0.315  | 0.316               | 0.305  | 0.297               | 0.288  | 0.306               | 3.6   |                     |  |                     |  |
| Bromodichloromethane           |  | 0.476               | 0.487  | 0.474               | 0.474  | 0.453               | 0.445  | 0.468               | 3.4   |                     |  |                     |  |
| 4-Methyl-2-Pentanone           |  | 0.156               | 0.170  | 0.173               | 0.181  | 0.181               | 0.173  | 0.172               | 5.3   |                     |  |                     |  |
| Toluene                        |  | 0.794               | 0.831  | 0.885               | 0.896  | 0.879               | 0.868  | 0.859               | 4.5   |                     |  |                     |  |

\* Compounds with required minimum RRF and maximum %RSD values.  
All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

# VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG No.: Q3058

Instrument ID: MSVOA\_Y

Calibration Date(s): 09/08/2025 09/08/2025

Heated Purge: (Y/N) Y

Calibration Time(s): 10:00 12:08

GC Column: RXI-624 ID: 0.25 (mm)

| LAB FILE ID:                |        |        |                     |        |        |                     |       |       |
|-----------------------------|--------|--------|---------------------|--------|--------|---------------------|-------|-------|
| RRF005 = VY023303.D         |        |        | RRF010 = VY023304.D |        |        | RRF020 = VY023305.D |       |       |
| RRF050 = VY023306.D         |        |        | RRF100 = VY023307.D |        |        | RRF150 = VY023308.D |       |       |
| COMPOUND                    | RRF005 | RRF010 | RRF020              | RRF050 | RRF100 | RRF150              | RRF   | % RSD |
| t-1,3-Dichloropropene       | 0.374  | 0.395  | 0.401               | 0.420  | 0.417  | 0.410               | 0.403 | 4.2   |
| cis-1,3-Dichloropropene     | 0.450  | 0.477  | 0.492               | 0.497  | 0.491  | 0.483               | 0.482 | 3.6   |
| 1,1,2-Trichloroethane       | 0.250  | 0.256  | 0.250               | 0.250  | 0.244  | 0.235               | 0.248 | 3     |
| 2-Hexanone                  | 0.106  | 0.116  | 0.120               | 0.126  | 0.127  | 0.120               | 0.119 | 6.6   |
| Dibromochloromethane        | 0.334  | 0.340  | 0.338               | 0.343  | 0.334  | 0.326               | 0.336 | 1.8   |
| 1,2-Dibromoethane           | 0.232  | 0.234  | 0.236               | 0.235  | 0.231  | 0.223               | 0.232 | 2     |
| Tetrachloroethene           | 0.534  | 0.483  | 0.546               | 0.512  | 0.472  | 0.442               | 0.498 | 7.9   |
| Chlorobenzene               | 1.113  | 1.087  | 1.124               | 1.106  | 1.081  | 1.061               | 1.095 | 2.1   |
| Ethyl Benzene               | 1.664  | 1.694  | 1.828               | 1.874  | 1.848  | 1.828               | 1.790 | 4.9   |
| m/p-Xylenes                 | 0.655  | 0.682  | 0.738               | 0.753  | 0.735  | 0.736               | 0.717 | 5.4   |
| o-Xylene                    | 0.587  | 0.613  | 0.676               | 0.703  | 0.699  | 0.701               | 0.663 | 7.6   |
| Styrene                     | 0.925  | 1.020  | 1.127               | 1.187  | 1.172  | 1.165               | 1.099 | 9.5   |
| Bromoform                   | 0.232  | 0.230  | 0.229               | 0.231  | 0.228  | 0.223               | 0.229 | 1.5   |
| Isopropylbenzene            | 3.112  | 3.173  | 3.434               | 3.542  | 3.466  | 3.482               | 3.368 | 5.3   |
| 1,1,2,2-Tetrachloroethane   | 0.564  | 0.627  | 0.549               | 0.557  | 0.554  | 0.541               | 0.565 | 5.6   |
| 1,3-Dichlorobenzene         | 1.712  | 1.672  | 1.704               | 1.742  | 1.688  | 1.699               | 1.703 | 1.4   |
| 1,4-Dichlorobenzene         | 1.775  | 1.681  | 1.711               | 1.695  | 1.632  | 1.624               | 1.686 | 3.3   |
| 1,2-Dichlorobenzene         | 1.508  | 1.480  | 1.491               | 1.513  | 1.467  | 1.448               | 1.485 | 1.7   |
| 1,2-Dibromo-3-Chloropropane | 0.096  | 0.089  | 0.089               | 0.087  | 0.087  | 0.084               | 0.089 | 4.4   |
| 1,2,4-Trichlorobenzene      | 0.786  | 0.797  | 0.844               | 0.905  | 0.924  | 0.940               | 0.866 | 7.7   |
| 1,2,3-Trichlorobenzene      | 0.697  | 0.704  | 0.733               | 0.787  | 0.786  | 0.805               | 0.752 | 6.2   |
| 1,2-Dichloroethane-d4       | 0.472  | 0.456  | 0.455               | 0.430  | 0.417  | 0.407               | 0.439 | 5.8   |
| Dibromofluoromethane        | 0.300  | 0.313  | 0.311               | 0.312  | 0.300  | 0.300               | 0.306 | 2.2   |
| Toluene-d8                  | 1.121  | 1.125  | 1.176               | 1.187  | 1.155  | 1.159               | 1.154 | 2.3   |
| 4-Bromofluorobenzene        | 0.355  | 0.339  | 0.363               | 0.378  | 0.373  | 0.371               | 0.363 | 4     |

\* Compounds with required minimum RRF and maximum %RSD values.  
All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

|                     |                   |                        |                              |
|---------------------|-------------------|------------------------|------------------------------|
| Lab Name:           | <u>Alliance</u>   | Contract:              | <u>GENV01</u>                |
| Lab Code:           | <u>ACE</u>        | SDG No.:               | <u>Q3058</u>                 |
| Instrument ID:      | <u>MSVOA_N</u>    | Calibration Date/Time: | <u>09/15/2025 08:39</u>      |
| Lab File ID:        | <u>VN087820.D</u> | Init. Calib. Date(s):  | <u>08/21/2025 08/21/2025</u> |
| Heated Purge: (Y/N) | <u>N</u>          | Init. Calib. Time(s):  | <u>12:09 14:44</u>           |
| GC Column:          | <u>RXI-624</u>    | ID:                    | <u>0.25</u> (mm)             |

| COMPOUND                       | RRF   | RRF050 | MIN RRF | %D     | MAX%D |
|--------------------------------|-------|--------|---------|--------|-------|
| Dichlorodifluoromethane        | 0.710 | 0.674  |         | -5.07  | 20    |
| Chloromethane                  | 0.730 | 0.685  | 0.1     | -6.16  | 20    |
| Vinyl Chloride                 | 0.846 | 0.784  |         | -7.33  | 20    |
| Bromomethane                   | 0.437 | 0.354  |         | -18.99 | 20    |
| Chloroethane                   | 0.595 | 0.533  |         | -10.42 | 20    |
| Trichlorofluoromethane         | 1.240 | 1.187  |         | -4.27  | 20    |
| 1,1,2-Trichlorotrifluoroethane | 0.701 | 0.651  |         | -7.13  | 20    |
| 1,1-Dichloroethene             | 0.721 | 0.665  |         | -7.77  | 20    |
| Acetone                        | 0.558 | 0.511  |         | -8.42  | 20    |
| Carbon Disulfide               | 1.985 | 1.956  |         | -1.46  | 20    |
| Methyl tert-butyl Ether        | 2.704 | 2.560  |         | -5.32  | 20    |
| Methyl Acetate                 | 1.168 | 1.112  |         | -4.8   | 20    |
| Methylene Chloride             | 0.948 | 0.796  |         | -16.03 | 20    |
| trans-1,2-Dichloroethene       | 0.781 | 0.739  |         | -5.38  | 20    |
| 1,1-Dichloroethane             | 1.546 | 1.420  | 0.1     | -8.15  | 20    |
| Cyclohexane                    | 1.289 | 1.117  |         | -13.34 | 20    |
| 2-Butanone                     | 0.734 | 0.692  |         | -5.72  | 20    |
| Carbon Tetrachloride           | 0.547 | 0.551  |         | 0.73   | 20    |
| cis-1,2-Dichloroethene         | 0.934 | 0.888  |         | -4.93  | 20    |
| Bromochloromethane             | 0.747 | 0.731  |         | -2.14  | 20    |
| Chloroform                     | 1.578 | 1.540  |         | -2.41  | 20    |
| 1,1,1-Trichloroethane          | 1.337 | 1.270  |         | -5.01  | 20    |
| Methylcyclohexane              | 0.610 | 0.567  |         | -7.05  | 20    |
| Benzene                        | 1.699 | 1.692  |         | -0.41  | 20    |
| 1,2-Dichloroethane             | 0.653 | 0.615  |         | -5.82  | 20    |
| Trichloroethene                | 0.388 | 0.380  |         | -2.06  | 20    |
| 1,2-Dichloropropane            | 0.448 | 0.427  |         | -4.69  | 20    |
| Bromodichloromethane           | 0.653 | 0.644  |         | -1.38  | 20    |
| 4-Methyl-2-Pentanone           | 0.713 | 0.714  |         | 0.14   | 20    |
| Toluene                        | 1.053 | 1.045  |         | -0.76  | 20    |
| t-1,3-Dichloropropene          | 0.676 | 0.681  |         | 0.74   | 20    |
| cis-1,3-Dichloropropene        | 0.702 | 0.706  |         | 0.57   | 20    |
| 1,1,2-Trichloroethane          | 0.407 | 0.423  |         | 3.93   | 20    |
| 2-Hexanone                     | 0.451 | 0.493  |         | 9.31   | 20    |
| Dibromochloromethane           | 0.472 | 0.494  |         | 4.66   | 20    |
| 1,2-Dibromoethane              | 0.430 | 0.445  |         | 3.49   | 20    |
| Tetrachloroethene              | 0.347 | 0.329  |         | -5.19  | 20    |
| Chlorobenzene                  | 1.244 | 1.210  | 0.3     | -2.73  | 20    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

|                     |                   |                        |                              |
|---------------------|-------------------|------------------------|------------------------------|
| Lab Name:           | <u>Alliance</u>   | Contract:              | <u>GENV01</u>                |
| Lab Code:           | <u>ACE</u>        | SDG No.:               | <u>Q3058</u>                 |
| Instrument ID:      | <u>MSVOA_N</u>    | Calibration Date/Time: | <u>09/15/2025 08:39</u>      |
| Lab File ID:        | <u>VN087820.D</u> | Init. Calib. Date(s):  | <u>08/21/2025 08/21/2025</u> |
| Heated Purge: (Y/N) | <u>N</u>          | Init. Calib. Time(s):  | <u>12:09 14:44</u>           |
| GC Column:          | <u>RXI-624</u>    | ID:                    | <u>0.25</u> (mm)             |

| COMPOUND                    | RRF   | RRF050 | MIN<br>RRF | %D     | MAX%D |
|-----------------------------|-------|--------|------------|--------|-------|
| Ethyl Benzene               | 2.154 | 2.096  |            | -2.69  | 20    |
| m/p-Xylenes                 | 0.790 | 0.800  |            | 1.27   | 20    |
| o-Xylene                    | 0.774 | 0.760  |            | -1.81  | 20    |
| Styrene                     | 1.297 | 1.352  |            | 4.24   | 20    |
| Bromoform                   | 0.317 | 0.354  | 0.1        | 11.67  | 20    |
| Isopropylbenzene            | 4.040 | 3.523  |            | -12.8  | 20    |
| 1,1,2,2-Tetrachloroethane   | 1.391 | 1.270  | 0.3        | -8.7   | 20    |
| 1,3-Dichlorobenzene         | 1.876 | 1.742  |            | -7.14  | 20    |
| 1,4-Dichlorobenzene         | 1.952 | 1.779  |            | -8.86  | 20    |
| 1,2-Dichlorobenzene         | 1.780 | 1.694  |            | -4.83  | 20    |
| 1,2-Dibromo-3-Chloropropane | 0.330 | 0.283  |            | -14.24 | 20    |
| 1,2,4-Trichlorobenzene      | 1.069 | 1.018  |            | -4.77  | 20    |
| 1,2,3-Trichlorobenzene      | 1.045 | 0.985  |            | -5.74  | 20    |
| 1,2-Dichloroethane-d4       | 1.024 | 1.010  |            | -1.37  | 20    |
| Dibromofluoromethane        | 0.361 | 0.390  |            | 8.03   | 20    |
| Toluene-d8                  | 1.351 | 1.382  |            | 2.3    | 20    |
| 4-Bromofluorobenzene        | 0.481 | 0.505  |            | 4.99   | 20    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

|                     |                   |                        |                                     |
|---------------------|-------------------|------------------------|-------------------------------------|
| Lab Name:           | <u>Alliance</u>   | Contract:              | <u>GENV01</u>                       |
| Lab Code:           | <u>ACE</u>        | SDG No.:               | <u>Q3058</u>                        |
| Instrument ID:      | <u>MSVOA_Y</u>    | Calibration Date/Time: | <u>09/10/2025</u> <u>09:09</u>      |
| Lab File ID:        | <u>VY023321.D</u> | Init. Calib. Date(s):  | <u>09/08/2025</u> <u>09/08/2025</u> |
| Heated Purge: (Y/N) | <u>Y</u>          | Init. Calib. Time(s):  | <u>10:00</u> <u>12:08</u>           |
| GC Column:          | <u>RXI-624</u>    | ID:                    | <u>0.25</u> (mm)                    |

| COMPOUND                       | RRF   | RRF050 | MIN RRF | %D    | MAX%D |
|--------------------------------|-------|--------|---------|-------|-------|
| Dichlorodifluoromethane        | 0.401 | 0.365  |         | -8.98 | 20    |
| Chloromethane                  | 0.574 | 0.562  | 0.1     | -2.09 | 20    |
| Vinyl Chloride                 | 0.723 | 0.711  |         | -1.66 | 20    |
| Bromomethane                   | 0.590 | 0.544  |         | -7.8  | 20    |
| Chloroethane                   | 0.489 | 0.487  |         | -0.41 | 20    |
| Trichlorofluoromethane         | 0.913 | 0.907  |         | -0.66 | 20    |
| 1,1,2-Trichlorotrifluoroethane | 0.491 | 0.482  |         | -1.83 | 20    |
| 1,1-Dichloroethene             | 0.472 | 0.460  |         | -2.54 | 20    |
| Acetone                        | 0.100 | 0.115  |         | 15    | 20    |
| Carbon Disulfide               | 1.495 | 1.461  |         | -2.27 | 20    |
| Methyl tert-butyl Ether        | 1.110 | 1.081  |         | -2.61 | 20    |
| Methyl Acetate                 | 0.245 | 0.230  |         | -6.12 | 20    |
| Methylene Chloride             | 0.535 | 0.538  |         | 0.56  | 20    |
| trans-1,2-Dichloroethene       | 0.528 | 0.524  |         | -0.76 | 20    |
| 1,1-Dichloroethane             | 0.869 | 0.854  | 0.1     | -1.73 | 20    |
| Cyclohexane                    | 0.769 | 0.736  |         | -4.29 | 20    |
| 2-Butanone                     | 0.123 | 0.128  |         | 4.07  | 20    |
| Carbon Tetrachloride           | 0.501 | 0.526  |         | 4.99  | 20    |
| cis-1,2-Dichloroethene         | 0.597 | 0.594  |         | -0.5  | 20    |
| Bromochloromethane             | 0.342 | 0.335  |         | -2.05 | 20    |
| Chloroform                     | 0.934 | 0.915  |         | -2.03 | 20    |
| 1,1,1-Trichloroethane          | 0.834 | 0.831  |         | -0.36 | 20    |
| Methylcyclohexane              | 0.557 | 0.600  |         | 7.72  | 20    |
| Benzene                        | 1.353 | 1.411  |         | 4.29  | 20    |
| 1,2-Dichloroethane             | 0.344 | 0.352  |         | 2.33  | 20    |
| Trichloroethene                | 0.377 | 0.392  |         | 3.98  | 20    |
| 1,2-Dichloropropane            | 0.306 | 0.314  |         | 2.61  | 20    |
| Bromodichloromethane           | 0.468 | 0.482  |         | 2.99  | 20    |
| 4-Methyl-2-Pentanone           | 0.172 | 0.179  |         | 4.07  | 20    |
| Toluene                        | 0.859 | 0.926  |         | 7.8   | 20    |
| t-1,3-Dichloropropene          | 0.403 | 0.417  |         | 3.47  | 20    |
| cis-1,3-Dichloropropene        | 0.482 | 0.499  |         | 3.53  | 20    |
| 1,1,2-Trichloroethane          | 0.248 | 0.253  |         | 2.02  | 20    |
| 2-Hexanone                     | 0.119 | 0.132  |         | 10.92 | 20    |
| Dibromochloromethane           | 0.336 | 0.353  |         | 5.06  | 20    |
| 1,2-Dibromoethane              | 0.232 | 0.240  |         | 3.45  | 20    |
| Tetrachloroethene              | 0.498 | 0.538  |         | 8.03  | 20    |
| Chlorobenzene                  | 1.095 | 1.134  | 0.3     | 3.56  | 20    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

|                     |                   |                        |                                     |
|---------------------|-------------------|------------------------|-------------------------------------|
| Lab Name:           | <u>Alliance</u>   | Contract:              | <u>GENV01</u>                       |
| Lab Code:           | <u>ACE</u>        | SDG No.:               | <u>Q3058</u>                        |
| Instrument ID:      | <u>MSVOA_Y</u>    | Calibration Date/Time: | <u>09/10/2025</u> <u>09:09</u>      |
| Lab File ID:        | <u>VY023321.D</u> | Init. Calib. Date(s):  | <u>09/08/2025</u> <u>09/08/2025</u> |
| Heated Purge: (Y/N) | <u>Y</u>          | Init. Calib. Time(s):  | <u>10:00</u> <u>12:08</u>           |
| GC Column:          | <u>RXI-624</u>    | ID:                    | <u>0.25</u> (mm)                    |

| COMPOUND                    | RRF   | RRF050 | MIN<br>RRF | %D    | MAX%D |
|-----------------------------|-------|--------|------------|-------|-------|
| Ethyl Benzene               | 1.790 | 1.942  |            | 8.49  | 20    |
| m/p-Xylenes                 | 0.717 | 0.782  |            | 9.07  | 20    |
| o-Xylene                    | 0.663 | 0.719  |            | 8.45  | 20    |
| Styrene                     | 1.099 | 1.196  |            | 8.83  | 20    |
| Bromoform                   | 0.229 | 0.234  | 0.1        | 2.18  | 20    |
| Isopropylbenzene            | 3.368 | 3.621  |            | 7.51  | 20    |
| 1,1,2,2-Tetrachloroethane   | 0.565 | 0.553  | 0.3        | -2.12 | 20    |
| 1,3-Dichlorobenzene         | 1.703 | 1.741  |            | 2.23  | 20    |
| 1,4-Dichlorobenzene         | 1.686 | 1.695  |            | 0.53  | 20    |
| 1,2-Dichlorobenzene         | 1.485 | 1.502  |            | 1.14  | 20    |
| 1,2-Dibromo-3-Chloropropane | 0.089 | 0.086  |            | -3.37 | 20    |
| 1,2,4-Trichlorobenzene      | 0.866 | 0.881  |            | 1.73  | 20    |
| 1,2,3-Trichlorobenzene      | 0.752 | 0.762  |            | 1.33  | 20    |
| 1,2-Dichloroethane-d4       | 0.439 | 0.427  |            | -2.73 | 20    |
| Dibromofluoromethane        | 0.306 | 0.315  |            | 2.94  | 20    |
| Toluene-d8                  | 1.154 | 1.218  |            | 5.55  | 20    |
| 4-Bromofluorobenzene        | 0.363 | 0.388  |            | 6.89  | 20    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.



# SAMPLE RAW DATA

5

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY091025\  
 Data File : VY023328.D  
 Acq On : 10 Sep 2025 12:49  
 Operator : SY/MD  
 Sample : Q3058-01  
 Misc : 6.55g/5.0mL/MSVOA\_Y/SOIL/A  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 RPXY5

### Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 09/11/2025  
 Supervised By :Semsettin Yesilyurt 09/11/2025

Quant Time: Sep 11 03:49:14 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y090825S.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 10 01:44:33 2025  
 Response via : Initial Calibration

| Compound                    | R.T.           | QIon | Response            | Conc    | Units  | Dev(Min) |
|-----------------------------|----------------|------|---------------------|---------|--------|----------|
| -----                       |                |      |                     |         |        |          |
| Internal Standards          |                |      |                     |         |        |          |
| 1) Pentafluorobenzene       | 7.713          | 168  | 542076              | 50.000  | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 8.615          | 114  | 1032321             | 50.000  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 11.414         | 117  | 977441              | 50.000  | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.346         | 152  | 418953              | 50.000  | ug/l   | 0.00     |
|                             |                |      |                     |         |        |          |
| System Monitoring Compounds |                |      |                     |         |        |          |
| 33) 1,2-Dichloroethane-d4   | 8.061          | 65   | 295336              | 62.002  | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 50 - 163 |      | Recovery = 124.000% |         |        |          |
| 35) Dibromofluoromethane    | 7.634          | 113  | 341940              | 54.188  | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 54 - 147 |      | Recovery = 108.380% |         |        |          |
| 50) Toluene-d8              | 10.109         | 98   | 1240309             | 52.053  | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 58 - 134 |      | Recovery = 104.100% |         |        |          |
| 62) 4-Bromofluorobenzene    | 12.407         | 95   | 409365              | 54.593  | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 30 - 143 |      | Recovery = 109.180% |         |        |          |
|                             |                |      |                     |         |        |          |
| Target Compounds            |                |      |                     |         | Qvalue |          |
| 16) Acetone                 | 3.866          | 43   | 12220m              | 11.263  | ug/l   |          |
| 20) Methylene Chloride      | 4.622          | 84   | 22696               | 3.910   | ug/l   | 91       |
| 52) Toluene                 | 10.170         | 92   | 45192               | 2.549   | ug/l   | 98       |
| 67) Ethyl Benzene           | 11.517         | 91   | 3721737             | 106.388 | ug/l   | 97       |
| 68) m/p-Xylenes             | 11.627         | 106  | 5242152             | 374.226 | ug/l   | 92       |
| 69) o-Xylene                | 11.956         | 106  | 1255978             | 96.856  | ug/l   | 98       |
| -----                       |                |      |                     |         |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY091025\  
Data File : VY023328.D  
Acq On : 10 Sep 2025 12:49  
Operator : SY/MD  
Sample : Q3058-01  
Misc : 6.55g/5.0mL/MSVOA\_Y/SOIL/A  
ALS Vial : 9 Sample Multiplier: 1


Instrument :  
MSVOA\_Y  
ClientSampleId :  
RPXY5

Manual Integrations  
APPROVED

Reviewed By :Mahesh Dadoda 09/11/2025  
Supervised By :Semsettin Yesilyurt 09/11/2025


Quant Time: Sep 11 03:49:14 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y090825S.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 10 01:44:33 2025  
Response via : Initial Calibration





#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 7.713 min Scan# 91  
Delta R.T. 0.006 min  
Lab File: VY023328.D  
Acq: 10 Sep 2025 12:49




Instrument :  
MSVOA\_Y  
ClientSampleId :  
RPXY5



Tgt Ion:168 Resp: 542070  
Ion Ratio Lower Upper  
168 100  
99 51.4 42.8 64.2




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 09/11/2025  
Supervised By :Semsettin Yesilyurt 09/11/2025



#16  
Acetone  
Concen: 11.263 ug/l m  
RT: 3.866 min Scan# 352  
Delta R.T. 0.006 min  
Lab File: VY023328.D  
Acq: 10 Sep 2025 12:49


Tgt Ion: 43 Resp: 12220  
Ion Ratio Lower Upper  
43 100  
58 35.9 27.1 40.7





#20  
Methylene Chloride  
Concen: 3.910 ug/l  
RT: 4.622 min Scan# 41  
Delta R.T. 0.018 min  
Lab File: VY023328.D  
Acq: 10 Sep 2025 12:49




Instrument :  
MSVOA\_Y  
ClientSampleId :  
RPXY5



Tgt Ion: 84 Resp: 22690  
Ion Ratio Lower Upper  
84 100  
49 105.1 93.0 139.6  
51 30.5 29.0 43.4  
86 59.2 50.2 75.4




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 09/11/2025  
Supervised By :Semsettin Yesilyurt 09/11/2025



#33  
1,2-Dichloroethane-d4  
Concen: 62.002 ug/l  
RT: 8.061 min Scan# 1040  
Delta R.T. 0.006 min  
Lab File: VY023328.D  
Acq: 10 Sep 2025 12:49

Tgt Ion: 65 Resp: 295336  
Ion Ratio Lower Upper  
65 100  
67 53.1 0.0 106.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.615 min Scan# 1131

Delta R.T. 0.006 min

Lab File: VY023328.D

Acq: 10 Sep 2025 12:49

Instrument :

MSVOA\_Y

ClientSampleId :

RPXY5

Tgt Ion:114 Resp: 103232

Ion Ratio Lower Upper

114 100

63 18.0 0.0 38.4





88 15.2 0.0 30.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/11/2025

Supervised By :Semsettin Yesilyurt 09/11/2025



#35

Dibromofluoromethane

Concen: 54.188 ug/l

RT: 7.634 min Scan# 970

Delta R.T. 0.006 min

Lab File: VY023328.D

Acq: 10 Sep 2025 12:49




Tgt Ion:113 Resp: 341940


Ion Ratio Lower Upper

113 100

111 101.7 81.1 121.7


192 21.1 16.2 24.4





#50  
Toluene-d8  
Concen: 52.053 ug/l  
RT: 10.109 min Scan# 1376  
Delta R.T. 0.006 min  
Lab File: VY023328.D  
Acq: 10 Sep 2025 12:49




Instrument :  
MSVOA\_Y  
ClientSampleId :  
RPXY5



Tgt Ion: 98 Resp: 1240309  
Ion Ratio Lower Upper  
98 100  
100 64.8 52.3 78.5




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 09/11/2025  
Supervised By :Semsettin Yesilyurt 09/11/2025



#52  
Toluene  
Concen: 2.549 ug/l  
RT: 10.170 min Scan# 1386  
Delta R.T. 0.006 min  
Lab File: VY023328.D  
Acq: 10 Sep 2025 12:49


Tgt Ion: 92 Resp: 45192  
Ion Ratio Lower Upper  
92 100  
91 168.3 136.4 204.6





#62  
4-Bromofluorobenzene  
Concen: 54.593 ug/l  
RT: 12.407 min Scan# 1753  
Delta R.T. 0.006 min  
Lab File: VY023328.D  
Acq: 10 Sep 2025 12:49




Instrument :  
MSVOA\_Y  
ClientSampleId :  
RPXY5



Tgt Ion: 95 Resp: 40936  
Ion Ratio Lower Upper  
95 100  
174 93.3 0.0 171.6  
176 89.7 0.0 167.6




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 09/11/2025  
Supervised By :Semsettin Yesilyurt 09/11/2025



#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.414 min Scan# 1590  
Delta R.T. 0.006 min  
Lab File: VY023328.D  
Acq: 10 Sep 2025 12:49

Tgt Ion:117 Resp: 977441  
Ion Ratio Lower Upper  
117 100  
82 52.7 43.4 65.0  
119 32.3 25.6 38.4





#67

Ethyl Benzene

Concen: 106.388 ug/l

RT: 11.517 min Scan# 1

Delta R.T. 0.006 min

Lab File: VY023328.D

Acq: 10 Sep 2025 12:49

Instrument :

MSVOA\_Y

ClientSampleId :

RPXY5

Tgt Ion: 91 Resp: 372173





Ion Ratio Lower Upper

91 100

106 32.6 25.0 37.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/11/2025  
Supervised By :Semsettin Yesilyurt 09/11/2025

#68

m/p-Xylenes

Concen: 374.226 ug/l

RT: 11.627 min Scan# 1625

Delta R.T. 0.006 min

Lab File: VY023328.D




Acq: 10 Sep 2025 12:49


Tgt Ion:106 Resp: 5242152

Ion Ratio Lower Upper

106 100


91 185.8 158.6 238.0





#69  
o-Xylene  
Concen: 96.856 ug/l  
RT: 11.956 min Scan# 1679  
Delta R.T. 0.006 min  
Lab File: VY023328.D  
Acq: 10 Sep 2025 12:49




Instrument :  
MSVOA\_Y  
ClientSampleId :  
RPXY5



Tgt Ion:106 Resp: 1255978  
Ion Ratio Lower Upper  
106 100  
91 202.8 103.2 309.4




Manual Integrations  
APPROVED

Reviewed By :Mahesh Dadoda 09/11/2025  
Supervised By :Semsettin Yesilyurt 09/11/2025



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 13.346 min Scan# 1907  
Delta R.T. 0.006 min  
Lab File: VY023328.D  
Acq: 10 Sep 2025 12:49

Tgt Ion:152 Resp: 418953  
Ion Ratio Lower Upper  
152 100  
115 55.7 28.2 84.6  
150 155.8 0.0 348.0



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY091025\  
Data File : VY023328.D  
Acq On : 10 Sep 2025 12:49  
Operator : SY/MD  
Sample : Q3058-01  
Misc : 6.55g/5.0mL/MSVOA\_Y/SOIL/A  
ALS Vial : 9 Sample Multiplier: 1

Instrument :  
MSVOA\_Y  
ClientSampleId :  
RPXY5

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Sampling : 1

Start Thrs: 0.2

Stop Thrs : 0

Filtering: 5

Min Area: 3 % of largest Peak

Max Peaks: 100

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y090825S.M

Title : SW846 8260

Signal : TIC: VY023328.D\data.ms

| peak<br># | R.T.<br>min | first<br>scan | max<br>scan | last<br>scan | PK<br>TY | peak<br>height | corr.<br>area | corr.<br>% max. | % of<br>total |
|-----------|-------------|---------------|-------------|--------------|----------|----------------|---------------|-----------------|---------------|
| 1         | 7.640       | 959           | 971         | 977          | rBV      | 462036         | 1148799       | 4.06%           | 1.898%        |
| 2         | 7.713       | 977           | 983         | 995          | rVB      | 674993         | 1516600       | 5.36%           | 2.506%        |
| 3         | 8.061       | 1031          | 1040        | 1057         | rBV      | 385458         | 849932        | 3.00%           | 1.405%        |
| 4         | 8.615       | 1122          | 1131        | 1146         | rBV      | 1172933        | 2275341       | 8.03%           | 3.760%        |
| 5         | 10.109      | 1365          | 1376        | 1382         | rBV      | 1887280        | 3165464       | 11.18%          | 5.231%        |
| 6         | 11.414      | 1582          | 1590        | 1600         | rBV      | 1730586        | 2839489       | 10.03%          | 4.692%        |
| 7         | 11.517      | 1600          | 1607        | 1617         | rVV      | 5632129        | 8348720       | 29.48%          | 13.796%       |
| 8         | 11.627      | 1617          | 1625        | 1643         | rVB      | 16528109       | 28318886      | 100.00%         | 46.797%       |
| 9         | 11.956      | 1669          | 1679        | 1692         | rBV      | 4518490        | 7016662       | 24.78%          | 11.595%       |
| 10        | 12.407      | 1744          | 1753        | 1766         | rBV2     | 1569912        | 2621072       | 9.26%           | 4.331%        |
| 11        | 13.346      | 1891          | 1907        | 1919         | rBV      | 1569534        | 2413879       | 8.52%           | 3.989%        |




Sum of corrected areas: 60514844

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY091025\  
Data File : VY023328.D  
Acq On : 10 Sep 2025 12:49  
Operator : SY/MD  
Sample : Q3058-01  
Misc : 6.55g/5.0mL/MSVOA\_Y/SOIL/A  
ALS Vial : 9 Sample Multiplier: 1

Instrument :  
MSVOA\_Y  
ClientSampleId :  
RPXY5

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y090825S.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY091025\  
Data File : VY023328.D  
Acq On : 10 Sep 2025 12:49  
Operator : SY/MD  
Sample : Q3058-01  
Misc : 6.55g/5.0mL/MSVOA\_Y/SOIL/A  
ALS Vial : 9 Sample Multiplier: 1

Instrument :  
MSVOA\_Y  
ClientSampleId :  
RPXY5

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y090825S.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

\*\*\*\*\*

5

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY091025\  
Data File : VY023328.D  
Acq On : 10 Sep 2025 12:49  
Operator : SY/MD  
Sample : Q3058-01  
Misc : 6.55g/5.0mL/MSVOA\_Y/SOIL/A  
ALS Vial : 9 Sample Multiplier: 1

Instrument :  
MSVOA\_Y  
ClientSampleId :  
RPXY5

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y090825S.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

| TIC Top Hit name | RT | EstConc | Units | Response | --Internal Standard-- |    |      |      |
|------------------|----|---------|-------|----------|-----------------------|----|------|------|
|                  |    |         |       |          | #                     | RT | Resp | Conc |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091525\  
Data File : VN087839.D  
Acq On : 15 Sep 2025 15:43  
Operator : JC\MD  
Sample : Q3058-01ME  
Misc : 6.28g/5mL/100uL/5.00mL/MSVOA\_N/MEOH  
ALS Vial : 22 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
RPXY5ME

Quant Time: Sep 16 01:41:32 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M  
Quant Title : SW846 8260  
QLast Update : Fri Aug 22 02:55:42 2025  
Response via : Initial Calibration


| Compound                    | R.T.   | QIon           | Response | Conc       | Units | Dev(Min) |
|-----------------------------|--------|----------------|----------|------------|-------|----------|
| Internal Standards          |        |                |          |            |       |          |
| 1) Pentafluorobenzene       | 8.200  | 168            | 191442   | 50.000     | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene     | 9.082  | 114            | 434787   | 50.000     | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 11.847 | 117            | 386645   | 50.000     | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.770 | 152            | 170499   | 50.000     | ug/l  | 0.00     |
| System Monitoring Compounds |        |                |          |            |       |          |
| 33) 1,2-Dichloroethane-d4   | 8.559  | 65             | 198573   | 50.625     | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range 74 - 125 | Recovery | = 101.260% |       |          |
| 35) Dibromofluoromethane    | 8.147  | 113            | 158865   | 50.608     | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range 75 - 124 | Recovery | = 101.220% |       |          |
| 50) Toluene-d8              | 10.541 | 98             | 549825   | 46.796     | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range 86 - 113 | Recovery | = 93.600%  |       |          |
| 62) 4-Bromofluorobenzene    | 12.829 | 95             | 166435   | 39.832     | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range 77 - 121 | Recovery | = 79.660%  |       |          |
| Target Compounds            |        |                |          |            |       | Qvalue   |
| 67) Ethyl Benzene           | 11.941 | 91             | 25828    | 1.551      | ug/l  | 96       |
| 68) m/p-Xylenes             | 12.047 | 106            | 28891    | 4.730      | ug/l  | 94       |
| 69) o-Xylene                | 12.370 | 106            | 7729     | 1.291      | ug/l  | 78       |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091525\  
Data File : VN087839.D  
Acq On : 15 Sep 2025 15:43  
Operator : JC\MD  
Sample : Q3058-01ME  
Misc : 6.28g/5mL/100uL/5.00mL/MSVOA\_N/MEOH  
ALS Vial : 22 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
RPXY5ME


Quant Time: Sep 16 01:41:32 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M  
Quant Title : SW846 8260  
QLast Update : Fri Aug 22 02:55:42 2025  
Response via : Initial Calibration








#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 8.200 min Scan# 1063  
Delta R.T. -0.006 min  
Lab File: VN087839.D  
Acq: 15 Sep 2025 15:43

Instrument :  
MSVOA\_N  
ClientSampleId :  
RPXY5ME







Tgt Ion:168 Resp: 191442  
Ion Ratio Lower Upper  
168 100  
99 60.3 57.2 85.8



#33  
1,2-Dichloroethane-d4  
Concen: 50.625 ug/l  
RT: 8.559 min Scan# 1123  
Delta R.T. -0.000 min  
Lab File: VN087839.D  
Acq: 15 Sep 2025 15:43

Tgt Ion: 65 Resp: 198573  
Ion Ratio Lower Upper  
65 100  
67 52.0 0.0 100.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.082 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN087839.D

Acq: 15 Sep 2025 15:43

Instrument :

MSVOA\_N

ClientSampleId :

RPXY5ME





Tgt Ion:114 Resp: 434787

Ion Ratio Lower Upper

114 100

63 23.0 0.0 45.2

88 15.2 0.0 33.4



#35

Dibromofluoromethane

Concen: 50.608 ug/l

RT: 8.147 min Scan# 1053

Delta R.T. -0.006 min

Lab File: VN087839.D

Acq: 15 Sep 2025 15:43




Tgt Ion:113 Resp: 158865


Ion Ratio Lower Upper

113 100

111 105.7 82.8 124.2

192 16.8 12.8 19.2









#50  
Toluene-d8  
Concen: 46.796 ug/l  
RT: 10.541 min Scan# 1460  
Delta R.T. -0.006 min  
Lab File: VN087839.D  
Acq: 15 Sep 2025 15:43




Instrument :  
MSVOA\_N  
ClientSampleId :  
RPXY5ME


Tgt Ion: 98 Resp: 549825  
Ion Ratio Lower Upper  
98 100  
100 63.7 52.8 79.2



#62  
4-Bromofluorobenzene  
Concen: 39.832 ug/l  
RT: 12.829 min Scan# 1849  
Delta R.T. 0.006 min  
Lab File: VN087839.D  
Acq: 15 Sep 2025 15:43

Tgt Ion: 95 Resp: 166435  
Ion Ratio Lower Upper  
95 100  
174 76.1 0.0 138.4  
176 69.0 0.0 132.6









#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.847 min Scan# 1681  
Delta R.T. 0.006 min  
Lab File: VN087839.D  
Acq: 15 Sep 2025 15:43




Instrument :  
MSVOA\_N  
ClientSampleId :  
RPXY5ME


| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 117     | 100   |       |       |
| 82      | 56.5  | 47.4  | 71.2  |
| 119     | 32.2  | 24.3  | 36.5  |



#67  
Ethyl Benzene  
Concen: 1.551 ug/l  
RT: 11.941 min Scan# 1698  
Delta R.T. -0.000 min  
Lab File: VN087839.D  
Acq: 15 Sep 2025 15:43

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 91      | 100   |       |       |
| 106     | 28.2  | 24.1  | 36.1  |





#68

m/p-Xylenes

Concen: 4.730 ug/l

RT: 12.047 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN087839.D

Acq: 15 Sep 2025 15:43

Instrument :

MSVOA\_N

ClientSampleId :





RPXY5ME

Tgt Ion:106 Resp: 28891

Ion Ratio Lower Upper

106 100

91 201.7 169.3 253.9



#69

o-Xylene

Concen: 1.291 ug/l

RT: 12.370 min Scan# 1771

Delta R.T. -0.006 min

Lab File: VN087839.D




Acq: 15 Sep 2025 15:43


Tgt Ion:106 Resp: 7729

Ion Ratio Lower Upper

106 100

91 186.6 111.4 334.2





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.770 min Scan# 20

Delta R.T. 0.006 min

Lab File: VN087839.D


Acq: 15 Sep 2025 15:43

Instrument :

MSVOA\_N

ClientSampleId :

RPXY5ME





Tgt Ion:152 Resp: 170499

Ion Ratio Lower Upper

152 100

115 62.1 32.3 96.8

150 158.1 0.0 351.0



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY091025\  
Data File : VY023329.D  
Acq On : 10 Sep 2025 13:12  
Operator : SY/MD  
Sample : Q3058-02  
Misc : 7.05g/5.0mL/MSVOA\_Y/SOIL/A  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_Y  
ClientSampleId :  
RPXY4

Quant Time: Sep 11 04:23:09 2025  
Quant Title :  
QLast Update : Wed Sep 10 01:44:33 2025  
Response via : Initial Calibration


| Compound                    | R.T.   | QIon  | Response | Conc     | Units      | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|------------|----------|
| Internal Standards          |        |       |          |          |            |          |
| 1) Pentafluorobenzene       | 7.713  | 168   | 539229   | 50.000   | ug/l       | 0.00     |
| 34) 1,4-Difluorobenzene     | 8.616  | 114   | 1018906  | 50.000   | ug/l       | 0.00     |
| 63) Chlorobenzene-d5        | 11.414 | 117   | 1004230  | 50.000   | ug/l       | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.346 | 152   | 442394   | 50.000   | ug/l       | 0.00     |
| System Monitoring Compounds |        |       |          |          |            |          |
| 33) 1,2-Dichloroethane-d4   | 8.067  | 65    | 321894   | 67.935   | ug/l       | 0.01     |
| Spiked Amount               | 50.000 | Range | 50 - 163 | Recovery | = 135.860% |          |
| 35) Dibromofluoromethane    | 7.634  | 113   | 360585   | 57.895   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 54 - 147 | Recovery | = 115.800% |          |
| 50) Toluene-d8              | 10.109 | 98    | 1196744  | 50.886   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 58 - 134 | Recovery | = 101.780% |          |
| 62) 4-Bromofluorobenzene    | 12.408 | 95    | 410835   | 55.510   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 30 - 143 | Recovery | = 111.020% |          |
| Target Compounds            |        |       |          |          |            | Qvalue   |
| 3) Chloromethane            | 2.074  | 50    | 879022   | 141.998  | ug/l       | 100      |
| 5) Bromomethane             | 2.604  | 94    | 104479   | 16.407   | ug/l       | 94       |
| 10) Methyl Iodide           | 4.013  | 142   | 22484    | 3.641    | ug/l       | 99       |
| 16) Acetone                 | 3.866  | 43    | 186911   | 173.178  | ug/l       | 97       |
| 20) Methylene Chloride      | 4.616  | 84    | 30418    | 5.268    | ug/l       | 95       |
| 25) 2-Butanone              | 6.896  | 43    | 7642     | 5.753    | ug/l       | 94       |
| 68) m/p-Xylenes             | 11.621 | 106   | 16972    | 1.179    | ug/l       | 99       |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY091025\  
Data File : VY023329.D  
Acq On : 10 Sep 2025 13:12  
Operator : SY/MD  
Sample : Q3058-02  
Misc : 7.05g/5.0mL/MSVOA\_Y/SOIL/A  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_Y  
ClientSampleId :  
RPXY4

Quant Time: Sep 11 04:23:09 2025  
Quant Title :  
QLast Update : Wed Sep 10 01:44:33 2025  
Response via : Initial Calibration









#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 7.713 min Scan# 983  
Delta R.T. 0.006 min  
Lab File: VY023329.D  
Acq: 10 Sep 2025 13:12




Instrument :  
MSVOA\_Y  
ClientSampleId :  
RPXY4


Tgt Ion:168 Resp: 539229  
Ion Ratio Lower Upper  
168 100  
99 50.9 42.8 64.2



#3  
Chloromethane  
Concen: 141.998 ug/l  
RT: 2.074 min Scan# 58  
Delta R.T. 0.006 min  
Lab File: VY023329.D  
Acq: 10 Sep 2025 13:12

Tgt Ion: 50 Resp: 879022  
Ion Ratio Lower Upper  
50 100  
52 33.4 26.7 40.1









#5  
Bromomethane  
Concen: 16.407 ug/l  
RT: 2.604 min Scan# 144  
Delta R.T. 0.012 min  
Lab File: VY023329.D  
Acq: 10 Sep 2025 13:12




Instrument :  
MSVOA\_Y  
ClientSampleId :  
RPXY4


Tgt Ion: 94 Resp: 104479  
Ion Ratio Lower Upper  
94 100  
96 98.7 74.6 112.0



#10  
Methyl Iodide  
Concen: 3.641 ug/l  
RT: 4.013 min Scan# 376  
Delta R.T. 0.018 min  
Lab File: VY023329.D  
Acq: 10 Sep 2025 13:12

Tgt Ion: 142 Resp: 22484  
Ion Ratio Lower Upper  
142 100  
127 42.9 34.9 52.3  
141 15.2 11.3 16.9









#16  
Acetone  
Concen: 173.178 ug/l  
RT: 3.866 min Scan# 31  
Delta R.T. 0.006 min  
Lab File: VY023329.D  
Acq: 10 Sep 2025 13:12




Instrument :  
MSVOA\_Y  
ClientSampleId :  
RPXY4


Tgt Ion: 43 Resp: 186911  
Ion Ratio Lower Upper  
43 100  
58 32.4 27.1 40.7



#20  
Methylene Chloride  
Concen: 5.268 ug/l  
RT: 4.616 min Scan# 475  
Delta R.T. 0.012 min  
Lab File: VY023329.D  
Acq: 10 Sep 2025 13:12

Tgt Ion: 84 Resp: 30418  
Ion Ratio Lower Upper  
84 100  
49 109.2 93.0 139.6  
51 32.3 29.0 43.4  
86 65.0 50.2 75.4





#25

2-Butanone

Concen: 5.753 ug/l

RT: 6.896 min Scan# 849

Delta R.T. 0.012 min

Lab File: VY023329.D

Acq: 10 Sep 2025 13:12

Instrument :

MSVOA\_Y

ClientSampleId :





RPXY4

Tgt Ion: 43 Resp: 7642

Ion Ratio Lower Upper

43 100

72 32.6 23.5 35.3



#33

1,2-Dichloroethane-d4

Concen: 67.935 ug/l

RT: 8.067 min Scan# 1041

Delta R.T. 0.012 min

Lab File: VY023329.D




Acq: 10 Sep 2025 13:12


Tgt Ion: 65 Resp: 321894

Ion Ratio Lower Upper

65 100

67 52.7 0.0 106.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.616 min Scan# 1131

Delta R.T. 0.006 min

Lab File: VY023329.D

Acq: 10 Sep 2025 13:12

Instrument :

MSVOA\_Y

ClientSampleId :

RPXY4





Tgt Ion:114 Resp: 1018906

Ion Ratio Lower Upper

114 100

63 16.8 0.0 38.4

88 14.9 0.0 30.0



#35

Dibromofluoromethane

Concen: 57.895 ug/l

RT: 7.634 min Scan# 970

Delta R.T. 0.006 min

Lab File: VY023329.D

Acq: 10 Sep 2025 13:12




Tgt Ion:113 Resp: 360585


Ion Ratio Lower Upper

113 100

111 101.7 81.1 121.7

192 20.9 16.2 24.4





#50

Toluene-d8

Concen: 50.886 ug/l

RT: 10.109 min Scan# 1376

Delta R.T. 0.006 min

Lab File: VY023329.D

Acq: 10 Sep 2025 13:12

Instrument :

MSVOA\_Y

ClientSampleId :





RPXY4

Tgt Ion: 98 Resp: 1196744

Ion Ratio Lower Upper

98 100

100 65.2 52.3 78.5



#62

4-Bromofluorobenzene

Concen: 55.510 ug/l

RT: 12.408 min Scan# 1753

Delta R.T. 0.006 min

Lab File: VY023329.D

Acq: 10 Sep 2025 13:12




Tgt Ion: 95 Resp: 410835


Ion Ratio Lower Upper

95 100

174 91.2 0.0 171.6

176 88.8 0.0 167.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.414 min Scan# 1591

Delta R.T. 0.006 min

Lab File: VY023329.D

Acq: 10 Sep 2025 13:12

Instrument :

MSVOA\_Y

ClientSampleId :

RPXY4





Tgt Ion:117 Resp: 1004230

Ion Ratio Lower Upper

117 100

82 52.4 43.4 65.0

119 31.9 25.6 38.4



#68

m/p-Xylenes

Concen: 1.179 ug/l

RT: 11.621 min Scan# 1624

Delta R.T. 0.000 min

Lab File: VY023329.D




Acq: 10 Sep 2025 13:12


Tgt Ion:106 Resp: 16972

Ion Ratio Lower Upper

106 100

91 196.8 158.6 238.0





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.346 min Scan# 1907

Delta R.T. 0.006 min

Lab File: VY023329.D


Acq: 10 Sep 2025 13:12

Instrument :

MSVOA\_Y

ClientSampleId :

RPXY4





Tgt Ion:152 Resp: 442394

Ion Ratio Lower Upper

152 100

115 56.1 28.2 84.6

150 155.5 0.0 348.0



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY091025\  
Data File : VY023329.D  
Acq On : 10 Sep 2025 13:12  
Operator : SY/MD  
Sample : Q3058-02  
Misc : 7.05g/5.0mL/MSVOA\_Y/SOIL/A  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_Y  
ClientSampleId :  
RPXY4

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Sampling : 1

Start Thrs: 0.2

Stop Thrs : 0

Filtering: 5

Min Area: 3 % of largest Peak

Max Peaks: 100

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y0908255.M  
Title : SW846 8260

Signal : TIC: VY023329.D\data.ms

| peak<br># | R.T.<br>min | first<br>scan | max<br>scan | last<br>scan | PK<br>TY | peak<br>height | corr.<br>area | corr.<br>% max. | % of<br>total |
|-----------|-------------|---------------|-------------|--------------|----------|----------------|---------------|-----------------|---------------|
| 1         | 1.855       | 15            | 22          | 25           | rBV      | 22900          | 32057         | 1.04%           | 0.160%        |
| 2         | 2.074       | 52            | 58          | 72           | rBV      | 1018169        | 1457672       | 47.28%          | 7.261%        |
| 3         | 2.336       | 96            | 101         | 106          | rBV      | 23304          | 40903         | 1.33%           | 0.204%        |
| 4         | 2.495       | 118           | 127         | 131          | rBV      | 10317          | 34255         | 1.11%           | 0.171%        |
| 5         | 2.604       | 139           | 145         | 159          | rVB      | 148187         | 299316        | 9.71%           | 1.491%        |
| 6         | 3.872       | 344           | 353         | 365          | rBV2     | 106828         | 295867        | 9.60%           | 1.474%        |
| 7         | 4.013       | 368           | 376         | 386          | rVB      | 13565          | 36301         | 1.18%           | 0.181%        |
| 8         | 4.616       | 465           | 475         | 491          | rVB      | 38218          | 126292        | 4.10%           | 0.629%        |
| 9         | 7.634       | 960           | 970         | 977          | rBV      | 489839         | 1204774       | 39.08%          | 6.001%        |
| 10        | 7.707       | 977           | 982         | 999          | rVB      | 655097         | 1512140       | 49.05%          | 7.532%        |
| 11        | 8.067       | 1029          | 1041        | 1052         | rBV      | 405957         | 921636        | 29.89%          | 4.591%        |
| 12        | 8.616       | 1120          | 1131        | 1145         | rBV      | 1151092        | 2251057       | 73.01%          | 11.213%       |
| 13        | 10.109      | 1368          | 1376        | 1392         | rBV      | 1823024        | 3083040       | 100.00%         | 15.357%       |
| 14        | 11.200      | 1553          | 1555        | 1564         | rVB3     | 19763          | 31725         | 1.03%           | 0.158%        |
| 15        | 11.310      | 1564          | 1573        | 1579         | rBV      | 26385          | 52083         | 1.69%           | 0.259%        |
| 16        | 11.414      | 1583          | 1590        | 1600         | rBV      | 1782757        | 2909012       | 94.36%          | 14.490%       |
| 17        | 11.517      | 1602          | 1607        | 1616         | rVB      | 23184          | 44640         | 1.45%           | 0.222%        |
| 18        | 11.621      | 1616          | 1624        | 1630         | rBV2     | 53673          | 111850        | 3.63%           | 0.557%        |
| 19        | 11.859      | 1658          | 1663        | 1667         | rVV2     | 20284          | 37901         | 1.23%           | 0.189%        |
| 20        | 11.938      | 1671          | 1676        | 1686         | rVV3     | 33894          | 93460         | 3.03%           | 0.466%        |
| 21        | 12.048      | 1686          | 1694        | 1700         | rVB2     | 19821          | 51461         | 1.67%           | 0.256%        |
| 22        | 12.408      | 1745          | 1753        | 1764         | rBV2     | 1624829        | 2854425       | 92.58%          | 14.218%       |
| 23        | 13.346      | 1898          | 1907        | 1920         | rVB      | 1700642        | 2547197       | 82.62%          | 12.688%       |
| 24        | 13.883      | 1989          | 1995        | 2001         | rVB2     | 26314          | 46519         | 1.51%           | 0.232%        |




Sum of corrected areas: 20075583

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY091025\  
Data File : VY023329.D  
Acq On : 10 Sep 2025 13:12  
Operator : SY/MD  
Sample : Q3058-02  
Misc : 7.05g/5.0mL/MSVOA\_Y/SOIL/A  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_Y  
ClientSampleId :  
RPXY4

Quant Title :

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY091025\  
Data File : VY023329.D  
Acq On : 10 Sep 2025 13:12  
Operator : SY/MD  
Sample : Q3058-02  
Misc : 7.05g/5.0mL/MSVOA\_Y/SOIL/A  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_Y  
ClientSampleId :  
RPXY4

Quant Title :

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

\*\*\*\*\*

5

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY091025\  
Data File : VY023329.D  
Acq On : 10 Sep 2025 13:12  
Operator : SY/MD  
Sample : Q3058-02  
Misc : 7.05g/5.0mL/MSVOA\_Y/SOIL/A  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_Y  
ClientSampleId :  
RPXY4

Quant Title :

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

| TIC Top Hit name | RT | EstConc | Units | Response | --Internal Standard-- |    |      |      |
|------------------|----|---------|-------|----------|-----------------------|----|------|------|
|                  |    |         |       |          | #                     | RT | Resp | Conc |

5

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091525\  
 Data File : VN087821.D  
 Acq On : 15 Sep 2025 09:14  
 Operator : JC\MD  
 Sample : VN0915MBL01  
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA\_N/MEOH  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0915MBL01

Quant Time: Sep 16 01:28:30 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Aug 22 02:55:42 2025  
 Response via : Initial Calibration

| Compound                   | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|--------|------|----------|--------|-------|----------|
| -----                      |        |      |          |        |       |          |
| Internal Standards         |        |      |          |        |       |          |
| 1) Pentafluorobenzene      | 8.206  | 168  | 198222   | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 9.083  | 114  | 425993   | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 11.847 | 117  | 408687   | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 13.771 | 152  | 193776   | 50.000 | ug/l  | 0.00     |

|                             |        |       |          |          |      |          |
|-----------------------------|--------|-------|----------|----------|------|----------|
| System Monitoring Compounds |        |       |          |          |      |          |
| 33) 1,2-Dichloroethane-d4   | 8.559  | 65    | 202603   | 49.886   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | =    | 99.780%  |
| 35) Dibromofluoromethane    | 8.147  | 113   | 160733   | 52.260   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =    | 104.520% |
| 50) Toluene-d8              | 10.547 | 98    | 556052   | 48.303   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | =    | 96.600%  |
| 62) 4-Bromofluorobenzene    | 12.829 | 95    | 186538   | 45.565   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 77 - 121 | Recovery | =    | 91.120%  |


| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091525\  
Data File : VN087821.D  
Acq On : 15 Sep 2025 09:14  
Operator : JC\MD  
Sample : VN0915MBL01  
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA\_N/MEOH  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0915MBL01

Quant Time: Sep 16 01:28:30 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M  
Quant Title : SW846 8260  
QLast Update : Fri Aug 22 02:55:42 2025  
Response via : Initial Calibration









#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 8.206 min Scan# 1063  
Delta R.T. 0.000 min  
Lab File: VN087821.D  
Acq: 15 Sep 2025 09:14




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0915MBL01


Tgt Ion:168 Resp: 198222  
Ion Ratio Lower Upper  
168 100  
99 62.9 57.2 85.8



#33  
1,2-Dichloroethane-d4  
Concen: 49.886 ug/l  
RT: 8.559 min Scan# 1123  
Delta R.T. 0.000 min  
Lab File: VN087821.D  
Acq: 15 Sep 2025 09:14

Tgt Ion: 65 Resp: 202603  
Ion Ratio Lower Upper  
65 100  
67 50.6 0.0 100.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.083 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN087821.D

Acq: 15 Sep 2025 09:14

Instrument :

MSVOA\_N

ClientSampleId :

VN0915MBL01





Tgt Ion:114 Resp: 425993

Ion Ratio Lower Upper

114 100

63 22.5 0.0 45.2

88 15.1 0.0 33.4



#35

Dibromofluoromethane

Concen: 52.260 ug/l

RT: 8.147 min Scan# 1053

Delta R.T. -0.006 min

Lab File: VN087821.D

Acq: 15 Sep 2025 09:14




Tgt Ion:113 Resp: 160733


Ion Ratio Lower Upper

113 100

111 104.1 82.8 124.2

192 17.5 12.8 19.2









#50  
Toluene-d8  
Concen: 48.303 ug/l  
RT: 10.547 min Scan# 1461  
Delta R.T. 0.000 min  
Lab File: VN087821.D  
Acq: 15 Sep 2025 09:14




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0915MBL01


Tgt Ion: 98 Resp: 556052  
Ion Ratio Lower Upper  
98 100  
100 63.5 52.8 79.2



#62  
4-Bromofluorobenzene  
Concen: 45.565 ug/l  
RT: 12.829 min Scan# 1849  
Delta R.T. 0.006 min  
Lab File: VN087821.D  
Acq: 15 Sep 2025 09:14

Tgt Ion: 95 Resp: 186538  
Ion Ratio Lower Upper  
95 100  
174 78.1 0.0 138.4  
176 70.0 0.0 132.6









#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.847 min Scan# 1  
Delta R.T. 0.006 min  
Lab File: VN087821.D  
Acq: 15 Sep 2025 09:14




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0915MBL01

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 117     | 100   |       |       |
| 82      | 57.6  | 47.4  | 71.2  |
| 119     | 30.7  | 24.3  | 36.5  |



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 13.771 min Scan# 2009  
Delta R.T. 0.006 min  
Lab File: VN087821.D  
Acq: 15 Sep 2025 09:14

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 152     | 100   |       |       |
| 115     | 59.5  | 32.3  | 96.8  |
| 150     | 157.0 | 0.0   | 351.0 |



5

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091525\  
Data File : VN087821.D  
Acq On : 15 Sep 2025 09:14  
Operator : JC\MD  
Sample : VN0915MBL01  
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA\_N/MEOH  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0915MBL01

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M

Title : SW846 8260

Signal : TIC: VN087821.D\data.ms

| peak<br># | R.T.<br>min | first<br>scan | max<br>scan | last<br>scan | PK<br>TY | peak<br>height | corr.<br>area | corr.<br>% max. | % of<br>total |
|-----------|-------------|---------------|-------------|--------------|----------|----------------|---------------|-----------------|---------------|
| 1         | 8.147       | 1043          | 1053        | 1058         | rBV      | 220382         | 544294        | 35.78%          | 6.853%        |
| 2         | 8.206       | 1058          | 1063        | 1076         | rVB      | 288762         | 655472        | 43.09%          | 8.253%        |
| 3         | 8.559       | 1113          | 1123        | 1134         | rBV      | 241365         | 557369        | 36.64%          | 7.018%        |
| 4         | 9.083       | 1202          | 1212        | 1224         | rBV      | 518177         | 1080687       | 71.04%          | 13.607%       |
| 5         | 10.547      | 1451          | 1461        | 1472         | rBV      | 828870         | 1521141       | 100.00%         | 19.153%       |
| 6         | 11.847      | 1673          | 1682        | 1694         | rBV      | 726992         | 1341753       | 88.21%          | 16.894%       |
| 7         | 12.829      | 1840          | 1849        | 1863         | rBV      | 523609         | 970978        | 63.83%          | 12.226%       |
| 8         | 13.771      | 2001          | 2009        | 2026         | rVB      | 700964         | 1270528       | 83.52%          | 15.997%       |


Sum of corrected areas: 7942222

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091525\  
Data File : VN087821.D  
Acq On : 15 Sep 2025 09:14  
Operator : JC\MD  
Sample : VN0915MBL01  
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA\_N/MEOH  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0915MBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091525\  
Data File : VN087821.D  
Acq On : 15 Sep 2025 09:14  
Operator : JC\MD  
Sample : VN0915MBL01  
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA\_N/MEOH  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0915MBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

\*\*\*\*\*

5

A

B

C

D

E

F

G

H

I

J

5

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091525\  
Data File : VN087821.D  
Acq On : 15 Sep 2025 09:14  
Operator : JC\MD  
Sample : VN0915MBL01  
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA\_N/MEOH  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0915MBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

| TIC Top Hit name | RT | EstConc | Units | Response | --Internal Standard-- |    |      |      |
|------------------|----|---------|-------|----------|-----------------------|----|------|------|
|                  |    |         |       |          | #                     | RT | Resp | Conc |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY091025\  
Data File : VY023322.D  
Acq On : 10 Sep 2025 09:39  
Operator : SY/MD  
Sample : VY0910SBL01  
Misc : 5.00g/5.0mL/MSVOA\_Y/SOIL  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
MSVOA\_Y  
ClientSampleId :  
VY0910SBL01

Quant Time: Sep 11 03:46:06 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y090825S.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 10 01:44:33 2025  
Response via : Initial Calibration

| Compound                   | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|--------|------|----------|--------|-------|----------|
| -----                      |        |      |          |        |       |          |
| Internal Standards         |        |      |          |        |       |          |
| 1) Pentafluorobenzene      | 7.713  | 168  | 583431   | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 8.615  | 114  | 1118169  | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 11.420 | 117  | 1034716  | 50.000 | ug/l  | 0.01     |
| 72) 1,4-Dichlorobenzene-d4 | 13.346 | 152  | 415042   | 50.000 | ug/l  | 0.00     |

|                             |        |       |          |          |      |          |
|-----------------------------|--------|-------|----------|----------|------|----------|
| System Monitoring Compounds |        |       |          |          |      |          |
| 33) 1,2-Dichloroethane-d4   | 8.067  | 65    | 313926   | 61.234   | ug/l | 0.01     |
| Spiked Amount               | 50.000 | Range | 50 - 163 | Recovery | =    | 122.460% |
| 35) Dibromofluoromethane    | 7.634  | 113   | 367281   | 53.735   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 54 - 147 | Recovery | =    | 107.480% |
| 50) Toluene-d8              | 10.109 | 98    | 1335188  | 51.733   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 58 - 134 | Recovery | =    | 103.460% |
| 62) 4-Bromofluorobenzene    | 12.407 | 95    | 400984   | 49.369   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 30 - 143 | Recovery | =    | 98.740%  |


| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY091025\  
Data File : VY023322.D  
Acq On : 10 Sep 2025 09:39  
Operator : SY/MD  
Sample : VY0910SBL01  
Misc : 5.00g/5.0mL/MSVOA\_Y/SOIL  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
MSVOA\_Y  
ClientSampleId :  
VY0910SBL01

Quant Time: Sep 11 03:46:06 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y090825S.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 10 01:44:33 2025  
Response via : Initial Calibration









#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 7.713 min Scan# 983  
Delta R.T. 0.006 min  
Lab File: VY023322.D  
Acq: 10 Sep 2025 09:39




Instrument :  
MSVOA\_Y  
ClientSampleId :  
VY0910SBL01


Tgt Ion:168 Resp: 583431  
Ion Ratio Lower Upper  
168 100  
99 49.9 42.8 64.2



#33  
1,2-Dichloroethane-d4  
Concen: 61.234 ug/l  
RT: 8.067 min Scan# 1041  
Delta R.T. 0.012 min  
Lab File: VY023322.D  
Acq: 10 Sep 2025 09:39

Tgt Ion: 65 Resp: 313926  
Ion Ratio Lower Upper  
65 100  
67 53.0 0.0 106.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.615 min Scan# 1131

Delta R.T. 0.006 min

Lab File: VY023322.D

Acq: 10 Sep 2025 09:39

Instrument :

MSVOA\_Y

ClientSampleId :

VY0910SBL01





Tgt Ion:114 Resp: 1118169

Ion Ratio Lower Upper

114 100

63 17.9 0.0 38.4

88 15.0 0.0 30.0



#35

Dibromofluoromethane

Concen: 53.735 ug/l

RT: 7.634 min Scan# 970

Delta R.T. 0.006 min

Lab File: VY023322.D

Acq: 10 Sep 2025 09:39




Tgt Ion:113 Resp: 367281


Ion Ratio Lower Upper

113 100

111 102.2 81.1 121.7

192 20.9 16.2 24.4





#50

Toluene-d8

Concen: 51.733 ug/l

RT: 10.109 min Scan# 1376

Delta R.T. 0.006 min

Lab File: VY023322.D

Acq: 10 Sep 2025 09:39

Instrument :

MSVOA\_Y

ClientSampleId :





VY0910SBL01

Tgt Ion: 98 Resp: 1335188

Ion Ratio Lower Upper

98 100

100 64.9 52.3 78.5



#62

4-Bromofluorobenzene

Concen: 49.369 ug/l

RT: 12.407 min Scan# 1753

Delta R.T. 0.006 min

Lab File: VY023322.D

Acq: 10 Sep 2025 09:39




Tgt Ion: 95 Resp: 400984


Ion Ratio Lower Upper

95 100

174 92.6 0.0 171.6

176 88.8 0.0 167.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.420 min Scan# 11

Delta R.T. 0.012 min

Lab File: VY023322.D

Acq: 10 Sep 2025 09:39

Instrument :

MSVOA\_Y

ClientSampleId :

VY0910SBL01





Tgt Ion:117 Resp: 1034716

Ion Ratio Lower Upper

117 100

82 50.3 43.4 65.0

119 32.0 25.6 38.4



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.346 min Scan# 1907

Delta R.T. 0.006 min

Lab File: VY023322.D

Acq: 10 Sep 2025 09:39




Tgt Ion:152 Resp: 415042

Ion Ratio Lower Upper

152 100

115 54.7 28.2 84.6

150 157.3 0.0 348.0



5

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY091025\  
Data File : VY023322.D  
Acq On : 10 Sep 2025 09:39  
Operator : SY/MD  
Sample : VY0910SBL01  
Misc : 5.00g/5.0mL/MSVOA\_Y/SOIL  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
MSVOA\_Y  
ClientSampleId :  
VY0910SBL01

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y090825S.M

Title : SW846 8260

Signal : TIC: VY023322.D\data.ms

| peak<br># | R.T.<br>min | first<br>scan | max<br>scan | last<br>scan | PK<br>TY | peak<br>height | corr.<br>area | corr.<br>% max. | % of<br>total |
|-----------|-------------|---------------|-------------|--------------|----------|----------------|---------------|-----------------|---------------|
| 1         | 4.616       | 466           | 475         | 486          | rBV4     | 14233          | 45669         | 1.34%           | 0.266%        |
| 2         | 7.634       | 957           | 970         | 977          | rBV      | 503299         | 1231717       | 36.07%          | 7.171%        |
| 3         | 7.713       | 977           | 983         | 997          | rVB      | 718615         | 1626925       | 47.65%          | 9.472%        |
| 4         | 8.067       | 1029          | 1041        | 1054         | rBV      | 404396         | 898640        | 26.32%          | 5.232%        |
| 5         | 8.615       | 1121          | 1131        | 1148         | rBV      | 1255460        | 2460690       | 72.07%          | 14.327%       |
| 6         | 10.109      | 1368          | 1376        | 1385         | rBV      | 2045478        | 3414527       | 100.00%         | 19.880%       |
| 7         | 11.420      | 1583          | 1591        | 1604         | rBV      | 1847973        | 3011233       | 88.19%          | 17.532%       |
| 8         | 12.407      | 1745          | 1753        | 1767         | rBV      | 1358622        | 2092320       | 61.28%          | 12.182%       |
| 9         | 13.346      | 1901          | 1907        | 1917         | rVB      | 1565415        | 2393558       | 70.10%          | 13.936%       |




Sum of corrected areas: 17175279

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY091025\  
Data File : VY023322.D  
Acq On : 10 Sep 2025 09:39  
Operator : SY/MD  
Sample : VY0910SBL01  
Misc : 5.00g/5.0mL/MSVOA\_Y/SOIL  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
MSVOA\_Y  
ClientSampleId :  
VY0910SBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y090825S.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY091025\  
Data File : VY023322.D  
Acq On : 10 Sep 2025 09:39  
Operator : SY/MD  
Sample : VY0910SBL01  
Misc : 5.00g/5.0mL/MSVOA\_Y/SOIL  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
MSVOA\_Y  
ClientSampleId :  
VY0910SBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y090825S.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

\*\*\*\*\*

5

A

B

C

D

E

F

G

H

I

J

5

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY091025\  
Data File : VY023322.D  
Acq On : 10 Sep 2025 09:39  
Operator : SY/MD  
Sample : VY0910SBL01  
Misc : 5.00g/5.0mL/MSVOA\_Y/SOIL  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
MSVOA\_Y  
ClientSampleId :  
VY0910SBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y090825S.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

| TIC Top Hit name | RT | EstConc | Units | Response | --Internal Standard-- |    |      |      |
|------------------|----|---------|-------|----------|-----------------------|----|------|------|
|                  |    |         |       |          | #                     | RT | Resp | Conc |

5

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091525\  
 Data File : VN087837.D  
 Acq On : 15 Sep 2025 15:02  
 Operator : JC\MD  
 Sample : VN0915MBS01  
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA\_N/MEOH  
 ALS Vial : 20 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0915MBS01

### Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 09/16/2025  
 Supervised By :Semsettin Yesilyurt 09/16/2025

Quant Time: Sep 16 01:39:16 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Aug 22 02:55:42 2025  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response   | Conc    | Units  | Dev(Min) |
|------------------------------|----------------|------|------------|---------|--------|----------|
| -----                        |                |      |            |         |        |          |
| Internal Standards           |                |      |            |         |        |          |
| 1) Pentafluorobenzene        | 8.212          | 168  | 204697     | 50.000  | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 9.082          | 114  | 405190     | 50.000  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 11.847         | 117  | 394337     | 50.000  | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 13.770         | 152  | 213436     | 50.000  | ug/l   | 0.00     |
| System Monitoring Compounds  |                |      |            |         |        |          |
| 33) 1,2-Dichloroethane-d4    | 8.565          | 65   | 203313     | 48.477  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 74 - 125 |      | Recovery = | 96.960% |        |          |
| 35) Dibromofluoromethane     | 8.153          | 113  | 144481     | 49.387  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 75 - 124 |      | Recovery = | 98.780% |        |          |
| 50) Toluene-d8               | 10.547         | 98   | 499460     | 45.615  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 86 - 113 |      | Recovery = | 91.220% |        |          |
| 62) 4-Bromofluorobenzene     | 12.829         | 95   | 180743     | 46.416  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 77 - 121 |      | Recovery = | 92.840% |        |          |
| Target Compounds             |                |      |            |         |        |          |
|                              |                |      |            |         | Qvalue |          |
| 2) Dichlorodifluoromethane   | 2.142          | 85   | 53195      | 18.297  | ug/l   | 95       |
| 3) Chloromethane             | 2.389          | 50   | 56288      | 18.840  | ug/l   | 90       |
| 4) Vinyl Chloride            | 2.536          | 62   | 64520      | 18.625  | ug/l   | 95       |
| 5) Bromomethane              | 2.983          | 94   | 28705      | 16.059  | ug/l   | 96       |
| 6) Chloroethane              | 3.142          | 64   | 44719      | 18.367  | ug/l   | 89       |
| 7) Trichlorofluoromethane    | 3.512          | 101  | 99089      | 19.524  | ug/l   | 92       |
| 8) Diethyl Ether             | 3.971          | 74   | 38411      | 17.378  | ug/l   | 97       |
| 9) 1,1,2-Trichlorotrifluo... | 4.371          | 101  | 51645      | 17.983  | ug/l   | 95       |
| 10) Methyl Iodide            | 4.583          | 142  | 19794      | 14.777  | ug/l # | 95       |
| 11) Tert butyl alcohol       | 5.530          | 59   | 62270      | 85.525  | ug/l   | 96       |
| 12) 1,1-Dichloroethene       | 4.341          | 96   | 53213      | 18.031  | ug/l   | 92       |
| 13) Acrolein                 | 4.177          | 56   | 78133      | 87.290  | ug/l   | 98       |
| 14) Allyl chloride           | 5.018          | 41   | 85840      | 16.051  | ug/l   | 98       |
| 15) Acrylonitrile            | 5.712          | 53   | 185996     | 90.579  | ug/l   | 98       |
| 16) Acetone                  | 4.424          | 43   | 179547     | 78.642  | ug/l   | 91       |
| 17) Carbon Disulfide         | 4.706          | 76   | 161646     | 19.896  | ug/l # | 95       |
| 18) Methyl Acetate           | 5.018          | 43   | 94490      | 19.766  | ug/l   | 94       |
| 19) Methyl tert-butyl Ether  | 5.788          | 73   | 198338     | 17.915  | ug/l   | 99       |
| 20) Methylene Chloride       | 5.271          | 84   | 66257      | 19.302  | ug/l   | 88       |
| 21) trans-1,2-Dichloroethene | 5.783          | 96   | 58706      | 18.354  | ug/l   | 86       |
| 22) Diisopropyl ether        | 6.659          | 45   | 206836     | 17.922  | ug/l   | 99       |
| 23) Vinyl Acetate            | 6.588          | 43   | 840944     | 80.090  | ug/l   | 99       |
| 24) 1,1-Dichloroethane       | 6.553          | 63   | 115355     | 18.230  | ug/l   | 97       |
| 25) 2-Butanone               | 7.471          | 43   | 259016     | 86.233  | ug/l   | 95       |
| 26) 2,2-Dichloropropane      | 7.477          | 77   | 86226      | 15.877  | ug/l   | 100      |
| 27) cis-1,2-Dichloroethene   | 7.477          | 96   | 70593      | 18.462  | ug/l   | 98       |
| 28) Bromochloromethane       | 7.800          | 49   | 66347      | 21.688  | ug/l   | 96       |
| 29) Tetrahydrofuran          | 7.824          | 42   | 165250     | 85.487  | ug/l   | 95       |
| 30) Chloroform               | 7.947          | 83   | 124164     | 19.224  | ug/l   | 99       |
| 31) Cyclohexane              | 8.241          | 56   | 88094      | 16.699  | ug/l   | 94       |
| 32) 1,1,1-Trichloroethane    | 8.153          | 97   | 101242     | 18.496  | ug/l   | 93       |
| 36) 1,1-Dichloropropene      | 8.353          | 75   | 77690      | 17.311  | ug/l   | 97       |
| 37) Ethyl Acetate            | 7.547          | 43   | 108388     | 17.674  | ug/l   | 97       |
| 38) Carbon Tetrachloride     | 8.347          | 117  | 86455      | 19.511  | ug/l   | 98       |
| 39) Methylcyclohexane        | 9.582          | 83   | 80456      | 16.283  | ug/l   | 98       |
| 40) Benzene                  | 8.594          | 78   | 256248     | 18.615  | ug/l   | 94       |

5

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091525\  
 Data File : VN087837.D  
 Acq On : 15 Sep 2025 15:02  
 Operator : JC\MD  
 Sample : VN0915MBS01  
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA\_N/MEOH  
 ALS Vial : 20 Sample Multiplier: 1

## Instrument :

MSVOA\_N

## ClientSampleId :

VN0915MBS01

## Manual Integrations

## APPROVED

Reviewed By :Mahesh Dadoda 09/16/2025  
 Supervised By :Semsettin Yesilyurt 09/16/2025

Quant Time: Sep 16 01:39:16 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M

Quant Title : SW846 8260

QLast Update : Fri Aug 22 02:55:42 2025

Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 41) Methacrylonitrile         | 7.765  | 41   | 56387    | 17.750  | ug/l   | 97       |
| 42) 1,2-Dichloroethane        | 8.653  | 62   | 95239    | 18.004  | ug/l   | 98       |
| 43) Isopropyl Acetate         | 8.671  | 43   | 169478   | 17.700  | ug/l   | 98       |
| 44) Trichloroethene           | 9.329  | 130  | 59459    | 18.919  | ug/l   | 84       |
| 45) 1,2-Dichloropropane       | 9.600  | 63   | 68179    | 18.790  | ug/l   | 98       |
| 46) Dibromomethane            | 9.688  | 93   | 49259    | 19.077  | ug/l   | 95       |
| 47) Bromodichloromethane      | 9.871  | 83   | 101576   | 19.195  | ug/l   | 99       |
| 48) Methyl methacrylate       | 9.665  | 41   | 73680    | 16.269  | ug/l   | 96       |
| 49) 1,4-Dioxane               | 9.676  | 88   | 22634    | 385.576 | ug/l # | 92       |
| 51) 4-Methyl-2-Pentanone      | 10.423 | 43   | 536906   | 92.940  | ug/l   | 99       |
| 52) Toluene                   | 10.612 | 92   | 161853   | 18.966  | ug/l   | 98       |
| 53) t-1,3-Dichloropropene     | 10.818 | 75   | 99488    | 18.163  | ug/l   | 95       |
| 54) cis-1,3-Dichloropropene   | 10.294 | 75   | 103336   | 18.169  | ug/l   | 99       |
| 55) 1,1,2-Trichloroethane     | 10.994 | 97   | 62857    | 19.040  | ug/l   | 94       |
| 56) Ethyl methacrylate        | 10.859 | 69   | 94548    | 17.921  | ug/l   | 99       |
| 57) 1,3-Dichloropropane       | 11.141 | 76   | 107237   | 18.355  | ug/l   | 99       |
| 58) 2-Chloroethyl Vinyl ether | 10.141 | 63   | 256924   | 89.974  | ug/l   | 98       |
| 59) 2-Hexanone                | 11.182 | 43   | 342435   | 93.792  | ug/l   | 100      |
| 60) Dibromochloromethane      | 11.335 | 129  | 71871    | 18.789  | ug/l   | 99       |
| 61) 1,2-Dibromoethane         | 11.447 | 107  | 66113    | 18.993  | ug/l   | 100      |
| 64) Tetrachloroethene         | 11.088 | 164  | 48378    | 17.683  | ug/l   | 88       |
| 65) Chlorobenzene             | 11.870 | 112  | 171204   | 17.451  | ug/l   | 94       |
| 66) 1,1,1,2-Tetrachloroethane | 11.941 | 131  | 65131    | 19.998  | ug/l   | 97       |
| 67) Ethyl Benzene             | 11.941 | 91   | 289282   | 17.030  | ug/l   | 98       |
| 68) m/p-Xylenes               | 12.047 | 106  | 217591   | 34.929  | ug/l   | 100      |
| 69) o-Xylene                  | 12.376 | 106  | 104520   | 17.121  | ug/l   | 97       |
| 70) Styrene                   | 12.394 | 104  | 184223   | 18.012  | ug/l   | 97       |
| 71) Bromoform                 | 12.559 | 173  | 46212    | 18.501  | ug/l # | 96       |
| 73) Isopropylbenzene          | 12.676 | 105  | 268813   | 15.586  | ug/l   | 100      |
| 74) N-amyl acetate            | 12.535 | 43   | 90375m   | 13.827  | ug/l   |          |
| 75) 1,1,2,2-Tetrachloroethane | 12.917 | 83   | 103264   | 17.391  | ug/l   | 98       |
| 76) 1,2,3-Trichloropropane    | 12.970 | 75   | 97106m   | 19.401  | ug/l   |          |
| 77) Bromobenzene              | 12.959 | 156  | 69647    | 17.127  | ug/l   | 94       |
| 78) n-propylbenzene           | 13.011 | 91   | 338010   | 15.911  | ug/l   | 99       |
| 79) 2-Chlorotoluene           | 13.106 | 91   | 203223   | 15.896  | ug/l   | 99       |
| 80) 1,3,5-Trimethylbenzene    | 13.153 | 105  | 237481   | 16.231  | ug/l   | 99       |
| 81) trans-1,4-Dichloro-2-b... | 12.711 | 75   | 27758    | 14.682  | ug/l # | 82       |
| 82) 4-Chlorotoluene           | 13.200 | 91   | 216369   | 16.097  | ug/l   | 98       |
| 83) tert-Butylbenzene         | 13.417 | 119  | 195076   | 16.000  | ug/l   | 97       |
| 84) 1,2,4-Trimethylbenzene    | 13.459 | 105  | 237351   | 16.205  | ug/l   | 99       |
| 85) sec-Butylbenzene          | 13.594 | 105  | 283340   | 15.660  | ug/l   | 98       |
| 86) p-Isopropyltoluene        | 13.706 | 119  | 240015   | 16.065  | ug/l   | 99       |
| 87) 1,3-Dichlorobenzene       | 13.711 | 146  | 135745   | 16.951  | ug/l   | 97       |
| 88) 1,4-Dichlorobenzene       | 13.788 | 146  | 140030   | 16.803  | ug/l   | 96       |
| 89) n-Butylbenzene            | 14.035 | 91   | 229284   | 15.452  | ug/l   | 99       |
| 90) Hexachloroethane          | 14.306 | 117  | 48471    | 17.029  | ug/l   | 96       |
| 91) 1,2-Dichlorobenzene       | 14.082 | 146  | 129184   | 17.004  | ug/l   | 98       |
| 92) 1,2-Dibromo-3-Chloropr... | 14.694 | 75   | 22560    | 15.994  | ug/l   | 96       |
| 93) 1,2,4-Trichlorobenzene    | 15.364 | 180  | 74092    | 16.240  | ug/l   | 99       |
| 94) Hexachlorobutadiene       | 15.476 | 225  | 29433    | 18.096  | ug/l   | 98       |
| 95) Naphthalene               | 15.611 | 128  | 246112   | 15.230  | ug/l   | 99       |
| 96) 1,2,3-Trichlorobenzene    | 15.811 | 180  | 72462    | 16.247  | ug/l   | 100      |

5

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091525\  
Data File : VN087837.D  
Acq On : 15 Sep 2025 15:02  
Operator : JC\MD  
Sample : VN0915MBS01  
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA\_N/MEOH  
ALS Vial : 20 Sample Multiplier: 1

**Instrument :**

MSVOA\_N

**ClientSampleId :**

VN0915MBS01

**Manual Integrations****APPROVED**

Reviewed By :Mahesh Dadoda 09/16/2025

Supervised By :Semsettin Yesilyurt 09/16/2025

Quant Time: Sep 16 01:39:16 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M  
Quant Title : SW846 8260  
QLast Update : Fri Aug 22 02:55:42 2025  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

-----  
(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091525\  
Data File : VN087837.D  
Acq On    : 15 Sep 2025  15:02  
Operator  : JC\MD  
Sample    : VN0915MBS01  
Misc      : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH  
ALS Vial  : 20   Sample Multiplier: 1
```

**Instrument :**  
MSVOA\_N  
**ClientSampleId :**  
VN0915MBS01

## Manual Integrations

Reviewed By :Mahesh Dadoda 09/16/2025  
Supervised By :Semsettin Yesilyurt 09/16/2025

Quant Time: Sep 16 01:39:16 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M  
Quant Title : SW846 8260  
QLast Update : Fri Aug 22 02:55:42 2025  
Response via : Initial Calibration



5

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY091025\  
 Data File : VY023323.D  
 Acq On : 10 Sep 2025 10:11  
 Operator : SY/MD  
 Sample : VY0910SBS01  
 Misc : 5.00g/5.0mL/MSVOA\_Y/SOIL  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 VY0910SBS01

### Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 09/11/2025  
 Supervised By :Semsettin Yesilyurt 09/11/2025

Quant Time: Sep 11 03:46:27 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y090825S.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 10 01:44:33 2025  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response   | Conc     | Units  | Dev(Min) |
|------------------------------|----------------|------|------------|----------|--------|----------|
| Internal Standards           |                |      |            |          |        |          |
| 1) Pentafluorobenzene        | 7.713          | 168  | 502093     | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 8.616          | 114  | 763457     | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 11.420         | 117  | 676997     | 50.000   | ug/l   | 0.01     |
| 72) 1,4-Dichlorobenzene-d4   | 13.346         | 152  | 350574     | 50.000   | ug/l   | 0.00     |
| System Monitoring Compounds  |                |      |            |          |        |          |
| 33) 1,2-Dichloroethane-d4    | 8.061          | 65   | 218744     | 49.580   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 50 - 163 |      | Recovery = | 99.160%  |        |          |
| 35) Dibromofluoromethane     | 7.634          | 113  | 230025     | 49.290   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 54 - 147 |      | Recovery = | 98.580%  |        |          |
| 50) Toluene-d8               | 10.109         | 98   | 890309     | 50.523   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 58 - 134 |      | Recovery = | 101.040% |        |          |
| 62) 4-Bromofluorobenzene     | 12.408         | 95   | 280578     | 50.595   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 30 - 143 |      | Recovery = | 101.200% |        |          |
| Target Compounds             |                |      |            |          |        |          |
|                              |                |      |            |          | Qvalue |          |
| 2) Dichlorodifluoromethane   | 1.867          | 85   | 83772      | 20.797   | ug/l   | 94       |
| 3) Chloromethane             | 2.074          | 50   | 124849     | 21.660   | ug/l   | 98       |
| 4) Vinyl Chloride            | 2.208          | 62   | 154360     | 21.257   | ug/l   | 98       |
| 5) Bromomethane              | 2.592          | 94   | 131063     | 22.105   | ug/l   | 98       |
| 6) Chloroethane              | 2.739          | 64   | 104502     | 21.266   | ug/l   | 99       |
| 7) Trichlorofluoromethane    | 3.056          | 101  | 191708     | 20.917   | ug/l   | 98       |
| 8) Diethyl Ether             | 3.458          | 74   | 48823      | 20.133   | ug/l   | 97       |
| 9) 1,1,2-Trichlorotrifluo... | 3.818          | 101  | 104265     | 21.154   | ug/l   | 97       |
| 10) Methyl Iodide            | 4.007          | 142  | 103901     | 18.071   | ug/l   | 99       |
| 11) Tert butyl alcohol       | 4.860          | 59   | 25296      | 99.227   | ug/l   | 100      |
| 12) 1,1-Dichloroethene       | 3.793          | 96   | 95363      | 20.112   | ug/l   | 94       |
| 13) Acrolein                 | 3.653          | 56   | 40337      | 93.593   | ug/l   | 98       |
| 14) Allyl chloride           | 4.385          | 41   | 122336     | 20.424   | ug/l   | 97       |
| 15) Acrylonitrile            | 5.055          | 53   | 93413      | 97.849   | ug/l   | 99       |
| 16) Acetone                  | 3.866          | 43   | 108202     | 107.667  | ug/l   | 99       |
| 17) Carbon Disulfide         | 4.104          | 76   | 307699     | 20.497   | ug/l   | 98       |
| 18) Methyl Acetate           | 4.385          | 43   | 47119      | 19.114   | ug/l   | 95       |
| 19) Methyl tert-butyl Ether  | 5.116          | 73   | 223833     | 20.077   | ug/l   | 99       |
| 20) Methylene Chloride       | 4.616          | 84   | 124829     | 23.217   | ug/l   | 91       |
| 21) trans-1,2-Dichloroethene | 5.116          | 96   | 109436     | 20.657   | ug/l   | 94       |
| 22) Diisopropyl ether        | 6.019          | 45   | 276095     | 20.698   | ug/l   | 95       |
| 23) Vinyl Acetate            | 5.964          | 43   | 734886     | 99.652   | ug/l   | 98       |
| 24) 1,1-Dichloroethane       | 5.915          | 63   | 180883     | 20.732   | ug/l   | 99       |
| 25) 2-Butanone               | 6.896          | 43   | 128844     | 104.167  | ug/l   | 96       |
| 26) 2,2-Dichloropropane      | 6.884          | 77   | 159633     | 20.721   | ug/l   | 98       |
| 27) cis-1,2-Dichloroethene   | 6.890          | 96   | 123443     | 20.576   | ug/l   | 93       |
| 28) Bromochloromethane       | 7.250          | 49   | 68387      | 19.920   | ug/l   | 87       |
| 29) Tetrahydrofuran          | 7.268          | 42   | 71592      | 97.583   | ug/l   | 96       |
| 30) Chloroform               | 7.421          | 83   | 194077     | 20.702   | ug/l   | 98       |
| 31) Cyclohexane              | 7.707          | 56   | 154305     | 19.986   | ug/l   | 93       |
| 32) 1,1,1-Trichloroethane    | 7.622          | 97   | 175185     | 20.916   | ug/l   | 98       |
| 36) 1,1-Dichloropropene      | 7.841          | 75   | 135903     | 20.447   | ug/l   | 98       |
| 37) Ethyl Acetate            | 6.988          | 43   | 50803      | 19.214   | ug/l   | 98       |
| 38) Carbon Tetrachloride     | 7.823          | 117  | 156951     | 20.522   | ug/l   | 96       |
| 39) Methylcyclohexane        | 9.109          | 83   | 171725     | 20.195   | ug/l   | 97       |
| 40) Benzene                  | 8.085          | 78   | 427299     | 20.682   | ug/l   | 100      |

5

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY091025\  
 Data File : VY023323.D  
 Acq On : 10 Sep 2025 10:11  
 Operator : SY/MD  
 Sample : VY0910SBS01  
 Misc : 5.00g/5.0mL/MSVOA\_Y/SOIL  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 VY0910SBS01

### Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 09/11/2025  
 Supervised By :Semsettin Yesilyurt 09/11/2025

Quant Time: Sep 11 03:46:27 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y090825S.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 10 01:44:33 2025  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 41) Methacrylonitrile         | 7.220  | 41   | 27118m   | 19.805  | ug/l   |          |
| 42) 1,2-Dichloroethane        | 8.158  | 62   | 107843   | 20.502  | ug/l   | 99       |
| 43) Isopropyl Acetate         | 8.201  | 43   | 100820   | 19.605  | ug/l   | 99       |
| 44) Trichloroethene           | 8.866  | 130  | 120223   | 20.892  | ug/l   | 94       |
| 45) 1,2-Dichloropropane       | 9.140  | 63   | 95935    | 20.561  | ug/l   | 100      |
| 46) Dibromomethane            | 9.231  | 93   | 58770    | 20.683  | ug/l   | 97       |
| 47) Bromodichloromethane      | 9.426  | 83   | 145234   | 20.316  | ug/l   | 98       |
| 48) Methyl methacrylate       | 9.219  | 41   | 45877    | 19.139  | ug/l   | 92       |
| 49) 1,4-Dioxane               | 9.231  | 88   | 11177    | 387.495 | ug/l   | 95       |
| 51) 4-Methyl-2-Pentanone      | 10.000 | 43   | 258579   | 98.282  | ug/l   | 99       |
| 52) Toluene                   | 10.170 | 92   | 276761   | 21.109  | ug/l   | 100      |
| 53) t-1,3-Dichloropropene     | 10.396 | 75   | 122968   | 19.980  | ug/l   | 97       |
| 54) cis-1,3-Dichloropropene   | 9.859  | 75   | 147845   | 20.103  | ug/l   | 97       |
| 55) 1,1,2-Trichloroethane     | 10.573 | 97   | 78045    | 20.648  | ug/l   | 98       |
| 56) Ethyl methacrylate        | 10.438 | 69   | 83357    | 19.035  | ug/l   | 97       |
| 57) 1,3-Dichloropropane       | 10.719 | 76   | 123908   | 20.193  | ug/l   | 99       |
| 58) 2-Chloroethyl Vinyl ether | 9.713  | 63   | 175044   | 91.713  | ug/l   | 94       |
| 59) 2-Hexanone                | 10.762 | 43   | 186389   | 102.479 | ug/l   | 99       |
| 60) Dibromochloromethane      | 10.914 | 129  | 104904   | 20.455  | ug/l   | 100      |
| 61) 1,2-Dibromoethane         | 11.018 | 107  | 71541    | 20.198  | ug/l   | 99       |
| 64) Tetrachloroethene         | 10.646 | 164  | 145016   | 21.492  | ug/l   | 99       |
| 65) Chlorobenzene             | 11.444 | 112  | 298856   | 20.151  | ug/l   | 97       |
| 66) 1,1,1,2-Tetrachloroethane | 11.518 | 131  | 106850   | 20.525  | ug/l   | 99       |
| 67) Ethyl Benzene             | 11.518 | 91   | 489602   | 20.207  | ug/l   | 98       |
| 68) m/p-Xylenes               | 11.627 | 106  | 400072   | 41.235  | ug/l   | 96       |
| 69) o-Xylene                  | 11.956 | 106  | 180809   | 20.131  | ug/l   | 98       |
| 70) Styrene                   | 11.969 | 104  | 304780   | 20.475  | ug/l   | 98       |
| 71) Bromoform                 | 12.133 | 173  | 62453    | 20.145  | ug/l # | 100      |
| 73) Isopropylbenzene          | 12.255 | 105  | 475777   | 20.147  | ug/l   | 100      |
| 74) N-amyl acetate            | 12.066 | 43   | 85132    | 19.072  | ug/l   | 96       |
| 75) 1,1,2,2-Tetrachloroethane | 12.505 | 83   | 76249    | 19.239  | ug/l   | 98       |
| 76) 1,2,3-Trichloropropane    | 12.554 | 75   | 71554m   | 20.155  | ug/l   |          |
| 77) Bromobenzene              | 12.530 | 156  | 121187   | 20.087  | ug/l   | 94       |
| 78) n-propylbenzene           | 12.597 | 91   | 571399   | 20.530  | ug/l   | 98       |
| 79) 2-Chlorotoluene           | 12.682 | 91   | 323583   | 20.185  | ug/l   | 96       |
| 80) 1,3,5-Trimethylbenzene    | 12.737 | 105  | 387970   | 20.200  | ug/l   | 97       |
| 81) trans-1,4-Dichloro-2-b... | 12.304 | 75   | 23944    | 18.393  | ug/l   | 97       |
| 82) 4-Chlorotoluene           | 12.780 | 91   | 339877   | 20.262  | ug/l   | 98       |
| 83) tert-Butylbenzene         | 12.999 | 119  | 345942   | 19.887  | ug/l   | 98       |
| 84) 1,2,4-Trimethylbenzene    | 13.042 | 105  | 390370   | 20.545  | ug/l   | 99       |
| 85) sec-Butylbenzene          | 13.176 | 105  | 511538   | 20.211  | ug/l   | 99       |
| 86) p-Isopropyltoluene        | 13.292 | 119  | 431225   | 20.170  | ug/l   | 99       |
| 87) 1,3-Dichlorobenzene       | 13.286 | 146  | 238915   | 20.012  | ug/l   | 98       |
| 88) 1,4-Dichlorobenzene       | 13.365 | 146  | 239570   | 20.260  | ug/l   | 99       |
| 89) n-Butylbenzene            | 13.615 | 91   | 377453   | 19.908  | ug/l   | 99       |
| 90) Hexachloroethane          | 13.877 | 117  | 94023    | 20.286  | ug/l   | 100      |
| 91) 1,2-Dichlorobenzene       | 13.657 | 146  | 210268   | 20.200  | ug/l   | 98       |
| 92) 1,2-Dibromo-3-Chloropr... | 14.273 | 75   | 11991    | 19.282  | ug/l   | 93       |
| 93) 1,2,4-Trichlorobenzene    | 14.919 | 180  | 120603   | 19.863  | ug/l   | 99       |
| 94) Hexachlorobutadiene       | 15.023 | 225  | 77088    | 20.534  | ug/l   | 99       |
| 95) Naphthalene               | 15.145 | 128  | 175490   | 17.982  | ug/l   | 100      |
| 96) 1,2,3-Trichlorobenzene    | 15.328 | 180  | 105368   | 19.982  | ug/l   | 98       |

5

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY091025\  
Data File : VY023323.D  
Acq On : 10 Sep 2025 10:11  
Operator : SY/MD  
Sample : VY0910SBS01  
Misc : 5.00g/5.0mL/MSVOA\_Y/SOIL  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_Y  
ClientSampleId :  
VY0910SBS01

Manual Integrations  
APPROVED

Reviewed By :Mahesh Dadoda 09/11/2025  
Supervised By :Semsettin Yesilyurt 09/11/2025

Quant Time: Sep 11 03:46:27 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y090825S.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 10 01:44:33 2025  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

-----  
(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY091025\  
 Data File : VY023323.D  
 Acq On : 10 Sep 2025 10:11  
 Operator : SY/MD  
 Sample : VY0910SBS01  
 Misc : 5.00g/5.0mL/MSVOA\_Y/SOIL  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 VY0910SBS01

### Manual Integrations APPROVED

Reviewed By : Mahesh Dadoda 09/11/2025  
 Supervised By : Semsettin Yesilyurt 09/11/2025

Quant Time: Sep 11 03:46:27 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y090825S.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 10 01:44:33 2025  
 Response via : Initial Calibration



5

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY091025\  
 Data File : VY023324.D  
 Acq On : 10 Sep 2025 10:33  
 Operator : SY/MD  
 Sample : VY0910SBS01  
 Misc : 5.00g/5.0mL/MSVOA\_Y/SOIL  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 VY0910SBS01

### Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 09/11/2025  
 Supervised By :Semsettin Yesilyurt 09/11/2025

Quant Time: Sep 11 03:47:17 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y090825S.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 10 01:44:33 2025  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response   | Conc    | Units | Dev(Min) |
|------------------------------|----------------|------|------------|---------|-------|----------|
| Internal Standards           |                |      |            |         |       |          |
| 1) Pentafluorobenzene        | 7.713          | 168  | 502400     | 50.000  | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene      | 8.616          | 114  | 765892     | 50.000  | ug/l  | 0.00     |
| 63) Chlorobenzene-d5         | 11.414         | 117  | 669720     | 50.000  | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 13.346         | 152  | 341290     | 50.000  | ug/l  | 0.00     |
| System Monitoring Compounds  |                |      |            |         |       |          |
| 33) 1,2-Dichloroethane-d4    | 8.067          | 65   | 213138     | 48.280  | ug/l  | 0.01     |
| Spiked Amount 50.000         | Range 50 - 163 |      | Recovery = | 96.560% |       |          |
| 35) Dibromofluoromethane     | 7.640          | 113  | 227052     | 48.498  | ug/l  | 0.01     |
| Spiked Amount 50.000         | Range 54 - 147 |      | Recovery = | 97.000% |       |          |
| 50) Toluene-d8               | 10.109         | 98   | 867003     | 49.044  | ug/l  | 0.00     |
| Spiked Amount 50.000         | Range 58 - 134 |      | Recovery = | 98.080% |       |          |
| 62) 4-Bromofluorobenzene     | 12.402         | 95   | 270726     | 48.663  | ug/l  | 0.00     |
| Spiked Amount 50.000         | Range 30 - 143 |      | Recovery = | 97.320% |       |          |
| Target Compounds             |                |      |            |         |       |          |
|                              |                |      |            |         |       | Qvalue   |
| 2) Dichlorodifluoromethane   | 1.867          | 85   | 84379      | 20.935  | ug/l  | 99       |
| 3) Chloromethane             | 2.074          | 50   | 121073     | 20.992  | ug/l  | 96       |
| 4) Vinyl Chloride            | 2.208          | 62   | 151147     | 20.802  | ug/l  | 100      |
| 5) Bromomethane              | 2.592          | 94   | 129401     | 21.811  | ug/l  | 99       |
| 6) Chloroethane              | 2.732          | 64   | 101661     | 20.675  | ug/l  | 100      |
| 7) Trichlorofluoromethane    | 3.062          | 101  | 188187     | 20.520  | ug/l  | 100      |
| 8) Diethyl Ether             | 3.458          | 74   | 49235      | 20.290  | ug/l  | 98       |
| 9) 1,1,2-Trichlorotrifluo... | 3.824          | 101  | 99410      | 20.156  | ug/l  | 98       |
| 10) Methyl Iodide            | 4.007          | 142  | 107253     | 18.642  | ug/l  | 99       |
| 11) Tert butyl alcohol       | 4.860          | 59   | 27385      | 107.356 | ug/l  | 99       |
| 12) 1,1-Dichloroethene       | 3.787          | 96   | 92477      | 19.492  | ug/l  | 92       |
| 13) Acrolein                 | 3.659          | 56   | 41537      | 96.319  | ug/l  | 96       |
| 14) Allyl chloride           | 4.385          | 41   | 118610     | 19.790  | ug/l  | 97       |
| 15) Acrylonitrile            | 5.061          | 53   | 93964      | 98.366  | ug/l  | 99       |
| 16) Acetone                  | 3.873          | 43   | 107417     | 106.821 | ug/l  | 99       |
| 17) Carbon Disulfide         | 4.110          | 76   | 296875     | 19.764  | ug/l  | 99       |
| 18) Methyl Acetate           | 4.385          | 43   | 50383      | 20.425  | ug/l  | 99       |
| 19) Methyl tert-butyl Ether  | 5.122          | 73   | 221043     | 19.814  | ug/l  | 95       |
| 20) Methylene Chloride       | 4.622          | 84   | 120067     | 22.317  | ug/l  | 93       |
| 21) trans-1,2-Dichloroethene | 5.116          | 96   | 106494     | 20.089  | ug/l  | 98       |
| 22) Diisopropyl ether        | 6.025          | 45   | 269535     | 20.194  | ug/l  | 94       |
| 23) Vinyl Acetate            | 5.964          | 43   | 728549     | 98.733  | ug/l  | 98       |
| 24) 1,1-Dichloroethane       | 5.921          | 63   | 175181     | 20.066  | ug/l  | 98       |
| 25) 2-Butanone               | 6.896          | 43   | 128969     | 104.205 | ug/l  | 98       |
| 26) 2,2-Dichloropropane      | 6.890          | 77   | 156397     | 20.289  | ug/l  | 100      |
| 27) cis-1,2-Dichloroethene   | 6.890          | 96   | 120207     | 20.024  | ug/l  | 94       |
| 28) Bromochloromethane       | 7.250          | 49   | 67872      | 19.758  | ug/l  | 88       |
| 29) Tetrahydrofuran          | 7.262          | 42   | 72731      | 99.075  | ug/l  | 96       |
| 30) Chloroform               | 7.421          | 83   | 188351     | 20.079  | ug/l  | 97       |
| 31) Cyclohexane              | 7.701          | 56   | 152437     | 19.732  | ug/l  | 97       |
| 32) 1,1,1-Trichloroethane    | 7.616          | 97   | 167942     | 20.039  | ug/l  | 98       |
| 36) 1,1-Dichloropropene      | 7.841          | 75   | 134914     | 20.233  | ug/l  | 99       |
| 37) Ethyl Acetate            | 6.988          | 43   | 50254      | 18.946  | ug/l  | 95       |
| 38) Carbon Tetrachloride     | 7.823          | 117  | 151104     | 19.694  | ug/l  | 94       |
| 39) Methylcyclohexane        | 9.109          | 83   | 167195     | 19.600  | ug/l  | 99       |
| 40) Benzene                  | 8.085          | 78   | 420339     | 20.281  | ug/l  | 98       |

5

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY091025\  
 Data File : VY023324.D  
 Acq On : 10 Sep 2025 10:33  
 Operator : SY/MD  
 Sample : VY0910SBS01  
 Misc : 5.00g/5.0mL/MSVOA\_Y/SOIL  
 ALS Vial : 5 Sample Multiplier: 1

## Instrument :

MSVOA\_Y

## ClientSampleId :

VY0910SBS01

## Manual Integrations

## APPROVED

Reviewed By :Mahesh Dadoda 09/11/2025

Supervised By :Semsettin Yesilyurt 09/11/2025

Quant Time: Sep 11 03:47:17 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y090825S.M

Quant Title : SW846 8260

QLast Update : Wed Sep 10 01:44:33 2025

Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 41) Methacrylonitrile         | 7.226  | 41   | 26112m   | 19.010  | ug/l   |          |
| 42) 1,2-Dichloroethane        | 8.158  | 62   | 107219   | 20.319  | ug/l   | 99       |
| 43) Isopropyl Acetate         | 8.201  | 43   | 99268    | 19.242  | ug/l   | 95       |
| 44) Trichloroethene           | 8.866  | 130  | 115823   | 20.063  | ug/l   | 99       |
| 45) 1,2-Dichloropropane       | 9.146  | 63   | 93560    | 19.988  | ug/l   | 97       |
| 46) Dibromomethane            | 9.231  | 93   | 57045    | 20.012  | ug/l   | 97       |
| 47) Bromodichloromethane      | 9.426  | 83   | 144019   | 20.082  | ug/l   | 96       |
| 48) Methyl methacrylate       | 9.225  | 41   | 46038    | 19.145  | ug/l   | 92       |
| 49) 1,4-Dioxane               | 9.231  | 88   | 12067    | 417.020 | ug/l   | 93       |
| 51) 4-Methyl-2-Pentanone      | 9.999  | 43   | 261316   | 99.007  | ug/l   | 99       |
| 52) Toluene                   | 10.170 | 92   | 264688   | 20.124  | ug/l   | 99       |
| 53) t-1,3-Dichloropropene     | 10.396 | 75   | 121598   | 19.694  | ug/l   | 95       |
| 54) cis-1,3-Dichloropropene   | 9.859  | 75   | 145341   | 19.700  | ug/l   | 98       |
| 55) 1,1,2-Trichloroethane     | 10.573 | 97   | 76175    | 20.089  | ug/l   | 99       |
| 56) Ethyl methacrylate        | 10.438 | 69   | 83419    | 18.989  | ug/l   | 96       |
| 57) 1,3-Dichloropropane       | 10.719 | 76   | 122160   | 19.844  | ug/l   | 99       |
| 58) 2-Chloroethyl Vinyl ether | 9.713  | 63   | 179310   | 93.649  | ug/l   | 97       |
| 59) 2-Hexanone                | 10.762 | 43   | 186046   | 101.965 | ug/l   | 100      |
| 60) Dibromochloromethane      | 10.914 | 129  | 103469   | 20.111  | ug/l   | 100      |
| 61) 1,2-Dibromoethane         | 11.018 | 107  | 71228    | 20.046  | ug/l   | 99       |
| 64) Tetrachloroethene         | 10.646 | 164  | 141550   | 21.206  | ug/l   | 98       |
| 65) Chlorobenzene             | 11.444 | 112  | 296013   | 20.176  | ug/l   | 95       |
| 66) 1,1,1,2-Tetrachloroethane | 11.518 | 131  | 103824   | 20.161  | ug/l   | 100      |
| 67) Ethyl Benzene             | 11.518 | 91   | 477904   | 19.938  | ug/l   | 96       |
| 68) m/p-Xylenes               | 11.627 | 106  | 388769   | 40.505  | ug/l   | 96       |
| 69) o-Xylene                  | 11.956 | 106  | 177226   | 19.947  | ug/l   | 99       |
| 70) Styrene                   | 11.969 | 104  | 298518   | 20.272  | ug/l   | 98       |
| 71) Bromoform                 | 12.133 | 173  | 62431    | 20.357  | ug/l # | 99       |
| 73) Isopropylbenzene          | 12.255 | 105  | 455493   | 19.813  | ug/l   | 98       |
| 74) N-amyl acetate            | 12.072 | 43   | 86950    | 20.009  | ug/l   | 97       |
| 75) 1,1,2,2-Tetrachloroethane | 12.505 | 83   | 74280    | 19.252  | ug/l   | 96       |
| 76) 1,2,3-Trichloropropane    | 12.554 | 75   | 68041m   | 19.687  | ug/l   |          |
| 77) Bromobenzene              | 12.530 | 156  | 118368   | 20.153  | ug/l   | 94       |
| 78) n-propylbenzene           | 12.597 | 91   | 553252   | 20.418  | ug/l   | 98       |
| 79) 2-Chlorotoluene           | 12.676 | 91   | 317780   | 20.362  | ug/l   | 97       |
| 80) 1,3,5-Trimethylbenzene    | 12.737 | 105  | 377848   | 20.208  | ug/l   | 99       |
| 81) trans-1,4-Dichloro-2-b... | 12.304 | 75   | 24546    | 19.368  | ug/l   | 98       |
| 82) 4-Chlorotoluene           | 12.773 | 91   | 329610   | 20.185  | ug/l   | 97       |
| 83) tert-Butylbenzene         | 12.993 | 119  | 337520   | 19.931  | ug/l   | 98       |
| 84) 1,2,4-Trimethylbenzene    | 13.042 | 105  | 379156   | 20.498  | ug/l   | 98       |
| 85) sec-Butylbenzene          | 13.176 | 105  | 503224   | 20.424  | ug/l   | 99       |
| 86) p-Isopropyltoluene        | 13.292 | 119  | 418859   | 20.125  | ug/l   | 99       |
| 87) 1,3-Dichlorobenzene       | 13.285 | 146  | 235675   | 20.278  | ug/l   | 99       |
| 88) 1,4-Dichlorobenzene       | 13.365 | 146  | 229951   | 19.976  | ug/l   | 98       |
| 89) n-Butylbenzene            | 13.615 | 91   | 368655   | 19.972  | ug/l   | 98       |
| 90) Hexachloroethane          | 13.877 | 117  | 89244    | 19.779  | ug/l   | 94       |
| 91) 1,2-Dichlorobenzene       | 13.657 | 146  | 204225   | 20.154  | ug/l   | 98       |
| 92) 1,2-Dibromo-3-Chloropr... | 14.273 | 75   | 12084    | 19.960  | ug/l   | 91       |
| 93) 1,2,4-Trichlorobenzene    | 14.919 | 180  | 119895   | 20.284  | ug/l   | 97       |
| 94) Hexachlorobutadiene       | 15.023 | 225  | 73735    | 20.176  | ug/l   | 98       |
| 95) Naphthalene               | 15.139 | 128  | 184113   | 19.379  | ug/l   | 99       |
| 96) 1,2,3-Trichlorobenzene    | 15.328 | 180  | 103579   | 20.177  | ug/l   | 99       |

5

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY091025\  
Data File : VY023324.D  
Acq On : 10 Sep 2025 10:33  
Operator : SY/MD  
Sample : VY0910SBSD01  
Misc : 5.00g/5.0mL/MSVOA\_Y/SOIL  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
MSVOA\_Y  
ClientSampleId :  
VY0910SBSD01

Manual Integrations  
APPROVED

Reviewed By :Mahesh Dadoda 09/11/2025  
Supervised By :Semsettin Yesilyurt 09/11/2025

Quant Time: Sep 11 03:47:17 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y090825S.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 10 01:44:33 2025  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

-----  
(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY091025\  
 Data File : VY023324.D  
 Acq On : 10 Sep 2025 10:33  
 Operator : SY/MD  
 Sample : VY0910SBSD01  
 Misc : 5.00g/5.0mL/MSVOA\_Y/soil  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 VY0910SBSD01

### Manual Integrations APPROVED

Reviewed By : Mahesh Dadoda 09/11/2025  
 Supervised By : Semsettin Yesilyurt 09/11/2025

Quant Time: Sep 11 03:47:17 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y090825S.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 10 01:44:33 2025  
 Response via : Initial Calibration



## Manual Integration Report

|           |          |            |         |
|-----------|----------|------------|---------|
| Sequence: | VN082125 | Instrument | MSVOA_n |
|-----------|----------|------------|---------|

| Sample ID   | File ID    | Parameter              | Review By | Review On            | Supervised By | Supervised On        | Reason                      |
|-------------|------------|------------------------|-----------|----------------------|---------------|----------------------|-----------------------------|
| VSTDICC001  | VN087619.D | 1,2,3-Trichloropropane | JOHN      | 8/22/2025 9:16:40 AM | MMDadoda      | 8/25/2025 8:11:19 AM | Peak Integrated by Software |
| VSTDICC001  | VN087619.D | 1,4-Dichlorobenzene    | JOHN      | 8/22/2025 9:16:40 AM | MMDadoda      | 8/25/2025 8:11:19 AM | Peak Integrated by Software |
| VSTDICC001  | VN087619.D | Diethyl Ether          | JOHN      | 8/22/2025 9:16:40 AM | MMDadoda      | 8/25/2025 8:11:19 AM | Peak Integrated by Software |
| VSTDICC001  | VN087619.D | Ethyl Acetate          | JOHN      | 8/22/2025 9:16:40 AM | MMDadoda      | 8/25/2025 8:11:19 AM | Peak Integrated by Software |
| VSTDICC001  | VN087619.D | Methyl Acetate         | JOHN      | 8/22/2025 9:16:40 AM | MMDadoda      | 8/25/2025 8:11:19 AM | Peak Integrated by Software |
| VSTDICC001  | VN087619.D | N-amyl acetate         | JOHN      | 8/22/2025 9:16:40 AM | MMDadoda      | 8/25/2025 8:11:19 AM | Peak Integrated by Software |
| VSTDICC005  | VN087620.D | 1,2,3-Trichloropropane | JOHN      | 8/22/2025 9:16:49 AM | MMDadoda      | 8/25/2025 8:11:20 AM | Peak Integrated by Software |
| VSTDICC005  | VN087620.D | Ethyl Acetate          | JOHN      | 8/22/2025 9:16:49 AM | MMDadoda      | 8/25/2025 8:11:20 AM | Peak Integrated by Software |
| VSTDICC005  | VN087620.D | N-amyl acetate         | JOHN      | 8/22/2025 9:16:49 AM | MMDadoda      | 8/25/2025 8:11:20 AM | Peak Integrated by Software |
| VSTDICC020  | VN087621.D | 1,2,3-Trichloropropane | JOHN      | 8/22/2025 9:17:20 AM | MMDadoda      | 8/25/2025 8:11:22 AM | Peak Integrated by Software |
| VSTDICC020  | VN087621.D | N-amyl acetate         | JOHN      | 8/22/2025 9:17:20 AM | MMDadoda      | 8/25/2025 8:11:22 AM | Peak Integrated by Software |
| VSTDICCC050 | VN087622.D | 1,2,3-Trichloropropane | JOHN      | 8/22/2025 9:17:24 AM | MMDadoda      | 8/25/2025 8:11:23 AM | Peak Integrated by Software |
| VSTDICCC050 | VN087622.D | N-amyl acetate         | JOHN      | 8/22/2025 9:17:24 AM | MMDadoda      | 8/25/2025 8:11:23 AM | Peak Integrated by Software |

## Manual Integration Report

Sequence:

VN082125

Instrument

MSVOA\_n

| Sample ID  | File ID    | Parameter              | Review By | Review On            | Supervised By | Supervised On        | Reason                      |
|------------|------------|------------------------|-----------|----------------------|---------------|----------------------|-----------------------------|
| VSTDICC100 | VN087623.D | 1,2,3-Trichloropropane | JOHN      | 8/22/2025 9:17:28 AM | MMDadoda      | 8/25/2025 8:11:24 AM | Peak Integrated by Software |
| VSTDICC150 | VN087624.D | 1,2,3-Trichloropropane | JOHN      | 8/22/2025 9:17:33 AM | MMDadoda      | 8/25/2025 8:11:26 AM | Peak Integrated by Software |
| VSTDICV050 | VN087626.D | 1,2,3-Trichloropropane | JOHN      | 8/22/2025 9:17:37 AM | MMDadoda      | 8/25/2025 8:11:28 AM | Peak Integrated by Software |
| VSTDICV050 | VN087626.D | N-amyl acetate         | JOHN      | 8/22/2025 9:17:37 AM | MMDadoda      | 8/25/2025 8:11:28 AM | Peak Integrated by Software |

## Manual Integration Report

|           |          |            |         |
|-----------|----------|------------|---------|
| Sequence: | VN091525 | Instrument | MSVOA_n |
|-----------|----------|------------|---------|

| Sample ID   | File ID    | Parameter              | Review By | Review On            | Supervised By | Supervised On        | Reason                      |
|-------------|------------|------------------------|-----------|----------------------|---------------|----------------------|-----------------------------|
| VSTDCCC050  | VN087820.D | 1,2,3-Trichloropropane | JOHN      | 9/16/2025 8:01:03 AM | MMDadoda      | 9/16/2025 8:01:46 AM | Peak Integrated by Software |
| VSTDCCC050  | VN087820.D | Bromomethane           | JOHN      | 9/16/2025 8:01:03 AM | MMDadoda      | 9/16/2025 8:01:46 AM | Peak Integrated by Software |
| VSTDCCC050  | VN087820.D | N-amyl acetate         | JOHN      | 9/16/2025 8:01:03 AM | MMDadoda      | 9/16/2025 8:01:46 AM | Peak Integrated by Software |
| VN0915MBS01 | VN087837.D | 1,2,3-Trichloropropane | MMDadoda  | 9/16/2025 8:02:11 AM | Sam           | 9/16/2025 8:05:40 AM | Peak Integrated by Software |
| VN0915MBS01 | VN087837.D | N-amyl acetate         | MMDadoda  | 9/16/2025 8:02:11 AM | Sam           | 9/16/2025 8:05:40 AM | Peak Integrated by Software |
| VSTDCCC050  | VN087845.D | 1,2,3-Trichloropropane | MMDadoda  | 9/16/2025 8:02:14 AM | Sam           | 9/16/2025 8:05:44 AM | Peak Integrated by Software |
| VSTDCCC050  | VN087845.D | N-amyl acetate         | MMDadoda  | 9/16/2025 8:02:14 AM | Sam           | 9/16/2025 8:05:44 AM | Peak Integrated by Software |

## Manual Integration Report

|           |          |            |         |
|-----------|----------|------------|---------|
| Sequence: | VY090825 | Instrument | MSVOA_y |
|-----------|----------|------------|---------|

| Sample ID   | File ID    | Parameter              | Review By    | Review On              | Supervised By | Supervised On          | Reason                      |
|-------------|------------|------------------------|--------------|------------------------|---------------|------------------------|-----------------------------|
| VSTDIC005   | VY023303.D | 1,2,3-Trichloropropane | MMDadod<br>a | 9/9/2025 3:12:20<br>PM | Sam           | 9/9/2025 3:14:24<br>PM | Peak Integrated by Software |
| VSTDIC005   | VY023303.D | Methacrylonitrile      | MMDadod<br>a | 9/9/2025 3:12:20<br>PM | Sam           | 9/9/2025 3:14:24<br>PM | Peak Integrated by Software |
| VSTDIC010   | VY023304.D | 1,2,3-Trichloropropane | MMDadod<br>a | 9/9/2025 3:12:22<br>PM | Sam           | 9/9/2025 3:14:25<br>PM | Peak Integrated by Software |
| VSTDIC010   | VY023304.D | Methacrylonitrile      | MMDadod<br>a | 9/9/2025 3:12:22<br>PM | Sam           | 9/9/2025 3:14:25<br>PM | Peak Integrated by Software |
| VSTDIC020   | VY023305.D | 1,2,3-Trichloropropane | MMDadod<br>a | 9/9/2025 3:12:23<br>PM | Sam           | 9/9/2025 3:14:30<br>PM | Peak Integrated by Software |
| VSTDIC020   | VY023305.D | Methacrylonitrile      | MMDadod<br>a | 9/9/2025 3:12:23<br>PM | Sam           | 9/9/2025 3:14:30<br>PM | Peak Integrated by Software |
| VSTDICCC050 | VY023306.D | 1,2,3-Trichloropropane | MMDadod<br>a | 9/9/2025 3:12:27<br>PM | Sam           | 9/9/2025 3:14:34<br>PM | Peak Integrated by Software |
| VSTDICCC050 | VY023306.D | Methacrylonitrile      | MMDadod<br>a | 9/9/2025 3:12:27<br>PM | Sam           | 9/9/2025 3:14:34<br>PM | Peak Integrated by Software |
| VSTDIC100   | VY023307.D | 1,2,3-Trichloropropane | MMDadod<br>a | 9/9/2025 3:12:48<br>PM | Sam           | 9/9/2025 3:14:36<br>PM | Peak Integrated by Software |
| VSTDIC150   | VY023308.D | 1,2,3-Trichloropropane | MMDadod<br>a | 9/9/2025 3:12:35<br>PM | Sam           | 9/9/2025 3:14:40<br>PM | Peak Integrated by Software |
| VSTDICV050  | VY023310.D | 1,2,3-Trichloropropane | MMDadod<br>a | 9/9/2025 3:12:36<br>PM | Sam           | 9/9/2025 3:14:46<br>PM | Peak Integrated by Software |

## Manual Integration Report

|           |          |            |         |
|-----------|----------|------------|---------|
| Sequence: | Vy091025 | Instrument | MSVOA_y |
|-----------|----------|------------|---------|

| Sample ID        | File ID    | Parameter              | Review By    | Review On               | Supervised By | Supervised On           | Reason                      |
|------------------|------------|------------------------|--------------|-------------------------|---------------|-------------------------|-----------------------------|
| VSTDCCC050       | VY023321.D | 1,2,3-Trichloropropane | MMDadod<br>a | 9/11/2025 5:30:08<br>PM | Sam           | 9/11/2025 5:40:46<br>PM | Peak Integrated by Software |
| VY0910SBS01      | VY023323.D | 1,2,3-Trichloropropane | MMDadod<br>a | 9/11/2025 5:30:10<br>PM | Sam           | 9/11/2025 5:39:42<br>PM | Peak Integrated by Software |
| VY0910SBS01      | VY023323.D | Methacrylonitrile      | MMDadod<br>a | 9/11/2025 5:30:10<br>PM | Sam           | 9/11/2025 5:39:42<br>PM | Peak Integrated by Software |
| VY0910SBSD0<br>1 | VY023324.D | 1,2,3-Trichloropropane | MMDadod<br>a | 9/11/2025 5:30:15<br>PM | Sam           | 9/11/2025 5:39:46<br>PM | Peak Integrated by Software |
| VY0910SBSD0<br>1 | VY023324.D | Methacrylonitrile      | MMDadod<br>a | 9/11/2025 5:30:15<br>PM | Sam           | 9/11/2025 5:39:46<br>PM | Peak Integrated by Software |
| Q3058-01         | VY023328.D | Acetone                | MMDadod<br>a | 9/11/2025 5:30:17<br>PM | Sam           | 9/11/2025 5:39:48<br>PM | Peak Integrated by Software |
| VSTDCCC050       | VY023334.D | 1,2,3-Trichloropropane | MMDadod<br>a | 9/11/2025 5:31:14<br>PM | Sam           | 9/11/2025 5:39:50<br>PM | Peak Integrated by Software |
| VSTDCCC050       | VY023334.D | Methacrylonitrile      | MMDadod<br>a | 9/11/2025 5:31:14<br>PM | Sam           | 9/11/2025 5:39:50<br>PM | Peak Integrated by Software |

Instrument ID: MSVOA\_N

**Daily Analysis Runlog For Sequence/QC Batch ID # VN082125**

|                          |                                                       |                   |                                   |
|--------------------------|-------------------------------------------------------|-------------------|-----------------------------------|
| Review By                | John Carlone                                          | Review On         | 8/22/2025 9:17:52 AM              |
| Supervise By             | Mahesh Dadoda                                         | Supervise On      | 8/25/2025 8:11:38 AM              |
| SubDirectory             | VN082125                                              | HP Acquire Method | HP Processing Method 82N082125W.M |
| <b>STD. NAME</b>         | <b>STD REF.#</b>                                      |                   |                                   |
| Tune/Reschk              | VP135214                                              |                   |                                   |
| Initial Calibration Stds | VP135215,VP135216,VP135217,VP135218,VP135219,VP135220 |                   |                                   |
| CCC                      |                                                       |                   |                                   |
| Internal Standard/PEM    |                                                       |                   |                                   |
| ICV/I.BLK                | VP135221                                              |                   |                                   |
| Surrogate Standard       |                                                       |                   |                                   |
| MS/MSD Standard          |                                                       |                   |                                   |
| LCS Standard             |                                                       |                   |                                   |

| Sr# | SampleId     | Data File Name | Date-Time         | Operator | Status |
|-----|--------------|----------------|-------------------|----------|--------|
| 1   | BFB          | VN087614.D     | 21 Aug 2025 09:30 | JC\MD    | Ok     |
| 2   | VSTDCCC020   | VN087615.D     | 21 Aug 2025 10:04 | JC\MD    | Not Ok |
| 3   | VN0821WBS01  | VN087616.D     | 21 Aug 2025 10:37 | JC\MD    | Not Ok |
| 4   | VN0821WBSD01 | VN087617.D     | 21 Aug 2025 11:11 | JC\MD    | Not Ok |
| 5   | VN0821WBL01  | VN087618.D     | 21 Aug 2025 11:36 | JC\MD    | Not Ok |
| 6   | VSTDICC001   | VN087619.D     | 21 Aug 2025 12:09 | JC\MD    | Ok,M   |
| 7   | VSTDICC005   | VN087620.D     | 21 Aug 2025 13:08 | JC\MD    | Ok,M   |
| 8   | VSTDICC020   | VN087621.D     | 21 Aug 2025 13:41 | JC\MD    | Ok,M   |
| 9   | VSTDICCC050  | VN087622.D     | 21 Aug 2025 14:02 | JC\MD    | Ok,M   |
| 10  | VSTDICC100   | VN087623.D     | 21 Aug 2025 14:23 | JC\MD    | Ok,M   |
| 11  | VSTDICC150   | VN087624.D     | 21 Aug 2025 14:44 | JC\MD    | Ok,M   |
| 12  | IBLK         | VN087625.D     | 21 Aug 2025 15:05 | JC\MD    | Ok     |
| 13  | VSTDICV050   | VN087626.D     | 21 Aug 2025 15:47 | JC\MD    | Ok,M   |

M : Manual Integration

Instrument ID: MSVOA\_N

## Daily Analysis Runlog For Sequence/QC Batch ID # VN091525

|                                                                                                    |                     |                   |                                   |
|----------------------------------------------------------------------------------------------------|---------------------|-------------------|-----------------------------------|
| Review By                                                                                          | Mahesh Dadoda       | Review On         | 9/16/2025 7:59:53 AM              |
| Supervise By                                                                                       | Semsettin Yesilyurt | Supervise On      | 9/16/2025 8:03:28 AM              |
| SubDirectory                                                                                       | VN091525            | HP Acquire Method | HP Processing Method 82N082125W.M |
| <b>STD. NAME</b>                                                                                   | <b>STD REF.#</b>    |                   |                                   |
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP135447            |                   |                                   |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP135448,VP135449   |                   |                                   |

| Sr# | SampleId     | Data File Name | Date-Time         | Operator | Status |
|-----|--------------|----------------|-------------------|----------|--------|
| 1   | BFB          | VN087819.D     | 15 Sep 2025 08:18 | JC\MD    | Ok     |
| 2   | VSTDCCC050   | VN087820.D     | 15 Sep 2025 08:39 | JC\MD    | Ok,M   |
| 3   | VN0915MBL01  | VN087821.D     | 15 Sep 2025 09:14 | JC\MD    | Ok     |
| 4   | VN0915WBL01  | VN087822.D     | 15 Sep 2025 09:35 | JC\MD    | Ok     |
| 5   | VN0915WBS01  | VN087823.D     | 15 Sep 2025 09:56 | JC\MD    | Ok,M   |
| 6   | VN0915WBSD01 | VN087824.D     | 15 Sep 2025 10:29 | JC\MD    | Ok,M   |
| 7   | Q3083-01     | VN087825.D     | 15 Sep 2025 10:50 | JC\MD    | Ok     |
| 8   | Q3088-02     | VN087826.D     | 15 Sep 2025 11:11 | JC\MD    | Ok     |
| 9   | PB169678TB   | VN087827.D     | 15 Sep 2025 11:32 | JC\MD    | Ok     |
| 10  | Q3082-02     | VN087828.D     | 15 Sep 2025 11:53 | JC\MD    | Ok     |
| 11  | Q3082-04     | VN087829.D     | 15 Sep 2025 12:14 | JC\MD    | Ok,M   |
| 12  | Q3082-06     | VN087830.D     | 15 Sep 2025 12:35 | JC\MD    | Ok     |
| 13  | Q3082-08     | VN087831.D     | 15 Sep 2025 12:56 | JC\MD    | Ok     |
| 14  | Q3082-10     | VN087832.D     | 15 Sep 2025 13:17 | JC\MD    | Ok,M   |
| 15  | Q3092-01     | VN087833.D     | 15 Sep 2025 13:38 | JC\MD    | Ok     |
| 16  | Q3092-02     | VN087834.D     | 15 Sep 2025 13:59 | JC\MD    | Ok     |
| 17  | Q3092-03     | VN087835.D     | 15 Sep 2025 14:20 | JC\MD    | Ok,M   |
| 18  | Q3092-04     | VN087836.D     | 15 Sep 2025 14:41 | JC\MD    | Ok     |
| 19  | VN0915MBS01  | VN087837.D     | 15 Sep 2025 15:02 | JC\MD    | Ok,M   |
| 20  | VN0915MBSD01 | VN087838.D     | 15 Sep 2025 15:22 | JC\MD    | Ok,M   |
| 21  | Q3058-01ME   | VN087839.D     | 15 Sep 2025 15:43 | JC\MD    | Ok     |

Instrument ID: MSVOA\_N

**Daily Analysis Runlog For Sequence/QCBatch ID # VN091525**

| Review By                                                                                          | Mahesh Dadoda       | Review On         | 9/16/2025 7:59:53 AM              |
|----------------------------------------------------------------------------------------------------|---------------------|-------------------|-----------------------------------|
| Supervise By                                                                                       | Semsettin Yesilyurt | Supervise On      | 9/16/2025 8:03:28 AM              |
| SubDirectory                                                                                       | VN091525            | HP Acquire Method | HP Processing Method 82N082125W.M |
| STD. NAME                                                                                          | STD REF.#           |                   |                                   |
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP135447            |                   |                                   |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP135448,VP135449   |                   |                                   |

|    |            |            |                   |       |        |
|----|------------|------------|-------------------|-------|--------|
| 22 | Q3075-04   | VN087840.D | 15 Sep 2025 16:04 | JC\MD | Ok     |
| 23 | Q3074-01   | VN087841.D | 15 Sep 2025 16:25 | JC\MD | Ok     |
| 24 | Q3073-02   | VN087842.D | 15 Sep 2025 16:46 | JC\MD | Ok     |
| 25 | Q3075-06   | VN087843.D | 15 Sep 2025 17:06 | JC\MD | Ok     |
| 26 | Q3077-01   | VN087844.D | 15 Sep 2025 17:27 | JC\MD | Not Ok |
| 27 | VSTDCCC050 | VN087845.D | 15 Sep 2025 18:09 | JC\MD | Ok,M   |

M : Manual Integration

Instrument ID: MSVOA\_Y

**Daily Analysis Runlog For Sequence/QCBatch ID # VY090825**

|                          |                     |                                                       |                      |                      |              |
|--------------------------|---------------------|-------------------------------------------------------|----------------------|----------------------|--------------|
| Review By                | Mahesh Dadoda       | Review On                                             | 9/9/2025 3:13:06 PM  |                      |              |
| Supervise By             | Semsettin Yesilyurt | Supervise On                                          | 9/19/2025 2:10:25 AM |                      |              |
| SubDirectory             | VY090825            | HP Acquire Method                                     | MSVOA_Y              | HP Processing Method | 82y090825s.m |
| STD. NAME                |                     | STD REF.#                                             |                      |                      |              |
| Tune/Reschk              |                     | VP135392                                              |                      |                      |              |
| Initial Calibration Stds |                     | VP135393,VP135394,VP135395,VP135396,VP135397,VP135398 |                      |                      |              |
| CCC                      |                     |                                                       |                      |                      |              |
| Internal Standard/PEM    |                     | VP133934                                              |                      |                      |              |
| ICV/I.BLK                |                     | VP135399                                              |                      |                      |              |
| Surrogate Standard       |                     |                                                       |                      |                      |              |
| MS/MSD Standard          |                     |                                                       |                      |                      |              |
| LCS Standard             |                     |                                                       |                      |                      |              |

| Sr# | SampleId   | Data File Name | Date-Time         | Operator | Status |
|-----|------------|----------------|-------------------|----------|--------|
| 1   | BFB        | VY023302.D     | 08 Sep 2025 09:07 | SY/MD    | Ok     |
| 2   | VSTDIC005  | VY023303.D     | 08 Sep 2025 10:00 | SY/MD    | Ok,M   |
| 3   | VSTDIC010  | VY023304.D     | 08 Sep 2025 10:23 | SY/MD    | Ok,M   |
| 4   | VSTDIC020  | VY023305.D     | 08 Sep 2025 11:01 | SY/MD    | Ok,M   |
| 5   | VSTDIC050  | VY023306.D     | 08 Sep 2025 11:23 | SY/MD    | Ok,M   |
| 6   | VSTDIC100  | VY023307.D     | 08 Sep 2025 11:46 | SY/MD    | Ok,M   |
| 7   | VSTDIC150  | VY023308.D     | 08 Sep 2025 12:08 | SY/MD    | Ok,M   |
| 8   | VIBLK      | VY023309.D     | 08 Sep 2025 12:32 | SY/MD    | Ok     |
| 9   | VSTDICV050 | VY023310.D     | 08 Sep 2025 12:55 | SY/MD    | Ok,M   |

M : Manual Integration

Instrument ID: MSVOA\_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY091025

|                                                                                                    |                               |                      |                      |
|----------------------------------------------------------------------------------------------------|-------------------------------|----------------------|----------------------|
| Review By                                                                                          | Semsettin Yesilyurt           | Review On            | 9/11/2025 5:40:36 PM |
| Supervise By                                                                                       | Maresh Dadoda                 | Supervise On         | 9/19/2025 2:06:55 AM |
| SubDirectory                                                                                       | VY091025                      | HP Acquire Method    | MSVOA_Y              |
|                                                                                                    |                               | HP Processing Method | 82y090825s.m         |
| <b>STD. NAME</b>                                                                                   | <b>STD REF.#</b>              |                      |                      |
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP135409                      |                      |                      |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP135410,VP135411<br>VP133934 |                      |                      |

| Sr# | SampleId     | Data File Name | Date-Time         | Operator | Status   |
|-----|--------------|----------------|-------------------|----------|----------|
| 1   | BFB          | VY023320.D     | 10 Sep 2025 08:38 | SY/MD    | Ok       |
| 2   | VSTDCCC050   | VY023321.D     | 10 Sep 2025 09:09 | SY/MD    | Ok,M     |
| 3   | VY0910SBL01  | VY023322.D     | 10 Sep 2025 09:39 | SY/MD    | Ok       |
| 4   | VY0910SBS01  | VY023323.D     | 10 Sep 2025 10:11 | SY/MD    | Ok,M     |
| 5   | VY0910SBSD01 | VY023324.D     | 10 Sep 2025 10:33 | SY/MD    | Ok,M     |
| 6   | Q3055-01     | VY023325.D     | 10 Sep 2025 11:39 | SY/MD    | Ok       |
| 7   | Q3056-01     | VY023326.D     | 10 Sep 2025 12:02 | SY/MD    | Ok       |
| 8   | Q3059-01     | VY023327.D     | 10 Sep 2025 12:25 | SY/MD    | Ok       |
| 9   | Q3058-01     | VY023328.D     | 10 Sep 2025 12:49 | SY/MD    | Dilution |
| 10  | Q3058-02     | VY023329.D     | 10 Sep 2025 13:12 | SY/MD    | Ok       |
| 11  | Q3066-01     | VY023330.D     | 10 Sep 2025 14:02 | SY/MD    | Ok       |
| 12  | Q3066-03     | VY023331.D     | 10 Sep 2025 14:26 | SY/MD    | Ok       |
| 13  | Q3065-01     | VY023332.D     | 10 Sep 2025 14:49 | SY/MD    | ReRun    |
| 14  | VIBLK        | VY023333.D     | 10 Sep 2025 15:13 | SY/MD    | Ok       |
| 15  | VSTDCCC050   | VY023334.D     | 10 Sep 2025 15:35 | SY/MD    | Ok,M     |

M : Manual Integration

Instrument ID: MSVOA\_N

### Daily Analysis Runlog For Sequence/QC Batch ID # VN082125

|                          |                                                       |                   |                                   |
|--------------------------|-------------------------------------------------------|-------------------|-----------------------------------|
| Review By                | John Carlone                                          | Review On         | 8/22/2025 9:17:52 AM              |
| Supervise By             | Maresh Dadoda                                         | Supervise On      | 8/25/2025 8:11:38 AM              |
| SubDirectory             | VN082125                                              | HP Acquire Method | HP Processing Method 82N082125W.M |
| <b>STD. NAME</b>         | <b>STD REF.#</b>                                      |                   |                                   |
| Tune/Reschk              | VP135214                                              |                   |                                   |
| Initial Calibration Stds | VP135215,VP135216,VP135217,VP135218,VP135219,VP135220 |                   |                                   |
| CCC                      |                                                       |                   |                                   |
| Internal Standard/PEM    |                                                       |                   |                                   |
| ICV/I.BLK                | VP135221                                              |                   |                                   |
| Surrogate Standard       |                                                       |                   |                                   |
| MS/MSD Standard          |                                                       |                   |                                   |
| LCS Standard             |                                                       |                   |                                   |

| Sr# | SampleID     | ClientID     | Data File Name | Date-Time         | Comment                              | Operator | Status |
|-----|--------------|--------------|----------------|-------------------|--------------------------------------|----------|--------|
| 1   | BFB          | BFB          | VN087614.D     | 21 Aug 2025 09:30 |                                      | JC\MD    | Ok     |
| 2   | VSTDCCC020   | VSTDCCC020   | VN087615.D     | 21 Aug 2025 10:04 | Need ICAL                            | JC\MD    | Not Ok |
| 3   | VN0821WBS01  | VN0821WBS01  | VN087616.D     | 21 Aug 2025 10:37 |                                      | JC\MD    | Not Ok |
| 4   | VN0821WBSD01 | VN0821WBSD01 | VN087617.D     | 21 Aug 2025 11:11 |                                      | JC\MD    | Not Ok |
| 5   | VN0821WBL01  | VN0821WBL01  | VN087618.D     | 21 Aug 2025 11:36 |                                      | JC\MD    | Not Ok |
| 6   | VSTDICC001   | VSTDICC001   | VN087619.D     | 21 Aug 2025 12:09 | LR-10,20                             | JC\MD    | Ok,M   |
| 7   | VSTDICC005   | VSTDICC005   | VN087620.D     | 21 Aug 2025 13:08 |                                      | JC\MD    | Ok,M   |
| 8   | VSTDICC020   | VSTDICC020   | VN087621.D     | 21 Aug 2025 13:41 | method not used for #N-amyle Acetate | JC\MD    | Ok,M   |
| 9   | VSTDICCC050  | VSTDICCC050  | VN087622.D     | 21 Aug 2025 14:02 |                                      | JC\MD    | Ok,M   |
| 10  | VSTDICC100   | VSTDICC100   | VN087623.D     | 21 Aug 2025 14:23 |                                      | JC\MD    | Ok,M   |
| 11  | VSTDICC150   | VSTDICC150   | VN087624.D     | 21 Aug 2025 14:44 |                                      | JC\MD    | Ok,M   |
| 12  | IBLK         | IBLK         | VN087625.D     | 21 Aug 2025 15:05 |                                      | JC\MD    | Ok     |
| 13  | VSTDICV050   | ICVVN082125  | VN087626.D     | 21 Aug 2025 15:47 | Icv Failed for Com.# 64 for DOD      | JC\MD    | Ok,M   |

M : Manual Integration

Instrument ID: MSVOA\_N

**Daily Analysis Runlog For Sequence/QC Batch ID # VN091525**

|                                                                                                    |                     |                   |                                   |
|----------------------------------------------------------------------------------------------------|---------------------|-------------------|-----------------------------------|
| Review By                                                                                          | Mahesh Dadoda       | Review On         | 9/16/2025 7:59:53 AM              |
| Supervise By                                                                                       | Semsettin Yesilyurt | Supervise On      | 9/16/2025 8:03:28 AM              |
| SubDirectory                                                                                       | VN091525            | HP Acquire Method | HP Processing Method 82N082125W.M |
| <b>STD. NAME</b>                                                                                   | <b>STD REF.#</b>    |                   |                                   |
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP135447            |                   |                                   |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP135448,VP135449   |                   |                                   |

| Sr# | SampleID     | ClientID     | Data File Name | Date-Time         | Comment                          | Operator | Status |
|-----|--------------|--------------|----------------|-------------------|----------------------------------|----------|--------|
| 1   | BFB          | BFB          | VN087819.D     | 15 Sep 2025 08:18 |                                  | JC\MD    | Ok     |
| 2   | VSTDCCC050   | VSTDCCC050   | VN087820.D     | 15 Sep 2025 08:39 | pH#Lot#V12668                    | JC\MD    | Ok,M   |
| 3   | VN0915MBL01  | VN0915MBL01  | VN087821.D     | 15 Sep 2025 09:14 |                                  | JC\MD    | Ok     |
| 4   | VN0915WBL01  | VN0915WBL01  | VN087822.D     | 15 Sep 2025 09:35 |                                  | JC\MD    | Ok     |
| 5   | VN0915WBS01  | VN0915WBS01  | VN087823.D     | 15 Sep 2025 09:56 |                                  | JC\MD    | Ok,M   |
| 6   | VN0915WBSD01 | VN0915WBSD01 | VN087824.D     | 15 Sep 2025 10:29 |                                  | JC\MD    | Ok,M   |
| 7   | Q3083-01     | MW4S         | VN087825.D     | 15 Sep 2025 10:50 | vial B pH<2                      | JC\MD    | Ok     |
| 8   | Q3088-02     | 90525-LQ     | VN087826.D     | 15 Sep 2025 11:11 | vial A pH<2 brown/ turbid sample | JC\MD    | Ok     |
| 9   | PB169678TB   | PB169678TB   | VN087827.D     | 15 Sep 2025 11:32 |                                  | JC\MD    | Ok     |
| 10  | Q3082-02     | WC-29        | VN087828.D     | 15 Sep 2025 11:53 | vial A pH#5.0                    | JC\MD    | Ok     |
| 11  | Q3082-04     | WC-30        | VN087829.D     | 15 Sep 2025 12:14 | vial A pH#5.0                    | JC\MD    | Ok,M   |
| 12  | Q3082-06     | WC-31        | VN087830.D     | 15 Sep 2025 12:35 | vial A pH#5.0                    | JC\MD    | Ok     |
| 13  | Q3082-08     | WC-32        | VN087831.D     | 15 Sep 2025 12:56 | vial A pH#5.0                    | JC\MD    | Ok     |
| 14  | Q3082-10     | WC-33        | VN087832.D     | 15 Sep 2025 13:17 | vial A pH#5.0                    | JC\MD    | Ok,M   |
| 15  | Q3092-01     | 291376       | VN087833.D     | 15 Sep 2025 13:38 | vial A pH#5.0                    | JC\MD    | Ok     |
| 16  | Q3092-02     | 291376       | VN087834.D     | 15 Sep 2025 13:59 | vial A pH#5.0                    | JC\MD    | Ok     |
| 17  | Q3092-03     | 279182       | VN087835.D     | 15 Sep 2025 14:20 | vial A pH#5.0                    | JC\MD    | Ok,M   |

Instrument ID: MSVOA\_N

**Daily Analysis Runlog For Sequence/QC Batch ID # VN091525**

| Review By                                                                                          | Maresh Dadoda       | Review On         | 9/16/2025 7:59:53 AM              |
|----------------------------------------------------------------------------------------------------|---------------------|-------------------|-----------------------------------|
| Supervise By                                                                                       | Semsettin Yesilyurt | Supervise On      | 9/16/2025 8:03:28 AM              |
| SubDirectory                                                                                       | VN091525            | HP Acquire Method | HP Processing Method 82N082125W.M |
| STD. NAME                                                                                          | STD REF.#           |                   |                                   |
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP135447            |                   |                                   |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP135448,VP135449   |                   |                                   |

|    |              |              |            |                   |                                  |       |        |
|----|--------------|--------------|------------|-------------------|----------------------------------|-------|--------|
| 18 | Q3092-04     | 279182       | VN087836.D | 15 Sep 2025 14:41 | vial A pH#5.0                    | JC\MD | Ok     |
| 19 | VN0915MBS01  | VN0915MBS01  | VN087837.D | 15 Sep 2025 15:02 |                                  | JC\MD | Ok,M   |
| 20 | VN0915MBSD01 | VN0915MBSD01 | VN087838.D | 15 Sep 2025 15:22 |                                  | JC\MD | Ok,M   |
| 21 | Q3058-01ME   | RPXY5ME      | VN087839.D | 15 Sep 2025 15:43 |                                  | JC\MD | Ok     |
| 22 | Q3075-04     | MOO-25-0258  | VN087840.D | 15 Sep 2025 16:04 | cloth material                   | JC\MD | Ok     |
| 23 | Q3074-01     | MOO-25-0141  | VN087841.D | 15 Sep 2025 16:25 | cloth material                   | JC\MD | Ok     |
| 24 | Q3073-02     | BEL-25-0023  | VN087842.D | 15 Sep 2025 16:46 | cloth material                   | JC\MD | Ok     |
| 25 | Q3075-06     | MOO-25-0259  | VN087843.D | 15 Sep 2025 17:06 | cloth material                   | JC\MD | Ok     |
| 26 | Q3077-01     | LAW-25-0137  | VN087844.D | 15 Sep 2025 17:27 | cloth material run straight MEOH | JC\MD | Not Ok |
| 27 | VSTDCCC050   | VSTDCCC050EC | VN087845.D | 15 Sep 2025 18:09 |                                  | JC\MD | Ok,M   |

M : Manual Integration

Instrument ID: MSVOA\_Y

**Daily Analysis Runlog For Sequence/QC Batch ID # VY090825**

|                          |                     |                                                       |                      |                      |              |
|--------------------------|---------------------|-------------------------------------------------------|----------------------|----------------------|--------------|
| Review By                | Mahesh Dadoda       | Review On                                             | 9/9/2025 3:13:06 PM  |                      |              |
| Supervise By             | Semsettin Yesilyurt | Supervise On                                          | 9/19/2025 2:10:25 AM |                      |              |
| SubDirectory             | VY090825            | HP Acquire Method                                     | MSVOA_Y              | HP Processing Method | 82y090825s.m |
| STD. NAME                |                     | STD REF.#                                             |                      |                      |              |
| Tune/Reschk              |                     | VP135392                                              |                      |                      |              |
| Initial Calibration Stds |                     | VP135393,VP135394,VP135395,VP135396,VP135397,VP135398 |                      |                      |              |
| CCC                      |                     |                                                       |                      |                      |              |
| Internal Standard/PEM    |                     | VP133934                                              |                      |                      |              |
| ICV/I.BLK                |                     | VP135399                                              |                      |                      |              |
| Surrogate Standard       |                     |                                                       |                      |                      |              |
| MS/MSD Standard          |                     |                                                       |                      |                      |              |
| LCS Standard             |                     |                                                       |                      |                      |              |

| Sr# | Sampleld   | ClientID    | Data File Name | Date-Time         | Comment | Operator | Status |
|-----|------------|-------------|----------------|-------------------|---------|----------|--------|
| 1   | BFB        | BFB         | VY023302.D     | 08 Sep 2025 09:07 |         | SY/MD    | Ok     |
| 2   | VSTDIC005  | VSTDIC005   | VY023303.D     | 08 Sep 2025 10:00 |         | SY/MD    | Ok,M   |
| 3   | VSTDIC010  | VSTDIC010   | VY023304.D     | 08 Sep 2025 10:23 |         | SY/MD    | Ok,M   |
| 4   | VSTDIC020  | VSTDIC020   | VY023305.D     | 08 Sep 2025 11:01 |         | SY/MD    | Ok,M   |
| 5   | VSTDIC050  | VSTDIC050   | VY023306.D     | 08 Sep 2025 11:23 |         | SY/MD    | Ok,M   |
| 6   | VSTDIC100  | VSTDIC100   | VY023307.D     | 08 Sep 2025 11:46 |         | SY/MD    | Ok,M   |
| 7   | VSTDIC150  | VSTDIC150   | VY023308.D     | 08 Sep 2025 12:08 |         | SY/MD    | Ok,M   |
| 8   | VIBLK      | VIBLK       | VY023309.D     | 08 Sep 2025 12:32 |         | SY/MD    | Ok     |
| 9   | VSTDICV050 | ICVVY090825 | VY023310.D     | 08 Sep 2025 12:55 |         | SY/MD    | Ok,M   |

M : Manual Integration

Instrument ID: MSVOA\_Y

### Daily Analysis Runlog For Sequence/QC Batch ID # VY091025

|                                                                                                    |                               |                   |                                           |
|----------------------------------------------------------------------------------------------------|-------------------------------|-------------------|-------------------------------------------|
| Review By                                                                                          | Semsettin Yesilyurt           | Review On         | 9/11/2025 5:40:36 PM                      |
| Supervise By                                                                                       | Maresh Dadoda                 | Supervise On      | 9/19/2025 2:06:55 AM                      |
| SubDirectory                                                                                       | VY091025                      | HP Acquire Method | MSVOA_Y HP Processing Method 82y090825s.m |
| <b>STD. NAME</b>                                                                                   | <b>STD REF.#</b>              |                   |                                           |
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP135409                      |                   |                                           |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP135410,VP135411<br>VP133934 |                   |                                           |

| Sr# | SampleID     | ClientID       | Data File Name | Date-Time         | Comment                       | Operator | Status   |
|-----|--------------|----------------|----------------|-------------------|-------------------------------|----------|----------|
| 1   | BFB          | BFB            | VY023320.D     | 10 Sep 2025 08:38 |                               | SY/MD    | Ok       |
| 2   | VSTDCCC050   | VSTDCCC050     | VY023321.D     | 10 Sep 2025 09:09 |                               | SY/MD    | Ok,M     |
| 3   | VY0910SBL01  | VY0910SBL01    | VY023322.D     | 10 Sep 2025 09:39 |                               | SY/MD    | Ok       |
| 4   | VY0910SBS01  | VY0910SBS01    | VY023323.D     | 10 Sep 2025 10:11 |                               | SY/MD    | Ok,M     |
| 5   | VY0910SBSD01 | VY0910SBSD01   | VY023324.D     | 10 Sep 2025 10:33 |                               | SY/MD    | Ok,M     |
| 6   | Q3055-01     | OR-03-090925   | VY023325.D     | 10 Sep 2025 11:39 | vial-A                        | SY/MD    | Ok       |
| 7   | Q3056-01     | OK-02-090925   | VY023326.D     | 10 Sep 2025 12:02 | vial-A                        | SY/MD    | Ok       |
| 8   | Q3059-01     | TR-06-09092025 | VY023327.D     | 10 Sep 2025 12:25 | vial-A                        | SY/MD    | Ok       |
| 9   | Q3058-01     | RPXY5          | VY023328.D     | 10 Sep 2025 12:49 | vial-A Need MeOH              | SY/MD    | Dilution |
| 10  | Q3058-02     | RPXY4          | VY023329.D     | 10 Sep 2025 13:12 | vial-A                        | SY/MD    | Ok       |
| 11  | Q3066-01     | 354            | VY023330.D     | 10 Sep 2025 14:02 | vial-A                        | SY/MD    | Ok       |
| 12  | Q3066-03     | VNJ-242        | VY023331.D     | 10 Sep 2025 14:26 | vial-A                        | SY/MD    | Ok       |
| 13  | Q3065-01     | VNJ-204        | VY023332.D     | 10 Sep 2025 14:49 | vial-A Internal Standard fail | SY/MD    | ReRun    |
| 14  | VIBLK        | VIBLK          | VY023333.D     | 10 Sep 2025 15:13 |                               | SY/MD    | Ok       |
| 15  | VSTDCCC050   | VSTDCCC050EC   | VY023334.D     | 10 Sep 2025 15:35 |                               | SY/MD    | Ok,M     |

M : Manual Integration

### LAB CHRONICLE

|                 |                 |                   |                     |
|-----------------|-----------------|-------------------|---------------------|
| <b>OrderID:</b> | Q3058           | <b>OrderDate:</b> | 9/9/2025 1:49:00 PM |
| <b>Client:</b>  | G Environmental | <b>Project:</b>   | Buffington          |
| <b>Contact:</b> | Gary Landis     | <b>Location:</b>  | J42,VOA Lab         |

| LabID             | ClientID       | Matrix       | Test          | Method   | Sample Date     | Prep Date | Anal Date | Received        |
|-------------------|----------------|--------------|---------------|----------|-----------------|-----------|-----------|-----------------|
| <b>Q3058-01</b>   | <b>RPXY5</b>   | <b>SOIL</b>  | VOC-TCLVOA-10 | 8260D    | <b>09/09/25</b> |           | 09/10/25  | <b>09/09/25</b> |
| <b>Q3058-01ME</b> | <b>RPXY5ME</b> | <b>SOIL</b>  | VOC-TCLVOA-10 | 8260D    | <b>09/09/25</b> |           | 09/15/25  | <b>09/09/25</b> |
| <b>Q3058-02</b>   | <b>RPXY4</b>   | <b>SOIL</b>  | VOC-TCLVOA-10 | 8260D    | <b>09/09/25</b> |           | 09/10/25  | <b>09/09/25</b> |
| <b>Q3058-03</b>   | <b>MW4</b>     | <b>Water</b> | VOCMS Group1  | 8260-Low | <b>09/09/25</b> |           | 09/10/25  | <b>09/09/25</b> |

**Hit Summary Sheet**  
**SW-846**

**SDG No.:** Q3058  
**Client:** G Environmental

| Sample ID         | Client ID  | Matrix | Parameter                   | Concentration | C | MDL  | RDL  | Units |
|-------------------|------------|--------|-----------------------------|---------------|---|------|------|-------|
| <b>Client ID:</b> | <b>MW4</b> |        |                             |               |   |      |      |       |
| Q3058-03          | MW4        | Water  | Chloromethane               | 24.3          |   | 0.32 | 1.00 | ug/L  |
| Q3058-03          | MW4        | Water  | Bromomethane                | 9.90          |   | 1.40 | 5.00 | ug/L  |
| Q3058-03          | MW4        | Water  | Acetone                     | 23.1          |   | 1.50 | 5.00 | ug/L  |
| Q3058-03          | MW4        | Water  | Carbon Disulfide            | 0.76          | J | 0.21 | 1.00 | ug/L  |
| Q3058-03          | MW4        | Water  | Chloroform                  | 3.00          |   | 0.25 | 1.00 | ug/L  |
| Q3058-03          | MW4        | Water  | Benzene                     | 0.79          | J | 0.15 | 1.00 | ug/L  |
| Q3058-03          | MW4        | Water  | Bromodichloromethane        | 0.37          | J | 0.22 | 1.00 | ug/L  |
| Q3058-03          | MW4        | Water  | Toluene                     | 1.10          |   | 0.14 | 1.00 | ug/L  |
| Q3058-03          | MW4        | Water  | Ethyl Benzene               | 0.82          | J | 0.13 | 1.00 | ug/L  |
| Q3058-03          | MW4        | Water  | m/p-Xylenes                 | 0.99          | J | 0.24 | 2.00 | ug/L  |
| Q3058-03          | MW4        | Water  | o-Xylene                    | 1.10          |   | 0.12 | 1.00 | ug/L  |
|                   |            |        | <b>Total Voc :</b>          |               |   | 66.2 |      |       |
| Q3058-03          | MW4        | Water  | Naphthalene                 | * 1.20        | J | 0.20 | 1.00 | ug/L  |
| Q3058-03          | MW4        | Water  | Methyl Iodide               | * 14.6        | J | 0.83 | 5.00 | ug/L  |
|                   |            |        | <b>Total Tics :</b>         |               |   | 15.8 |      |       |
|                   |            |        | <b>Total Concentration:</b> |               |   | 82.0 |      |       |



# SAMPLE DATA

### Report of Analysis

|                    |                 |           |                 |              |    |
|--------------------|-----------------|-----------|-----------------|--------------|----|
| Client:            | G Environmental |           | Date Collected: | 09/09/25     |    |
| Project:           | Buffington      |           | Date Received:  | 09/09/25     |    |
| Client Sample ID:  | MW4             |           | SDG No.:        | Q3058        |    |
| Lab Sample ID:     | Q3058-03        |           | Matrix:         | Water        |    |
| Analytical Method: | 8260D           |           | % Solid:        | 0            |    |
| Sample Wt/Vol:     | 5               | Units: mL | Final Vol:      | 5000         | uL |
| Soil Aliquot Vol:  |                 | uL        | Test:           | VOCMS Group1 |    |
| GC Column:         | RXI-624         | ID : 0.25 | Level :         | LOW          |    |
| Prep Method :      |                 |           |                 |              |    |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VN087782.D        | 1         | 09/10/25 15:47 | VN091025      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL   | LOQ / CRQL | Units |
|----------------|--------------------------------|-------|-----------|-------|------------|-------|
| <b>TARGETS</b> |                                |       |           |       |            |       |
| 75-71-8        | Dichlorodifluoromethane        | 0.22  | U         | 0.22  | 1.00       | ug/L  |
| 74-87-3        | Chloromethane                  | 24.3  |           | 0.32  | 1.00       | ug/L  |
| 75-01-4        | Vinyl Chloride                 | 0.26  | U         | 0.26  | 1.00       | ug/L  |
| 74-83-9        | Bromomethane                   | 9.90  |           | 1.40  | 5.00       | ug/L  |
| 75-00-3        | Chloroethane                   | 0.47  | U         | 0.47  | 1.00       | ug/L  |
| 75-69-4        | Trichlorofluoromethane         | 0.33  | U         | 0.33  | 1.00       | ug/L  |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 0.25  | U         | 0.25  | 1.00       | ug/L  |
| 75-35-4        | 1,1-Dichloroethene             | 0.23  | U         | 0.23  | 1.00       | ug/L  |
| 67-64-1        | Acetone                        | 23.1  |           | 1.50  | 5.00       | ug/L  |
| 75-15-0        | Carbon Disulfide               | 0.76  | J         | 0.21  | 1.00       | ug/L  |
| 1634-04-4      | Methyl tert-butyl Ether        | 0.16  | U         | 0.16  | 1.00       | ug/L  |
| 79-20-9        | Methyl Acetate                 | 0.27  | U         | 0.27  | 1.00       | ug/L  |
| 75-09-2        | Methylene Chloride             | 0.28  | U         | 0.28  | 1.00       | ug/L  |
| 156-60-5       | trans-1,2-Dichloroethene       | 0.23  | U         | 0.23  | 1.00       | ug/L  |
| 75-34-3        | 1,1-Dichloroethane             | 0.23  | U         | 0.23  | 1.00       | ug/L  |
| 110-82-7       | Cyclohexane                    | 1.50  | U         | 1.50  | 5.00       | ug/L  |
| 78-93-3        | 2-Butanone                     | 0.98  | U         | 0.98  | 5.00       | ug/L  |
| 56-23-5        | Carbon Tetrachloride           | 0.25  | U         | 0.25  | 1.00       | ug/L  |
| 156-59-2       | cis-1,2-Dichloroethene         | 0.19  | U         | 0.19  | 1.00       | ug/L  |
| 74-97-5        | Bromochloromethane             | 0.22  | U         | 0.22  | 1.00       | ug/L  |
| 67-66-3        | Chloroform                     | 3.00  |           | 0.25  | 1.00       | ug/L  |
| 71-55-6        | 1,1,1-Trichloroethane          | 0.20  | U         | 0.20  | 1.00       | ug/L  |
| 108-87-2       | Methylcyclohexane              | 0.16  | U         | 0.16  | 1.00       | ug/L  |
| 71-43-2        | Benzene                        | 0.79  | J         | 0.15  | 1.00       | ug/L  |
| 107-06-2       | 1,2-Dichloroethane             | 0.22  | U         | 0.22  | 1.00       | ug/L  |
| 79-01-6        | Trichloroethene                | 0.090 | U         | 0.090 | 1.00       | ug/L  |
| 78-87-5        | 1,2-Dichloropropane            | 0.20  | U         | 0.20  | 1.00       | ug/L  |
| 74-95-3        | Dibromomethane                 | 0.25  | U         | 0.25  | 1.00       | ug/L  |
| 75-27-4        | Bromodichloromethane           | 0.37  | J         | 0.22  | 1.00       | ug/L  |
| 108-10-1       | 4-Methyl-2-Pentanone           | 0.68  | U         | 0.68  | 5.00       | ug/L  |

### Report of Analysis

|                    |                 |           |                 |              |    |
|--------------------|-----------------|-----------|-----------------|--------------|----|
| Client:            | G Environmental |           | Date Collected: | 09/09/25     |    |
| Project:           | Buffington      |           | Date Received:  | 09/09/25     |    |
| Client Sample ID:  | MW4             |           | SDG No.:        | Q3058        |    |
| Lab Sample ID:     | Q3058-03        |           | Matrix:         | Water        |    |
| Analytical Method: | 8260D           |           | % Solid:        | 0            |    |
| Sample Wt/Vol:     | 5               | Units: mL | Final Vol:      | 5000         | uL |
| Soil Aliquot Vol:  |                 | uL        | Test:           | VOCMS Group1 |    |
| GC Column:         | RXI-624         | ID : 0.25 | Level :         | LOW          |    |
| Prep Method :      |                 |           |                 |              |    |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VN087782.D        | 1         | 09/10/25 15:47 | VN091025      |

| CAS Number                            | Parameter                   | Conc.  | Qualifier | MDL                 | LOQ / CRQL | Units   |
|---------------------------------------|-----------------------------|--------|-----------|---------------------|------------|---------|
| 108-88-3                              | Toluene                     | 1.10   |           | 0.14                | 1.00       | ug/L    |
| 10061-02-6                            | t-1,3-Dichloropropene       | 0.17   | U         | 0.17                | 1.00       | ug/L    |
| 10061-01-5                            | cis-1,3-Dichloropropene     | 0.16   | U         | 0.16                | 1.00       | ug/L    |
| 79-00-5                               | 1,1,2-Trichloroethane       | 0.21   | U         | 0.21                | 1.00       | ug/L    |
| 591-78-6                              | 2-Hexanone                  | 0.89   | U         | 0.89                | 5.00       | ug/L    |
| 106-93-4                              | 1,2-Dibromoethane           | 0.15   | U         | 0.15                | 1.00       | ug/L    |
| 127-18-4                              | Tetrachloroethene           | 0.23   | U         | 0.23                | 1.00       | ug/L    |
| 108-90-7                              | Chlorobenzene               | 0.12   | U         | 0.12                | 1.00       | ug/L    |
| 100-41-4                              | Ethyl Benzene               | 0.82   | J         | 0.13                | 1.00       | ug/L    |
| 179601-23-1                           | m/p-Xylenes                 | 0.99   | J         | 0.24                | 2.00       | ug/L    |
| 95-47-6                               | o-Xylene                    | 1.10   |           | 0.12                | 1.00       | ug/L    |
| 100-42-5                              | Styrene                     | 0.15   | U         | 0.15                | 1.00       | ug/L    |
| 75-25-2                               | Bromoform                   | 0.19   | U         | 0.19                | 1.00       | ug/L    |
| 98-82-8                               | Isopropylbenzene            | 0.12   | U         | 0.12                | 1.00       | ug/L    |
| 79-34-5                               | 1,1,2,2-Tetrachloroethane   | 0.26   | U         | 0.26                | 1.00       | ug/L    |
| 541-73-1                              | 1,3-Dichlorobenzene         | 0.16   | U         | 0.16                | 1.00       | ug/L    |
| 106-46-7                              | 1,4-Dichlorobenzene         | 0.19   | U         | 0.19                | 1.00       | ug/L    |
| 95-50-1                               | 1,2-Dichlorobenzene         | 0.16   | U         | 0.16                | 1.00       | ug/L    |
| 96-12-8                               | 1,2-Dibromo-3-Chloropropane | 0.53   | U         | 0.53                | 1.00       | ug/L    |
| 120-82-1                              | 1,2,4-Trichlorobenzene      | 0.20   | U         | 0.20                | 1.00       | ug/L    |
| 87-61-6                               | 1,2,3-Trichlorobenzene      | 0.20   | U         | 0.20                | 1.00       | ug/L    |
| <b>SURROGATES</b>                     |                             |        |           |                     |            |         |
| 17060-07-0                            | 1,2-Dichloroethane-d4       | 51.6   |           | 70 (74) - 130 (125) | 103%       | SPK: 50 |
| 1868-53-7                             | Dibromofluoromethane        | 50.3   |           | 70 (75) - 130 (124) | 101%       | SPK: 50 |
| 2037-26-5                             | Toluene-d8                  | 46.8   |           | 70 (86) - 130 (113) | 94%        | SPK: 50 |
| 460-00-4                              | 4-Bromofluorobenzene        | 47.1   |           | 70 (77) - 130 (121) | 94%        | SPK: 50 |
| <b>INTERNAL STANDARDS</b>             |                             |        |           |                     |            |         |
| 363-72-4                              | Pentafluorobenzene          | 197000 | 8.206     |                     |            |         |
| 540-36-3                              | 1,4-Difluorobenzene         | 454000 | 9.082     |                     |            |         |
| 3114-55-4                             | Chlorobenzene-d5            | 435000 | 11.841    |                     |            |         |
| 3855-82-1                             | 1,4-Dichlorobenzene-d4      | 206000 | 13.77     |                     |            |         |
| <b>TENTATIVE IDENTIFIED COMPOUNDS</b> |                             |        |           |                     |            |         |

## Report of Analysis

|                    |                 |           |                 |              |    |
|--------------------|-----------------|-----------|-----------------|--------------|----|
| Client:            | G Environmental |           | Date Collected: | 09/09/25     |    |
| Project:           | Buffington      |           | Date Received:  | 09/09/25     |    |
| Client Sample ID:  | MW4             |           | SDG No.:        | Q3058        |    |
| Lab Sample ID:     | Q3058-03        |           | Matrix:         | Water        |    |
| Analytical Method: | 8260D           |           | % Solid:        | 0            |    |
| Sample Wt/Vol:     | 5               | Units: mL | Final Vol:      | 5000         | uL |
| Soil Aliquot Vol:  |                 | uL        | Test:           | VOCMS Group1 |    |
| GC Column:         | RXI-624         | ID : 0.25 | Level :         | LOW          |    |
| Prep Method :      |                 |           |                 |              |    |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VN087782.D        | 1         | 09/10/25 15:47 | VN091025      |

| CAS Number | Parameter     | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|---------------|-------|-----------|-----|------------|-------|
| 74-88-4    | Methyl Iodide | 14.6  | J         |     | 4.58       | ug/L  |
| 91-20-3    | Naphthalene   | 1.20  | J         |     | 15.6       | ug/L  |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



# QC SUMMARY

### Surrogate Summary

SDG No.: **Q3058**

Client: **G Environmental**

Analytical Method: **SW8260-Low**

| Lab Sample ID | Client ID    | Parameter             | Spike | Result | Recovery (%) | Qual | Limits (%) |           |
|---------------|--------------|-----------------------|-------|--------|--------------|------|------------|-----------|
|               |              |                       |       |        |              |      | Low        | High      |
| Q3058-03      | MW4          | 1,2-Dichloroethane-d4 | 50    | 51.6   | 103          |      | 70 (74)    | 130 (125) |
|               |              | Dibromofluoromethane  | 50    | 50.3   | 101          |      | 70 (75)    | 130 (124) |
|               |              | Toluene-d8            | 50    | 46.8   | 94           |      | 70 (86)    | 130 (113) |
|               |              | 4-Bromofluorobenzene  | 50    | 47.1   | 94           |      | 70 (77)    | 130 (121) |
| VN0910WBL01   | VN0910WBL01  | 1,2-Dichloroethane-d4 | 50    | 55.0   | 110          |      | 70 (74)    | 130 (125) |
|               |              | Dibromofluoromethane  | 50    | 52.2   | 104          |      | 70 (75)    | 130 (124) |
|               |              | Toluene-d8            | 50    | 47.9   | 96           |      | 70 (86)    | 130 (113) |
|               |              | 4-Bromofluorobenzene  | 50    | 44.2   | 88           |      | 70 (77)    | 130 (121) |
| VN0910WBS01   | VN0910WBS01  | 1,2-Dichloroethane-d4 | 50    | 46.3   | 93           |      | 70 (74)    | 130 (125) |
|               |              | Dibromofluoromethane  | 50    | 49.1   | 98           |      | 70 (75)    | 130 (124) |
|               |              | Toluene-d8            | 50    | 46.1   | 92           |      | 70 (86)    | 130 (113) |
|               |              | 4-Bromofluorobenzene  | 50    | 47.1   | 94           |      | 70 (77)    | 130 (121) |
| VN0910WBSD0   | VN0910WBSD01 | 1,2-Dichloroethane-d4 | 50    | 45.1   | 90           |      | 70 (74)    | 130 (125) |
|               |              | Dibromofluoromethane  | 50    | 49.6   | 99           |      | 70 (75)    | 130 (124) |
|               |              | Toluene-d8            | 50    | 48.0   | 96           |      | 70 (86)    | 130 (113) |
|               |              | 4-Bromofluorobenzene  | 50    | 47.0   | 94           |      | 70 (77)    | 130 (121) |

( ) = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

|          |                        |                    |                   |
|----------|------------------------|--------------------|-------------------|
| SDG No.: | <u>Q3058</u>           | Analytical Method: | <u>SW8260-Low</u> |
| Client:  | <u>G Environmental</u> | Datafile :         | <u>VN087777.D</u> |

| Lab Sample ID | Parameter                      | Spike | Result | Unit | Rec | RPD | Qual | Low     | Limits High | RPD |
|---------------|--------------------------------|-------|--------|------|-----|-----|------|---------|-------------|-----|
| VN0910WBS01   | Dichlorodifluoromethane        | 20    | 19.0   | ug/L | 95  |     |      | 40 (69) | 160 (116)   |     |
|               | Chloromethane                  | 20    | 19.9   | ug/L | 100 |     |      | 40 (65) | 160 (116)   |     |
|               | Vinyl chloride                 | 20    | 19.7   | ug/L | 99  |     |      | 70 (65) | 130 (117)   |     |
|               | Bromomethane                   | 20    | 21.0   | ug/L | 105 |     |      | 40 (58) | 160 (125)   |     |
|               | Chloroethane                   | 20    | 19.4   | ug/L | 97  |     |      | 40 (56) | 160 (128)   |     |
|               | Trichlorofluoromethane         | 20    | 20.2   | ug/L | 101 |     |      | 40 (73) | 160 (115)   |     |
|               | 1,1,2-Trichlorotrifluoroethane | 20    | 19.6   | ug/L | 98  |     |      | 70 (80) | 130 (112)   |     |
|               | 1,1-Dichloroethene             | 20    | 19.5   | ug/L | 98  |     |      | 70 (74) | 130 (110)   |     |
|               | Acetone                        | 100   | 97.1   | ug/L | 97  |     |      | 40 (60) | 160 (125)   |     |
|               | Carbon disulfide               | 20    | 17.6   | ug/L | 88  |     |      | 40 (64) | 160 (112)   |     |
|               | Methyl tert-butyl Ether        | 20    | 19.7   | ug/L | 99  |     |      | 70 (78) | 130 (114)   |     |
|               | Methyl Acetate                 | 20    | 23.8   | ug/L | 119 |     |      | 70 (67) | 130 (125)   |     |
|               | Methylene Chloride             | 20    | 20.0   | ug/L | 100 |     |      | 70 (72) | 130 (114)   |     |
|               | trans-1,2-Dichloroethene       | 20    | 19.0   | ug/L | 95  |     |      | 70 (75) | 130 (108)   |     |
|               | 1,1-Dichloroethane             | 20    | 20.3   | ug/L | 102 |     |      | 70 (78) | 130 (112)   |     |
|               | Cyclohexane                    | 20    | 17.9   | ug/L | 90  |     |      | 70 (75) | 130 (110)   |     |
|               | 2-Butanone                     | 100   | 97.4   | ug/L | 97  |     |      | 40 (65) | 160 (122)   |     |
|               | Carbon Tetrachloride           | 20    | 20.5   | ug/L | 103 |     |      | 70 (77) | 130 (113)   |     |
|               | cis-1,2-Dichloroethene         | 20    | 19.8   | ug/L | 99  |     |      | 70 (77) | 130 (110)   |     |
|               | Bromochloromethane             | 20    | 18.5   | ug/L | 93  |     |      | 70 (70) | 130 (124)   |     |
|               | Chloroform                     | 20    | 20.8   | ug/L | 104 |     |      | 70 (79) | 130 (113)   |     |
|               | 1,1,1-Trichloroethane          | 20    | 20.2   | ug/L | 101 |     |      | 70 (80) | 130 (108)   |     |
|               | Methylcyclohexane              | 20    | 17.5   | ug/L | 88  |     |      | 70 (72) | 130 (115)   |     |
|               | Benzene                        | 20    | 19.8   | ug/L | 99  |     |      | 70 (82) | 130 (109)   |     |
|               | 1,2-Dichloroethane             | 20    | 20.2   | ug/L | 101 |     |      | 70 (80) | 130 (115)   |     |
|               | Trichloroethene                | 20    | 20.0   | ug/L | 100 |     |      | 70 (77) | 130 (113)   |     |
|               | 1,2-Dichloropropane            | 20    | 19.9   | ug/L | 100 |     |      | 70 (83) | 130 (111)   |     |
|               | Dibromomethane                 | 20    | 20.9   | ug/L | 104 |     |      | 70 (82) | 130 (110)   |     |
|               | Bromodichloromethane           | 20    | 21.2   | ug/L | 106 |     |      | 70 (83) | 130 (110)   |     |
|               | 4-Methyl-2-Pentanone           | 100   | 110    | ug/L | 110 |     |      | 40 (74) | 160 (118)   |     |
|               | Toluene                        | 20    | 20.1   | ug/L | 101 |     |      | 70 (82) | 130 (110)   |     |
|               | t-1,3-Dichloropropene          | 20    | 20.9   | ug/L | 104 |     |      | 70 (79) | 130 (110)   |     |
|               | cis-1,3-Dichloropropene        | 20    | 20.3   | ug/L | 102 |     |      | 70 (82) | 130 (110)   |     |
|               | 1,1,2-Trichloroethane          | 20    | 21.6   | ug/L | 108 |     |      | 70 (83) | 130 (112)   |     |
|               | 2-Hexanone                     | 100   | 110    | ug/L | 110 |     |      | 40 (73) | 160 (117)   |     |
|               | 1,2-Dibromoethane              | 20    | 20.9   | ug/L | 104 |     |      | 70 (81) | 130 (110)   |     |
|               | Tetrachloroethene              | 20    | 19.4   | ug/L | 97  |     |      | 70 (67) | 130 (123)   |     |
|               | Chlorobenzene                  | 20    | 20.3   | ug/L | 102 |     |      | 70 (82) | 130 (109)   |     |
|               | Ethyl Benzene                  | 20    | 20.1   | ug/L | 101 |     |      | 70 (83) | 130 (109)   |     |
|               | m/p-Xylenes                    | 40    | 40.3   | ug/L | 101 |     |      | 70 (82) | 130 (110)   |     |
|               | o-Xylene                       | 20    | 20.2   | ug/L | 101 |     |      | 70 (83) | 130 (109)   |     |
|               | Styrene                        | 20    | 21.1   | ug/L | 106 |     |      | 70 (80) | 130 (111)   |     |
|               | Bromoform                      | 20    | 23.2   | ug/L | 116 |     |      | 70 (79) | 130 (109)   |     |
|               | Isopropylbenzene               | 20    | 18.6   | ug/L | 93  |     |      | 70 (83) | 130 (112)   |     |
|               | 1,1,2,2-Tetrachloroethane      | 20    | 20.1   | ug/L | 101 |     |      | 70 (76) | 130 (118)   |     |
|               | 1,3-Dichlorobenzene            | 20    | 20.1   | ug/L | 101 |     |      | 70 (82) | 130 (108)   |     |
|               | 1,4-Dichlorobenzene            | 20    | 19.8   | ug/L | 99  |     |      | 70 (82) | 130 (107)   |     |
|               | 1,2-Dichlorobenzene            | 20    | 20.3   | ug/L | 102 |     |      | 70 (82) | 130 (109)   |     |

() = LABORATORY INHOUSE LIMIT

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

|                 |                               |                           |                          |
|-----------------|-------------------------------|---------------------------|--------------------------|
| <b>SDG No.:</b> | <u><b>Q3058</b></u>           | <b>Analytical Method:</b> | <u><b>SW8260-Low</b></u> |
| <b>Client:</b>  | <u><b>G Environmental</b></u> | <b>Datafile :</b>         | <u><b>VN087777.D</b></u> |

| Lab Sample ID | Parameter                   | Spike | Result | Unit | Rec | RPD | Qual | Low     | Limits<br>High | RPD |
|---------------|-----------------------------|-------|--------|------|-----|-----|------|---------|----------------|-----|
| VN0910WBS01   | 1,2-Dibromo-3-Chloropropane | 20    | 19.6   | ug/L | 98  |     |      | 40 (68) | 160 (112)      |     |
|               | 1,2,4-Trichlorobenzene      | 20    | 19.1   | ug/L | 96  |     |      | 70 (75) | 130 (113)      |     |
|               | 1,2,3-Trichlorobenzene      | 20    | 19.5   | ug/L | 98  |     |      | 70 (76) | 130 (114)      |     |

() = LABORATORY INHOUSE LIMIT

## Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

|          |                        |                    |                   |
|----------|------------------------|--------------------|-------------------|
| SDG No.: | <u>Q3058</u>           | Analytical Method: | <u>SW8260-Low</u> |
| Client:  | <u>G Environmental</u> | Datafile :         | <u>VN087778.D</u> |

| Lab Sample ID | Parameter                      | Spike | Result | Unit | Rec | RPD | Qual | Low     | Limits High | RPD     |
|---------------|--------------------------------|-------|--------|------|-----|-----|------|---------|-------------|---------|
| VN0910WBSD01  | Dichlorodifluoromethane        | 20    | 19.5   | ug/L | 98  | 3   |      | 40 (69) | 160 (116)   | 20 (19) |
|               | Chloromethane                  | 20    | 20.1   | ug/L | 101 | 1   |      | 40 (65) | 160 (116)   | 20 (21) |
|               | Vinyl chloride                 | 20    | 18.9   | ug/L | 95  | 4   |      | 70 (65) | 130 (117)   | 20 (19) |
|               | Bromomethane                   | 20    | 21.3   | ug/L | 106 | 1   |      | 40 (58) | 160 (125)   | 20 (20) |
|               | Chloroethane                   | 20    | 18.7   | ug/L | 94  | 3   |      | 40 (56) | 160 (128)   | 20 (20) |
|               | Trichlorofluoromethane         | 20    | 20.8   | ug/L | 104 | 3   |      | 40 (73) | 160 (115)   | 20 (16) |
|               | 1,1,2-Trichlorotrifluoroethane | 20    | 19.9   | ug/L | 100 | 2   |      | 70 (80) | 130 (112)   | 20 (15) |
|               | 1,1-Dichloroethene             | 20    | 19.2   | ug/L | 96  | 2   |      | 70 (74) | 130 (110)   | 20 (20) |
|               | Acetone                        | 100   | 93.6   | ug/L | 94  | 3   |      | 40 (60) | 160 (125)   | 20 (20) |
|               | Carbon disulfide               | 20    | 17.0   | ug/L | 85  | 3   |      | 40 (64) | 160 (112)   | 20 (20) |
|               | Methyl tert-butyl Ether        | 20    | 19.2   | ug/L | 96  | 3   |      | 70 (78) | 130 (114)   | 20 (20) |
|               | Methyl Acetate                 | 20    | 23.7   | ug/L | 119 | 0   |      | 70 (67) | 130 (125)   | 20 (20) |
|               | Methylene Chloride             | 20    | 19.0   | ug/L | 95  | 5   |      | 70 (72) | 130 (114)   | 20 (20) |
|               | trans-1,2-Dichloroethene       | 20    | 18.4   | ug/L | 92  | 3   |      | 70 (75) | 130 (108)   | 20 (16) |
|               | 1,1-Dichloroethane             | 20    | 19.0   | ug/L | 95  | 7   |      | 70 (78) | 130 (112)   | 20 (20) |
|               | Cyclohexane                    | 20    | 18.0   | ug/L | 90  | 0   |      | 70 (75) | 130 (110)   | 20 (20) |
|               | 2-Butanone                     | 100   | 96.7   | ug/L | 97  | 0   |      | 40 (65) | 160 (122)   | 20 (26) |
|               | Carbon Tetrachloride           | 20    | 21.5   | ug/L | 108 | 5   |      | 70 (77) | 130 (113)   | 20 (15) |
|               | cis-1,2-Dichloroethene         | 20    | 19.1   | ug/L | 96  | 3   |      | 70 (77) | 130 (110)   | 20 (20) |
|               | Bromochloromethane             | 20    | 17.8   | ug/L | 89  | 4   |      | 70 (70) | 130 (124)   | 20 (20) |
|               | Chloroform                     | 20    | 20.3   | ug/L | 102 | 2   |      | 70 (79) | 130 (113)   | 20 (20) |
|               | 1,1,1-Trichloroethane          | 20    | 19.8   | ug/L | 99  | 2   |      | 70 (80) | 130 (108)   | 20 (20) |
|               | Methylcyclohexane              | 20    | 19.0   | ug/L | 95  | 8   |      | 70 (72) | 130 (115)   | 20 (20) |
|               | Benzene                        | 20    | 20.4   | ug/L | 102 | 3   |      | 70 (82) | 130 (109)   | 20 (15) |
|               | 1,2-Dichloroethane             | 20    | 20.6   | ug/L | 103 | 2   |      | 70 (80) | 130 (115)   | 20 (20) |
|               | Trichloroethene                | 20    | 20.4   | ug/L | 102 | 2   |      | 70 (77) | 130 (113)   | 20 (15) |
|               | 1,2-Dichloropropane            | 20    | 19.9   | ug/L | 100 | 0   |      | 70 (83) | 130 (111)   | 20 (16) |
|               | Dibromomethane                 | 20    | 21.2   | ug/L | 106 | 2   |      | 70 (82) | 130 (110)   | 20 (20) |
|               | Bromodichloromethane           | 20    | 21.5   | ug/L | 108 | 2   |      | 70 (83) | 130 (110)   | 20 (16) |
|               | 4-Methyl-2-Pentanone           | 100   | 110    | ug/L | 110 | 0   |      | 40 (74) | 160 (118)   | 20 (25) |
|               | Toluene                        | 20    | 20.4   | ug/L | 102 | 1   |      | 70 (82) | 130 (110)   | 20 (16) |
|               | t-1,3-Dichloropropene          | 20    | 21.3   | ug/L | 106 | 2   |      | 70 (79) | 130 (110)   | 20 (20) |
|               | cis-1,3-Dichloropropene        | 20    | 21.0   | ug/L | 105 | 3   |      | 70 (82) | 130 (110)   | 20 (16) |
|               | 1,1,2-Trichloroethane          | 20    | 22.1   | ug/L | 111 | 3   |      | 70 (83) | 130 (112)   | 20 (20) |
|               | 2-Hexanone                     | 100   | 110    | ug/L | 110 | 0   |      | 40 (73) | 160 (117)   | 20 (25) |
|               | 1,2-Dibromoethane              | 20    | 21.7   | ug/L | 109 | 5   |      | 70 (81) | 130 (110)   | 20 (20) |
|               | Tetrachloroethene              | 20    | 19.8   | ug/L | 99  | 2   |      | 70 (67) | 130 (123)   | 20 (15) |
|               | Chlorobenzene                  | 20    | 20.6   | ug/L | 103 | 1   |      | 70 (82) | 130 (109)   | 20 (15) |
|               | Ethyl Benzene                  | 20    | 20.2   | ug/L | 101 | 0   |      | 70 (83) | 130 (109)   | 20 (16) |
|               | m/p-Xylenes                    | 40    | 42.0   | ug/L | 105 | 4   |      | 70 (82) | 130 (110)   | 20 (15) |
|               | o-Xylene                       | 20    | 19.6   | ug/L | 98  | 3   |      | 70 (83) | 130 (109)   | 20 (20) |
|               | Styrene                        | 20    | 21.0   | ug/L | 105 | 1   |      | 70 (80) | 130 (111)   | 20 (17) |
|               | Bromoform                      | 20    | 22.7   | ug/L | 114 | 2   |      | 70 (79) | 130 (109)   | 20 (20) |
|               | Isopropylbenzene               | 20    | 19.2   | ug/L | 96  | 3   |      | 70 (83) | 130 (112)   | 20 (29) |
|               | 1,1,2,2-Tetrachloroethane      | 20    | 21.3   | ug/L | 106 | 5   |      | 70 (76) | 130 (118)   | 20 (20) |
|               | 1,3-Dichlorobenzene            | 20    | 19.7   | ug/L | 99  | 2   |      | 70 (82) | 130 (108)   | 20 (20) |
|               | 1,4-Dichlorobenzene            | 20    | 19.8   | ug/L | 99  | 0   |      | 70 (82) | 130 (107)   | 20 (15) |
|               | 1,2-Dichlorobenzene            | 20    | 20.2   | ug/L | 101 | 1   |      | 70 (82) | 130 (109)   | 20 (20) |

() = LABORATORY INHOUSE LIMIT

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

|                 |                               |                           |                          |
|-----------------|-------------------------------|---------------------------|--------------------------|
| <b>SDG No.:</b> | <u><b>Q3058</b></u>           | <b>Analytical Method:</b> | <u><b>SW8260-Low</b></u> |
| <b>Client:</b>  | <u><b>G Environmental</b></u> | <b>Datafile :</b>         | <u><b>VN087778.D</b></u> |

| Lab Sample ID | Parameter                   | Spike | Result | Unit | Rec | RPD | Qual | Low     | Limits<br>High | RPD     |
|---------------|-----------------------------|-------|--------|------|-----|-----|------|---------|----------------|---------|
| VN0910WBSD01  | 1,2-Dibromo-3-Chloropropane | 20    | 20.2   | ug/L | 101 | 3   |      | 40 (68) | 160 (112)      | 20 (20) |
|               | 1,2,4-Trichlorobenzene      | 20    | 19.7   | ug/L | 99  | 3   |      | 70 (75) | 130 (113)      | 20 (29) |
|               | 1,2,3-Trichlorobenzene      | 20    | 19.8   | ug/L | 99  | 1   |      | 70 (76) | 130 (114)      | 20 (29) |

() = LABORATORY INHOUSE LIMIT

VOLATILE METHOD BLANK SUMMARY

Client ID

VN0910WBL01

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG NO.: Q3058

Lab File ID: VN087776.D

Lab Sample ID: VN0910WBL01

Date Analyzed: 09/10/2025

Time Analyzed: 12:46

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA\_N

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

| EPA<br>SAMPLE NO. | LAB<br>SAMPLE ID | LAB<br>FILE ID | DATE<br>ANALYZED |
|-------------------|------------------|----------------|------------------|
| VN0910WBS01       | VN0910WBS01      | VN087777.D     | 09/10/2025       |
| VN0910WBSD01      | VN0910WBSD01     | VN087778.D     | 09/10/2025       |
| MW4               | Q3058-03         | VN087782.D     | 09/10/2025       |

COMMENTS: \_\_\_\_\_

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG NO.: Q3058

Lab File ID: VN087614.D

BFB Injection Date: 08/21/2025

Instrument ID: MSVOA\_N

BFB Injection Time: 09:30

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N N

| m/e | ION ABUNDANCE CRITERIA             | % RELATIVE ABUNDANCE |
|-----|------------------------------------|----------------------|
| 50  | 15.0 - 40.0% of mass 95            | 25.8                 |
| 75  | 30.0 - 60.0% of mass 95            | 58.1                 |
| 95  | Base Peak, 100% relative abundance | 100                  |
| 96  | 5.0 - 9.0% of mass 95              | 7.3                  |
| 173 | Less than 2.0% of mass 174         | 1.1 ( 1.6 ) 1        |
| 174 | 50.0 - 100.0% of mass 95           | 69                   |
| 175 | 5.0 - 9.0% of mass 174             | 5.6 ( 8.1 ) 1        |
| 176 | 95.0 - 101.0% of mass 174          | 67.7 ( 98.1 ) 1      |
| 177 | 5.0 - 9.0% of mass 176             | 4.8 ( 7.1 ) 2        |

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| CLIENT ID | LAB SAMPLE ID | LAB FILE ID | DATE ANALYZED | TIME ANALYZED |
|-----------|---------------|-------------|---------------|---------------|
| VSTDIC001 | VSTDIC001     | VN087619.D  | 08/21/2025    | 12:09         |
| VSTDIC005 | VSTDIC005     | VN087620.D  | 08/21/2025    | 13:08         |
| VSTDIC020 | VSTDIC020     | VN087621.D  | 08/21/2025    | 13:41         |
| VSTDIC050 | VSTDIC050     | VN087622.D  | 08/21/2025    | 14:02         |
| VSTDIC100 | VSTDIC100     | VN087623.D  | 08/21/2025    | 14:23         |
| VSTDIC150 | VSTDIC150     | VN087624.D  | 08/21/2025    | 14:44         |

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance  
Lab Code: ACE  
Lab File ID: VN087773.D  
Instrument ID: MSVOA\_N  
GC Column: RXI-624 ID: 0.25 (mm)

Contract: GENV01  
SDG NO.: Q3058  
BFB Injection Date: 09/10/2025  
BFB Injection Time: 11:13  
Heated Purge: Y/N N

| m/e | ION ABUNDANCE CRITERIA             | % RELATIVE ABUNDANCE |
|-----|------------------------------------|----------------------|
| 50  | 15.0 - 40.0% of mass 95            | 22                   |
| 75  | 30.0 - 60.0% of mass 95            | 56.3                 |
| 95  | Base Peak, 100% relative abundance | 100                  |
| 96  | 5.0 - 9.0% of mass 95              | 7.4                  |
| 173 | Less than 2.0% of mass 174         | 0.5 ( 0.7 ) 1        |
| 174 | 50.0 - 100.0% of mass 95           | 72.4                 |
| 175 | 5.0 - 9.0% of mass 174             | 5.9 ( 8.1 ) 1        |
| 176 | 95.0 - 101.0% of mass 174          | 69.2 ( 95.5 ) 1      |
| 177 | 5.0 - 9.0% of mass 176             | 4.9 ( 7.1 ) 2        |

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| CLIENT ID    | LAB SAMPLE ID | LAB FILE ID | DATE ANALYZED | TIME ANALYZED |
|--------------|---------------|-------------|---------------|---------------|
| VSTDCCC050   | VSTDCCC050    | VN087774.D  | 09/10/2025    | 11:47         |
| VN0910WBL01  | VN0910WBL01   | VN087776.D  | 09/10/2025    | 12:46         |
| VN0910WBS01  | VN0910WBS01   | VN087777.D  | 09/10/2025    | 13:29         |
| VN0910WBSD01 | VN0910WBSD01  | VN087778.D  | 09/10/2025    | 13:50         |
| MW4          | Q3058-03      | VN087782.D  | 09/10/2025    | 15:47         |

## VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01  
Lab Code: ACE SDG NO.: Q3058  
Lab File ID: VN087774.D Date Analyzed: 09/10/2025  
Instrument ID: MSVOA\_N Time Analyzed: 11:47  
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #   |
|----------------|---------------|-------|---------------|-------|---------------|--------|
| 12 HOUR STD    | 241352        | 8.21  | 467783        | 9.08  | 458675        | 11.84  |
| UPPER LIMIT    | 482704        | 8.706 | 935566        | 9.582 | 917350        | 12.341 |
| LOWER LIMIT    | 120676        | 7.706 | 233892        | 8.582 | 229338        | 11.341 |
| EPA SAMPLE NO. |               |       |               |       |               |        |
| MW4            | 196528        | 8.21  | 453885        | 9.08  | 434796        | 11.84  |
| VN0910WBL01    | 177881        | 8.21  | 413161        | 9.08  | 392580        | 11.84  |
| VN0910WBS01    | 247948        | 8.21  | 481179        | 9.08  | 443678        | 11.85  |
| VN0910WBSD01   | 258583        | 8.21  | 480224        | 9.08  | 451507        | 11.84  |

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01  
 Lab Code: ACE SDG NO.: Q3058  
 Lab File ID: VN087774.D Date Analyzed: 09/10/2025  
 Instrument ID: MSVOA\_N Time Analyzed: 11:47  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                | IS4<br>AREA # | RT #  |  |  |  |  |
|----------------|---------------|-------|--|--|--|--|
| 12 HOUR STD    | 238791        | 13.77 |  |  |  |  |
| UPPER LIMIT    | 477582        | 14.27 |  |  |  |  |
| LOWER LIMIT    | 119396        | 13.27 |  |  |  |  |
| EPA SAMPLE NO. |               |       |  |  |  |  |
| MW4            | 206220        | 13.77 |  |  |  |  |
| VN0910WBL01    | 179388        | 13.77 |  |  |  |  |
| VN0910WBS01    | 236791        | 13.77 |  |  |  |  |
| VN0910WBSD01   | 236702        | 13.77 |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.



# QC SAMPLE DATA

### Report of Analysis

|                    |                 |           |                 |              |
|--------------------|-----------------|-----------|-----------------|--------------|
| Client:            | G Environmental |           | Date Collected: |              |
| Project:           | Buffington      |           | Date Received:  |              |
| Client Sample ID:  | VN0910WBL01     |           | SDG No.:        | Q3058        |
| Lab Sample ID:     | VN0910WBL01     |           | Matrix:         | Water        |
| Analytical Method: | 8260D           |           | % Solid:        | 0            |
| Sample Wt/Vol:     | 5               | Units: mL | Final Vol:      | 5000 uL      |
| Soil Aliquot Vol:  |                 | uL        | Test:           | VOCMS Group1 |
| GC Column:         | RXI-624         | ID : 0.25 | Level :         | LOW          |
| Prep Method :      |                 |           |                 |              |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VN087776.D        | 1         | 09/10/25 12:46 | VN091025      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL   | LOQ / CRQL | Units |
|----------------|--------------------------------|-------|-----------|-------|------------|-------|
| <b>TARGETS</b> |                                |       |           |       |            |       |
| 75-71-8        | Dichlorodifluoromethane        | 0.22  | U         | 0.22  | 1.00       | ug/L  |
| 74-87-3        | Chloromethane                  | 0.32  | U         | 0.32  | 1.00       | ug/L  |
| 75-01-4        | Vinyl Chloride                 | 0.26  | U         | 0.26  | 1.00       | ug/L  |
| 74-83-9        | Bromomethane                   | 1.40  | U         | 1.40  | 5.00       | ug/L  |
| 75-00-3        | Chloroethane                   | 0.47  | U         | 0.47  | 1.00       | ug/L  |
| 75-69-4        | Trichlorofluoromethane         | 0.33  | U         | 0.33  | 1.00       | ug/L  |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 0.25  | U         | 0.25  | 1.00       | ug/L  |
| 75-35-4        | 1,1-Dichloroethene             | 0.23  | U         | 0.23  | 1.00       | ug/L  |
| 67-64-1        | Acetone                        | 1.50  | U         | 1.50  | 5.00       | ug/L  |
| 75-15-0        | Carbon Disulfide               | 0.21  | U         | 0.21  | 1.00       | ug/L  |
| 1634-04-4      | Methyl tert-butyl Ether        | 0.16  | U         | 0.16  | 1.00       | ug/L  |
| 79-20-9        | Methyl Acetate                 | 0.27  | U         | 0.27  | 1.00       | ug/L  |
| 75-09-2        | Methylene Chloride             | 0.28  | U         | 0.28  | 1.00       | ug/L  |
| 156-60-5       | trans-1,2-Dichloroethene       | 0.23  | U         | 0.23  | 1.00       | ug/L  |
| 75-34-3        | 1,1-Dichloroethane             | 0.23  | U         | 0.23  | 1.00       | ug/L  |
| 110-82-7       | Cyclohexane                    | 1.50  | U         | 1.50  | 5.00       | ug/L  |
| 78-93-3        | 2-Butanone                     | 0.98  | U         | 0.98  | 5.00       | ug/L  |
| 56-23-5        | Carbon Tetrachloride           | 0.25  | U         | 0.25  | 1.00       | ug/L  |
| 156-59-2       | cis-1,2-Dichloroethene         | 0.19  | U         | 0.19  | 1.00       | ug/L  |
| 74-97-5        | Bromochloromethane             | 0.22  | U         | 0.22  | 1.00       | ug/L  |
| 67-66-3        | Chloroform                     | 0.25  | U         | 0.25  | 1.00       | ug/L  |
| 71-55-6        | 1,1,1-Trichloroethane          | 0.20  | U         | 0.20  | 1.00       | ug/L  |
| 108-87-2       | Methylcyclohexane              | 0.16  | U         | 0.16  | 1.00       | ug/L  |
| 71-43-2        | Benzene                        | 0.15  | U         | 0.15  | 1.00       | ug/L  |
| 107-06-2       | 1,2-Dichloroethane             | 0.22  | U         | 0.22  | 1.00       | ug/L  |
| 79-01-6        | Trichloroethene                | 0.090 | U         | 0.090 | 1.00       | ug/L  |
| 78-87-5        | 1,2-Dichloropropane            | 0.20  | U         | 0.20  | 1.00       | ug/L  |
| 74-95-3        | Dibromomethane                 | 0.25  | U         | 0.25  | 1.00       | ug/L  |
| 75-27-4        | Bromodichloromethane           | 0.22  | U         | 0.22  | 1.00       | ug/L  |
| 108-10-1       | 4-Methyl-2-Pentanone           | 0.68  | U         | 0.68  | 5.00       | ug/L  |

### Report of Analysis

|                    |                 |           |                 |              |    |
|--------------------|-----------------|-----------|-----------------|--------------|----|
| Client:            | G Environmental |           | Date Collected: |              |    |
| Project:           | Buffington      |           | Date Received:  |              |    |
| Client Sample ID:  | VN0910WBL01     |           | SDG No.:        | Q3058        |    |
| Lab Sample ID:     | VN0910WBL01     |           | Matrix:         | Water        |    |
| Analytical Method: | 8260D           |           | % Solid:        | 0            |    |
| Sample Wt/Vol:     | 5               | Units: mL | Final Vol:      | 5000         | uL |
| Soil Aliquot Vol:  |                 | uL        | Test:           | VOCMS Group1 |    |
| GC Column:         | RXI-624         | ID : 0.25 | Level :         | LOW          |    |
| Prep Method :      |                 |           |                 |              |    |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VN087776.D        | 1         | 09/10/25 12:46 | VN091025      |

| CAS Number                | Parameter                   | Conc.  | Qualifier | MDL                 | LOQ / CRQL | Units   |
|---------------------------|-----------------------------|--------|-----------|---------------------|------------|---------|
| 108-88-3                  | Toluene                     | 0.14   | U         | 0.14                | 1.00       | ug/L    |
| 10061-02-6                | t-1,3-Dichloropropene       | 0.17   | U         | 0.17                | 1.00       | ug/L    |
| 10061-01-5                | cis-1,3-Dichloropropene     | 0.16   | U         | 0.16                | 1.00       | ug/L    |
| 79-00-5                   | 1,1,2-Trichloroethane       | 0.21   | U         | 0.21                | 1.00       | ug/L    |
| 591-78-6                  | 2-Hexanone                  | 0.89   | U         | 0.89                | 5.00       | ug/L    |
| 106-93-4                  | 1,2-Dibromoethane           | 0.15   | U         | 0.15                | 1.00       | ug/L    |
| 127-18-4                  | Tetrachloroethene           | 0.23   | U         | 0.23                | 1.00       | ug/L    |
| 108-90-7                  | Chlorobenzene               | 0.12   | U         | 0.12                | 1.00       | ug/L    |
| 100-41-4                  | Ethyl Benzene               | 0.13   | U         | 0.13                | 1.00       | ug/L    |
| 179601-23-1               | m/p-Xylenes                 | 0.24   | U         | 0.24                | 2.00       | ug/L    |
| 95-47-6                   | o-Xylene                    | 0.12   | U         | 0.12                | 1.00       | ug/L    |
| 100-42-5                  | Styrene                     | 0.15   | U         | 0.15                | 1.00       | ug/L    |
| 75-25-2                   | Bromoform                   | 0.19   | U         | 0.19                | 1.00       | ug/L    |
| 98-82-8                   | Isopropylbenzene            | 0.12   | U         | 0.12                | 1.00       | ug/L    |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 0.26   | U         | 0.26                | 1.00       | ug/L    |
| 541-73-1                  | 1,3-Dichlorobenzene         | 0.16   | U         | 0.16                | 1.00       | ug/L    |
| 106-46-7                  | 1,4-Dichlorobenzene         | 0.19   | U         | 0.19                | 1.00       | ug/L    |
| 95-50-1                   | 1,2-Dichlorobenzene         | 0.16   | U         | 0.16                | 1.00       | ug/L    |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 0.53   | U         | 0.53                | 1.00       | ug/L    |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 0.20   | U         | 0.20                | 1.00       | ug/L    |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 0.20   | U         | 0.20                | 1.00       | ug/L    |
| <b>SURROGATES</b>         |                             |        |           |                     |            |         |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 55.0   |           | 70 (74) - 130 (125) | 110%       | SPK: 50 |
| 1868-53-7                 | Dibromofluoromethane        | 52.2   |           | 70 (75) - 130 (124) | 104%       | SPK: 50 |
| 2037-26-5                 | Toluene-d8                  | 47.9   |           | 70 (86) - 130 (113) | 96%        | SPK: 50 |
| 460-00-4                  | 4-Bromofluorobenzene        | 44.2   |           | 70 (77) - 130 (121) | 88%        | SPK: 50 |
| <b>INTERNAL STANDARDS</b> |                             |        |           |                     |            |         |
| 363-72-4                  | Pentafluorobenzene          | 178000 | 8.206     |                     |            |         |
| 540-36-3                  | 1,4-Difluorobenzene         | 413000 | 9.082     |                     |            |         |
| 3114-55-4                 | Chlorobenzene-d5            | 393000 | 11.841    |                     |            |         |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 179000 | 13.77     |                     |            |         |



### Report of Analysis

|                    |                 |           |                 |              |
|--------------------|-----------------|-----------|-----------------|--------------|
| Client:            | G Environmental |           | Date Collected: |              |
| Project:           | Buffington      |           | Date Received:  |              |
| Client Sample ID:  | VN0910WBS01     |           | SDG No.:        | Q3058        |
| Lab Sample ID:     | VN0910WBS01     |           | Matrix:         | Water        |
| Analytical Method: | 8260D           |           | % Solid:        | 0            |
| Sample Wt/Vol:     | 5               | Units: mL | Final Vol:      | 5000 uL      |
| Soil Aliquot Vol:  |                 | uL        | Test:           | VOCMS Group1 |
| GC Column:         | RXI-624         | ID : 0.25 | Level :         | LOW          |
| Prep Method :      |                 |           |                 |              |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VN087777.D        | 1         | 09/10/25 13:29 | VN091025      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL   | LOQ / CRQL | Units |
|----------------|--------------------------------|-------|-----------|-------|------------|-------|
| <b>TARGETS</b> |                                |       |           |       |            |       |
| 75-71-8        | Dichlorodifluoromethane        | 19.0  |           | 0.22  | 1.00       | ug/L  |
| 74-87-3        | Chloromethane                  | 19.9  |           | 0.32  | 1.00       | ug/L  |
| 75-01-4        | Vinyl Chloride                 | 19.7  |           | 0.26  | 1.00       | ug/L  |
| 74-83-9        | Bromomethane                   | 21.0  |           | 1.40  | 5.00       | ug/L  |
| 75-00-3        | Chloroethane                   | 19.4  |           | 0.47  | 1.00       | ug/L  |
| 75-69-4        | Trichlorofluoromethane         | 20.2  |           | 0.33  | 1.00       | ug/L  |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 19.6  |           | 0.25  | 1.00       | ug/L  |
| 75-35-4        | 1,1-Dichloroethene             | 19.5  |           | 0.23  | 1.00       | ug/L  |
| 67-64-1        | Acetone                        | 97.1  |           | 1.50  | 5.00       | ug/L  |
| 75-15-0        | Carbon Disulfide               | 17.6  |           | 0.21  | 1.00       | ug/L  |
| 1634-04-4      | Methyl tert-butyl Ether        | 19.7  |           | 0.16  | 1.00       | ug/L  |
| 79-20-9        | Methyl Acetate                 | 23.8  |           | 0.27  | 1.00       | ug/L  |
| 75-09-2        | Methylene Chloride             | 20.0  |           | 0.28  | 1.00       | ug/L  |
| 156-60-5       | trans-1,2-Dichloroethene       | 19.0  |           | 0.23  | 1.00       | ug/L  |
| 75-34-3        | 1,1-Dichloroethane             | 20.3  |           | 0.23  | 1.00       | ug/L  |
| 110-82-7       | Cyclohexane                    | 17.9  |           | 1.50  | 5.00       | ug/L  |
| 78-93-3        | 2-Butanone                     | 97.4  |           | 0.98  | 5.00       | ug/L  |
| 56-23-5        | Carbon Tetrachloride           | 20.5  |           | 0.25  | 1.00       | ug/L  |
| 156-59-2       | cis-1,2-Dichloroethene         | 19.8  |           | 0.19  | 1.00       | ug/L  |
| 74-97-5        | Bromochloromethane             | 18.5  |           | 0.22  | 1.00       | ug/L  |
| 67-66-3        | Chloroform                     | 20.8  |           | 0.25  | 1.00       | ug/L  |
| 71-55-6        | 1,1,1-Trichloroethane          | 20.2  |           | 0.20  | 1.00       | ug/L  |
| 108-87-2       | Methylcyclohexane              | 17.5  |           | 0.16  | 1.00       | ug/L  |
| 71-43-2        | Benzene                        | 19.8  |           | 0.15  | 1.00       | ug/L  |
| 107-06-2       | 1,2-Dichloroethane             | 20.2  |           | 0.22  | 1.00       | ug/L  |
| 79-01-6        | Trichloroethene                | 20.0  |           | 0.090 | 1.00       | ug/L  |
| 78-87-5        | 1,2-Dichloropropane            | 19.9  |           | 0.20  | 1.00       | ug/L  |
| 74-95-3        | Dibromomethane                 | 20.9  |           | 0.25  | 1.00       | ug/L  |
| 75-27-4        | Bromodichloromethane           | 21.2  |           | 0.22  | 1.00       | ug/L  |
| 108-10-1       | 4-Methyl-2-Pentanone           | 110   |           | 0.68  | 5.00       | ug/L  |

### Report of Analysis

|                    |                 |           |                 |              |
|--------------------|-----------------|-----------|-----------------|--------------|
| Client:            | G Environmental |           | Date Collected: |              |
| Project:           | Buffington      |           | Date Received:  |              |
| Client Sample ID:  | VN0910WBS01     |           | SDG No.:        | Q3058        |
| Lab Sample ID:     | VN0910WBS01     |           | Matrix:         | Water        |
| Analytical Method: | 8260D           |           | % Solid:        | 0            |
| Sample Wt/Vol:     | 5               | Units: mL | Final Vol:      | 5000 uL      |
| Soil Aliquot Vol:  |                 | uL        | Test:           | VOCMS Group1 |
| GC Column:         | RXI-624         | ID : 0.25 | Level :         | LOW          |
| Prep Method :      |                 |           |                 |              |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VN087777.D        | 1         | 09/10/25 13:29 | VN091025      |

| CAS Number                | Parameter                   | Conc.  | Qualifier | MDL                 | LOQ / CRQL | Units   |
|---------------------------|-----------------------------|--------|-----------|---------------------|------------|---------|
| 108-88-3                  | Toluene                     | 20.1   |           | 0.14                | 1.00       | ug/L    |
| 10061-02-6                | t-1,3-Dichloropropene       | 20.9   |           | 0.17                | 1.00       | ug/L    |
| 10061-01-5                | cis-1,3-Dichloropropene     | 20.3   |           | 0.16                | 1.00       | ug/L    |
| 79-00-5                   | 1,1,2-Trichloroethane       | 21.6   |           | 0.21                | 1.00       | ug/L    |
| 591-78-6                  | 2-Hexanone                  | 110    |           | 0.89                | 5.00       | ug/L    |
| 106-93-4                  | 1,2-Dibromoethane           | 20.9   |           | 0.15                | 1.00       | ug/L    |
| 127-18-4                  | Tetrachloroethene           | 19.4   |           | 0.23                | 1.00       | ug/L    |
| 108-90-7                  | Chlorobenzene               | 20.3   |           | 0.12                | 1.00       | ug/L    |
| 100-41-4                  | Ethyl Benzene               | 20.1   |           | 0.13                | 1.00       | ug/L    |
| 179601-23-1               | m/p-Xylenes                 | 40.3   |           | 0.24                | 2.00       | ug/L    |
| 95-47-6                   | o-Xylene                    | 20.2   |           | 0.12                | 1.00       | ug/L    |
| 100-42-5                  | Styrene                     | 21.1   |           | 0.15                | 1.00       | ug/L    |
| 75-25-2                   | Bromoform                   | 23.2   |           | 0.19                | 1.00       | ug/L    |
| 98-82-8                   | Isopropylbenzene            | 18.6   |           | 0.12                | 1.00       | ug/L    |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 20.1   |           | 0.26                | 1.00       | ug/L    |
| 541-73-1                  | 1,3-Dichlorobenzene         | 20.1   |           | 0.16                | 1.00       | ug/L    |
| 106-46-7                  | 1,4-Dichlorobenzene         | 19.8   |           | 0.19                | 1.00       | ug/L    |
| 95-50-1                   | 1,2-Dichlorobenzene         | 20.3   |           | 0.16                | 1.00       | ug/L    |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 19.6   |           | 0.53                | 1.00       | ug/L    |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 19.1   |           | 0.20                | 1.00       | ug/L    |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 19.5   |           | 0.20                | 1.00       | ug/L    |
| <b>SURROGATES</b>         |                             |        |           |                     |            |         |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 46.3   |           | 70 (74) - 130 (125) | 93%        | SPK: 50 |
| 1868-53-7                 | Dibromofluoromethane        | 49.1   |           | 70 (75) - 130 (124) | 98%        | SPK: 50 |
| 2037-26-5                 | Toluene-d8                  | 46.1   |           | 70 (86) - 130 (113) | 92%        | SPK: 50 |
| 460-00-4                  | 4-Bromofluorobenzene        | 47.1   |           | 70 (77) - 130 (121) | 94%        | SPK: 50 |
| <b>INTERNAL STANDARDS</b> |                             |        |           |                     |            |         |
| 363-72-4                  | Pentafluorobenzene          | 248000 | 8.206     |                     |            |         |
| 540-36-3                  | 1,4-Difluorobenzene         | 481000 | 9.083     |                     |            |         |
| 3114-55-4                 | Chlorobenzene-d5            | 444000 | 11.847    |                     |            |         |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 237000 | 13.77     |                     |            |         |

## Report of Analysis

|                    |                                           |                 |                              |
|--------------------|-------------------------------------------|-----------------|------------------------------|
| Client:            | G Environmental                           | Date Collected: |                              |
| Project:           | Buffington                                | Date Received:  |                              |
| Client Sample ID:  | VN0910WBS01                               | SDG No.:        | Q3058                        |
| Lab Sample ID:     | VN0910WBS01                               | Matrix:         | Water                        |
| Analytical Method: | 8260D                                     | % Solid:        | 0                            |
| Sample Wt/Vol:     | 5                      Units:    mL       | Final Vol:      | 5000                      uL |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOCMS Group1                 |
| GC Column:         | RXI-624                      ID :    0.25 | Level :         | LOW                          |
| Prep Method :      |                                           |                 |                              |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VN087777.D        | 1         | 09/10/25 13:29 | VN091025      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

### Report of Analysis

|                    |                 |           |                 |              |
|--------------------|-----------------|-----------|-----------------|--------------|
| Client:            | G Environmental |           | Date Collected: |              |
| Project:           | Buffington      |           | Date Received:  |              |
| Client Sample ID:  | VN0910WBSD01    |           | SDG No.:        | Q3058        |
| Lab Sample ID:     | VN0910WBSD01    |           | Matrix:         | Water        |
| Analytical Method: | 8260D           |           | % Solid:        | 0            |
| Sample Wt/Vol:     | 5               | Units: mL | Final Vol:      | 5000 uL      |
| Soil Aliquot Vol:  |                 | uL        | Test:           | VOCMS Group1 |
| GC Column:         | RXI-624         | ID : 0.25 | Level :         | LOW          |
| Prep Method :      |                 |           |                 |              |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VN087778.D        | 1         | 09/10/25 13:50 | VN091025      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL   | LOQ / CRQL | Units |
|----------------|--------------------------------|-------|-----------|-------|------------|-------|
| <b>TARGETS</b> |                                |       |           |       |            |       |
| 75-71-8        | Dichlorodifluoromethane        | 19.5  |           | 0.22  | 1.00       | ug/L  |
| 74-87-3        | Chloromethane                  | 20.1  |           | 0.32  | 1.00       | ug/L  |
| 75-01-4        | Vinyl Chloride                 | 18.9  |           | 0.26  | 1.00       | ug/L  |
| 74-83-9        | Bromomethane                   | 21.3  |           | 1.40  | 5.00       | ug/L  |
| 75-00-3        | Chloroethane                   | 18.7  |           | 0.47  | 1.00       | ug/L  |
| 75-69-4        | Trichlorofluoromethane         | 20.8  |           | 0.33  | 1.00       | ug/L  |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 19.9  |           | 0.25  | 1.00       | ug/L  |
| 75-35-4        | 1,1-Dichloroethene             | 19.2  |           | 0.23  | 1.00       | ug/L  |
| 67-64-1        | Acetone                        | 93.6  |           | 1.50  | 5.00       | ug/L  |
| 75-15-0        | Carbon Disulfide               | 17.0  |           | 0.21  | 1.00       | ug/L  |
| 1634-04-4      | Methyl tert-butyl Ether        | 19.2  |           | 0.16  | 1.00       | ug/L  |
| 79-20-9        | Methyl Acetate                 | 23.7  |           | 0.27  | 1.00       | ug/L  |
| 75-09-2        | Methylene Chloride             | 19.0  |           | 0.28  | 1.00       | ug/L  |
| 156-60-5       | trans-1,2-Dichloroethene       | 18.4  |           | 0.23  | 1.00       | ug/L  |
| 75-34-3        | 1,1-Dichloroethane             | 19.0  |           | 0.23  | 1.00       | ug/L  |
| 110-82-7       | Cyclohexane                    | 18.0  |           | 1.50  | 5.00       | ug/L  |
| 78-93-3        | 2-Butanone                     | 96.7  |           | 0.98  | 5.00       | ug/L  |
| 56-23-5        | Carbon Tetrachloride           | 21.5  |           | 0.25  | 1.00       | ug/L  |
| 156-59-2       | cis-1,2-Dichloroethene         | 19.1  |           | 0.19  | 1.00       | ug/L  |
| 74-97-5        | Bromochloromethane             | 17.8  |           | 0.22  | 1.00       | ug/L  |
| 67-66-3        | Chloroform                     | 20.3  |           | 0.25  | 1.00       | ug/L  |
| 71-55-6        | 1,1,1-Trichloroethane          | 19.8  |           | 0.20  | 1.00       | ug/L  |
| 108-87-2       | Methylcyclohexane              | 19.0  |           | 0.16  | 1.00       | ug/L  |
| 71-43-2        | Benzene                        | 20.4  |           | 0.15  | 1.00       | ug/L  |
| 107-06-2       | 1,2-Dichloroethane             | 20.6  |           | 0.22  | 1.00       | ug/L  |
| 79-01-6        | Trichloroethene                | 20.4  |           | 0.090 | 1.00       | ug/L  |
| 78-87-5        | 1,2-Dichloropropane            | 19.9  |           | 0.20  | 1.00       | ug/L  |
| 74-95-3        | Dibromomethane                 | 21.2  |           | 0.25  | 1.00       | ug/L  |
| 75-27-4        | Bromodichloromethane           | 21.5  |           | 0.22  | 1.00       | ug/L  |
| 108-10-1       | 4-Methyl-2-Pentanone           | 110   |           | 0.68  | 5.00       | ug/L  |

### Report of Analysis

|                    |                 |           |                 |              |    |
|--------------------|-----------------|-----------|-----------------|--------------|----|
| Client:            | G Environmental |           | Date Collected: |              |    |
| Project:           | Buffington      |           | Date Received:  |              |    |
| Client Sample ID:  | VN0910WBSD01    |           | SDG No.:        | Q3058        |    |
| Lab Sample ID:     | VN0910WBSD01    |           | Matrix:         | Water        |    |
| Analytical Method: | 8260D           |           | % Solid:        | 0            |    |
| Sample Wt/Vol:     | 5               | Units: mL | Final Vol:      | 5000         | uL |
| Soil Aliquot Vol:  |                 | uL        | Test:           | VOCMS Group1 |    |
| GC Column:         | RXI-624         | ID : 0.25 | Level :         | LOW          |    |
| Prep Method :      |                 |           |                 |              |    |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VN087778.D        | 1         | 09/10/25 13:50 | VN091025      |

| CAS Number                | Parameter                   | Conc.  | Qualifier | MDL                 | LOQ / CRQL | Units   |
|---------------------------|-----------------------------|--------|-----------|---------------------|------------|---------|
| 108-88-3                  | Toluene                     | 20.4   |           | 0.14                | 1.00       | ug/L    |
| 10061-02-6                | t-1,3-Dichloropropene       | 21.3   |           | 0.17                | 1.00       | ug/L    |
| 10061-01-5                | cis-1,3-Dichloropropene     | 21.0   |           | 0.16                | 1.00       | ug/L    |
| 79-00-5                   | 1,1,2-Trichloroethane       | 22.1   |           | 0.21                | 1.00       | ug/L    |
| 591-78-6                  | 2-Hexanone                  | 110    |           | 0.89                | 5.00       | ug/L    |
| 106-93-4                  | 1,2-Dibromoethane           | 21.7   |           | 0.15                | 1.00       | ug/L    |
| 127-18-4                  | Tetrachloroethene           | 19.8   |           | 0.23                | 1.00       | ug/L    |
| 108-90-7                  | Chlorobenzene               | 20.6   |           | 0.12                | 1.00       | ug/L    |
| 100-41-4                  | Ethyl Benzene               | 20.2   |           | 0.13                | 1.00       | ug/L    |
| 179601-23-1               | m/p-Xylenes                 | 42.0   |           | 0.24                | 2.00       | ug/L    |
| 95-47-6                   | o-Xylene                    | 19.6   |           | 0.12                | 1.00       | ug/L    |
| 100-42-5                  | Styrene                     | 21.0   |           | 0.15                | 1.00       | ug/L    |
| 75-25-2                   | Bromoform                   | 22.7   |           | 0.19                | 1.00       | ug/L    |
| 98-82-8                   | Isopropylbenzene            | 19.2   |           | 0.12                | 1.00       | ug/L    |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 21.3   |           | 0.26                | 1.00       | ug/L    |
| 541-73-1                  | 1,3-Dichlorobenzene         | 19.7   |           | 0.16                | 1.00       | ug/L    |
| 106-46-7                  | 1,4-Dichlorobenzene         | 19.8   |           | 0.19                | 1.00       | ug/L    |
| 95-50-1                   | 1,2-Dichlorobenzene         | 20.2   |           | 0.16                | 1.00       | ug/L    |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 20.2   |           | 0.53                | 1.00       | ug/L    |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 19.7   |           | 0.20                | 1.00       | ug/L    |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 19.8   |           | 0.20                | 1.00       | ug/L    |
| <b>SURROGATES</b>         |                             |        |           |                     |            |         |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 45.1   |           | 70 (74) - 130 (125) | 90%        | SPK: 50 |
| 1868-53-7                 | Dibromofluoromethane        | 49.6   |           | 70 (75) - 130 (124) | 99%        | SPK: 50 |
| 2037-26-5                 | Toluene-d8                  | 48.0   |           | 70 (86) - 130 (113) | 96%        | SPK: 50 |
| 460-00-4                  | 4-Bromofluorobenzene        | 47.0   |           | 70 (77) - 130 (121) | 94%        | SPK: 50 |
| <b>INTERNAL STANDARDS</b> |                             |        |           |                     |            |         |
| 363-72-4                  | Pentafluorobenzene          | 259000 | 8.206     |                     |            |         |
| 540-36-3                  | 1,4-Difluorobenzene         | 480000 | 9.083     |                     |            |         |
| 3114-55-4                 | Chlorobenzene-d5            | 452000 | 11.841    |                     |            |         |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 237000 | 13.771    |                     |            |         |

## Report of Analysis

|                    |                 |           |                 |              |
|--------------------|-----------------|-----------|-----------------|--------------|
| Client:            | G Environmental |           | Date Collected: |              |
| Project:           | Buffington      |           | Date Received:  |              |
| Client Sample ID:  | VN0910WBSD01    |           | SDG No.:        | Q3058        |
| Lab Sample ID:     | VN0910WBSD01    |           | Matrix:         | Water        |
| Analytical Method: | 8260D           |           | % Solid:        | 0            |
| Sample Wt/Vol:     | 5               | Units: mL | Final Vol:      | 5000 uL      |
| Soil Aliquot Vol:  |                 | uL        | Test:           | VOCMS Group1 |
| GC Column:         | RXI-624         | ID : 0.25 | Level :         | LOW          |
| Prep Method :      |                 |           |                 |              |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VN087778.D        | 1         | 09/10/25 13:50 | VN091025      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected  
 LOQ = Limit of Quantitation  
 MDL = Method Detection Limit  
 LOD = Limit of Detection  
 E = Value Exceeds Calibration Range  
 Q = indicates LCS control criteria did not meet requirements  
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value  
 B = Analyte Found in Associated Method Blank  
 N = Presumptive Evidence of a Compound  
 \* = Values outside of QC limits  
 D = Dilution  
 () = Laboratory InHouse Limit  
 A = Aldol-Condensation Reaction Products



# CALIBRATION SUMMARY

# VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG No.: Q3058

Instrument ID: MSVOA\_N

Calibration Date(s): 08/21/2025 08/21/2025

Heated Purge: (Y/N) N

Calibration Time(s): 12:09 14:44

GC Column: RXI-624 ID: 0.25 (mm)

| LAB FILE ID:                   |        |        |                     |        |        |                     |       |       |
|--------------------------------|--------|--------|---------------------|--------|--------|---------------------|-------|-------|
| RRF001 = VN087619.D            |        |        | RRF005 = VN087620.D |        |        | RRF020 = VN087621.D |       |       |
| RRF050 = VN087622.D            |        |        | RRF100 = VN087623.D |        |        | RRF150 = VN087624.D |       |       |
| COMPOUND                       | RRF001 | RRF005 | RRF020              | RRF050 | RRF100 | RRF150              | RRF   | % RSD |
| Dichlorodifluoromethane        | 0.707  | 0.579  | 0.606               | 0.810  | 0.777  | 0.782               | 0.710 | 13.7  |
| Chloromethane                  | 0.739  | 0.637  | 0.710               | 0.795  | 0.752  | 0.747               | 0.730 | 7.3   |
| Vinyl Chloride                 | 0.909  | 0.739  | 0.832               | 0.903  | 0.847  | 0.846               | 0.846 | 7.3   |
| Bromomethane                   |        | 0.388  | 0.393               | 0.448  | 0.466  | 0.488               | 0.437 | 10.2  |
| Chloroethane                   | 0.727  | 0.528  | 0.599               | 0.603  | 0.558  | 0.554               | 0.595 | 11.9  |
| Trichlorofluoromethane         | 1.323  | 1.139  | 1.242               | 1.274  | 1.227  | 1.233               | 1.240 | 4.9   |
| 1,1,2-Trichlorotrifluoroethane | 0.879  | 0.648  | 0.677               | 0.695  | 0.654  | 0.656               | 0.701 | 12.6  |
| 1,1-Dichloroethene             | 0.892  | 0.649  | 0.730               | 0.700  | 0.667  | 0.687               | 0.721 | 12.3  |
| Acetone                        | 0.633  | 0.634  | 0.584               | 0.522  | 0.484  | 0.489               | 0.558 | 12.3  |
| Carbon Disulfide               | 2.272  | 1.746  | 1.970               | 2.003  | 1.937  | 1.979               | 1.985 | 8.5   |
| Methyl tert-butyl Ether        | 2.859  | 2.508  | 2.740               | 2.733  | 2.643  | 2.742               | 2.704 | 4.4   |
| Methyl Acetate                 | 1.450  | 1.043  | 1.097               | 1.157  | 1.117  | 1.141               | 1.168 | 12.3  |
| Methylene Chloride             | 1.494  | 0.916  | 0.875               | 0.823  | 0.779  | 0.799               | 0.948 | 28.8  |
| trans-1,2-Dichloroethene       | 0.952  | 0.734  | 0.768               | 0.750  | 0.737  | 0.745               | 0.781 | 10.8  |
| 1,1-Dichloroethane             | 1.798  | 1.500  | 1.549               | 1.501  | 1.454  | 1.471               | 1.546 | 8.3   |
| Cyclohexane                    |        | 1.490  | 1.310               | 1.234  | 1.201  | 1.207               | 1.289 | 9.4   |
| 2-Butanone                     | 0.783  | 0.722  | 0.757               | 0.732  | 0.700  | 0.707               | 0.734 | 4.3   |
| Carbon Tetrachloride           | 0.571  | 0.509  | 0.568               | 0.542  | 0.549  | 0.542               | 0.547 | 4.1   |
| cis-1,2-Dichloroethene         | 0.997  | 0.916  | 0.937               | 0.941  | 0.908  | 0.905               | 0.934 | 3.7   |
| Bromochloromethane             | 0.800  | 0.781  | 0.722               | 0.701  | 0.748  | 0.731               | 0.747 | 5     |
| Chloroform                     | 1.677  | 1.535  | 1.629               | 1.579  | 1.519  | 1.528               | 1.578 | 4     |
| 1,1,1-Trichloroethane          | 1.510  | 1.297  | 1.359               | 1.303  | 1.265  | 1.288               | 1.337 | 6.8   |
| Methylcyclohexane              | 0.686  | 0.580  | 0.610               | 0.590  | 0.597  | 0.596               | 0.610 | 6.3   |
| Benzene                        | 1.853  | 1.637  | 1.727               | 1.684  | 1.639  | 1.652               | 1.699 | 4.9   |
| 1,2-Dichloroethane             | 0.714  | 0.627  | 0.675               | 0.650  | 0.624  | 0.626               | 0.653 | 5.5   |
| Trichloroethene                | 0.435  | 0.387  | 0.390               | 0.377  | 0.367  | 0.370               | 0.388 | 6.5   |
| 1,2-Dichloropropane            | 0.544  | 0.435  | 0.447               | 0.433  | 0.415  | 0.412               | 0.448 | 11    |
| Dibromomethane                 | 0.351  | 0.294  | 0.327               | 0.318  | 0.310  | 0.312               | 0.319 | 6.1   |
| Bromodichloromethane           | 0.686  | 0.649  | 0.667               | 0.638  | 0.638  | 0.639               | 0.653 | 3     |
| 4-Methyl-2-Pentanone           | 0.701  | 0.675  | 0.744               | 0.736  | 0.718  | 0.704               | 0.713 | 3.5   |

\* Compounds with required minimum RRF and maximum %RSD values.  
All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

# VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG No.: Q3058

Instrument ID: MSVOA\_N

Calibration Date(s): 08/21/2025 08/21/2025

Heated Purge: (Y/N) N

Calibration Time(s): 12:09 14:44

GC Column: RXI-624 ID: 0.25 (mm)

| LAB FILE ID:                |        | RRF001 = VN087619.D |        | RRF005 = VN087620.D |        | RRF020 = VN087621.D |       |       |  |
|-----------------------------|--------|---------------------|--------|---------------------|--------|---------------------|-------|-------|--|
|                             |        | RRF050 = VN087622.D |        | RRF100 = VN087623.D |        | RRF150 = VN087624.D |       |       |  |
| COMPOUND                    | RRF001 | RRF005              | RRF020 | RRF050              | RRF100 | RRF150              | RRF   | % RSD |  |
| Toluene                     | 1.148  | 0.987               | 1.062  | 1.050               | 1.037  | 1.033               | 1.053 | 5     |  |
| t-1,3-Dichloropropene       | 0.620  | 0.628               | 0.691  | 0.703               | 0.699  | 0.715               | 0.676 | 6.1   |  |
| cis-1,3-Dichloropropene     | 0.717  | 0.640               | 0.712  | 0.718               | 0.714  | 0.710               | 0.702 | 4.3   |  |
| 1,1,2-Trichloroethane       | 0.439  | 0.383               | 0.417  | 0.408               | 0.400  | 0.397               | 0.407 | 4.7   |  |
| 2-Hexanone                  | 0.295  | 0.390               | 0.493  | 0.510               | 0.508  | 0.507               | 0.451 | 19.8  |  |
| 1,2-Dibromoethane           | 0.412  | 0.433               | 0.444  | 0.432               | 0.429  | 0.427               | 0.430 | 2.4   |  |
| Tetrachloroethene           | 0.447  | 0.353               | 0.340  | 0.313               | 0.315  | 0.313               | 0.347 | 14.9  |  |
| Chlorobenzene               | 1.408  | 1.164               | 1.253  | 1.210               | 1.216  | 1.212               | 1.244 | 6.9   |  |
| Ethyl Benzene               | 2.232  | 1.993               | 2.224  | 2.144               | 2.152  | 2.178               | 2.154 | 4     |  |
| m/p-Xylenes                 | 0.759  | 0.735               | 0.811  | 0.812               | 0.808  | 0.814               | 0.790 | 4.3   |  |
| o-Xylene                    | 0.777  | 0.707               | 0.794  | 0.796               | 0.788  | 0.784               | 0.774 | 4.3   |  |
| Styrene                     | 1.142  | 1.086               | 1.391  | 1.382               | 1.386  | 1.393               | 1.297 | 11    |  |
| Bromoform                   | 0.303  | 0.285               | 0.320  | 0.320               | 0.336  | 0.336               | 0.317 | 6.3   |  |
| Isopropylbenzene            | 4.328  | 3.674               | 4.096  | 4.026               | 3.998  | 4.120               | 4.040 | 5.3   |  |
| 1,1,2,2-Tetrachloroethane   | 1.488  | 1.381               | 1.421  | 1.363               | 1.322  | 1.371               | 1.391 | 4.1   |  |
| 1,3-Dichlorobenzene         | 2.087  | 1.770               | 1.940  | 1.827               | 1.806  | 1.826               | 1.876 | 6.3   |  |
| 1,4-Dichlorobenzene         | 2.348  | 1.884               | 1.956  | 1.867               | 1.818  | 1.841               | 1.952 | 10.2  |  |
| 1,2-Dichlorobenzene         | 1.942  | 1.662               | 1.853  | 1.755               | 1.723  | 1.743               | 1.780 | 5.7   |  |
| 1,2-Dibromo-3-Chloropropane | 0.370  | 0.323               | 0.329  | 0.317               | 0.313  | 0.330               | 0.330 | 6.2   |  |
| 1,2,4-Trichlorobenzene      | 1.192  | 0.995               | 1.081  | 1.026               | 1.040  | 1.079               | 1.069 | 6.4   |  |
| 1,2,3-Trichlorobenzene      | 1.140  | 0.972               | 1.059  | 1.006               | 1.019  | 1.072               | 1.045 | 5.7   |  |
| 1,2-Dichloroethane-d4       |        | 1.025               | 1.016  | 0.959               | 1.035  | 1.089               | 1.024 | 4.5   |  |
| Dibromofluoromethane        |        | 0.369               | 0.344  | 0.334               | 0.373  | 0.386               | 0.361 | 6     |  |
| Toluene-d8                  |        | 1.384               | 1.266  | 1.223               | 1.408  | 1.475               | 1.351 | 7.7   |  |
| 4-Bromofluorobenzene        |        | 0.446               | 0.454  | 0.450               | 0.514  | 0.540               | 0.481 | 9     |  |

\* Compounds with required minimum RRF and maximum %RSD values.  
All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

|                     |                   |                        |                              |
|---------------------|-------------------|------------------------|------------------------------|
| Lab Name:           | <u>Alliance</u>   | Contract:              | <u>GENV01</u>                |
| Lab Code:           | <u>ACE</u>        | SDG No.:               | <u>Q3058</u>                 |
| Instrument ID:      | <u>MSVOA_N</u>    | Calibration Date/Time: | <u>09/10/2025 11:47</u>      |
| Lab File ID:        | <u>VN087774.D</u> | Init. Calib. Date(s):  | <u>08/21/2025 08/21/2025</u> |
| Heated Purge: (Y/N) | <u>N</u>          | Init. Calib. Time(s):  | <u>12:09 14:44</u>           |
| GC Column:          | <u>RXI-624</u>    | ID:                    | <u>0.25</u> (mm)             |

| COMPOUND                       | RRF   | RRF050 | MIN RRF | %D     | MAX%D |
|--------------------------------|-------|--------|---------|--------|-------|
| Dichlorodifluoromethane        | 0.710 | 0.610  |         | -14.09 | 20    |
| Chloromethane                  | 0.730 | 0.651  | 0.1     | -10.82 | 20    |
| Vinyl Chloride                 | 0.846 | 0.733  |         | -13.36 | 20    |
| Bromomethane                   | 0.437 | 0.399  |         | -8.7   | 20    |
| Chloroethane                   | 0.595 | 0.515  |         | -13.44 | 20    |
| Trichlorofluoromethane         | 1.240 | 1.154  |         | -6.93  | 20    |
| 1,1,2-Trichlorotrifluoroethane | 0.701 | 0.632  |         | -9.84  | 20    |
| 1,1-Dichloroethene             | 0.721 | 0.641  |         | -11.1  | 20    |
| Acetone                        | 0.558 | 0.461  |         | -17.38 | 20    |
| Carbon Disulfide               | 1.985 | 1.593  |         | -19.75 | 20    |
| Methyl tert-butyl Ether        | 2.704 | 2.578  |         | -4.66  | 20    |
| Methyl Acetate                 | 1.168 | 1.280  |         | 9.59   | 20    |
| Methylene Chloride             | 0.948 | 0.752  |         | -20.67 | 20    |
| trans-1,2-Dichloroethene       | 0.781 | 0.691  |         | -11.52 | 20    |
| 1,1-Dichloroethane             | 1.546 | 1.426  | 0.1     | -7.76  | 20    |
| Cyclohexane                    | 1.289 | 1.032  |         | -19.94 | 20    |
| 2-Butanone                     | 0.734 | 0.668  |         | -8.99  | 20    |
| Carbon Tetrachloride           | 0.547 | 0.536  |         | -2.01  | 20    |
| cis-1,2-Dichloroethene         | 0.934 | 0.870  |         | -6.85  | 20    |
| Bromochloromethane             | 0.747 | 0.716  |         | -4.15  | 20    |
| Chloroform                     | 1.578 | 1.527  |         | -3.23  | 20    |
| 1,1,1-Trichloroethane          | 1.337 | 1.246  |         | -6.81  | 20    |
| Methylcyclohexane              | 0.610 | 0.519  |         | -14.92 | 20    |
| Benzene                        | 1.699 | 1.610  |         | -5.24  | 20    |
| 1,2-Dichloroethane             | 0.653 | 0.618  |         | -5.36  | 20    |
| Trichloroethene                | 0.388 | 0.359  |         | -7.47  | 20    |
| 1,2-Dichloropropane            | 0.448 | 0.419  |         | -6.47  | 20    |
| Dibromomethane                 | 0.319 | 0.315  |         | -1.25  | 20    |
| Bromodichloromethane           | 0.653 | 0.651  |         | -0.31  | 20    |
| 4-Methyl-2-Pentanone           | 0.713 | 0.712  |         | -0.14  | 20    |
| Toluene                        | 1.053 | 1.008  |         | -4.27  | 20    |
| t-1,3-Dichloropropene          | 0.676 | 0.671  |         | -0.74  | 20    |
| cis-1,3-Dichloropropene        | 0.702 | 0.684  |         | -2.56  | 20    |
| 1,1,2-Trichloroethane          | 0.407 | 0.420  |         | 3.19   | 20    |
| 2-Hexanone                     | 0.451 | 0.481  |         | 6.65   | 20    |
| 1,2-Dibromoethane              | 0.430 | 0.430  |         | 0      | 20    |
| Tetrachloroethene              | 0.347 | 0.301  |         | -13.26 | 20    |
| Chlorobenzene                  | 1.244 | 1.159  | 0.3     | -6.83  | 20    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

|                     |                   |                        |                              |
|---------------------|-------------------|------------------------|------------------------------|
| Lab Name:           | <u>Alliance</u>   | Contract:              | <u>GENV01</u>                |
| Lab Code:           | <u>ACE</u>        | SDG No.:               | <u>Q3058</u>                 |
| Instrument ID:      | <u>MSVOA_N</u>    | Calibration Date/Time: | <u>09/10/2025 11:47</u>      |
| Lab File ID:        | <u>VN087774.D</u> | Init. Calib. Date(s):  | <u>08/21/2025 08/21/2025</u> |
| Heated Purge: (Y/N) | <u>N</u>          | Init. Calib. Time(s):  | <u>12:09 14:44</u>           |
| GC Column:          | <u>RXI-624</u>    | ID:                    | <u>0.25</u> (mm)             |

| COMPOUND                    | RRF   | RRF050 | MIN<br>RRF | %D     | MAX%D |
|-----------------------------|-------|--------|------------|--------|-------|
| Ethyl Benzene               | 2.154 | 1.995  |            | -7.38  | 20    |
| m/p-Xylenes                 | 0.790 | 0.761  |            | -3.67  | 20    |
| o-Xylene                    | 0.774 | 0.730  |            | -5.68  | 20    |
| Styrene                     | 1.297 | 1.300  |            | 0.23   | 20    |
| Bromoform                   | 0.317 | 0.332  | 0.1        | 4.73   | 20    |
| Isopropylbenzene            | 4.040 | 3.630  |            | -10.15 | 20    |
| 1,1,2,2-Tetrachloroethane   | 1.391 | 1.305  | 0.3        | -6.18  | 20    |
| 1,3-Dichlorobenzene         | 1.876 | 1.776  |            | -5.33  | 20    |
| 1,4-Dichlorobenzene         | 1.952 | 1.790  |            | -8.3   | 20    |
| 1,2-Dichlorobenzene         | 1.780 | 1.690  |            | -5.06  | 20    |
| 1,2-Dibromo-3-Chloropropane | 0.330 | 0.291  |            | -11.82 | 20    |
| 1,2,4-Trichlorobenzene      | 1.069 | 1.007  |            | -5.8   | 20    |
| 1,2,3-Trichlorobenzene      | 1.045 | 0.990  |            | -5.26  | 20    |
| 1,2-Dichloroethane-d4       | 1.024 | 0.986  |            | -3.71  | 20    |
| Dibromofluoromethane        | 0.361 | 0.374  |            | 3.6    | 20    |
| Toluene-d8                  | 1.351 | 1.346  |            | -0.37  | 20    |
| 4-Bromofluorobenzene        | 0.481 | 0.487  |            | 1.25   | 20    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.



# SAMPLE RAW DATA

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091025\  
 Data File : VN087782.D  
 Acq On : 10 Sep 2025 15:47  
 Operator : JC\MD  
 Sample : Q3058-03  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 MW4

Quant Time: Sep 11 02:05:50 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Aug 22 02:55:42 2025  
 Response via : Initial Calibration


| Compound                    | R.T.   | QIon           | Response | Conc       | Units  | Dev(Min) |
|-----------------------------|--------|----------------|----------|------------|--------|----------|
| -----                       |        |                |          |            |        |          |
| Internal Standards          |        |                |          |            |        |          |
| 1) Pentafluorobenzene       | 8.206  | 168            | 196528   | 50.000     | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 9.082  | 114            | 453885   | 50.000     | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 11.841 | 117            | 434796   | 50.000     | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.770 | 152            | 206220   | 50.000     | ug/l   | 0.00     |
| System Monitoring Compounds |        |                |          |            |        |          |
| 33) 1,2-Dichloroethane-d4   | 8.559  | 65             | 207855   | 51.620     | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range 74 - 125 | Recovery | = 103.240% |        |          |
| 35) Dibromofluoromethane    | 8.147  | 113            | 164718   | 50.264     | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range 75 - 124 | Recovery | = 100.520% |        |          |
| 50) Toluene-d8              | 10.547 | 98             | 573709   | 46.775     | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range 86 - 113 | Recovery | = 93.540%  |        |          |
| 62) 4-Bromofluorobenzene    | 12.829 | 95             | 205625   | 47.141     | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range 77 - 121 | Recovery | = 94.280%  |        |          |
| Target Compounds            |        |                |          |            |        |          |
|                             |        |                |          |            | Qvalue |          |
| 3) Chloromethane            | 2.389  | 50             | 69637    | 24.277     | ug/l   | 95       |
| 5) Bromomethane             | 2.983  | 94             | 16998    | 9.905      | ug/l   | 89       |
| 10) Methyl Iodide           | 4.577  | 142            | 18472    | 14.554     | ug/l # | 96       |
| 16) Acetone                 | 4.424  | 43             | 50618    | 23.092     | ug/l   | 97       |
| 17) Carbon Disulfide        | 4.700  | 76             | 5929     | 0.760      | ug/l # | 88       |
| 30) Chloroform              | 7.953  | 83             | 18426    | 2.971      | ug/l   | 87       |
| 40) Benzene                 | 8.588  | 78             | 12142    | 0.787      | ug/l   | 97       |
| 47) Bromodichloromethane    | 9.871  | 83             | 2163     | 0.365      | ug/l # | 70       |
| 52) Toluene                 | 10.612 | 92             | 10092    | 1.056      | ug/l   | 85       |
| 67) Ethyl Benzene           | 11.941 | 91             | 15321    | 0.818      | ug/l   | 95       |
| 68) m/p-Xylenes             | 12.047 | 106            | 6766     | 0.985      | ug/l   | 99       |
| 69) o-Xylene                | 12.376 | 106            | 7363     | 1.094      | ug/l   | 96       |
| 95) Naphthalene             | 15.611 | 128            | 18706    | 1.198      | ug/l # | 92       |
| -----                       |        |                |          |            |        |          |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091025\  
Data File : VN087782.D  
Acq On : 10 Sep 2025 15:47  
Operator : JC\MD  
Sample : Q3058-03  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
MW4

Quant Time: Sep 11 02:05:50 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M  
Quant Title : SW846 8260  
QLast Update : Fri Aug 22 02:55:42 2025  
Response via : Initial Calibration









#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 8.206 min Scan# 1063  
Delta R.T. 0.000 min  
Lab File: VN087782.D  
Acq: 10 Sep 2025 15:47




Instrument :  
MSVOA\_N  
ClientSampleId :  
MW4


Tgt Ion:168 Resp: 196528  
Ion Ratio Lower Upper  
168 100  
99 64.8 57.2 85.8



#3  
Chloromethane  
Concen: 24.277 ug/l  
RT: 2.389 min Scan# 74  
Delta R.T. 0.006 min  
Lab File: VN087782.D  
Acq: 10 Sep 2025 15:47

Tgt Ion: 50 Resp: 69637  
Ion Ratio Lower Upper  
50 100  
52 30.2 26.2 39.4









#5  
Bromomethane  
Concen: 9.905 ug/l  
RT: 2.983 min Scan# 11  
Delta R.T. 0.012 min  
Lab File: VN087782.D  
Acq: 10 Sep 2025 15:47




Instrument :  
MSVOA\_N  
ClientSampleId :  
MW4


Tgt Ion: 94 Resp: 16998  
Ion Ratio Lower Upper  
94 100  
96 105.1 75.4 113.2



#10  
Methyl Iodide  
Concen: 14.554 ug/l  
RT: 4.577 min Scan# 446  
Delta R.T. -0.000 min  
Lab File: VN087782.D  
Acq: 10 Sep 2025 15:47

Tgt Ion:142 Resp: 18472  
Ion Ratio Lower Upper  
142 100  
127 46.8 38.0 57.0  
141 8.8 12.0 18.0#









#16  
Acetone  
Concen: 23.092 ug/l  
RT: 4.424 min Scan# 41  
Delta R.T. -0.000 min  
Lab File: VN087782.D  
Acq: 10 Sep 2025 15:47




Instrument :  
MSVOA\_N  
ClientSampleId :  
MW4


Tgt Ion: 43 Resp: 50618  
Ion Ratio Lower Upper  
43 100  
58 26.8 22.6 33.8



#17  
Carbon Disulfide  
Concen: 0.760 ug/l  
RT: 4.700 min Scan# 467  
Delta R.T. -0.000 min  
Lab File: VN087782.D  
Acq: 10 Sep 2025 15:47

Tgt Ion: 76 Resp: 5929  
Ion Ratio Lower Upper  
76 100  
78 13.8 7.6 11.4#









#30  
Chloroform  
Concen: 2.971 ug/l  
RT: 7.953 min Scan# 1020  
Delta R.T. 0.006 min  
Lab File: VN087782.D  
Acq: 10 Sep 2025 15:47




Instrument :  
MSVOA\_N  
ClientSampleId :  
MW4


Tgt Ion: 83 Resp: 18426  
Ion Ratio Lower Upper  
83 100  
85 73.8 50.7 76.1



#33  
1,2-Dichloroethane-d4  
Concen: 51.620 ug/l  
RT: 8.559 min Scan# 1123  
Delta R.T. -0.000 min  
Lab File: VN087782.D  
Acq: 10 Sep 2025 15:47

Tgt Ion: 65 Resp: 207855  
Ion Ratio Lower Upper  
65 100  
67 51.5 0.0 100.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.082 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN087782.D

Acq: 10 Sep 2025 15:47

Instrument :

MSVOA\_N

ClientSampleId :

MW4





Tgt Ion:114 Resp: 453885

Ion Ratio Lower Upper

114 100

63 21.0 0.0 45.2

88 15.7 0.0 33.4



#35

Dibromofluoromethane

Concen: 50.264 ug/l

RT: 8.147 min Scan# 1053

Delta R.T. -0.006 min

Lab File: VN087782.D

Acq: 10 Sep 2025 15:47




Tgt Ion:113 Resp: 164718


Ion Ratio Lower Upper

113 100

111 103.7 82.8 124.2

192 16.5 12.8 19.2









#40  
Benzene  
Concen: 0.787 ug/l  
RT: 8.588 min Scan# 1128  
Delta R.T. -0.000 min  
Lab File: VN087782.D  
Acq: 10 Sep 2025 15:47




Instrument :  
MSVOA\_N  
ClientSampleId :  
MW4


Tgt Ion: 78 Resp: 12142  
Ion Ratio Lower Upper  
78 100  
77 25.9 19.7 29.5



#47  
Bromodichloromethane  
Concen: 0.365 ug/l  
RT: 9.871 min Scan# 1346  
Delta R.T. -0.000 min  
Lab File: VN087782.D  
Acq: 10 Sep 2025 15:47

Tgt Ion: 83 Resp: 2163  
Ion Ratio Lower Upper  
83 100  
85 40.8 52.5 78.7#  
127 0.0 5.8 8.6#









#50  
Toluene-d8  
Concen: 46.775 ug/l  
RT: 10.547 min Scan# 1461  
Delta R.T. -0.000 min  
Lab File: VN087782.D  
Acq: 10 Sep 2025 15:47




Instrument :  
MSVOA\_N  
ClientSampleId :  
MW4


Tgt Ion: 98 Resp: 573709  
Ion Ratio Lower Upper  
98 100  
100 62.6 52.8 79.2



#52  
Toluene  
Concen: 1.056 ug/l  
RT: 10.612 min Scan# 1472  
Delta R.T. 0.006 min  
Lab File: VN087782.D  
Acq: 10 Sep 2025 15:47

Tgt Ion: 92 Resp: 10092  
Ion Ratio Lower Upper  
92 100  
91 155.0 140.4 210.6





#62

4-Bromofluorobenzene

Concen: 47.141 ug/l

RT: 12.829 min Scan# 1849

Delta R.T. 0.006 min

Lab File: VN087782.D

Acq: 10 Sep 2025 15:47

Instrument :

MSVOA\_N

ClientSampleId :

MW4





Tgt Ion: 95 Resp: 205625

Ion Ratio Lower Upper

95 100

174 73.3 0.0 138.4

176 69.8 0.0 132.6



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.841 min Scan# 1681

Delta R.T. -0.000 min

Lab File: VN087782.D

Acq: 10 Sep 2025 15:47




Tgt Ion: 117 Resp: 434796


Ion Ratio Lower Upper

117 100

82 59.1 47.4 71.2

119 32.8 24.3 36.5





#67

Ethyl Benzene

Concen: 0.818 ug/l

RT: 11.941 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN087782.D

Acq: 10 Sep 2025 15:47

Instrument :

MSVOA\_N

ClientSampleId :





MW4

Tgt Ion: 91 Resp: 15321

Ion Ratio Lower Upper

91 100

106 27.3 24.1 36.1



#68

m/p-Xylenes

Concen: 0.985 ug/l

RT: 12.047 min Scan# 1716

Delta R.T. -0.000 min

Lab File: VN087782.D




Acq: 10 Sep 2025 15:47


Tgt Ion: 106 Resp: 6766

Ion Ratio Lower Upper

106 100

91 210.8 169.3 253.9









#69  
o-Xylene  
Concen: 1.094 ug/l  
RT: 12.376 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN087782.D  
Acq: 10 Sep 2025 15:47




Instrument :  
MSVOA\_N  
ClientSampleId :  
MW4


Tgt Ion:106 Resp: 7363  
Ion Ratio Lower Upper  
106 100  
91 216.7 111.4 334.2



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 13.770 min Scan# 2009  
Delta R.T. 0.006 min  
Lab File: VN087782.D  
Acq: 10 Sep 2025 15:47

Tgt Ion:152 Resp: 206220  
Ion Ratio Lower Upper  
152 100  
115 61.9 32.3 96.8  
150 157.2 0.0 351.0





#95

Naphthalene

Concen: 1.198 ug/l

RT: 15.611 min Scan# 21

Delta R.T. -0.000 min

Lab File: VN087782.D


Acq: 10 Sep 2025 15:47

Instrument :



MSVOA\_N

ClientSampleId :

MW4



|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 128   | Resp: | 18706 |
| Ion      | Ratio | Lower | Upper |
| 128      | 100   |       |       |
| 127      | 17.3  | 10.6  | 15.8# |
| 129      | 12.7  | 8.6   | 12.8  |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091025\  
Data File : VN087782.D  
Acq On : 10 Sep 2025 15:47  
Operator : JC\MD  
Sample : Q3058-03  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
MW4

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M  
Title : SW846 8260

Signal : TIC: VN087782.D\data.ms

| peak<br># | R.T.<br>min | first<br>scan | max<br>scan | last<br>scan | PK<br>TY | peak<br>height | corr.<br>area | corr.<br>% max. | % of<br>total |
|-----------|-------------|---------------|-------------|--------------|----------|----------------|---------------|-----------------|---------------|
| 1         | 2.383       | 67            | 73          | 84           | rVB      | 67977          | 117052        | 7.46%           | 1.315%        |
| 2         | 2.983       | 169           | 175         | 183          | rBV2     | 25594          | 55171         | 3.51%           | 0.620%        |
| 3         | 4.424       | 410           | 420         | 429          | rBV2     | 25812          | 76083         | 4.85%           | 0.854%        |
| 4         | 4.577       | 438           | 446         | 457          | rVB2     | 11021          | 32540         | 2.07%           | 0.365%        |
| 5         | 7.953       | 1012          | 1020        | 1030         | rVB2     | 21198          | 51995         | 3.31%           | 0.584%        |
| 6         | 8.147       | 1044          | 1053        | 1058         | rBV2     | 227605         | 558173        | 35.56%          | 6.269%        |
| 7         | 8.206       | 1058          | 1063        | 1074         | rVB2     | 299647         | 660529        | 42.08%          | 7.418%        |
| 8         | 8.559       | 1113          | 1123        | 1134         | rBV      | 255118         | 610124        | 38.87%          | 6.852%        |
| 9         | 9.082       | 1203          | 1212        | 1223         | rBV      | 549472         | 1118036       | 71.22%          | 12.557%       |
| 10        | 10.547      | 1450          | 1461        | 1468         | rBV      | 861017         | 1569828       | 100.00%         | 17.631%       |
| 11        | 10.612      | 1468          | 1472        | 1478         | rVB      | 24161          | 42762         | 2.72%           | 0.480%        |
| 12        | 11.841      | 1672          | 1681        | 1693         | rBV      | 772412         | 1416933       | 90.26%          | 15.914%       |
| 13        | 11.941      | 1693          | 1698        | 1704         | rVV2     | 23311          | 46928         | 2.99%           | 0.527%        |
| 14        | 12.047      | 1708          | 1716        | 1723         | rVB4     | 19154          | 43171         | 2.75%           | 0.485%        |
| 15        | 12.376      | 1766          | 1772        | 1781         | rBV4     | 24950          | 47870         | 3.05%           | 0.538%        |
| 16        | 12.829      | 1839          | 1849        | 1861         | rBV      | 582221         | 1050009       | 66.89%          | 11.793%       |
| 17        | 13.770      | 2001          | 2009        | 2023         | rBV      | 756859         | 1364026       | 86.89%          | 15.319%       |
| 18        | 15.611      | 2317          | 2322        | 2330         | rBV2     | 19947          | 42677         | 2.72%           | 0.479%        |




Sum of corrected areas: 8903907

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091025\  
Data File : VN087782.D  
Acq On : 10 Sep 2025 15:47  
Operator : JC\MD  
Sample : Q3058-03  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
MW4

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091025\  
Data File : VN087782.D  
Acq On : 10 Sep 2025 15:47  
Operator : JC\MD  
Sample : Q3058-03  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
MW4

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

\*\*\*\*\*

6

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091025\  
Data File : VN087782.D  
Acq On : 10 Sep 2025 15:47  
Operator : JC\MD  
Sample : Q3058-03  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
MW4

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

| TIC Top Hit name | RT | EstConc | Units | Response | --Internal Standard-- |    |      |      |
|------------------|----|---------|-------|----------|-----------------------|----|------|------|
|                  |    |         |       |          | #                     | RT | Resp | Conc |

6

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091025\  
Data File : VN087776.D  
Acq On : 10 Sep 2025 12:46  
Operator : JC\MD  
Sample : VN0910WBL01  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0910WBL01

Quant Time: Sep 11 02:01:55 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M  
Quant Title : SW846 8260  
QLast Update : Fri Aug 22 02:55:42 2025  
Response via : Initial Calibration

| Compound                   | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|--------|------|----------|--------|-------|----------|
| -----                      |        |      |          |        |       |          |
| Internal Standards         |        |      |          |        |       |          |
| 1) Pentafluorobenzene      | 8.206  | 168  | 177881   | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 9.082  | 114  | 413161   | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 11.841 | 117  | 392580   | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 13.770 | 152  | 179388   | 50.000 | ug/l  | 0.00     |

|                             |        |       |          |          |      |          |
|-----------------------------|--------|-------|----------|----------|------|----------|
| System Monitoring Compounds |        |       |          |          |      |          |
| 33) 1,2-Dichloroethane-d4   | 8.559  | 65    | 200486   | 55.009   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | =    | 110.020% |
| 35) Dibromofluoromethane    | 8.153  | 113   | 155658   | 52.181   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =    | 104.360% |
| 50) Toluene-d8              | 10.547 | 98    | 534676   | 47.889   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | =    | 95.780%  |
| 62) 4-Bromofluorobenzene    | 12.823 | 95    | 175489   | 44.197   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 77 - 121 | Recovery | =    | 88.400%  |


| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091025\  
Data File : VN087776.D  
Acq On : 10 Sep 2025 12:46  
Operator : JC\MD  
Sample : VN0910WBL01  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0910WBL01

Quant Time: Sep 11 02:01:55 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M  
Quant Title : SW846 8260  
QLast Update : Fri Aug 22 02:55:42 2025  
Response via : Initial Calibration









#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 8.206 min Scan# 1063  
Delta R.T. 0.000 min  
Lab File: VN087776.D  
Acq: 10 Sep 2025 12:46




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0910WBL01


Tgt Ion:168 Resp: 177881  
Ion Ratio Lower Upper  
168 100  
99 65.5 57.2 85.8



#33  
1,2-Dichloroethane-d4  
Concen: 55.009 ug/l  
RT: 8.559 min Scan# 1123  
Delta R.T. -0.000 min  
Lab File: VN087776.D  
Acq: 10 Sep 2025 12:46

Tgt Ion: 65 Resp: 200486  
Ion Ratio Lower Upper  
65 100  
67 49.1 0.0 100.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.082 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN087776.D

Acq: 10 Sep 2025 12:46

Instrument :

MSVOA\_N

ClientSampleId :

VN0910WBL01





Tgt Ion:114 Resp: 413161

Ion Ratio Lower Upper

114 100

63 22.1 0.0 45.2

88 13.8 0.0 33.4



#35

Dibromofluoromethane

Concen: 52.181 ug/l

RT: 8.153 min Scan# 1054

Delta R.T. -0.000 min

Lab File: VN087776.D

Acq: 10 Sep 2025 12:46




Tgt Ion:113 Resp: 155658


Ion Ratio Lower Upper

113 100

111 102.0 82.8 124.2

192 16.2 12.8 19.2









#50  
Toluene-d8  
Concen: 47.889 ug/l  
RT: 10.547 min Scan# 1461  
Delta R.T. -0.000 min  
Lab File: VN087776.D  
Acq: 10 Sep 2025 12:46




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0910WBL01


Tgt Ion: 98 Resp: 534676  
Ion Ratio Lower Upper  
98 100  
100 63.5 52.8 79.2



#62  
4-Bromofluorobenzene  
Concen: 44.197 ug/l  
RT: 12.823 min Scan# 1848  
Delta R.T. -0.000 min  
Lab File: VN087776.D  
Acq: 10 Sep 2025 12:46

Tgt Ion: 95 Resp: 175489  
Ion Ratio Lower Upper  
95 100  
174 73.1 0.0 138.4  
176 70.1 0.0 132.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.841 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN087776.D

Acq: 10 Sep 2025 12:46

Instrument :

MSVOA\_N

ClientSampleId :

VN0910WBL01





Tgt Ion:117 Resp: 392580

Ion Ratio Lower Upper

117 100

82 57.6 47.4 71.2

119 32.8 24.3 36.5



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.770 min Scan# 2009

Delta R.T. 0.006 min

Lab File: VN087776.D

Acq: 10 Sep 2025 12:46




Tgt Ion:152 Resp: 179388

Ion Ratio Lower Upper

152 100

115 62.6 32.3 96.8

150 156.4 0.0 351.0



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091025\  
Data File : VN087776.D  
Acq On : 10 Sep 2025 12:46  
Operator : JC\MD  
Sample : VN0910WBL01  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0910WBL01

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M

Title : SW846 8260

Signal : TIC: VN087776.D\data.ms

| peak<br># | R.T.<br>min | first<br>scan | max<br>scan | last<br>scan | PK<br>TY | peak<br>height | corr.<br>area | corr.<br>% max. | % of<br>total |
|-----------|-------------|---------------|-------------|--------------|----------|----------------|---------------|-----------------|---------------|
| 1         | 8.147       | 1044          | 1053        | 1058         | rBV2     | 211778         | 527773        | 35.76%          | 6.930%        |
| 2         | 8.206       | 1058          | 1063        | 1074         | rVB      | 260328         | 594890        | 40.31%          | 7.811%        |
| 3         | 8.559       | 1114          | 1123        | 1138         | rBV      | 240530         | 554261        | 37.56%          | 7.278%        |
| 4         | 9.082       | 1201          | 1212        | 1225         | rBV      | 503228         | 1046443       | 70.91%          | 13.741%       |
| 5         | 10.547      | 1452          | 1461        | 1472         | rBV      | 783829         | 1475718       | 100.00%         | 19.378%       |
| 6         | 11.841      | 1673          | 1681        | 1695         | rBV      | 705296         | 1290176       | 87.43%          | 16.941%       |
| 7         | 12.829      | 1840          | 1849        | 1867         | rBV      | 487126         | 937673        | 63.54%          | 12.313%       |
| 8         | 13.770      | 2001          | 2009        | 2022         | rBV      | 661104         | 1188668       | 80.55%          | 15.608%       |




Sum of corrected areas: 7615602

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091025\  
Data File : VN087776.D  
Acq On : 10 Sep 2025 12:46  
Operator : JC\MD  
Sample : VN0910WBL01  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0910WBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091025\  
Data File : VN087776.D  
Acq On : 10 Sep 2025 12:46  
Operator : JC\MD  
Sample : VN0910WBL01  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0910WBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

\*\*\*\*\*

6

A

B

C

D

E

F

G

H

I

J

6

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091025\  
Data File : VN087776.D  
Acq On : 10 Sep 2025 12:46  
Operator : JC\MD  
Sample : VN0910WBL01  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0910WBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

| TIC Top Hit name | RT | EstConc | Units | Response | --Internal Standard-- |    |      |      |
|------------------|----|---------|-------|----------|-----------------------|----|------|------|
|                  |    |         |       |          | #                     | RT | Resp | Conc |

6

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091025\  
 Data File : VN087777.D  
 Acq On : 10 Sep 2025 13:29  
 Operator : JC\MD  
 Sample : VN0910WBS01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0910WBS01

### Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 09/11/2025  
 Supervised By :Mahesh Dadoda 09/11/2025

Quant Time: Sep 11 02:02:18 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Aug 22 02:55:42 2025  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response   | Conc    | Units  | Dev(Min) |
|------------------------------|----------------|------|------------|---------|--------|----------|
| Internal Standards           |                |      |            |         |        |          |
| 1) Pentafluorobenzene        | 8.206          | 168  | 247948     | 50.000  | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 9.083          | 114  | 481179     | 50.000  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 11.847         | 117  | 443678     | 50.000  | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 13.770         | 152  | 236791     | 50.000  | ug/l   | 0.00     |
| System Monitoring Compounds  |                |      |            |         |        |          |
| 33) 1,2-Dichloroethane-d4    | 8.559          | 65   | 234962     | 46.251  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 74 - 125 |      | Recovery = | 92.500% |        |          |
| 35) Dibromofluoromethane     | 8.147          | 113  | 170515     | 49.082  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 75 - 124 |      | Recovery = | 98.160% |        |          |
| 50) Toluene-d8               | 10.547         | 98   | 598958     | 46.063  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 86 - 113 |      | Recovery = | 92.120% |        |          |
| 62) 4-Bromofluorobenzene     | 12.829         | 95   | 217795     | 47.098  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 77 - 121 |      | Recovery = | 94.200% |        |          |
| Target Compounds             |                |      |            |         |        |          |
|                              |                |      |            |         | Qvalue |          |
| 2) Dichlorodifluoromethane   | 2.142          | 85   | 66905      | 18.999  | ug/l   | 99       |
| 3) Chloromethane             | 2.383          | 50   | 72113      | 19.926  | ug/l   | 95       |
| 4) Vinyl Chloride            | 2.542          | 62   | 82564      | 19.676  | ug/l   | 98       |
| 5) Bromomethane              | 2.983          | 94   | 45461      | 20.997  | ug/l   | 91       |
| 6) Chloroethane              | 3.142          | 64   | 57277      | 19.421  | ug/l   | 97       |
| 7) Trichlorofluoromethane    | 3.512          | 101  | 124353     | 20.227  | ug/l   | 90       |
| 8) Diethyl Ether             | 3.965          | 74   | 54751      | 20.449  | ug/l   | 99       |
| 9) 1,1,2-Trichlorotrifluo... | 4.365          | 101  | 68198      | 19.605  | ug/l   | 96       |
| 10) Methyl Iodide            | 4.583          | 142  | 42240      | 20.845  | ug/l   | 98       |
| 11) Tert butyl alcohol       | 5.518          | 59   | 88895      | 100.795 | ug/l   | 97       |
| 12) 1,1-Dichloroethene       | 4.342          | 96   | 69558      | 19.458  | ug/l   | 88       |
| 13) Acrolein                 | 4.177          | 56   | 83930      | 77.410  | ug/l   | 94       |
| 14) Allyl chloride           | 5.006          | 41   | 115073     | 17.763  | ug/l   | 97       |
| 15) Acrylonitrile            | 5.706          | 53   | 249028     | 100.121 | ug/l   | 99       |
| 16) Acetone                  | 4.424          | 43   | 268456     | 97.074  | ug/l   | 98       |
| 17) Carbon Disulfide         | 4.700          | 76   | 172758     | 17.554  | ug/l   | 100      |
| 18) Methyl Acetate           | 5.018          | 43   | 137712     | 23.783  | ug/l   | 99       |
| 19) Methyl tert-butyl Ether  | 5.783          | 73   | 263760     | 19.668  | ug/l   | 97       |
| 20) Methylene Chloride       | 5.271          | 84   | 82792      | 19.952  | ug/l   | 97       |
| 21) trans-1,2-Dichloroethene | 5.771          | 96   | 73778      | 19.043  | ug/l   | 98       |
| 22) Diisopropyl ether        | 6.653          | 45   | 282609     | 20.216  | ug/l   | 97       |
| 23) Vinyl Acetate            | 6.589          | 43   | 1234043    | 97.027  | ug/l   | 98       |
| 24) 1,1-Dichloroethane       | 6.553          | 63   | 155898     | 20.339  | ug/l   | 98       |
| 25) 2-Butanone               | 7.465          | 43   | 354230     | 97.360  | ug/l   | 100      |
| 26) 2,2-Dichloropropane      | 7.471          | 77   | 127484     | 19.379  | ug/l   | 99       |
| 27) cis-1,2-Dichloroethene   | 7.465          | 96   | 91844      | 19.830  | ug/l   | 99       |
| 28) Bromochloromethane       | 7.794          | 49   | 68600      | 18.513  | ug/l   | 98       |
| 29) Tetrahydrofuran          | 7.824          | 42   | 228318     | 97.510  | ug/l   | 100      |
| 30) Chloroform               | 7.947          | 83   | 162362     | 20.753  | ug/l   | 99       |
| 31) Cyclohexane              | 8.236          | 56   | 114288     | 17.886  | ug/l   | 99       |
| 32) 1,1,1-Trichloroethane    | 8.147          | 97   | 133678     | 20.162  | ug/l   | 93       |
| 36) 1,1-Dichloropropene      | 8.359          | 75   | 98694      | 18.518  | ug/l   | 99       |
| 37) Ethyl Acetate            | 7.547          | 43   | 144354     | 19.821  | ug/l   | 99       |
| 38) Carbon Tetrachloride     | 8.341          | 117  | 107894     | 20.504  | ug/l   | 100      |
| 39) Methylcyclohexane        | 9.582          | 83   | 102544     | 17.476  | ug/l   | 97       |
| 40) Benzene                  | 8.588          | 78   | 323696     | 19.802  | ug/l   | 99       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091025\  
 Data File : VN087777.D  
 Acq On : 10 Sep 2025 13:29  
 Operator : JC\MD  
 Sample : VN0910WBS01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0910WBS01

### Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 09/11/2025  
 Supervised By : Mahesh Dadoda 09/11/2025

Quant Time: Sep 11 02:02:18 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Aug 22 02:55:42 2025  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 41) Methacrylonitrile         | 7.765  | 41   | 77810    | 20.625  | ug/l   | 96       |
| 42) 1,2-Dichloroethane        | 8.647  | 62   | 127177   | 20.245  | ug/l   | 99       |
| 43) Isopropyl Acetate         | 8.671  | 43   | 231457   | 20.356  | ug/l   | 100      |
| 44) Trichloroethene           | 9.335  | 130  | 74553    | 19.976  | ug/l   | 87       |
| 45) 1,2-Dichloropropane       | 9.594  | 63   | 85809    | 19.914  | ug/l   | 98       |
| 46) Dibromomethane            | 9.688  | 93   | 64080    | 20.898  | ug/l   | 97       |
| 47) Bromodichloromethane      | 9.865  | 83   | 133142   | 21.187  | ug/l   | 97       |
| 48) Methyl methacrylate       | 9.659  | 41   | 105144   | 19.550  | ug/l   | 99       |
| 49) 1,4-Dioxane               | 9.677  | 88   | 31057    | 445.513 | ug/l # | 99       |
| 51) 4-Methyl-2-Pentanone      | 10.424 | 43   | 727298   | 106.015 | ug/l   | 100      |
| 52) Toluene                   | 10.606 | 92   | 203676   | 20.098  | ug/l   | 98       |
| 53) t-1,3-Dichloropropene     | 10.818 | 75   | 136007   | 20.909  | ug/l   | 97       |
| 54) cis-1,3-Dichloropropene   | 10.288 | 75   | 136995   | 20.283  | ug/l   | 98       |
| 55) 1,1,2-Trichloroethane     | 10.994 | 97   | 84642    | 21.590  | ug/l   | 97       |
| 56) Ethyl methacrylate        | 10.853 | 69   | 123475   | 19.708  | ug/l   | 96       |
| 57) 1,3-Dichloropropane       | 11.141 | 76   | 147355   | 21.238  | ug/l   | 99       |
| 58) 2-Chloroethyl Vinyl ether | 10.141 | 63   | 302321   | 89.152  | ug/l   | 98       |
| 59) 2-Hexanone                | 11.177 | 43   | 468254   | 107.999 | ug/l   | 98       |
| 60) Dibromochloromethane      | 11.341 | 129  | 96636    | 21.274  | ug/l   | 99       |
| 61) 1,2-Dibromoethane         | 11.447 | 107  | 86421    | 20.906  | ug/l   | 100      |
| 64) Tetrachloroethene         | 11.082 | 164  | 59854    | 19.445  | ug/l   | 96       |
| 65) Chlorobenzene             | 11.871 | 112  | 224483   | 20.337  | ug/l   | 96       |
| 66) 1,1,1,2-Tetrachloroethane | 11.941 | 131  | 80324    | 21.920  | ug/l   | 98       |
| 67) Ethyl Benzene             | 11.941 | 91   | 384763   | 20.133  | ug/l   | 95       |
| 68) m/p-Xylenes               | 12.047 | 106  | 282361   | 40.285  | ug/l   | 98       |
| 69) o-Xylene                  | 12.376 | 106  | 138993   | 20.235  | ug/l   | 100      |
| 70) Styrene                   | 12.388 | 104  | 243299   | 21.143  | ug/l   | 98       |
| 71) Bromoform                 | 12.559 | 173  | 65248    | 23.217  | ug/l # | 98       |
| 73) Isopropylbenzene          | 12.676 | 105  | 356780   | 18.646  | ug/l   | 100      |
| 74) N-amyl acetate            | 12.518 | 43   | 128873m  | 17.773  | ug/l   |          |
| 75) 1,1,2,2-Tetrachloroethane | 12.918 | 83   | 132478   | 20.110  | ug/l   | 100      |
| 76) 1,2,3-Trichloropropane    | 12.970 | 75   | 125262m  | 22.558  | ug/l   |          |
| 77) Bromobenzene              | 12.959 | 156  | 88727    | 19.667  | ug/l   | 97       |
| 78) n-propylbenzene           | 13.012 | 91   | 456054   | 19.350  | ug/l   | 100      |
| 79) 2-Chlorotoluene           | 13.100 | 91   | 272010   | 19.178  | ug/l   | 99       |
| 80) 1,3,5-Trimethylbenzene    | 13.153 | 105  | 307840   | 18.965  | ug/l   | 98       |
| 81) trans-1,4-Dichloro-2-b... | 12.718 | 75   | 40685    | 19.397  | ug/l   | 88       |
| 82) 4-Chlorotoluene           | 13.200 | 91   | 280156   | 18.787  | ug/l   | 98       |
| 83) tert-Butylbenzene         | 13.412 | 119  | 261058   | 19.300  | ug/l   | 98       |
| 84) 1,2,4-Trimethylbenzene    | 13.459 | 105  | 313317   | 19.282  | ug/l   | 98       |
| 85) sec-Butylbenzene          | 13.594 | 105  | 389426   | 19.401  | ug/l   | 99       |
| 86) p-Isopropyltoluene        | 13.706 | 119  | 324109   | 19.554  | ug/l   | 100      |
| 87) 1,3-Dichlorobenzene       | 13.712 | 146  | 178936   | 20.140  | ug/l   | 96       |
| 88) 1,4-Dichlorobenzene       | 13.788 | 146  | 182845   | 19.776  | ug/l   | 99       |
| 89) n-Butylbenzene            | 14.035 | 91   | 310752   | 18.877  | ug/l   | 99       |
| 90) Hexachloroethane          | 14.306 | 117  | 66252    | 20.980  | ug/l   | 94       |
| 91) 1,2-Dichlorobenzene       | 14.082 | 146  | 170771   | 20.261  | ug/l   | 98       |
| 92) 1,2-Dibromo-3-Chloropr... | 14.694 | 75   | 30727    | 19.635  | ug/l   | 94       |
| 93) 1,2,4-Trichlorobenzene    | 15.364 | 180  | 96639    | 19.093  | ug/l   | 100      |
| 94) Hexachlorobutadiene       | 15.476 | 225  | 38772    | 21.487  | ug/l   | 97       |
| 95) Naphthalene               | 15.612 | 128  | 337020   | 18.799  | ug/l   | 99       |
| 96) 1,2,3-Trichlorobenzene    | 15.812 | 180  | 96427    | 19.488  | ug/l   | 98       |

6

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091025\  
Data File : VN087777.D  
Acq On : 10 Sep 2025 13:29  
Operator : JC\MD  
Sample : VN0910WBS01  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0910WBS01

Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 09/11/2025  
Supervised By :Mahesh Dadoda 09/11/2025

Quant Time: Sep 11 02:02:18 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M  
Quant Title : SW846 8260  
QLast Update : Fri Aug 22 02:55:42 2025  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

-----  
(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091025\  
Data File : VN087777.D  
Acq On : 10 Sep 2025 13:29  
Operator : JC\MD  
Sample : VN0910WBS01  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 1 Sample Multiplier: 1

**Instrument :**  
MSVOA\_N  
**ClientSampleId :**  
VN0910WBS01

## Manual Integrations

Reviewed By :Semsettin Yesilyurt 09/11/2025  
Supervised By :Mahesh Dadoda 09/11/2025

Quant Time: Sep 11 02:02:18 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M  
Quant Title : SW846 8260  
QLast Update : Fri Aug 22 02:55:42 2025  
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091025\  
 Data File : VN087778.D  
 Acq On : 10 Sep 2025 13:50  
 Operator : JC\MD  
 Sample : VN0910WBSD01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0910WBSD01

### Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 09/11/2025  
 Supervised By :Mahesh Dadoda 09/11/2025

Quant Time: Sep 11 02:03:13 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Aug 22 02:55:42 2025  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response   | Conc    | Units  | Dev(Min) |
|------------------------------|----------------|------|------------|---------|--------|----------|
| Internal Standards           |                |      |            |         |        |          |
| 1) Pentafluorobenzene        | 8.206          | 168  | 258583     | 50.000  | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 9.083          | 114  | 480224     | 50.000  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 11.841         | 117  | 451507     | 50.000  | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 13.771         | 152  | 236702     | 50.000  | ug/l   | 0.00     |
| System Monitoring Compounds  |                |      |            |         |        |          |
| 33) 1,2-Dichloroethane-d4    | 8.559          | 65   | 238905     | 45.093  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 74 - 125 |      | Recovery = | 90.180% |        |          |
| 35) Dibromofluoromethane     | 8.147          | 113  | 171837     | 49.561  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 75 - 124 |      | Recovery = | 99.120% |        |          |
| 50) Toluene-d8               | 10.547         | 98   | 623165     | 48.020  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 86 - 113 |      | Recovery = | 96.040% |        |          |
| 62) 4-Bromofluorobenzene     | 12.823         | 95   | 216720     | 46.959  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 77 - 121 |      | Recovery = | 93.920% |        |          |
| Target Compounds             |                |      |            |         |        |          |
|                              |                |      |            |         | Qvalue |          |
| 2) Dichlorodifluoromethane   | 2.142          | 85   | 71486      | 19.465  | ug/l   | 97       |
| 3) Chloromethane             | 2.383          | 50   | 75689      | 20.054  | ug/l   | 97       |
| 4) Vinyl Chloride            | 2.536          | 62   | 82893      | 18.942  | ug/l   | 96       |
| 5) Bromomethane              | 2.983          | 94   | 48062      | 21.285  | ug/l   | 99       |
| 6) Chloroethane              | 3.142          | 64   | 57557      | 18.713  | ug/l   | 100      |
| 7) Trichlorofluoromethane    | 3.512          | 101  | 133185     | 20.773  | ug/l   | 93       |
| 8) Diethyl Ether             | 3.965          | 74   | 55136      | 19.746  | ug/l   | 94       |
| 9) 1,1,2-Trichlorotrifluo... | 4.365          | 101  | 72141      | 19.885  | ug/l   | 99       |
| 10) Methyl Iodide            | 4.577          | 142  | 50321      | 22.843  | ug/l   | 93       |
| 11) Tert butyl alcohol       | 5.518          | 59   | 89083      | 96.854  | ug/l   | 99       |
| 12) 1,1-Dichloroethene       | 4.336          | 96   | 71585      | 19.201  | ug/l   | 87       |
| 13) Acrolein                 | 4.177          | 56   | 87390      | 77.286  | ug/l   | 97       |
| 14) Allyl chloride           | 5.012          | 41   | 116475     | 17.240  | ug/l   | 97       |
| 15) Acrylonitrile            | 5.706          | 53   | 259106     | 99.888  | ug/l   | 98       |
| 16) Acetone                  | 4.418          | 43   | 269935     | 93.594  | ug/l   | 98       |
| 17) Carbon Disulfide         | 4.700          | 76   | 174184     | 16.971  | ug/l   | 98       |
| 18) Methyl Acetate           | 5.018          | 43   | 143215     | 23.716  | ug/l   | 95       |
| 19) Methyl tert-butyl Ether  | 5.783          | 73   | 268923     | 19.228  | ug/l   | 98       |
| 20) Methylene Chloride       | 5.265          | 84   | 82321      | 18.963  | ug/l   | 96       |
| 21) trans-1,2-Dichloroethene | 5.783          | 96   | 74353      | 18.402  | ug/l   | 92       |
| 22) Diisopropyl ether        | 6.659          | 45   | 284767     | 19.532  | ug/l   | 99       |
| 23) Vinyl Acetate            | 6.589          | 43   | 1263475    | 95.256  | ug/l   | 98       |
| 24) 1,1-Dichloroethane       | 6.559          | 63   | 151717     | 18.980  | ug/l   | 99       |
| 25) 2-Butanone               | 7.471          | 43   | 366930     | 96.703  | ug/l   | 99       |
| 26) 2,2-Dichloropropane      | 7.471          | 77   | 131518     | 19.170  | ug/l   | 99       |
| 27) cis-1,2-Dichloroethene   | 7.477          | 96   | 92241      | 19.096  | ug/l   | 98       |
| 28) Bromochloromethane       | 7.794          | 49   | 68719      | 17.782  | ug/l   | 100      |
| 29) Tetrahydrofuran          | 7.830          | 42   | 234760     | 96.138  | ug/l   | 99       |
| 30) Chloroform               | 7.947          | 83   | 165649     | 20.302  | ug/l   | 94       |
| 31) Cyclohexane              | 8.241          | 56   | 119732     | 17.967  | ug/l   | 94       |
| 32) 1,1,1-Trichloroethane    | 8.153          | 97   | 136846     | 19.791  | ug/l   | 97       |
| 36) 1,1-Dichloropropene      | 8.353          | 75   | 100769     | 18.945  | ug/l   | 98       |
| 37) Ethyl Acetate            | 7.547          | 43   | 150409     | 20.693  | ug/l   | 99       |
| 38) Carbon Tetrachloride     | 8.341          | 117  | 112680     | 21.456  | ug/l   | 93       |
| 39) Methylcyclohexane        | 9.583          | 83   | 110999     | 18.954  | ug/l   | 97       |
| 40) Benzene                  | 8.588          | 78   | 332807     | 20.399  | ug/l   | 100      |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091025\  
 Data File : VN087778.D  
 Acq On : 10 Sep 2025 13:50  
 Operator : JC\MD  
 Sample : VN0910WBSD01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0910WBSD01

Manual Integrations  
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/11/2025  
 Supervised By :Mahesh Dadoda 09/11/2025

Quant Time: Sep 11 02:03:13 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Aug 22 02:55:42 2025  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 41) Methacrylonitrile         | 7.765  | 41   | 82456    | 21.900  | ug/l   | 95       |
| 42) 1,2-Dichloroethane        | 8.653  | 62   | 129154   | 20.601  | ug/l   | 98       |
| 43) Isopropyl Acetate         | 8.677  | 43   | 235768   | 20.776  | ug/l   | 99       |
| 44) Trichloroethene           | 9.330  | 130  | 76012    | 20.407  | ug/l   | 98       |
| 45) 1,2-Dichloropropane       | 9.600  | 63   | 85499    | 19.882  | ug/l   | 98       |
| 46) Dibromomethane            | 9.683  | 93   | 64916    | 21.213  | ug/l   | 98       |
| 47) Bromodichloromethane      | 9.871  | 83   | 134925   | 21.513  | ug/l   | 94       |
| 48) Methyl methacrylate       | 9.665  | 41   | 109874   | 20.470  | ug/l   | 97       |
| 49) 1,4-Dioxane               | 9.671  | 88   | 28068    | 403.436 | ug/l # | 92       |
| 51) 4-Methyl-2-Pentanone      | 10.424 | 43   | 744444   | 108.730 | ug/l   | 100      |
| 52) Toluene                   | 10.606 | 92   | 206022   | 20.370  | ug/l   | 99       |
| 53) t-1,3-Dichloropropene     | 10.812 | 75   | 138171   | 21.284  | ug/l   | 98       |
| 54) cis-1,3-Dichloropropene   | 10.288 | 75   | 141735   | 21.027  | ug/l   | 98       |
| 55) 1,1,2-Trichloroethane     | 10.994 | 97   | 86280    | 22.052  | ug/l   | 96       |
| 56) Ethyl methacrylate        | 10.859 | 69   | 130062   | 20.801  | ug/l   | 97       |
| 57) 1,3-Dichloropropane       | 11.141 | 76   | 146172   | 21.110  | ug/l   | 98       |
| 58) 2-Chloroethyl Vinyl ether | 10.141 | 63   | 308927   | 91.281  | ug/l   | 99       |
| 59) 2-Hexanone                | 11.177 | 43   | 481127   | 111.189 | ug/l   | 98       |
| 60) Dibromochloromethane      | 11.335 | 129  | 98766    | 21.786  | ug/l   | 100      |
| 61) 1,2-Dibromoethane         | 11.447 | 107  | 89587    | 21.715  | ug/l   | 97       |
| 64) Tetrachloroethene         | 11.082 | 164  | 61912    | 19.764  | ug/l   | 94       |
| 65) Chlorobenzene             | 11.871 | 112  | 231898   | 20.645  | ug/l   | 99       |
| 66) 1,1,1,2-Tetrachloroethane | 11.935 | 131  | 82123    | 22.022  | ug/l   | 97       |
| 67) Ethyl Benzene             | 11.941 | 91   | 393494   | 20.232  | ug/l   | 100      |
| 68) m/p-Xylenes               | 12.047 | 106  | 299338   | 41.967  | ug/l   | 95       |
| 69) o-Xylene                  | 12.376 | 106  | 136754   | 19.564  | ug/l   | 99       |
| 70) Styrene                   | 12.388 | 104  | 245673   | 20.979  | ug/l   | 98       |
| 71) Bromoform                 | 12.553 | 173  | 64901    | 22.693  | ug/l # | 98       |
| 73) Isopropylbenzene          | 12.671 | 105  | 367905   | 19.235  | ug/l   | 100      |
| 74) N-amyl acetate            | 12.541 | 43   | 124884m  | 17.229  | ug/l   |          |
| 75) 1,1,2,2-Tetrachloroethane | 12.918 | 83   | 140037   | 21.265  | ug/l   | 100      |
| 76) 1,2,3-Trichloropropane    | 12.971 | 75   | 120754m  | 21.754  | ug/l   |          |
| 77) Bromobenzene              | 12.959 | 156  | 90349    | 20.034  | ug/l   | 98       |
| 78) n-propylbenzene           | 13.012 | 91   | 460998   | 19.567  | ug/l   | 99       |
| 79) 2-Chlorotoluene           | 13.100 | 91   | 274914   | 19.390  | ug/l   | 100      |
| 80) 1,3,5-Trimethylbenzene    | 13.147 | 105  | 316481   | 19.505  | ug/l   | 97       |
| 81) trans-1,4-Dichloro-2-b... | 12.718 | 75   | 40528    | 19.329  | ug/l   | 90       |
| 82) 4-Chlorotoluene           | 13.200 | 91   | 292510   | 19.623  | ug/l   | 98       |
| 83) tert-Butylbenzene         | 13.412 | 119  | 270855   | 20.032  | ug/l   | 99       |
| 84) 1,2,4-Trimethylbenzene    | 13.459 | 105  | 317922   | 19.573  | ug/l   | 99       |
| 85) sec-Butylbenzene          | 13.594 | 105  | 410735   | 20.470  | ug/l   | 100      |
| 86) p-Isopropyltoluene        | 13.706 | 119  | 337508   | 20.370  | ug/l   | 99       |
| 87) 1,3-Dichlorobenzene       | 13.712 | 146  | 175185   | 19.726  | ug/l   | 99       |
| 88) 1,4-Dichlorobenzene       | 13.788 | 146  | 183301   | 19.833  | ug/l   | 99       |
| 89) n-Butylbenzene            | 14.035 | 91   | 328447   | 19.960  | ug/l   | 98       |
| 90) Hexachloroethane          | 14.312 | 117  | 69127    | 21.899  | ug/l   | 98       |
| 91) 1,2-Dichlorobenzene       | 14.082 | 146  | 170592   | 20.247  | ug/l   | 98       |
| 92) 1,2-Dibromo-3-Chloropr... | 14.694 | 75   | 31540    | 20.162  | ug/l   | 97       |
| 93) 1,2,4-Trichlorobenzene    | 15.365 | 180  | 99886    | 19.742  | ug/l   | 99       |
| 94) Hexachlorobutadiene       | 15.470 | 225  | 40181    | 22.276  | ug/l   | 98       |
| 95) Naphthalene               | 15.612 | 128  | 344659   | 19.232  | ug/l   | 99       |
| 96) 1,2,3-Trichlorobenzene    | 15.812 | 180  | 98128    | 19.839  | ug/l   | 98       |

6

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091025\  
Data File : VN087778.D  
Acq On : 10 Sep 2025 13:50  
Operator : JC\MD  
Sample : VN0910WBSD01  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0910WBSD01

Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 09/11/2025  
Supervised By :Mahesh Dadoda 09/11/2025

Quant Time: Sep 11 02:03:13 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M  
Quant Title : SW846 8260  
QLast Update : Fri Aug 22 02:55:42 2025  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

-----  
(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091025\  
Data File : VN087778.D  
Acq On : 10 Sep 2025 13:50  
Operator : JC\MD  
Sample : VN0910WBSD01  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 1 Sample Multiplier: 1

**Instrument :**  
MSVOA\_N  
**ClientSampleId :**  
VN0910WBSD01

## Manual Integrations

Reviewed By :Semsettin Yesilyurt 09/11/2025  
Supervised By :Mahesh Dadoda 09/11/2025

Quant Time: Sep 11 02:03:13 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N082125W.M  
Quant Title : SW846 8260  
QLast Update : Fri Aug 22 02:55:42 2025  
Response via : Initial Calibration



## Manual Integration Report

|           |          |            |         |
|-----------|----------|------------|---------|
| Sequence: | VN082125 | Instrument | MSVOA_n |
|-----------|----------|------------|---------|

| Sample ID   | File ID    | Parameter              | Review By | Review On            | Supervised By | Supervised On        | Reason                      |
|-------------|------------|------------------------|-----------|----------------------|---------------|----------------------|-----------------------------|
| VSTDICC001  | VN087619.D | 1,2,3-Trichloropropane | JOHN      | 8/22/2025 9:16:40 AM | MMDadoda      | 8/25/2025 8:11:19 AM | Peak Integrated by Software |
| VSTDICC001  | VN087619.D | 1,4-Dichlorobenzene    | JOHN      | 8/22/2025 9:16:40 AM | MMDadoda      | 8/25/2025 8:11:19 AM | Peak Integrated by Software |
| VSTDICC001  | VN087619.D | Diethyl Ether          | JOHN      | 8/22/2025 9:16:40 AM | MMDadoda      | 8/25/2025 8:11:19 AM | Peak Integrated by Software |
| VSTDICC001  | VN087619.D | Ethyl Acetate          | JOHN      | 8/22/2025 9:16:40 AM | MMDadoda      | 8/25/2025 8:11:19 AM | Peak Integrated by Software |
| VSTDICC001  | VN087619.D | Methyl Acetate         | JOHN      | 8/22/2025 9:16:40 AM | MMDadoda      | 8/25/2025 8:11:19 AM | Peak Integrated by Software |
| VSTDICC001  | VN087619.D | N-amyl acetate         | JOHN      | 8/22/2025 9:16:40 AM | MMDadoda      | 8/25/2025 8:11:19 AM | Peak Integrated by Software |
| VSTDICC005  | VN087620.D | 1,2,3-Trichloropropane | JOHN      | 8/22/2025 9:16:49 AM | MMDadoda      | 8/25/2025 8:11:20 AM | Peak Integrated by Software |
| VSTDICC005  | VN087620.D | Ethyl Acetate          | JOHN      | 8/22/2025 9:16:49 AM | MMDadoda      | 8/25/2025 8:11:20 AM | Peak Integrated by Software |
| VSTDICC005  | VN087620.D | N-amyl acetate         | JOHN      | 8/22/2025 9:16:49 AM | MMDadoda      | 8/25/2025 8:11:20 AM | Peak Integrated by Software |
| VSTDICC020  | VN087621.D | 1,2,3-Trichloropropane | JOHN      | 8/22/2025 9:17:20 AM | MMDadoda      | 8/25/2025 8:11:22 AM | Peak Integrated by Software |
| VSTDICC020  | VN087621.D | N-amyl acetate         | JOHN      | 8/22/2025 9:17:20 AM | MMDadoda      | 8/25/2025 8:11:22 AM | Peak Integrated by Software |
| VSTDICCC050 | VN087622.D | 1,2,3-Trichloropropane | JOHN      | 8/22/2025 9:17:24 AM | MMDadoda      | 8/25/2025 8:11:23 AM | Peak Integrated by Software |
| VSTDICCC050 | VN087622.D | N-amyl acetate         | JOHN      | 8/22/2025 9:17:24 AM | MMDadoda      | 8/25/2025 8:11:23 AM | Peak Integrated by Software |

## Manual Integration Report

|           |          |            |         |
|-----------|----------|------------|---------|
| Sequence: | VN082125 | Instrument | MSVOA_n |
|-----------|----------|------------|---------|

| Sample ID  | File ID    | Parameter              | Review By | Review On            | Supervised By | Supervised On        | Reason                      |
|------------|------------|------------------------|-----------|----------------------|---------------|----------------------|-----------------------------|
| VSTDICC100 | VN087623.D | 1,2,3-Trichloropropane | JOHN      | 8/22/2025 9:17:28 AM | MMDadoda      | 8/25/2025 8:11:24 AM | Peak Integrated by Software |
| VSTDICC150 | VN087624.D | 1,2,3-Trichloropropane | JOHN      | 8/22/2025 9:17:33 AM | MMDadoda      | 8/25/2025 8:11:26 AM | Peak Integrated by Software |
| VSTDICV050 | VN087626.D | 1,2,3-Trichloropropane | JOHN      | 8/22/2025 9:17:37 AM | MMDadoda      | 8/25/2025 8:11:28 AM | Peak Integrated by Software |
| VSTDICV050 | VN087626.D | N-amyl acetate         | JOHN      | 8/22/2025 9:17:37 AM | MMDadoda      | 8/25/2025 8:11:28 AM | Peak Integrated by Software |

## Manual Integration Report

|           |          |            |         |
|-----------|----------|------------|---------|
| Sequence: | VN091025 | Instrument | MSVOA_n |
|-----------|----------|------------|---------|

| Sample ID        | File ID    | Parameter              | Review By | Review On            | Supervised By | Supervised On        | Reason                      |
|------------------|------------|------------------------|-----------|----------------------|---------------|----------------------|-----------------------------|
| VSTDCCC050       | VN087774.D | 1,2,3-Trichloropropane | sam       | 9/11/2025 9:11:57 AM | MMDadoda      | 9/11/2025 5:36:45 PM | Peak Integrated by Software |
| VSTDCCC050       | VN087774.D | N-amyl acetate         | sam       | 9/11/2025 9:11:57 AM | MMDadoda      | 9/11/2025 5:36:45 PM | Peak Integrated by Software |
| VN0910WBS01      | VN087777.D | 1,2,3-Trichloropropane | sam       | 9/11/2025 9:12:02 AM | MMDadoda      | 9/11/2025 5:36:39 PM | Peak Integrated by Software |
| VN0910WBS01      | VN087777.D | N-amyl acetate         | sam       | 9/11/2025 9:12:02 AM | MMDadoda      | 9/11/2025 5:36:39 PM | Peak Integrated by Software |
| VN0910WBSD0<br>1 | VN087778.D | 1,2,3-Trichloropropane | sam       | 9/11/2025 9:12:06 AM | MMDadoda      | 9/11/2025 5:36:36 PM | Peak Integrated by Software |
| VN0910WBSD0<br>1 | VN087778.D | N-amyl acetate         | sam       | 9/11/2025 9:12:06 AM | MMDadoda      | 9/11/2025 5:36:36 PM | Peak Integrated by Software |
| VSTDCCC050       | VN087790.D | 1,2,3-Trichloropropane | sam       | 9/11/2025 9:12:33 AM | MMDadoda      | 9/11/2025 5:36:15 PM | Peak Integrated by Software |
| VSTDCCC050       | VN087790.D | N-amyl acetate         | sam       | 9/11/2025 9:12:33 AM | MMDadoda      | 9/11/2025 5:36:15 PM | Peak Integrated by Software |

Instrument ID: MSVOA\_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN082125

|                          |                                                       |                   |                                   |
|--------------------------|-------------------------------------------------------|-------------------|-----------------------------------|
| Review By                | John Carlone                                          | Review On         | 8/22/2025 9:17:52 AM              |
| Supervise By             | Mahesh Dadoda                                         | Supervise On      | 8/25/2025 8:11:38 AM              |
| SubDirectory             | VN082125                                              | HP Acquire Method | HP Processing Method 82N082125W.M |
| <b>STD. NAME</b>         | <b>STD REF.#</b>                                      |                   |                                   |
| Tune/Reschk              | VP135214                                              |                   |                                   |
| Initial Calibration Stds | VP135215,VP135216,VP135217,VP135218,VP135219,VP135220 |                   |                                   |
| CCC                      |                                                       |                   |                                   |
| Internal Standard/PEM    |                                                       |                   |                                   |
| ICV/I.BLK                | VP135221                                              |                   |                                   |
| Surrogate Standard       |                                                       |                   |                                   |
| MS/MSD Standard          |                                                       |                   |                                   |
| LCS Standard             |                                                       |                   |                                   |

| Sr# | SampleId     | Data File Name | Date-Time         | Operator | Status |
|-----|--------------|----------------|-------------------|----------|--------|
| 1   | BFB          | VN087614.D     | 21 Aug 2025 09:30 | JC\MD    | Ok     |
| 2   | VSTDCCC020   | VN087615.D     | 21 Aug 2025 10:04 | JC\MD    | Not Ok |
| 3   | VN0821WBS01  | VN087616.D     | 21 Aug 2025 10:37 | JC\MD    | Not Ok |
| 4   | VN0821WBSD01 | VN087617.D     | 21 Aug 2025 11:11 | JC\MD    | Not Ok |
| 5   | VN0821WBL01  | VN087618.D     | 21 Aug 2025 11:36 | JC\MD    | Not Ok |
| 6   | VSTDICC001   | VN087619.D     | 21 Aug 2025 12:09 | JC\MD    | Ok,M   |
| 7   | VSTDICC005   | VN087620.D     | 21 Aug 2025 13:08 | JC\MD    | Ok,M   |
| 8   | VSTDICC020   | VN087621.D     | 21 Aug 2025 13:41 | JC\MD    | Ok,M   |
| 9   | VSTDICCC050  | VN087622.D     | 21 Aug 2025 14:02 | JC\MD    | Ok,M   |
| 10  | VSTDICC100   | VN087623.D     | 21 Aug 2025 14:23 | JC\MD    | Ok,M   |
| 11  | VSTDICC150   | VN087624.D     | 21 Aug 2025 14:44 | JC\MD    | Ok,M   |
| 12  | IBLK         | VN087625.D     | 21 Aug 2025 15:05 | JC\MD    | Ok     |
| 13  | VSTDICV050   | VN087626.D     | 21 Aug 2025 15:47 | JC\MD    | Ok,M   |

M : Manual Integration

Instrument ID: MSVOA\_N

**Daily Analysis Runlog For Sequence/QCBatch ID # VN091025**

|                                                                                                    |                               |                      |                      |
|----------------------------------------------------------------------------------------------------|-------------------------------|----------------------|----------------------|
| Review By                                                                                          | Mahesh Dadoda                 | Review On            | 9/11/2025 5:36:58 PM |
| Supervise By                                                                                       | Semsettin Yesilyurt           | Supervise On         | 9/11/2025 5:39:28 PM |
| SubDirectory                                                                                       | VN091025                      | HP Acquire Method    | MSVOA_N              |
|                                                                                                    |                               | HP Processing Method | VN082125W.M          |
| <b>STD. NAME</b>                                                                                   | <b>STD REF.#</b>              |                      |                      |
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP135412                      |                      |                      |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP135413,VP135414<br>VP134956 |                      |                      |

| Sr# | SampleId     | Data File Name | Date-Time         | Operator | Status   |
|-----|--------------|----------------|-------------------|----------|----------|
| 1   | BFB          | VN087773.D     | 10 Sep 2025 11:13 | JC\MD    | Ok       |
| 2   | VSTDCCC050   | VN087774.D     | 10 Sep 2025 11:47 | JC\MD    | Ok,M     |
| 3   | VN0910MBL01  | VN087775.D     | 10 Sep 2025 12:25 | JC\MD    | Ok       |
| 4   | VN0910WBL01  | VN087776.D     | 10 Sep 2025 12:46 | JC\MD    | Ok       |
| 5   | VN0910WBS01  | VN087777.D     | 10 Sep 2025 13:29 | JC\MD    | Ok,M     |
| 6   | VN0910WBSD01 | VN087778.D     | 10 Sep 2025 13:50 | JC\MD    | Ok,M     |
| 7   | Q3045-06     | VN087779.D     | 10 Sep 2025 14:24 | JC\MD    | Dilution |
| 8   | Q3045-06DL   | VN087780.D     | 10 Sep 2025 15:06 | JC\MD    | Ok       |
| 9   | VIBLK        | VN087781.D     | 10 Sep 2025 15:26 | JC\MD    | Ok       |
| 10  | Q3058-03     | VN087782.D     | 10 Sep 2025 15:47 | JC\MD    | Ok       |
| 11  | VIBLK        | VN087783.D     | 10 Sep 2025 16:08 | JC\MD    | Ok       |
| 12  | Q3045-05     | VN087784.D     | 10 Sep 2025 16:29 | JC\MD    | Dilution |
| 13  | Q3045-05DL   | VN087785.D     | 10 Sep 2025 16:50 | JC\MD    | Ok,M     |
| 14  | Q3045-05DL   | VN087786.D     | 10 Sep 2025 17:11 | JC\MD    | Not Ok   |
| 15  | VIBLK        | VN087787.D     | 10 Sep 2025 17:32 | JC\MD    | Ok       |
| 16  | PB169608TB   | VN087788.D     | 10 Sep 2025 17:53 | JC\MD    | Ok,M     |
| 17  | PB169609TB   | VN087789.D     | 10 Sep 2025 18:14 | JC\MD    | Ok,M     |
| 18  | VSTDCCC050   | VN087790.D     | 10 Sep 2025 18:35 | JC\MD    | Ok,M     |

M : Manual Integration

Instrument ID: MSVOA\_N

### Daily Analysis Runlog For Sequence/QC Batch ID # VN082125

|                          |                                                       |                   |                                   |
|--------------------------|-------------------------------------------------------|-------------------|-----------------------------------|
| Review By                | John Carlone                                          | Review On         | 8/22/2025 9:17:52 AM              |
| Supervise By             | Maresh Dadoda                                         | Supervise On      | 8/25/2025 8:11:38 AM              |
| SubDirectory             | VN082125                                              | HP Acquire Method | HP Processing Method 82N082125W.M |
| <b>STD. NAME</b>         | <b>STD REF.#</b>                                      |                   |                                   |
| Tune/Reschk              | VP135214                                              |                   |                                   |
| Initial Calibration Stds | VP135215,VP135216,VP135217,VP135218,VP135219,VP135220 |                   |                                   |
| CCC                      |                                                       |                   |                                   |
| Internal Standard/PEM    |                                                       |                   |                                   |
| ICV/I.BLK                | VP135221                                              |                   |                                   |
| Surrogate Standard       |                                                       |                   |                                   |
| MS/MSD Standard          |                                                       |                   |                                   |
| LCS Standard             |                                                       |                   |                                   |

| Sr# | Sampled      | ClientID     | Data File Name | Date-Time         | Comment                              | Operator | Status |
|-----|--------------|--------------|----------------|-------------------|--------------------------------------|----------|--------|
| 1   | BFB          | BFB          | VN087614.D     | 21 Aug 2025 09:30 |                                      | JC\MD    | Ok     |
| 2   | VSTDCCC020   | VSTDCCC020   | VN087615.D     | 21 Aug 2025 10:04 | Need ICAL                            | JC\MD    | Not Ok |
| 3   | VN0821WBS01  | VN0821WBS01  | VN087616.D     | 21 Aug 2025 10:37 |                                      | JC\MD    | Not Ok |
| 4   | VN0821WBSD01 | VN0821WBSD01 | VN087617.D     | 21 Aug 2025 11:11 |                                      | JC\MD    | Not Ok |
| 5   | VN0821WBL01  | VN0821WBL01  | VN087618.D     | 21 Aug 2025 11:36 |                                      | JC\MD    | Not Ok |
| 6   | VSTDICC001   | VSTDICC001   | VN087619.D     | 21 Aug 2025 12:09 | LR-10,20                             | JC\MD    | Ok,M   |
| 7   | VSTDICC005   | VSTDICC005   | VN087620.D     | 21 Aug 2025 13:08 |                                      | JC\MD    | Ok,M   |
| 8   | VSTDICC020   | VSTDICC020   | VN087621.D     | 21 Aug 2025 13:41 | method not used for #N-amyle Acetate | JC\MD    | Ok,M   |
| 9   | VSTDICCC050  | VSTDICCC050  | VN087622.D     | 21 Aug 2025 14:02 |                                      | JC\MD    | Ok,M   |
| 10  | VSTDICC100   | VSTDICC100   | VN087623.D     | 21 Aug 2025 14:23 |                                      | JC\MD    | Ok,M   |
| 11  | VSTDICC150   | VSTDICC150   | VN087624.D     | 21 Aug 2025 14:44 |                                      | JC\MD    | Ok,M   |
| 12  | IBLK         | IBLK         | VN087625.D     | 21 Aug 2025 15:05 |                                      | JC\MD    | Ok     |
| 13  | VSTDICV050   | ICVVN082125  | VN087626.D     | 21 Aug 2025 15:47 | Icv Failed for Com.# 64 for DOD      | JC\MD    | Ok,M   |

M : Manual Integration

Instrument ID: MSVOA\_N

**Daily Analysis Runlog For Sequence/QC Batch ID # VN091025**

|                                                                                                    |                               |                   |                                          |
|----------------------------------------------------------------------------------------------------|-------------------------------|-------------------|------------------------------------------|
| Review By                                                                                          | Maresh Dadoda                 | Review On         | 9/11/2025 5:36:58 PM                     |
| Supervise By                                                                                       | Semsettin Yesilyurt           | Supervise On      | 9/11/2025 5:39:28 PM                     |
| SubDirectory                                                                                       | VN091025                      | HP Acquire Method | MSVOA_N HP Processing Method VN082125W.M |
| <b>STD. NAME</b>                                                                                   | <b>STD REF.#</b>              |                   |                                          |
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP135412                      |                   |                                          |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP135413,VP135414<br>VP134956 |                   |                                          |

| Sr# | SampleID     | ClientID     | Data File Name | Date-Time         | Comment                | Operator | Status   |
|-----|--------------|--------------|----------------|-------------------|------------------------|----------|----------|
| 1   | BFB          | BFB          | VN087773.D     | 10 Sep 2025 11:13 |                        | JC\MD    | Ok       |
| 2   | VSTDCCC050   | VSTDCCC050   | VN087774.D     | 10 Sep 2025 11:47 |                        | JC\MD    | Ok,M     |
| 3   | VN0910MBL01  | VN0910MBL01  | VN087775.D     | 10 Sep 2025 12:25 |                        | JC\MD    | Ok       |
| 4   | VN0910WBL01  | VN0910WBL01  | VN087776.D     | 10 Sep 2025 12:46 |                        | JC\MD    | Ok       |
| 5   | VN0910WBS01  | VN0910WBS01  | VN087777.D     | 10 Sep 2025 13:29 |                        | JC\MD    | Ok,M     |
| 6   | VN0910WBSD01 | VN0910WBSD01 | VN087778.D     | 10 Sep 2025 13:50 |                        | JC\MD    | Ok,M     |
| 7   | Q3045-06     | DNAPL        | VN087779.D     | 10 Sep 2025 14:24 | Need 200X,vial A pH# 9 | JC\MD    | Dilution |
| 8   | Q3045-06DL   | DNAPL DL     | VN087780.D     | 10 Sep 2025 15:06 |                        | JC\MD    | Ok       |
| 9   | VIBLK        | VIBLK        | VN087781.D     | 10 Sep 2025 15:26 |                        | JC\MD    | Ok       |
| 10  | Q3058-03     | MW4          | VN087782.D     | 10 Sep 2025 15:47 | vial A pH<2            | JC\MD    | Ok       |
| 11  | VIBLK        | VIBLK        | VN087783.D     | 10 Sep 2025 16:08 |                        | JC\MD    | Ok       |
| 12  | Q3045-05     | DNAPL-OIL    | VN087784.D     | 10 Sep 2025 16:29 | Need 50X,vial A pH# 5  | JC\MD    | Dilution |
| 13  | Q3045-05DL   | DNAPL-OIL DL | VN087785.D     | 10 Sep 2025 16:50 |                        | JC\MD    | Ok,M     |
| 14  | Q3045-05DL   | DNAPL-OIL DL | VN087786.D     | 10 Sep 2025 17:11 | Not required           | JC\MD    | Not Ok   |
| 15  | VIBLK        | VIBLK        | VN087787.D     | 10 Sep 2025 17:32 |                        | JC\MD    | Ok       |
| 16  | PB169608TB   | PB169608TB   | VN087788.D     | 10 Sep 2025 17:53 | vial A pH# 5           | JC\MD    | Ok,M     |
| 17  | PB169609TB   | PB169609TB   | VN087789.D     | 10 Sep 2025 18:14 | vial A pH# 5           | JC\MD    | Ok,M     |
| 18  | VSTDCCC050   | VSTDCCC050EC | VN087790.D     | 10 Sep 2025 18:35 |                        | JC\MD    | Ok,M     |

Instrument ID: MSVOA\_N

**Daily Analysis Runlog For Sequence/QC Batch ID # VN091025**

|                                         |                     |                   |                      |                      |             |
|-----------------------------------------|---------------------|-------------------|----------------------|----------------------|-------------|
| Review By                               | Mahesh Dadoda       | Review On         | 9/11/2025 5:36:58 PM |                      |             |
| Supervise By                            | Semsettin Yesilyurt | Supervise On      | 9/11/2025 5:39:28 PM |                      |             |
| SubDirectory                            | VN091025            | HP Acquire Method | MSVOA_N              | HP Processing Method | VN082125W.M |
| STD. NAME                               |                     | STD REF.#         |                      |                      |             |
| Tune/Reschk<br>Initial Calibration Stds |                     | VP135412          |                      |                      |             |
| CCC                                     |                     | VP135413,VP135414 |                      |                      |             |
| Internal Standard/PEM                   |                     | VP134956          |                      |                      |             |
| ICV/I.BLK                               |                     |                   |                      |                      |             |
| Surrogate Standard                      |                     |                   |                      |                      |             |
| MS/MSD Standard                         |                     |                   |                      |                      |             |
| LCS Standard                            |                     |                   |                      |                      |             |

M : Manual Integration

LAB CHRONICLE

|          |                 |            |                     |
|----------|-----------------|------------|---------------------|
| OrderID: | Q3058           | OrderDate: | 9/9/2025 1:49:00 PM |
| Client:  | G Environmental | Project:   | Buffington          |
| Contact: | Gary Landis     | Location:  | J42,VOA Lab         |

| LabID    | ClientID | Matrix | Test         | Method   | Sample Date | Prep Date | Anal Date | Received |
|----------|----------|--------|--------------|----------|-------------|-----------|-----------|----------|
| Q3058-03 | MW4      | Water  | VOCMS Group1 | 8260-Low | 09/09/25    |           | 09/10/25  | 09/09/25 |



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**Hit Summary Sheet**  
**SW-846**

**SDG No.:** Q3058  
**Client:** G Environmental

**Order ID:** Q3058  
**Project ID:** Buffington

| Sample ID              | Client ID | Matrix | Parameter | Concentration | C | MDL  | RDL  | Units |
|------------------------|-----------|--------|-----------|---------------|---|------|------|-------|
| <b>Client ID : MW4</b> |           |        |           |               |   |      |      |       |
| Q3058-03               | MW4       | Water  | Iron      | 5320          |   | 11.7 | 50.0 | ug/L  |
| Q3058-03               | MW4       | Water  | Manganese | 303           |   | 2.97 | 10.0 | ug/L  |
| Q3058-03               | MW4       | Water  | Sodium    | 56100         |   | 434  | 1000 | ug/L  |



# SAMPLE DATA

## Report of Analysis

|                   |                 |                 |          |
|-------------------|-----------------|-----------------|----------|
| Client:           | G Environmental | Date Collected: | 09/09/25 |
| Project:          | Buffington      | Date Received:  | 09/09/25 |
| Client Sample ID: | MW4             | SDG No.:        | Q3058    |
| Lab Sample ID:    | Q3058-03        | Matrix:         | Water    |
| Level (low/med):  | low             | % Solid:        | 0        |

| Cas       | Parameter | Conc. | Qua. | DF | MDL  | LOQ / CRQL | Units | Prep Date      | Date Ana.      | Ana Met. | Prep Met. |
|-----------|-----------|-------|------|----|------|------------|-------|----------------|----------------|----------|-----------|
| 7439-89-6 | Iron      | 5320  | N*   | 1  | 11.7 | 50.0       | ug/L  | 09/12/25 10:05 | 09/12/25 18:29 | 6010D    | SW3010    |
| 7439-96-5 | Manganese | 303   | *    | 1  | 2.97 | 10.0       | ug/L  | 09/12/25 10:05 | 09/12/25 18:29 | 6010D    | SW3010    |
| 7440-23-5 | Sodium    | 56100 | *    | 1  | 434  | 1000       | ug/L  | 09/12/25 10:05 | 09/12/25 18:29 | 6010D    | SW3010    |

|               |               |                 |       |            |
|---------------|---------------|-----------------|-------|------------|
| Color Before: | Colorless     | Clarity Before: | Clear | Texture:   |
| Color After:  | Colorless     | Clarity After:  | Clear | Artifacts: |
| Comments:     | Metals Group4 |                 |       |            |

U = Not Detected  
 LOQ = Limit of Quantitation  
 MDL = Method Detection Limit  
 LOD = Limit of Detection  
 D = Dilution  
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value  
 B = Analyte Found in Associated Method Blank  
 \* = indicates the duplicate analysis is not within control limits.  
 E = Indicates the reported value is estimated because of the presence of interference.  
 OR = Over Range  
 N =Spiked sample recovery not within control limits



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**Metals**

**- 3a -**

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

|                  |                        |                  |              |
|------------------|------------------------|------------------|--------------|
| <b>Client:</b>   | <u>G Environmental</u> | <b>SDG No.:</b>  | <u>Q3058</u> |
| <b>Contract:</b> | <u>GENV01</u>          | <b>Lab Code:</b> | <u>ACE</u>   |

| Sample ID | Analyte   | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|-----------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| ICB01     | Iron      | 23.4           | +/-50               | U            | 100  | P | 09/12/2025       | 13:42            | LB137179      |
|           | Manganese | 5.94           | +/-10               | U            | 20.0 | P | 09/12/2025       | 13:42            | LB137179      |
|           | Sodium    | 868            | +/-1000             | U            | 2000 | P | 09/12/2025       | 13:42            | LB137179      |

**Metals**

- 3a -

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** G Environmental

**SDG No.:** Q3058

**Contract:** GENV01

**Lab Code:** ACE

| Sample ID | Analyte   | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|-----------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| CCB01     | Iron      | 23.4           | +/-50               | U            | 100  | P | 09/12/2025       | 14:14            | LB137179      |
|           | Manganese | 5.94           | +/-10               | U            | 20.0 | P | 09/12/2025       | 14:14            | LB137179      |
|           | Sodium    | 868            | +/-1000             | U            | 2000 | P | 09/12/2025       | 14:14            | LB137179      |
| CCB02     | Iron      | 23.4           | +/-50               | U            | 100  | P | 09/12/2025       | 15:12            | LB137179      |
|           | Manganese | 5.94           | +/-10               | U            | 20.0 | P | 09/12/2025       | 15:12            | LB137179      |
|           | Sodium    | 868            | +/-1000             | U            | 2000 | P | 09/12/2025       | 15:12            | LB137179      |
| CCB03     | Iron      | 23.4           | +/-50               | U            | 100  | P | 09/12/2025       | 16:03            | LB137179      |
|           | Manganese | 5.94           | +/-10               | U            | 20.0 | P | 09/12/2025       | 16:03            | LB137179      |
|           | Sodium    | 868            | +/-1000             | U            | 2000 | P | 09/12/2025       | 16:03            | LB137179      |
| CCB04     | Iron      | 23.4           | +/-50               | U            | 100  | P | 09/12/2025       | 17:14            | LB137179      |
|           | Manganese | 5.94           | +/-10               | U            | 20.0 | P | 09/12/2025       | 17:14            | LB137179      |
|           | Sodium    | 868            | +/-1000             | U            | 2000 | P | 09/12/2025       | 17:14            | LB137179      |
| CCB05     | Iron      | 23.4           | +/-50               | U            | 100  | P | 09/12/2025       | 18:04            | LB137179      |
|           | Manganese | 5.94           | +/-10               | U            | 20.0 | P | 09/12/2025       | 18:04            | LB137179      |
|           | Sodium    | 868            | +/-1000             | U            | 2000 | P | 09/12/2025       | 18:04            | LB137179      |
| CCB06     | Iron      | 23.4           | +/-50               | U            | 100  | P | 09/12/2025       | 18:51            | LB137179      |
|           | Manganese | 5.94           | +/-10               | U            | 20.0 | P | 09/12/2025       | 18:51            | LB137179      |
|           | Sodium    | 868            | +/-1000             | U            | 2000 | P | 09/12/2025       | 18:51            | LB137179      |

**Metals**

- 3a -

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** G Environmental **SDG No.:** Q3058  
**Contract:** GENV01 **Lab Code:** ACE

| Sample ID | Analyte   | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|-----------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| ICB01     | Iron      | 23.4           | +/-50               | U            | 100  | P | 09/15/2025       | 11:43            | LB137198      |
|           | Manganese | 5.94           | +/-10               | U            | 20.0 | P | 09/15/2025       | 11:43            | LB137198      |
|           | Sodium    | 868            | +/-1000             | U            | 2000 | P | 09/15/2025       | 11:43            | LB137198      |

A  
B  
C  
D  
E  
F  
G  
H  
I  
J

**Metals**

- 3a -

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** G Environmental **SDG No.:** Q3058  
**Contract:** GENV01 **Lab Code:** ACE

| Sample ID | Analyte   | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|-----------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| CCB01     | Iron      | 23.4           | +/-50               | U            | 100  | P | 09/15/2025       | 12:39            | LB137198      |
|           | Manganese | 5.94           | +/-10               | U            | 20.0 | P | 09/15/2025       | 12:39            | LB137198      |
|           | Sodium    | 868            | +/-1000             | U            | 2000 | P | 09/15/2025       | 12:39            | LB137198      |
| CCB02     | Iron      | 23.4           | +/-50               | U            | 100  | P | 09/15/2025       | 13:47            | LB137198      |
|           | Manganese | 5.94           | +/-10               | U            | 20.0 | P | 09/15/2025       | 13:47            | LB137198      |
|           | Sodium    | 868            | +/-1000             | U            | 2000 | P | 09/15/2025       | 13:47            | LB137198      |
| CCB03     | Iron      | 23.4           | +/-50               | U            | 100  | P | 09/15/2025       | 15:15            | LB137198      |
|           | Manganese | 5.94           | +/-10               | U            | 20.0 | P | 09/15/2025       | 15:15            | LB137198      |
|           | Sodium    | 868            | +/-1000             | U            | 2000 | P | 09/15/2025       | 15:15            | LB137198      |
| CCB04     | Iron      | 23.4           | +/-50               | U            | 100  | P | 09/15/2025       | 16:16            | LB137198      |
|           | Manganese | 5.94           | +/-10               | U            | 20.0 | P | 09/15/2025       | 16:16            | LB137198      |
|           | Sodium    | 868            | +/-1000             | U            | 2000 | P | 09/15/2025       | 16:16            | LB137198      |
| CCB05     | Iron      | 23.4           | +/-50               | U            | 100  | P | 09/15/2025       | 16:59            | LB137198      |
|           | Manganese | 5.94           | +/-10               | U            | 20.0 | P | 09/15/2025       | 16:59            | LB137198      |
|           | Sodium    | 868            | +/-1000             | U            | 2000 | P | 09/15/2025       | 16:59            | LB137198      |

**Metals**  
**- 3b -**  
**PREPARATION BLANK SUMMARY**

**Client:** G Environmental

**SDG No.:** Q3058

**Instrument:** P4

| Sample ID         | Analyte      | Result<br>(ug/L) | Acceptance<br>Limit | Conc<br>Qual         | CRQL<br>ug/L    | M | Analysis<br>Date  | Analysis<br>Time  | Run      |
|-------------------|--------------|------------------|---------------------|----------------------|-----------------|---|-------------------|-------------------|----------|
| <b>PB169680BL</b> | <b>WATER</b> |                  |                     | <b>Batch Number:</b> | <b>PB169680</b> |   | <b>Prep Date:</b> | <b>09/12/2025</b> |          |
|                   | Iron         | 11.7             | <25                 | U                    | 50.0            | P | 09/15/2025        | 12:58             | LB137198 |
|                   | Manganese    | 2.97             | <5                  | U                    | 10.0            | P | 09/15/2025        | 12:58             | LB137198 |
|                   | Sodium       | 434              | <500                | U                    | 1000            | P | 09/15/2025        | 12:58             | LB137198 |

A

B

C

D

E

F

G

H

I

J



# METAL CALIBRATION DATA

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** G Environmental

**SDG No.:** Q3058

**Contract:** GENV01

**Lab Code:** ACE

**Initial Calibration Source:** EPA

**Continuing Calibration Source:** Inorganic Ventures

| Sample ID | Analyte   | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|-----------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| ICV01     | Iron      | 3780           | 4000       | 95            | 90 - 110                  | P | 09/12/2025       | 13:34            | LB137179      |
|           | Manganese | 1850           | 2000       | 92            | 90 - 110                  | P | 09/12/2025       | 13:34            | LB137179      |
|           | Sodium    | 18700          | 20000      | 94            | 90 - 110                  | P | 09/12/2025       | 13:34            | LB137179      |

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** G Environmental  
**Contract:** GENV01  
**Initial Calibration Source:** EPA  
**Continuing Calibration Source:** Inorganic Ventures

**SDG No.:** Q3058  
**Lab Code:** ACE

| Sample ID | Analyte   | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|-----------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| LLICV01   | Iron      | 101            | 100        | 101           | 80 - 120                  | P | 09/12/2025       | 13:38            | LB137179      |
|           | Manganese | 19.4           | 20.0       | 97            | 80 - 120                  | P | 09/12/2025       | 13:38            | LB137179      |
|           | Sodium    | 1890           | 2000       | 95            | 80 - 120                  | P | 09/12/2025       | 13:38            | LB137179      |

# Metals

- 2a -

## INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3058

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

| Sample ID | Analyte   | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|-----------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| CCV01     | Iron      | 4910           | 5000       | 98            | 90 - 110                  | P | 09/12/2025       | 14:10            | LB137179      |
|           | Manganese | 2410           | 2500       | 96            | 90 - 110                  | P | 09/12/2025       | 14:10            | LB137179      |
|           | Sodium    | 24600          | 25000      | 98            | 90 - 110                  | P | 09/12/2025       | 14:10            | LB137179      |
| CCV02     | Iron      | 4880           | 5000       | 98            | 90 - 110                  | P | 09/12/2025       | 15:08            | LB137179      |
|           | Manganese | 2330           | 2500       | 93            | 90 - 110                  | P | 09/12/2025       | 15:08            | LB137179      |
|           | Sodium    | 24500          | 25000      | 98            | 90 - 110                  | P | 09/12/2025       | 15:08            | LB137179      |
| CCV03     | Iron      | 4880           | 5000       | 98            | 90 - 110                  | P | 09/12/2025       | 15:59            | LB137179      |
|           | Manganese | 2380           | 2500       | 95            | 90 - 110                  | P | 09/12/2025       | 15:59            | LB137179      |
|           | Sodium    | 24500          | 25000      | 98            | 90 - 110                  | P | 09/12/2025       | 15:59            | LB137179      |
| CCV04     | Iron      | 5020           | 5000       | 100           | 90 - 110                  | P | 09/12/2025       | 16:51            | LB137179      |
|           | Manganese | 2410           | 2500       | 96            | 90 - 110                  | P | 09/12/2025       | 16:51            | LB137179      |
|           | Sodium    | 25800          | 25000      | 103           | 90 - 110                  | P | 09/12/2025       | 16:51            | LB137179      |
| CCV05     | Iron      | 4740           | 5000       | 95            | 90 - 110                  | P | 09/12/2025       | 18:00            | LB137179      |
|           | Manganese | 2340           | 2500       | 94            | 90 - 110                  | P | 09/12/2025       | 18:00            | LB137179      |
|           | Sodium    | 23400          | 25000      | 94            | 90 - 110                  | P | 09/12/2025       | 18:00            | LB137179      |
| CCV06     | Iron      | 4710           | 5000       | 94            | 90 - 110                  | P | 09/12/2025       | 18:47            | LB137179      |
|           | Manganese | 2380           | 2500       | 95            | 90 - 110                  | P | 09/12/2025       | 18:47            | LB137179      |
|           | Sodium    | 23300          | 25000      | 93            | 90 - 110                  | P | 09/12/2025       | 18:47            | LB137179      |

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** G Environmental

**SDG No.:** Q3058

**Contract:** GENV01

**Lab Code:** ACE

**Initial Calibration Source:** EPA

**Continuing Calibration Source:** Inorganic Ventures

| Sample ID | Analyte   | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|-----------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| ICV01     | Iron      | 3930           | 4000       | 98            | 90 - 110                  | P | 09/15/2025       | 11:31            | LB137198      |
|           | Manganese | 1910           | 2000       | 96            | 90 - 110                  | P | 09/15/2025       | 11:31            | LB137198      |
|           | Sodium    | 19200          | 20000      | 96            | 90 - 110                  | P | 09/15/2025       | 11:31            | LB137198      |

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** G Environmental  
**Contract:** GENV01  
**Initial Calibration Source:** EPA  
**Continuing Calibration Source:** Inorganic Ventures

**SDG No.:** Q3058  
**Lab Code:** ACE

| Sample ID | Analyte   | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|-----------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| LLICV01   | Iron      | 104            | 100        | 104           | 80 - 120                  | P | 09/15/2025       | 11:36            | LB137198      |
|           | Manganese | 20.6           | 20.0       | 103           | 80 - 120                  | P | 09/15/2025       | 11:36            | LB137198      |
|           | Sodium    | 1900           | 2000       | 95            | 80 - 120                  | P | 09/15/2025       | 11:36            | LB137198      |

## Metals

- 2a -

### INITIAL AND CONTINUING CALIBRATION VERIFICATION

**Client:** G Environmental

**Contract:** GENV01

**Initial Calibration Source:** EPA

**Continuing Calibration Source:** Inorganic Ventures

**SDG No.:** Q3058

**Lab Code:** ACE

| Sample ID | Analyte   | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|-----------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| CCV01     | Iron      | 4930           | 5000       | 99            | 90 - 110                  | P | 09/15/2025       | 12:35            | LB137198      |
|           | Manganese | 2480           | 2500       | 99            | 90 - 110                  | P | 09/15/2025       | 12:35            | LB137198      |
|           | Sodium    | 24200          | 25000      | 97            | 90 - 110                  | P | 09/15/2025       | 12:35            | LB137198      |
| CCV02     | Iron      | 4720           | 5000       | 94            | 90 - 110                  | P | 09/15/2025       | 13:42            | LB137198      |
|           | Manganese | 2370           | 2500       | 95            | 90 - 110                  | P | 09/15/2025       | 13:42            | LB137198      |
|           | Sodium    | 23300          | 25000      | 93            | 90 - 110                  | P | 09/15/2025       | 13:42            | LB137198      |
| CCV03     | Iron      | 4770           | 5000       | 95            | 90 - 110                  | P | 09/15/2025       | 15:11            | LB137198      |
|           | Manganese | 2340           | 2500       | 94            | 90 - 110                  | P | 09/15/2025       | 15:11            | LB137198      |
|           | Sodium    | 22600          | 25000      | 91            | 90 - 110                  | P | 09/15/2025       | 15:11            | LB137198      |
| CCV04     | Iron      | 4840           | 5000       | 97            | 90 - 110                  | P | 09/15/2025       | 16:12            | LB137198      |
|           | Manganese | 2360           | 2500       | 94            | 90 - 110                  | P | 09/15/2025       | 16:12            | LB137198      |
|           | Sodium    | 23200          | 25000      | 93            | 90 - 110                  | P | 09/15/2025       | 16:12            | LB137198      |
| CCV05     | Iron      | 4740           | 5000       | 95            | 90 - 110                  | P | 09/15/2025       | 16:55            | LB137198      |
|           | Manganese | 2360           | 2500       | 95            | 90 - 110                  | P | 09/15/2025       | 16:55            | LB137198      |
|           | Sodium    | 23000          | 25000      | 92            | 90 - 110                  | P | 09/15/2025       | 16:55            | LB137198      |



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**Metals**

**- 2b -**

**CRDL STANDARD FOR AA & ICP**

**Client:** G Environmental

**SDG No.:** Q3058

**Contract:** GENV01

**Lab Code:** ACE

**Initial Calibration Source:** \_\_\_\_\_

**Continuing Calibration Source:** \_\_\_\_\_

| Sample ID | Analyte   | Result<br>ug/L | True Value<br>ug/L | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|-----------|----------------|--------------------|---------------|---------------------------|---|------------------|------------------|---------------|
| CRI01     | Iron      | 92.9           | 100                | 93            | 65 - 135                  | P | 09/12/2025       | 13:47            | LB137179      |
|           | Manganese | 19.5           | 20.0               | 98            | 65 - 135                  | P | 09/12/2025       | 13:47            | LB137179      |
|           | Sodium    | 1870           | 2000               | 94            | 65 - 135                  | P | 09/12/2025       | 13:47            | LB137179      |
| CRI01     | Iron      | 104            | 100                | 104           | 65 - 135                  | P | 09/15/2025       | 11:47            | LB137198      |
|           | Manganese | 19.6           | 20.0               | 98            | 65 - 135                  | P | 09/15/2025       | 11:47            | LB137198      |
|           | Sodium    | 1840           | 2000               | 92            | 65 - 135                  | P | 09/15/2025       | 11:47            | LB137198      |

**Metals**  
**- 4 -**  
**INTERFERENCE CHECK SAMPLE**

|                                       |                                 |
|---------------------------------------|---------------------------------|
| <b>Client:</b> <u>G Environmental</u> | <b>SDG No.:</b> <u>Q3058</u>    |
| <b>Contract:</b> <u>GENV01</u>        | <b>Lab Code:</b> <u>ACE</u>     |
| <b>ICS Source:</b> <u>EPA</u>         | <b>Instrument ID:</b> <u>P4</u> |

| Sample ID      | Analyte   | Result<br>ug/L | True Value<br>ug/L | %<br>Recovery | Low<br>Limit<br>(ug/L) | High<br>Limit<br>(ug/L) | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|----------------|-----------|----------------|--------------------|---------------|------------------------|-------------------------|------------------|------------------|---------------|
| <b>ICSA01</b>  | Iron      | 98100          | 101000             | 97            | 85600                  | 116500                  | 09/12/2025       | 13:53            | LB137179      |
|                | Manganese | -0.47          | 7.0                | 7             | -13                    | 27                      | 09/12/2025       | 13:53            | LB137179      |
|                | Sodium    | 125            |                    |               | 0                      | 0                       | 09/12/2025       | 13:53            | LB137179      |
| <b>ICSAB01</b> | Iron      | 101000         | 99300              | 102           | 84400                  | 114500                  | 09/12/2025       | 13:58            | LB137179      |
|                | Manganese | 477            | 507                | 94            | 430                    | 584                     | 09/12/2025       | 13:58            | LB137179      |
|                | Sodium    | 30.4           |                    |               | 0                      | 0                       | 09/12/2025       | 13:58            | LB137179      |
| <b>ICSA01</b>  | Iron      | 98700          | 101000             | 98            | 85600                  | 116500                  | 09/15/2025       | 12:12            | LB137198      |
|                | Manganese | 9.55           | 7.0                | 136           | -13                    | 27                      | 09/15/2025       | 12:12            | LB137198      |
|                | Sodium    | 109            |                    |               | 0                      | 0                       | 09/15/2025       | 12:12            | LB137198      |
| <b>ICSAB01</b> | Iron      | 95600          | 99300              | 96            | 84400                  | 114500                  | 09/15/2025       | 12:16            | LB137198      |
|                | Manganese | 473            | 507                | 93            | 430                    | 584                     | 09/15/2025       | 12:16            | LB137198      |
|                | Sodium    | 116            |                    |               | 0                      | 0                       | 09/15/2025       | 12:16            | LB137198      |



# METAL QC DATA

**metals**  
**- 5a -**  
**MATRIX SPIKE SUMMARY**

|                                             |                                     |                                                   |
|---------------------------------------------|-------------------------------------|---------------------------------------------------|
| <b>client:</b> <u>G Environmental</u>       | <b>level:</b> <u>low</u>            | <b>sdg no.:</b> <u>Q3058</u>                      |
| <b>contract:</b> <u>GENV01</u>              |                                     | <b>lab code:</b> <u>ACE</u>                       |
| <b>matrix:</b> <u>Water</u>                 | <b>sample id:</b> <u>Q3071-01</u>   | <b>client id:</b> <u>4089MS</u>                   |
| <b>Percent Solids for Sample:</b> <u>NA</u> | <b>Spiked ID:</b> <u>Q3071-01MS</u> | <b>Percent Solids for Spike Sample:</b> <u>NA</u> |

| Analyte   | Units | Acceptance<br>Limit %R | Spiked<br>Result | C | Sample<br>Result | C | Spike<br>Added | %<br>Recovery | Qual | M |
|-----------|-------|------------------------|------------------|---|------------------|---|----------------|---------------|------|---|
| Iron      | ug/L  | 75 - 125               | 1590             |   | 1860             |   | 1500           | -17           | N    | P |
| Manganese | ug/L  | 75 - 125               | 108              |   | 33.1             |   | 100            | 75            |      | P |
| Sodium    | ug/L  | 75 - 125               | 31500            |   | 21100            |   | 1500           | 690           |      | P |

**metals**  
**- 5a -**  
**MATRIX SPIKE DUPLICATE SUMMARY**

|                                             |                                      |                                                   |
|---------------------------------------------|--------------------------------------|---------------------------------------------------|
| <b>client:</b> <u>G Environmental</u>       | <b>level:</b> <u>low</u>             | <b>sdg no.:</b> <u>Q3058</u>                      |
| <b>contract:</b> <u>GENV01</u>              |                                      | <b>lab code:</b> <u>ACE</u>                       |
| <b>matrix:</b> <u>Water</u>                 | <b>sample id:</b> <u>Q3071-01</u>    | <b>client id:</b> <u>4089MSD</u>                  |
| <b>Percent Solids for Sample:</b> <u>NA</u> | <b>Spiked ID:</b> <u>Q3071-01MSD</u> | <b>Percent Solids for Spike Sample:</b> <u>NA</u> |

| Analyte   | Units | Acceptance<br>Limit %R | MSD<br>Result | C | Sample<br>Result | C | Spike<br>Added | %<br>Recovery | Qual | M |
|-----------|-------|------------------------|---------------|---|------------------|---|----------------|---------------|------|---|
| Iron      | ug/L  | 75 - 125               | 1510          |   | 1860             |   | 1500           | -23           | N    | P |
| Manganese | ug/L  | 75 - 125               | 109           |   | 33.1             |   | 100            | 76            |      | P |
| Sodium    | ug/L  | 75 - 125               | 29800         |   | 21100            |   | 1500           | 580           |      | P |

**Metals**  
**- 5b -**  
**POST DIGEST SPIKE SUMMARY**

|                   |                        |                   |              |
|-------------------|------------------------|-------------------|--------------|
| <b>Client:</b>    | <u>G Environmental</u> | <b>SDG No.:</b>   | <u>Q3058</u> |
| <b>Contract:</b>  | <u>GENV01</u>          | <b>Lab Code:</b>  | <u>ACE</u>   |
| <b>Matrix:</b>    | <u>Water</u>           | <b>Level:</b>     | <u>LOW</u>   |
| <b>Sample ID:</b> | <u>Q3071-01</u>        | <b>Client ID:</b> | <u>4089A</u> |
|                   | <b>Spiked ID:</b>      | <u>Q3071-01A</u>  |              |

| Analyte | Units | Acceptance<br>Limit %R | Spiked<br>Result | C | Sample<br>Result | C | Spike<br>Added | %<br>Recovery | Qual | M |
|---------|-------|------------------------|------------------|---|------------------|---|----------------|---------------|------|---|
| Iron    | ug/L  | 75 - 125               | 3540             |   | 1860             |   | 1500           | 112           |      | P |

A

B

C

D

E

F

G

H

I

J

**Metals**

- 6 -

**DUPLICATE SAMPLE SUMMARY**

|                                   |                        |                     |                 |                                         |                |
|-----------------------------------|------------------------|---------------------|-----------------|-----------------------------------------|----------------|
| <b>Client:</b>                    | <u>G Environmental</u> | <b>Level:</b>       | <u>LOW</u>      | <b>SDG No.:</b>                         | <u>Q3058</u>   |
| <b>Contract:</b>                  | <u>GENV01</u>          |                     |                 | <b>Lab Code:</b>                        | <u>ACE</u>     |
| <b>Matrix:</b>                    | <u>Water</u>           | <b>Sample ID:</b>   | <u>Q3071-01</u> | <b>Client ID:</b>                       | <u>4089DUP</u> |
| <b>Percent Solids for Sample:</b> | NA                     | <b>Duplicate ID</b> | Q3071-01DUP     | <b>Percent Solids for Spike Sample:</b> | NA             |

| Analyte   | Units | Acceptance Limit | Sample Result | C | Duplicate Result | C | RPD | Qual | M |
|-----------|-------|------------------|---------------|---|------------------|---|-----|------|---|
| Iron      | ug/L  | 20               | 1860          |   | 2320             |   | 22  | *    | P |
| Manganese | ug/L  | 20               | 33.1          |   | 44.4             |   | 29  | *    | P |
| Sodium    | ug/L  | 20               | 21100         |   | 26900            |   | 24  | *    | P |

“A control limit of  $\pm 20\%$  RPD for each matrix applies for sample values greater than 10 times Detection Limit”

**Metals**

- 6 -

**DUPLICATE SAMPLE SUMMARY**

|                                   |                        |                     |                   |                                         |                |
|-----------------------------------|------------------------|---------------------|-------------------|-----------------------------------------|----------------|
| <b>Client:</b>                    | <u>G Environmental</u> | <b>Level:</b>       | <u>LOW</u>        | <b>SDG No.:</b>                         | <u>Q3058</u>   |
| <b>Contract:</b>                  | <u>GENV01</u>          |                     |                   | <b>Lab Code:</b>                        | <u>ACE</u>     |
| <b>Matrix:</b>                    | <u>Water</u>           | <b>Sample ID:</b>   | <u>Q3071-01MS</u> | <b>Client ID:</b>                       | <u>4089MSD</u> |
| <b>Percent Solids for Sample:</b> | NA                     | <b>Duplicate ID</b> | Q3071-01MSD       | <b>Percent Solids for Spike Sample:</b> | NA             |

| Analyte   | Units | Acceptance Limit | Sample Result | C | Duplicate Result | C | RPD | Qual | M |
|-----------|-------|------------------|---------------|---|------------------|---|-----|------|---|
| Iron      | ug/L  | 20               | 1590          |   | 1510             |   | 5   |      | P |
| Manganese | ug/L  | 20               | 108           |   | 109              |   | 1   |      | P |
| Sodium    | ug/L  | 20               | 31500         |   | 29800            |   | 6   |      | P |

“A control limit of  $\pm 20\%$  RPD for each matrix applies for sample values greater than 10 times Detection Limit”

**Metals**

- 7 -

**LABORATORY CONTROL SAMPLE SUMMARY**

|                  |                        |                  |              |
|------------------|------------------------|------------------|--------------|
| <b>Client:</b>   | <u>G Environmental</u> | <b>SDG No.:</b>  | <u>Q3058</u> |
| <b>Contract:</b> | <u>GENV01</u>          | <b>Lab Code:</b> | <u>ACE</u>   |

| Analyte    | Units | True Value | Result | C | % Recovery | Acceptance Limits | M |
|------------|-------|------------|--------|---|------------|-------------------|---|
| PB169680BS |       |            |        |   |            |                   |   |
| Iron       | ug/L  | 1500       | 1480   |   | 99         | 80 - 120          | P |
| Manganese  | ug/L  | 100        | 100    |   | 100        | 80 - 120          | P |
| Sodium     | ug/L  | 1500       | 1490   |   | 99         | 80 - 120          | P |

**Metals**  
**-9 -**  
**ICP SERIAL DILUTIONS**

**SAMPLE NO.**

4089L

**Lab Name:** Alliance **Contract:** GENV01  
**Lab Code:** ACE **Lb No.:** lb137179 **Lab Sample ID :** Q3071-01L **SDG No.:** Q3058  
**Matrix (soil/water):** Water **Level (low/med):** LOW  
**Concentration Units:** ug/L

| Analyte   | Initial Sample<br>Result (I)<br>C | Serial Dilution<br>Result (S)<br>C | % Differ-<br>ence | Q | M |
|-----------|-----------------------------------|------------------------------------|-------------------|---|---|
| Iron      | 1860                              | 2170                               | 17                |   | P |
| Manganese | 33.1                              | 38.5 J                             | 16                |   | P |
| Sodium    | 21100                             | 24900                              | 18                |   | P |

A

B

C

D

E

F

G

H

I

J

**metals**  
**- 14 -**  
**ANALYSIS RUN LOG**

**Client:** G Environmental

**Contract:** GENV01

**Lab code:** ACE

**Sdg no.:** Q3058

**Instrument id number:** \_\_\_\_\_

**Method:** \_\_\_\_\_

**Run number:** LB137179

**Start date:** 09/12/2025

**End date:** 09/12/2025

| Lab sample id. | Client Sample Id | d/f | Time | Parameter list |
|----------------|------------------|-----|------|----------------|
| S0             | S0               | 1   | 1309 | Fe,Mn,Na       |
| S1             | S1               | 1   | 1313 | Fe,Mn,Na       |
| S2             | S2               | 1   | 1318 | Fe,Mn,Na       |
| S3             | S3               | 1   | 1322 | Fe,Mn,Na       |
| S4             | S4               | 1   | 1326 | Fe,Mn,Na       |
| S5             | S5               | 1   | 1330 | Fe,Mn,Na       |
| ICV01          | ICV01            | 1   | 1334 | Fe,Mn,Na       |
| LLICV01        | LLICV01          | 1   | 1338 | Fe,Mn,Na       |
| ICB01          | ICB01            | 1   | 1342 | Fe,Mn,Na       |
| CRI01          | CRI01            | 1   | 1347 | Fe,Mn,Na       |
| ICSA01         | ICSA01           | 1   | 1353 | Fe,Mn,Na       |
| ICSAB01        | ICSAB01          | 1   | 1358 | Fe,Mn,Na       |
| CCV01          | CCV01            | 1   | 1410 | Fe,Mn,Na       |
| CCB01          | CCB01            | 1   | 1414 | Fe,Mn,Na       |
| CCV02          | CCV02            | 1   | 1508 | Fe,Mn,Na       |
| CCB02          | CCB02            | 1   | 1512 | Fe,Mn,Na       |
| CCV03          | CCV03            | 1   | 1559 | Fe,Mn,Na       |
| CCB03          | CCB03            | 1   | 1603 | Fe,Mn,Na       |
| CCV04          | CCV04            | 1   | 1651 | Fe,Mn,Na       |
| CCB04          | CCB04            | 1   | 1714 | Fe,Mn,Na       |
| CCV05          | CCV05            | 1   | 1800 | Fe,Mn,Na       |
| CCB05          | CCB05            | 1   | 1804 | Fe,Mn,Na       |
| Q3071-01DUP    | 4089DUP          | 1   | 1809 | Fe,Mn,Na       |
| Q3071-01L      | 4089L            | 5   | 1813 | Fe,Mn,Na       |
| Q3071-01MS     | 4089MS           | 1   | 1817 | Fe,Mn,Na       |
| Q3071-01MSD    | 4089MSD          | 1   | 1821 | Fe,Mn,Na       |
| Q3071-01A      | 4089A            | 1   | 1825 | Fe             |
| Q3058-03       | MW4              | 1   | 1829 | Fe,Mn,Na       |
| CCV06          | CCV06            | 1   | 1847 | Fe,Mn,Na       |
| CCB06          | CCB06            | 1   | 1851 | Fe,Mn,Na       |

**metals**  
**- 14 -**  
**ANALYSIS RUN LOG**

**Client:** G Environmental

**Contract:** GENV01

**Lab code:** ACE

**Sdg no.:** Q3058

**Instrument id number:** \_\_\_\_\_

**Method:** \_\_\_\_\_

**Run number:** LB137198

**Start date:** 09/15/2025

**End date:** 09/15/2025

| Lab sample id. | Client Sample Id | d/f | Time | Parameter list |
|----------------|------------------|-----|------|----------------|
| S0             | S0               | 1   | 1106 | Fe,Mn,Na       |
| S1             | S1               | 1   | 1110 | Fe,Mn,Na       |
| S2             | S2               | 1   | 1114 | Fe,Mn,Na       |
| S3             | S3               | 1   | 1118 | Fe,Mn,Na       |
| S4             | S4               | 1   | 1122 | Fe,Mn,Na       |
| S5             | S5               | 1   | 1126 | Fe,Mn,Na       |
| ICV01          | ICV01            | 1   | 1131 | Fe,Mn,Na       |
| LLICV01        | LLICV01          | 1   | 1136 | Fe,Mn,Na       |
| ICB01          | ICB01            | 1   | 1143 | Fe,Mn,Na       |
| CRI01          | CRI01            | 1   | 1147 | Fe,Mn,Na       |
| ICSA01         | ICSA01           | 1   | 1212 | Fe,Mn,Na       |
| ICSAB01        | ICSAB01          | 1   | 1216 | Fe,Mn,Na       |
| CCV01          | CCV01            | 1   | 1235 | Fe,Mn,Na       |
| CCB01          | CCB01            | 1   | 1239 | Fe,Mn,Na       |
| PB169680BL     | PB169680BL       | 1   | 1258 | Fe,Mn,Na       |
| PB169680BS     | PB169680BS       | 1   | 1302 | Fe,Mn,Na       |
| CCV02          | CCV02            | 1   | 1342 | Fe,Mn,Na       |
| CCB02          | CCB02            | 1   | 1347 | Fe,Mn,Na       |
| CCV03          | CCV03            | 1   | 1511 | Fe,Mn,Na       |
| CCB03          | CCB03            | 1   | 1515 | Fe,Mn,Na       |
| CCV04          | CCV04            | 1   | 1612 | Fe,Mn,Na       |
| CCB04          | CCB04            | 1   | 1616 | Fe,Mn,Na       |
| CCV05          | CCV05            | 1   | 1655 | Fe,Mn,Na       |
| CCB05          | CCB05            | 1   | 1659 | Fe,Mn,Na       |



# METAL PREPARATION & INSTRUMENT DATA

**Metals**

- 11 -

**ICP INTERELEMEN CORRECTION FACTORS**

**Client:** G Environmental

**SDG No.:** Q3058

**Contract:** GENV01

**Lab Code:** ACE

**Instrument ID:** \_\_\_\_\_

**Date:** \_\_\_\_\_

**Interement Correction Factors (apparent ppb analyte/ppm interferent )**

| Analyte   | Wave-<br>Length<br>(nm) | ICP Interement Correction Factors For: |           |           |           |           |
|-----------|-------------------------|----------------------------------------|-----------|-----------|-----------|-----------|
|           |                         | Al                                     | Ca        | Fe        | Mg        | Ag        |
| Iron      | 240.488                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000 |
| Manganese | 257.610                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000 |
| Sodium    | 589.592                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000 |

**Metals**

- 11 -

**ICP INTERELEMEN CORRECTION FACTORS**

**Client:** G Environmental

**SDG No.:** Q3058

**Contract:** GENV01

**Lab Code:** ACE

**Instrument ID:** \_\_\_\_\_

**Date:** \_\_\_\_\_

**Interement Correction Factors (apparent ppb analyte/ppm interferent )**

| Analyte   | Wave-<br>Length<br>(nm) | ICP Interement Correction Factors For: |           |           |           |            |
|-----------|-------------------------|----------------------------------------|-----------|-----------|-----------|------------|
|           |                         | As                                     | Ba        | Be        | Cd        | Co         |
| Iron      | 240.488                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | -0.0039600 |
| Manganese | 257.610                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |
| Sodium    | 589.592                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |

A

B

C

D

E

F

G

H

I

J

**Metals**

- 11 -

**ICP INTERELEMEN CORRECTION FACTORS**

**Client:** G Environmental

**SDG No.:** Q3058

**Contract:** GENV01

**Lab Code:** ACE

**Instrument ID:**

**Date:**

**Interement Correction Factors (apparent ppb analyte/ppm interferent )**

| Analyte   | Wave-<br>Length<br>(nm) | ICP Interement Correction Factors For: |           |           |           |            |
|-----------|-------------------------|----------------------------------------|-----------|-----------|-----------|------------|
|           |                         | Cr                                     | Cu        | K         | Mn        | Mo         |
| Iron      | 240.488                 | 0.0000000                              | 0.0000000 | 0.0000730 | 0.0000000 | -0.0015250 |
| Manganese | 257.610                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |
| Sodium    | 589.592                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |

**Metals**

- 11 -

**ICP INTERELEMENT CORRECTION FACTORS**

**Client:** G Environmental

**SDG No.:** Q3058

**Contract:** GENV01

**Lab Code:** ACE

**Instrument ID:** \_\_\_\_\_

**Date:** \_\_\_\_\_

**Interelement Correction Factors (apparent ppb analyte/ppm interferent )**

| Analyte   | Wave-<br>Length<br>(nm) | ICP Interelement Correction Factors For: |            |           |           |           |
|-----------|-------------------------|------------------------------------------|------------|-----------|-----------|-----------|
|           |                         | Na                                       | Ni         | Pb        | Sb        | Se        |
| Iron      | 240.488                 | 0.0000000                                | -0.0017000 | 0.0000000 | 0.0000000 | 0.0000000 |
| Manganese | 257.610                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Sodium    | 589.592                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |

**Metals**

- 11 -

**ICP INTERELEMEN CORRECTION FACTORS**

**Client:** G Environmental

**SDG No.:** Q3058

**Contract:** GENV01

**Lab Code:** ACE

**Instrument ID:** \_\_\_\_\_

**Date:** \_\_\_\_\_

**Interement Correction Factors (apparent ppb analyte/ppm interferent )**

| Analyte   | Wave-<br>Length<br>(nm) | ICP Interement Correction Factors For: |           |           |           |           |
|-----------|-------------------------|----------------------------------------|-----------|-----------|-----------|-----------|
|           |                         | Sn                                     | Ti        | Tl        | V         | Zn        |
| Iron      | 240.488                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000 |
| Manganese | 257.610                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000 |
| Sodium    | 589.592                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000 |

LAB CHRONICLE

|          |                 |            |                     |
|----------|-----------------|------------|---------------------|
| OrderID: | Q3058           | OrderDate: | 9/9/2025 1:49:00 PM |
| Client:  | G Environmental | Project:   | Buffington          |
| Contact: | Gary Landis     | Location:  | J42,VOA Lab         |

| LabID    | ClientID | Matrix | Test          | Method | Sample Date | Prep Date | Anal Date | Received |
|----------|----------|--------|---------------|--------|-------------|-----------|-----------|----------|
| Q3058-03 | MW4      | Water  | Metals Group4 | 6010D  | 09/09/25    | 09/12/25  | 09/12/25  | 09/09/25 |



# METAL PREPARATION & ANALYICAL SUMMARY

**Metals**  
**- 13 -**

**SAMPLE PREPARATION SUMMARY**

**Client:** G Environmental **SDG No.:** Q3058  
**Contract:** GENV01 **Lab Code:** ACE **Method:**

| Sample ID                     | Client ID  | Sample Type | Matrix | Prep Date  | Initial Sample Size(mL) | Final Sample Volume (mL) | Percent Solids |
|-------------------------------|------------|-------------|--------|------------|-------------------------|--------------------------|----------------|
| <b>Batch Number: PB169680</b> |            |             |        |            |                         |                          |                |
| PB169680BL                    | PB169680BL | MB          | WATER  | 09/12/2025 | 50.0                    | 25.0                     |                |
| PB169680BS                    | PB169680BS | LCS         | WATER  | 09/12/2025 | 50.0                    | 25.0                     |                |
| Q3058-03                      | MW4        | SAM         | WATER  | 09/12/2025 | 50.0                    | 25.0                     |                |
| Q3071-01DUP                   | 4089DUP    | DUP         | WATER  | 09/12/2025 | 50.0                    | 25.0                     |                |
| Q3071-01MS                    | 4089MS     | MS          | WATER  | 09/12/2025 | 50.0                    | 25.0                     |                |
| Q3071-01MSD                   | 4089MSD    | MSD         | WATER  | 09/12/2025 | 50.0                    | 25.0                     |                |

Instrument ID: P4

**Daily Analysis Runlog For Sequence/QC Batch ID # LB137179**

|                  |                                                 |              |                       |
|------------------|-------------------------------------------------|--------------|-----------------------|
| Review By        | Janvi                                           | Review On    | 9/15/2025 10:21:14 AM |
| Supervise By     | jaswal                                          | Supervise On | 9/16/2025 11:52:32 AM |
| <b>STD. NAME</b> | <b>STD REF.#</b>                                |              |                       |
| ICAL Standard    | MP87031,MP87038,MP87035,MP87034,MP87033,MP87032 |              |                       |
| ICV Standard     | MP87047,MP87038                                 |              |                       |
| CCV Standard     | MP87042                                         |              |                       |
| ICSA Standard    | MP87040,MP87048                                 |              |                       |
| CRI Standard     | MP87038                                         |              |                       |
| LCS Standard     |                                                 |              |                       |
| Chk Standard     | MP87043,MP87044,MP87045,MP87046,MP87049         |              |                       |

| Sr# | SampleId | ClientID | QcType   | Date           | Comment | Operator | Status |
|-----|----------|----------|----------|----------------|---------|----------|--------|
| 1   | S0       | S0       | CAL1     | 09/12/25 13:09 |         | Jaswal   | OK     |
| 2   | S1       | S1       | CAL2     | 09/12/25 13:13 |         | Jaswal   | OK     |
| 3   | S2       | S2       | CAL3     | 09/12/25 13:18 |         | Jaswal   | OK     |
| 4   | S3       | S3       | CAL4     | 09/12/25 13:22 |         | Jaswal   | OK     |
| 5   | S4       | S4       | CAL5     | 09/12/25 13:26 |         | Jaswal   | OK     |
| 6   | S5       | S5       | CAL6     | 09/12/25 13:30 |         | Jaswal   | OK     |
| 7   | ICV01    | ICV01    | ICV      | 09/12/25 13:34 |         | Jaswal   | OK     |
| 8   | LLICV01  | LLICV01  | LLICV    | 09/12/25 13:38 |         | Jaswal   | OK     |
| 9   | ICB01    | ICB01    | ICB      | 09/12/25 13:42 |         | Jaswal   | OK     |
| 10  | CRI01    | CRI01    | CRDL     | 09/12/25 13:47 |         | Jaswal   | OK     |
| 11  | ICSA01   | ICSA01   | ICSA     | 09/12/25 13:53 |         | Jaswal   | OK     |
| 12  | ICSAB01  | ICSAB01  | ICSAB    | 09/12/25 13:58 |         | Jaswal   | OK     |
| 13  | ICSADL   | ICSADL   | ICSA     | 09/12/25 14:02 |         | Jaswal   | OK     |
| 14  | ICSABDL  | ICSABDL  | ICSAB    | 09/12/25 14:06 |         | Jaswal   | OK     |
| 15  | CCV01    | CCV01    | CCV      | 09/12/25 14:10 |         | Jaswal   | OK     |
| 16  | CCB01    | CCB01    | CCB      | 09/12/25 14:14 |         | Jaswal   | OK     |
| 17  | LR-1     | LR-1     | HIGH STD | 09/12/25 14:19 |         | Jaswal   | OK     |
| 18  | LR-2     | LR-2     | HIGH STD | 09/12/25 14:23 |         | Jaswal   | OK     |

Instrument ID: P4

**Daily Analysis Runlog For Sequence/QC Batch ID # LB137179**

| Review By     | Janvi                                           | Review On    | 9/15/2025 10:21:14 AM |
|---------------|-------------------------------------------------|--------------|-----------------------|
| Supervise By  | jaswal                                          | Supervise On | 9/16/2025 11:52:32 AM |
| STD. NAME     | STD REF.#                                       |              |                       |
| ICAL Standard | MP87031,MP87038,MP87035,MP87034,MP87033,MP87032 |              |                       |
| ICV Standard  | MP87047,MP87038                                 |              |                       |
| CCV Standard  | MP87042                                         |              |                       |
| ICSA Standard | MP87040,MP87048                                 |              |                       |
| CRI Standard  | MP87038                                         |              |                       |
| LCS Standard  |                                                 |              |                       |
| Chk Standard  | MP87043,MP87044,MP87045,MP87046,MP87049         |              |                       |

|    |             |                   |     |                |                                  |        |    |
|----|-------------|-------------------|-----|----------------|----------------------------------|--------|----|
| 19 | PB169652BS  | PB169652BS        | LCS | 09/12/25 14:47 |                                  | Jaswal | OK |
| 20 | Q3065-01    | VNJ-204           | SAM | 09/12/25 14:51 |                                  | Jaswal | OK |
| 21 | Q3066-01    | 354               | SAM | 09/12/25 14:55 |                                  | Jaswal | OK |
| 22 | Q3066-03    | VNJ-242           | SAM | 09/12/25 14:59 |                                  | Jaswal | OK |
| 23 | PB169652BL  | PB169652BL        | MB  | 09/12/25 15:03 |                                  | Jaswal | OK |
| 24 | CCV02       | CCV02             | CCV | 09/12/25 15:08 |                                  | Jaswal | OK |
| 25 | CCB02       | CCB02             | CCB | 09/12/25 15:12 |                                  | Jaswal | OK |
| 26 | Q3072-01DUP | 30-GARFIELD-PL-CO | DUP | 09/12/25 15:17 |                                  | Jaswal | OK |
| 27 | Q3072-01L   | 30-GARFIELD-PL-CO | SD  | 09/12/25 15:21 |                                  | Jaswal | OK |
| 28 | Q3072-01MS  | 30-GARFIELD-PL-CO | MS  | 09/12/25 15:25 |                                  | Jaswal | OK |
| 29 | Q3072-01MSD | 30-GARFIELD-PL-CO | MSD | 09/12/25 15:29 |                                  | Jaswal | OK |
| 30 | Q3072-01A   | 30-GARFIELD-PL-CO | PS  | 09/12/25 15:33 | 0.1ml M6008,M6017-10ML<br>SAMPLE | Jaswal | OK |
| 31 | PB169660BL  | PB169660BL        | MB  | 09/12/25 15:38 |                                  | Jaswal | OK |
| 32 | PB169660BS  | PB169660BS        | LCS | 09/12/25 15:42 |                                  | Jaswal | OK |
| 33 | Q3072-01    | 30-GARFIELD-PL-CO | SAM | 09/12/25 15:46 |                                  | Jaswal | OK |
| 34 | Q3054-08    | MH-384            | SAM | 09/12/25 15:50 |                                  | Jaswal | OK |
| 35 | Q3066-06    | 354               | SAM | 09/12/25 15:55 |                                  | Jaswal | OK |
| 36 | CCV03       | CCV03             | CCV | 09/12/25 15:59 |                                  | Jaswal | OK |
| 37 | CCB03       | CCB03             | CCB | 09/12/25 16:03 |                                  | Jaswal | OK |

Instrument ID: P4

**Daily Analysis Runlog For Sequence/QC Batch ID # LB137179**

|                  |                                                 |              |                       |
|------------------|-------------------------------------------------|--------------|-----------------------|
| Review By        | Janvi                                           | Review On    | 9/15/2025 10:21:14 AM |
| Supervise By     | jaswal                                          | Supervise On | 9/16/2025 11:52:32 AM |
| <b>STD. NAME</b> | <b>STD REF.#</b>                                |              |                       |
| ICAL Standard    | MP87031,MP87038,MP87035,MP87034,MP87033,MP87032 |              |                       |
| ICV Standard     | MP87047,MP87038                                 |              |                       |
| CCV Standard     | MP87042                                         |              |                       |
| ICSA Standard    | MP87040,MP87048                                 |              |                       |
| CRI Standard     | MP87038                                         |              |                       |
| LCS Standard     |                                                 |              |                       |
| Chk Standard     | MP87043,MP87044,MP87045,MP87046,MP87049         |              |                       |

|    |             |                   |     |                |                               |        |    |
|----|-------------|-------------------|-----|----------------|-------------------------------|--------|----|
| 38 | Q3066-07    | VNJ-242           | SAM | 09/12/25 16:08 |                               | Jaswal | OK |
| 39 | Q3072-02    | 30-GARFIELD-PL-CO | SAM | 09/12/25 16:12 |                               | Jaswal | OK |
| 40 | Q3073-03    | BEL-25-0023       | SAM | 09/12/25 16:16 |                               | Jaswal | OK |
| 41 | Q3074-05    | MOO-25-0141       | SAM | 09/12/25 16:21 |                               | Jaswal | OK |
| 42 | Q3075-03    | MOO-25-0254-0256  | SAM | 09/12/25 16:25 |                               | Jaswal | OK |
| 43 | Q3075-05    | MOO-25-0258       | SAM | 09/12/25 16:29 |                               | Jaswal | OK |
| 44 | Q3075-07    | MOO-25-0259       | SAM | 09/12/25 16:34 |                               | Jaswal | OK |
| 45 | Q3075-09    | MOO-25-0260       | SAM | 09/12/25 16:38 |                               | Jaswal | OK |
| 46 | Q3075-12    | MOO-25-0261-0263  | SAM | 09/12/25 16:42 |                               | Jaswal | OK |
| 47 | Q3077-02    | LAW-25-0137       | SAM | 09/12/25 16:47 |                               | Jaswal | OK |
| 48 | CCV04       | CCV04             | CCV | 09/12/25 16:51 |                               | Jaswal | OK |
| 49 | CCB04       | CCB04             | CCB | 09/12/25 17:14 |                               | Jaswal | OK |
| 50 | Q3077-04    | LAW-25-0139       | SAM | 09/12/25 17:18 |                               | Jaswal | OK |
| 51 | Q3077-04DUP | LAW-25-0139DUP    | DUP | 09/12/25 17:22 |                               | Jaswal | OK |
| 52 | Q3077-04L   | LAW-25-0139L      | SD  | 09/12/25 17:27 |                               | Jaswal | OK |
| 53 | Q3077-04MS  | LAW-25-0139MS     | MS  | 09/12/25 17:31 |                               | Jaswal | OK |
| 54 | Q3077-04MSD | LAW-25-0139MSD    | MSD | 09/12/25 17:35 |                               | Jaswal | OK |
| 55 | Q3077-04A   | LAW-25-0139A      | PS  | 09/12/25 17:39 | 0.1ml M6008,M6017-10ML SAMPLE | Jaswal | OK |
| 56 | PB169640TB  | PB169640TB        | MB  | 09/12/25 17:43 |                               | Jaswal | OK |

Instrument ID: P4

**Daily Analysis Runlog For Sequence/QC Batch ID # LB137179**

|                  |                                                 |              |                       |
|------------------|-------------------------------------------------|--------------|-----------------------|
| Review By        | Janvi                                           | Review On    | 9/15/2025 10:21:14 AM |
| Supervise By     | jaswal                                          | Supervise On | 9/16/2025 11:52:32 AM |
| <b>STD. NAME</b> | <b>STD REF.#</b>                                |              |                       |
| ICAL Standard    | MP87031,MP87038,MP87035,MP87034,MP87033,MP87032 |              |                       |
| ICV Standard     | MP87047,MP87038                                 |              |                       |
| CCV Standard     | MP87042                                         |              |                       |
| ICSA Standard    | MP87040,MP87048                                 |              |                       |
| CRI Standard     | MP87038                                         |              |                       |
| LCS Standard     |                                                 |              |                       |
| Chk Standard     | MP87043,MP87044,MP87045,MP87046,MP87049         |              |                       |

|    |             |              |          |                |                                  |        |    |
|----|-------------|--------------|----------|----------------|----------------------------------|--------|----|
| 57 | Q3071-01    | 4089         | SAM      | 09/12/25 17:56 |                                  | Jaswal | OK |
| 58 | CCV05       | CCV05        | CCV      | 09/12/25 18:00 |                                  | Jaswal | OK |
| 59 | CCB05       | CCB05        | CCB      | 09/12/25 18:04 |                                  | Jaswal | OK |
| 60 | Q3071-01DUP | 4089DUP      | DUP      | 09/12/25 18:09 |                                  | Jaswal | OK |
| 61 | Q3071-01L   | 4089L        | SD       | 09/12/25 18:13 |                                  | Jaswal | OK |
| 62 | Q3071-01MS  | 4089MS       | MS       | 09/12/25 18:17 |                                  | Jaswal | OK |
| 63 | Q3071-01MSD | 4089MSD      | MSD      | 09/12/25 18:21 |                                  | Jaswal | OK |
| 64 | Q3071-01A   | 4089A        | PS       | 09/12/25 18:25 | 0.1ml M6008,M6017-10ML<br>SAMPLE | Jaswal | OK |
| 65 | Q3058-03    | MW4          | SAM      | 09/12/25 18:29 |                                  | Jaswal | OK |
| 66 | LR-3        | LR-3         | HIGH STD | 09/12/25 18:33 |                                  | Jaswal | OK |
| 67 | Q3054-04    | MH-385       | SAM      | 09/12/25 18:38 |                                  | Jaswal | OK |
| 68 | Q3080-01    | EO-03-091125 | SAM      | 09/12/25 18:43 |                                  | Jaswal | OK |
| 69 | CCV06       | CCV06        | CCV      | 09/12/25 18:47 |                                  | Jaswal | OK |
| 70 | CCB06       | CCB06        | CCB      | 09/12/25 18:51 |                                  | Jaswal | OK |

Instrument ID: P4

**Daily Analysis Runlog For Sequence/QC Batch ID # LB137198**

|                  |                                                 |              |                       |
|------------------|-------------------------------------------------|--------------|-----------------------|
| Review By        | Janvi                                           | Review On    | 9/16/2025 11:58:31 AM |
| Supervise By     | jaswal                                          | Supervise On | 9/16/2025 12:00:47 PM |
| <b>STD. NAME</b> | <b>STD REF.#</b>                                |              |                       |
| ICAL Standard    | MP87031,MP87038,MP87035,MP87034,MP87033,MP87032 |              |                       |
| ICV Standard     | MP87047,MP87038                                 |              |                       |
| CCV Standard     | MP87042                                         |              |                       |
| ICSA Standard    | MP87040,MP87048                                 |              |                       |
| CRI Standard     | MP87038                                         |              |                       |
| LCS Standard     |                                                 |              |                       |
| Chk Standard     | MP87043,MP87044,MP87045,MP87046,MP87049         |              |                       |

| Sr# | SampleId   | ClientID   | QcType   | Date           | Comment | Operator | Status |
|-----|------------|------------|----------|----------------|---------|----------|--------|
| 1   | S0         | S0         | CAL1     | 09/15/25 11:06 |         | Jaswal   | OK     |
| 2   | S1         | S1         | CAL2     | 09/15/25 11:10 |         | Jaswal   | OK     |
| 3   | S2         | S2         | CAL3     | 09/15/25 11:14 |         | Jaswal   | OK     |
| 4   | S3         | S3         | CAL4     | 09/15/25 11:18 |         | Jaswal   | OK     |
| 5   | S4         | S4         | CAL5     | 09/15/25 11:22 |         | Jaswal   | OK     |
| 6   | S5         | S5         | CAL6     | 09/15/25 11:26 |         | Jaswal   | OK     |
| 7   | ICV01      | ICV01      | ICV      | 09/15/25 11:31 |         | Jaswal   | OK     |
| 8   | LLICV01    | LLICV01    | LLICV    | 09/15/25 11:36 |         | Jaswal   | OK     |
| 9   | ICB01      | ICB01      | ICB      | 09/15/25 11:43 |         | Jaswal   | OK     |
| 10  | CRI01      | CRI01      | CRDL     | 09/15/25 11:47 |         | Jaswal   | OK     |
| 11  | ICSA01     | ICSA01     | ICSA     | 09/15/25 12:12 |         | Jaswal   | OK     |
| 12  | ICSAB01    | ICSAB01    | ICSAB    | 09/15/25 12:16 |         | Jaswal   | OK     |
| 13  | ICSADL     | ICSADL     | ICSA     | 09/15/25 12:26 |         | Jaswal   | OK     |
| 14  | ICSABDL    | ICSABDL    | ICSAB    | 09/15/25 12:30 |         | Jaswal   | OK     |
| 15  | CCV01      | CCV01      | CCV      | 09/15/25 12:35 |         | Jaswal   | OK     |
| 16  | CCB01      | CCB01      | CCB      | 09/15/25 12:39 |         | Jaswal   | OK     |
| 17  | LR-1       | LR-1       | HIGH STD | 09/15/25 12:44 |         | Jaswal   | OK     |
| 18  | PB169680BL | PB169680BL | MB       | 09/15/25 12:58 |         | Jaswal   | OK     |

Instrument ID: P4

**Daily Analysis Runlog For Sequence/QC Batch ID # LB137198**

|                  |                                                 |              |                       |
|------------------|-------------------------------------------------|--------------|-----------------------|
| Review By        | Janvi                                           | Review On    | 9/16/2025 11:58:31 AM |
| Supervise By     | jaswal                                          | Supervise On | 9/16/2025 12:00:47 PM |
| <b>STD. NAME</b> | <b>STD REF.#</b>                                |              |                       |
| ICAL Standard    | MP87031,MP87038,MP87035,MP87034,MP87033,MP87032 |              |                       |
| ICV Standard     | MP87047,MP87038                                 |              |                       |
| CCV Standard     | MP87042                                         |              |                       |
| ICSA Standard    | MP87040,MP87048                                 |              |                       |
| CRI Standard     | MP87038                                         |              |                       |
| LCS Standard     |                                                 |              |                       |
| Chk Standard     | MP87043,MP87044,MP87045,MP87046,MP87049         |              |                       |

|    |             |            |          |                |                                           |        |    |
|----|-------------|------------|----------|----------------|-------------------------------------------|--------|----|
| 19 | PB169680BS  | PB169680BS | LCS      | 09/15/25 13:02 |                                           | Jaswal | OK |
| 20 | PB169682BL  | PB169682BL | MB       | 09/15/25 13:06 |                                           | Jaswal | OK |
| 21 | PB169682BS  | PB169682BS | LCS      | 09/15/25 13:10 |                                           | Jaswal | OK |
| 22 | Q3082-01    | WC-29      | SAM      | 09/15/25 13:14 |                                           | Jaswal | OK |
| 23 | Q3082-03    | WC-30      | SAM      | 09/15/25 13:18 |                                           | Jaswal | OK |
| 24 | Q3082-05    | WC-31      | SAM      | 09/15/25 13:22 |                                           | Jaswal | OK |
| 25 | CCV02       | CCV02      | CCV      | 09/15/25 13:42 |                                           | Jaswal | OK |
| 26 | CCB02       | CCB02      | CCB      | 09/15/25 13:47 |                                           | Jaswal | OK |
| 27 | Q3082-07    | WC-32      | SAM      | 09/15/25 13:51 |                                           | Jaswal | OK |
| 28 | Q3082-09    | WC-33      | SAM      | 09/15/25 13:55 |                                           | Jaswal | OK |
| 29 | Q3082-01DUP | WC-29DUP   | DUP      | 09/15/25 13:59 |                                           | Jaswal | OK |
| 30 | Q3082-01L   | WC-29L     | SD       | 09/15/25 14:03 |                                           | Jaswal | OK |
| 31 | Q3082-01MS  | WC-29MS    | MS       | 09/15/25 14:07 |                                           | Jaswal | OK |
| 32 | Q3082-01MSD | WC-29MSD   | MSD      | 09/15/25 14:11 |                                           | Jaswal | OK |
| 33 | Q3082-01A   | WC-29A     | PS       | 09/15/25 14:15 | 0.1 ML OF M6008 AND M6017 IN 10ML SAMPLE. | Jaswal | OK |
| 34 | LR-2        | LR-2       | HIGH STD | 09/15/25 14:40 |                                           | Jaswal | OK |
| 35 | LR-3        | LR-3       | HIGH STD | 09/15/25 14:48 |                                           | Jaswal | OK |
| 36 | CCV03       | CCV03      | CCV      | 09/15/25 15:11 |                                           | Jaswal | OK |
| 37 | CCB03       | CCB03      | CCB      | 09/15/25 15:15 |                                           | Jaswal | OK |

Instrument ID: P4

**Daily Analysis Runlog For Sequence/QC Batch ID # LB137198**

|                  |                                                 |              |                       |
|------------------|-------------------------------------------------|--------------|-----------------------|
| Review By        | Janvi                                           | Review On    | 9/16/2025 11:58:31 AM |
| Supervise By     | jaswal                                          | Supervise On | 9/16/2025 12:00:47 PM |
| <b>STD. NAME</b> | <b>STD REF.#</b>                                |              |                       |
| ICAL Standard    | MP87031,MP87038,MP87035,MP87034,MP87033,MP87032 |              |                       |
| ICV Standard     | MP87047,MP87038                                 |              |                       |
| CCV Standard     | MP87042                                         |              |                       |
| ICSA Standard    | MP87040,MP87048                                 |              |                       |
| CRI Standard     | MP87038                                         |              |                       |
| LCS Standard     |                                                 |              |                       |
| Chk Standard     | MP87043,MP87044,MP87045,MP87046,MP87049         |              |                       |

|    |             |              |     |                |                                           |        |    |
|----|-------------|--------------|-----|----------------|-------------------------------------------|--------|----|
| 38 | PB169685BL  | PB169685BL   | MB  | 09/15/25 15:19 |                                           | Jaswal | OK |
| 39 | PB169685BS  | PB169685BS   | LCS | 09/15/25 15:23 |                                           | Jaswal | OK |
| 40 | Q3082-02    | WC-29        | SAM | 09/15/25 15:27 |                                           | Jaswal | OK |
| 41 | Q3082-04    | WC-30        | SAM | 09/15/25 15:32 |                                           | Jaswal | OK |
| 42 | Q3082-06    | WC-31        | SAM | 09/15/25 15:36 |                                           | Jaswal | OK |
| 43 | Q3082-08    | WC-32        | SAM | 09/15/25 15:41 |                                           | Jaswal | OK |
| 44 | Q3082-10    | WC-33        | SAM | 09/15/25 15:45 |                                           | Jaswal | OK |
| 45 | Q3085-01    | PAULDING-WC1 | SAM | 09/15/25 15:49 |                                           | Jaswal | OK |
| 46 | Q3091-07    | 12/9C-COMP   | SAM | 09/15/25 15:54 |                                           | Jaswal | OK |
| 47 | Q3091-14    | 12/9B-COMP   | SAM | 09/15/25 15:58 |                                           | Jaswal | OK |
| 48 | CCV04       | CCV04        | CCV | 09/15/25 16:12 |                                           | Jaswal | OK |
| 49 | CCB04       | CCB04        | CCB | 09/15/25 16:16 |                                           | Jaswal | OK |
| 50 | Q3092-02    | 291376       | SAM | 09/15/25 16:21 |                                           | Jaswal | OK |
| 51 | Q3092-04    | 279182       | SAM | 09/15/25 16:25 |                                           | Jaswal | OK |
| 52 | Q3092-04DUP | 279182DUP    | DUP | 09/15/25 16:29 |                                           | Jaswal | OK |
| 53 | Q3092-04L   | 279182L      | SD  | 09/15/25 16:34 |                                           | Jaswal | OK |
| 54 | Q3092-04MS  | 279182MS     | MS  | 09/15/25 16:38 |                                           | Jaswal | OK |
| 55 | Q3092-04MSD | 279182MSD    | MSD | 09/15/25 16:42 |                                           | Jaswal | OK |
| 56 | Q3092-04A   | 279182A      | PS  | 09/15/25 16:46 | 0.1 ML OF M6008 AND M6017 IN 10ML SAMPLE. | Jaswal | OK |

Instrument ID: P4

**Daily Analysis Runlog For Sequence/QC Batch ID # LB137198**

|              |        |              |                       |
|--------------|--------|--------------|-----------------------|
| Review By    | Janvi  | Review On    | 9/16/2025 11:58:31 AM |
| Supervise By | jaswal | Supervise On | 9/16/2025 12:00:47 PM |

| STD. NAME     | STD REF.#                                       |
|---------------|-------------------------------------------------|
| ICAL Standard | MP87031,MP87038,MP87035,MP87034,MP87033,MP87032 |
| ICV Standard  | MP87047,MP87038                                 |
| CCV Standard  | MP87042                                         |
| ICSA Standard | MP87040,MP87048                                 |
| CRI Standard  | MP87038                                         |
| LCS Standard  |                                                 |
| Chk Standard  | MP87043,MP87044,MP87045,MP87046,MP87049         |

|    |            |            |     |                |  |        |    |
|----|------------|------------|-----|----------------|--|--------|----|
| 57 | PB169677TB | PB169677TB | MB  | 09/15/25 16:50 |  | Jaswal | OK |
| 58 | CCV05      | CCV05      | CCV | 09/15/25 16:55 |  | Jaswal | OK |
| 59 | CCB05      | CCB05      | CCB | 09/15/25 16:59 |  | Jaswal | OK |

SOP ID : M3010A-Digestion-17

SDG No : N/A

Matrix : WATER

Pipette ID : ICP A

Balance ID : N/A

Filter paper ID : N/A

pH Strip ID : M6069

Hood ID : #3

Block ID : 1. HOT BLOCK #1 2. N/A

Start Digest Date: 09/12/2025 Time : 10:05 Temp : 96 °C

End Digest Date: 09/12/2025 Time : 13:10 Temp : 96 °C

Digestion tube ID: M5595

Block thermometer ID: MET-DIG. #1

Dig Technician Signature: SKG.

Supervisor Signature: jpp

Temp : 1. 96°C 2. N/A

| Standard Name | MLS USED | STD REF. # FROM LOG |
|---------------|----------|---------------------|
| LFS-1         | 1.00     | M6180               |
| LFS-2         | 1.00     | M6181               |
| N/A           | N/A      | N/A                 |
| N/A           | N/A      | N/A                 |
| N/A           | N/A      | N/A                 |

| Chemical Used | ML/SAMPLE USED | Lot Number |
|---------------|----------------|------------|
| Conc. HNO3    | 3.00           | M6158      |
| 1:1 HCL       | 5.00           | MP85156    |
| N/A           | N/A            | N/A        |
| N/A           | N/A            | N/A        |
| N/A           | N/A            | N/A        |
| N/A           | N/A            | N/A        |
| N/A           | N/A            | N/A        |
| N/A           | N/A            | N/A        |
| N/A           | N/A            | N/A        |
| N/A           | N/A            | N/A        |

Extraction Conformance/Non-Conformance Comments:

HOT BLOCK#1 CELL#50 96C

| Date / Time    | Prepped Sample Relinquished By/Location | Received By/Location |
|----------------|-----------------------------------------|----------------------|
| 09/12/25 14:10 | SKG.mct di9                             | jpp   MetLab         |
|                | Preparation Group                       | Analysis Group       |

| Lab Sample ID | Client Sample ID | pH | Initial Vol (ml) | Final Vol (ml) | Color Before | Color After | Clarity Before | Clarity After | Comment     | Prep Pos |
|---------------|------------------|----|------------------|----------------|--------------|-------------|----------------|---------------|-------------|----------|
| PB169680BL    | PBW680           | <2 | 50               | 25             | Colorless    | Colorless   | Clear          | Clear         | N/A         | 8        |
| PB169680BS    | LCS680           | <2 | 50               | 25             | Colorless    | Colorless   | Clear          | Clear         | M6180,M6181 | 9        |
| Q3058-03      | MW4              | <2 | 50               | 25             | Colorless    | Colorless   | Clear          | Clear         | N/A         | 10       |
| Q3071-01      | 4089             | <2 | 50               | 25             | Colorless    | Colorless   | Clear          | Clear         | N/A         | 11       |
| Q3071-01MS    | 4089MS           | <2 | 50               | 25             | Colorless    | Colorless   | Clear          | Clear         | M6180,M6181 | 13       |
| Q3071-01MSD   | 4089MSD          | <2 | 50               | 25             | Colorless    | Colorless   | Clear          | Clear         | M6180,M6181 | 14       |
| Q3071-01DUP   | 4089DUP          | <2 | 50               | 25             | Colorless    | Colorless   | Clear          | Clear         | N/A         | 12       |



# SHIPPING DOCUMENTS

CLIENT INFORMATION

COMPANY: **G Environmental**  
ADDRESS:  
CITY: **8 CARRO** STATE: **NJ** ZIP:  
ATTENTION: **Shirley**  
PHONE: FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: **Buffington**  
PROJECT NO.: LOCATION:  
PROJECT MANAGER: **GL**  
e-mail:  
PHONE: FAX:

CLIENT BILLING INFORMATION

BILL TO: **G Environmental**  
ADDRESS: **8 CARRO**  
CITY: **Shirley** STATE: **NJ** ZIP:  
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) **5 day** DAYS\*  
HARDCOPY (DATA PACKAGE): DAYS\*  
EDD: DAYS\*  
\*TO BE APPROVED BY CHEMTECH  
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☒ Level 4 (QC + Full Raw Data)  
☐ Level 2 (Results + QC) ☒ NJ Reduced ☐ US EPA CLP  
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B  
+ Raw Data ☐ Other **Excel**  
EDD FORMAT **HAZOP NJ DEP**

PRESERVATIVES

COMMENTS

| ALLIANCE<br>SAMPLE<br>ID | PROJECT<br>SAMPLE IDENTIFICATION | SAMPLE<br>MATRIX | SAMPLE<br>TYPE |      | SAMPLE<br>COLLECTION |      | # OF BOTTLES | PRESERVATIVES |   |   |   |   |   |   |   |   | COMMENTS |
|--------------------------|----------------------------------|------------------|----------------|------|----------------------|------|--------------|---------------|---|---|---|---|---|---|---|---|----------|
|                          |                                  |                  | COMP           | GRAB | DATE                 | TIME |              | 1             | 2 | 3 | 4 | 5 | 6 | 7 | 8 | 9 |          |
| 1.                       | RPXY5                            | soil             | X              |      | 9/25                 | 0154 |              |               | X |   |   |   |   |   |   |   |          |
| 2.                       | RPXY4                            | soil             | X              |      | 9/25                 | 1034 |              |               | X |   |   |   |   |   |   |   |          |
| 3.                       | MW4                              | GW               | X              |      | 9/25                 | 1253 |              |               |   |   | X |   | X |   |   |   |          |
| 4.                       |                                  |                  |                |      |                      |      |              |               |   |   |   |   |   |   |   |   |          |
| 5.                       |                                  |                  |                |      |                      |      |              |               |   |   |   |   |   |   |   |   |          |
| 6.                       |                                  |                  |                |      |                      |      |              |               |   |   |   |   |   |   |   |   |          |
| 7.                       |                                  |                  |                |      |                      |      |              |               |   |   |   |   |   |   |   |   |          |
| 8.                       |                                  |                  |                |      |                      |      |              |               |   |   |   |   |   |   |   |   |          |
| 9.                       |                                  |                  |                |      |                      |      |              |               |   |   |   |   |   |   |   |   |          |
| 10.                      |                                  |                  |                |      |                      |      |              |               |   |   |   |   |   |   |   |   |          |

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

|                          |                             |                        |                                                                                                                                                                         |
|--------------------------|-----------------------------|------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| RELINQUISHED BY SAMPLER: | DATE/TIME: <b>9/25 1343</b> | RECEIVED BY: <b>CD</b> | Conditions of bottles or coolers at receipt: <input type="checkbox"/> COMPLIANT <input type="checkbox"/> NON COMPLIANT <input type="checkbox"/> COOLER TEMP <b>23°C</b> |
| 1. <b>Rel</b>            |                             | 1.                     | Comments: <b>It Got #1</b>                                                                                                                                              |
| RELINQUISHED BY SAMPLER: | DATE/TIME:                  | RECEIVED BY:           |                                                                                                                                                                         |
| 2.                       |                             | 2.                     |                                                                                                                                                                         |
| RELINQUISHED BY SAMPLER: | DATE/TIME:                  | RECEIVED BY:           |                                                                                                                                                                         |
| 3.                       |                             | 3.                     |                                                                                                                                                                         |

Page \_\_\_\_ of \_\_\_\_

CLIENT: ☐ Hand Delivered ☐ Other

Shipment Complete

☐ YES ☐ NO

### Laboratory Certification

| Certified By    | License No.      |
|-----------------|------------------|
|                 |                  |
| Connecticut     | PH-0830          |
|                 |                  |
| DOD ELAP (ANAB) | L2219            |
|                 |                  |
| Maine           | 2024021          |
|                 |                  |
| Maryland        | 296              |
|                 |                  |
| New Hampshire   | 255425           |
|                 |                  |
| New Jersey      | 20012            |
|                 |                  |
| New York        | 11376            |
|                 |                  |
| Pennsylvania    | 68-00548         |
|                 |                  |
| Soil Permit     | 525-24-234-08441 |
|                 |                  |
| Texas           | TX-C25-00189     |
|                 |                  |
| Virginia        | 460312           |

## LOGIN REPORT/SAMPLE TRANSFER

|                                       |        |                                               |                                 |
|---------------------------------------|--------|-----------------------------------------------|---------------------------------|
| <b>Order ID :</b> Q3058               | GENV01 | <b>Order Date :</b> 9/9/2025 1:49:00 PM       | <b>Project Mgr :</b>            |
| <b>Client Name :</b> G Environmental  |        | <b>Project Name :</b> Buffington              | <b>Report Type :</b> NJ Reduced |
| <b>Client Contact :</b> Gary Landis   |        | <b>Receive DateTime :</b> 9/9/2025 1:43:00 PM | <b>EDD Type :</b> NJ HAZSITE    |
| <b>Invoice Name :</b> G Environmental |        | <b>Purchase Order :</b>                       | <b>Hard Copy Date :</b>         |
| <b>Invoice Contact :</b> Gary Landis  |        |                                               | <b>Date Signoff :</b>           |

| LAB ID   | CLIENT ID | MATRIX | SAMPLE DATE | SAMPLE TIME | TEST          | TEST GROUP | METHOD   | FAX DATE | DUE DATES   |
|----------|-----------|--------|-------------|-------------|---------------|------------|----------|----------|-------------|
| Q3058-01 | RPXY5     | Solid  | 09/09/2025  | 09:45       |               |            |          |          |             |
|          |           |        |             |             | VOC-TCLVOA-10 |            | 8260D    |          | 5 Bus. Days |
| Q3058-02 | RPXY4     | Solid  | 09/09/2025  | 10:30       |               |            |          |          |             |
|          |           |        |             |             | VOC-TCLVOA-10 |            | 8260D    |          | 5 Bus. Days |
| Q3058-03 | MW4       | Water  | 09/09/2025  | 12:15       |               |            |          |          |             |
|          |           |        |             |             | VOCMS Group1  |            | 8260-Low |          | 5 Bus. Days |

**Relinquished By :** af  
**Date / Time :** 9/9/25 15:03

**Received By :** Sam  
**Date / Time :** 09/09/25 15:03

**Storage Area :** VOA Refridgerator Room

*Handwritten notes:*  
Ry H 4  
Ref #6  
P 72